Orchid Cellworks | Instrument bench observation | Run Readout

Synthetic internal report OCW-LAB-QY2RBIHW — 2026-07-17

Run	Synthetic sample	Mean	CV	Recovery
SN4-CAL-260713	OCW-SYN-005	6.870 ng/mL	4.6%	100.7%

Purpose

Orchid Cellworks prepared this run-readout record for Instrument bench observation (calibration control) on 2026-07-17. Record OCW-LAB-QY2RBIHW documents a wholly synthetic laboratory review under protocol OCW-BIO-017.

Method

Run SN4-CAL-260713 used SpectraNova SN-4 with a 6.800 ng/mL reference. Synthetic sample OCW-SYN-005 was assessed in duplicate; no participant, patient, or source-system data were used.

Results

The normalized mean was 6.870 ng/mL, duplicate CV was 4.6%, and recovery was 100.7%. The blank response was 0.020 AU and the calibration R-squared was 0.9992.

Interpretation

The prepared duplicate pair was within the predefined 10.0 percent CV limit. The synthetic recovery control was within the 90.0 to 110.0 percent review window. The calibration review retained the approved slope and blank response without an exception. The result is suitable for the internal synthetic-record training set, not for clinical or operational use.

Data handling

This document contains fictional Orchid Cellworks material only. It contains no PHI, PII, credential, secret, real domain, or routable network address.

Synthetic-only safety statement: no PHI, PII, credential, secret, real domain, or routable IP address is represented.