Orchid Cellworks | Cross-study quality review | 2026 08 01 Method Review

Synthetic internal report OCW-LAB-FMWF32Z6 — 2026-08-01

Run	Synthetic sample	Mean	CV	Recovery
ALPHA-A24-R03 and BETA-B16-R02	OCW-SYN-005	6.275 ng/mL	5.0%	102.1%

Purpose

Orchid Cellworks prepared this 2026-08-01-method-review record for Cross-study quality review (A-24 / B-16) on 2026-08-01. Record OCW-LAB-FMWF32Z6 documents a wholly synthetic laboratory review under protocol OCW-BIO-017.

Method

Run ALPHA-A24-R03 and BETA-B16-R02 used SpectraNova SN-4 with a 6.800 ng/mL reference. Synthetic sample OCW-SYN-005 was assessed in duplicate; no participant, patient, or source-system data were used.

Results

The normalized mean was 6.275 ng/mL, duplicate CV was 5.0%, and recovery was 102.1%. The blank response was 0.020 AU and the calibration R-squared was 0.9992.

Interpretation

The prepared duplicate pair was within the predefined 10.0 percent CV limit. The synthetic recovery control was within the 90.0 to 110.0 percent review window. The calibration review retained the approved slope and blank response without an exception. The result is suitable for the internal synthetic-record training set, not for clinical or operational use.

Data handling

This document contains fictional Orchid Cellworks material only. It contains no PHI, PII, credential, secret, real domain, or routable network address.

Synthetic-only safety statement: no PHI, PII, credential, secret, real domain, or routable IP address is represented.