

Project: Orchid Cellworks

Date: 2026-07-13

Study: Study Alpha

Page: 1 of 1

Cohort/ID: A-24

Timepoint: Baseline

Objective: Baseline biomarker assay (Sample Set A)

Assay: Immunoassay (Sandwich ELISA)

Plate ID: ELISA-20260713-A

Operator Initials: PB


Reagents: Capture Ab (Lot 0726C) Detection Ab (Lot 0726D)

Standard (Lot 0726S) Diluent (Lot 0726B)

Pre-Run Checks:

- ☒ Reagents within expiry
- ☒ Plate reader check (450 nm)
- ☒ Pipettes calibrated
- ☒ Standards prepared
- ☒ Wash buffer ready
- ☒ Workspace clean

Pipette Sketch (P200):



Procedure Summary:

1. Added 100 µL standards, controls, samples
2. Added 100 µL detection Ab
3. Incubated 60 min @ RT
4. Washed 3x
5. Added 100 µL substrate
6. Incubated 15 min (dark)
7. Added 50 µL stop solution
8. Read plate @ 450 nm

Time Log (24h):

Start Assay	08:42
End Incubation	09:42
End Wash	09:50
Read Start	09:55
Read End	10:05

Result / Observation:

Calculated concentration (Sample A-24):

6.8 ng/mL

QC Notes:

All QC within range. Standard curve $R^2 = 0.998$.
No issues observed.

QC Status:

- ☒ Pass
- ☐ Review
- ☐ Fail

Additional Notes: Plate stored at 4°C. Data uploaded to study drive.