

# Instrument Calibration Summary

Orchid Cellworks | Synthetic instrument operations | Reference date 2026-07-13

## Scope

SpectraNova SN-4, PipetSure P200, ThermoNest IN-2; lot CAL-260713-A.

## Acceptance criteria

Plate-reader slope 0.998; R2 0.9992; blank 0.02 AU.

## Study linkage

Study Alpha / cohort A-24 / assay result 6.8 ng/mL recorded only as synthetic QA context.

# Instrument Calibration Summary

Orchid Cellworks | Synthetic instrument operations | Reference date 2026-07-13

## Results

All six calibration forms completed on 2026-07-13.

## Review

Instrument lead review scheduled 17:30 UTC; no deviations requiring escalation.

## Controls

Synthetic IDs and values only; not derived from human samples.