

Instrument Maintenance History

Orchid Cellworks | Synthetic instrument operations | Reference date 2026-07-13

SpectraNova SN-4

2026-07-13: lamp output, wavelength, and linearity checks accepted.

PipetSure P200

2026-07-13: gravimetric mean 199.7 uL; accepted.

ThermoNest IN-2

2026-07-13: chamber map within 36.8-37.2 C.

Instrument Maintenance History

Orchid Cellworks | Synthetic instrument operations | Reference date 2026-07-13

Traceability

Calibration lot CAL-260713-A retained as a synthetic operational reference.

Study context

Study Alpha cohort A-24 checks supported assay 6.8 ng/mL control review.

Disposition

No maintenance hold; next synthetic review due 2026-08-13.