

SpectraNova wavelength verification

Wavelength / OCW-WL-012

Orchid Cellworks


Wavelength 450 nm verified

2026-07-13

Reference filter: W-450

Deviation: 0.3 nm

Status: ACCEPT



Synthetic calibration record - no clinical or participant data

