

Study Alpha assay run 03 readout

Orchid Cellworks · wholly synthetic method-training record · 2026-07-13

Control and acceptance record for ALPHA-A24-R03, cohort A-24.

Control	Observed	Acceptance	Decision
Blank	0.02 AU	≤0.05 AU	Pass
Calibration slope	0.998	0.995–1.005	Pass
Duplicate CV	8.0%	≤10.0%	Pass
Recovery	96%	90–110%	Pass
Median assay	6.8 ng/mL	6.0–7.5 ng/mL	Accept

All listed observations and values are synthetic; no external specimens, participants, or patient data were used.