

Assay acceptance criteria

Orchid Cellworks | Study Alpha | 2026-07-13 | cohort A-24 | assay 6.8 ng/mL

Blank control: 0.02 AU.

Calibration slope target: 0.998; R squared: 0.9992.

Duplicate CV limit: 10 percent.

Recovery band: 90 to 110 percent.

Result release is a synthetic documentation exercise.