
Study Alpha A-24 Assay Method Review

A synthetic protocol-readiness review for a plate-based quantification workflow.

This original Orchid Cellworks review is anchored to Study Alpha on 2026-07-13 for cohort A-24. It describes a fictional planning target of assay 6.8 ng/mL; it is not a result about people, specimens, patients, or an external study.

Run frame

Element	Synthetic planning record	Review implication
Protocol	OCW-BIO-017	Lock method intent before reviewing exports
Run lineage	ALPHA-A24-R03	Keep plate and review notes connected
Duplicate check	CV <=10%	Treat higher variability as a rerun discussion
Recovery window	90-110%	Use only as a fictional acceptance convention
Controls	blank, low, mid, high	Require all controls before interpretation

Control-first interpretation

The workflow makes control behavior visible before any summary value is discussed. Duplicate agreement supports repeatability of the planned handling step, while recovery supports a narrow check of the fictional spiking procedure. Neither measure establishes biological truth.

Assay review pathway

Decision sequence for OCW-BIO-017 / ALPHA-A24-R03

Review point	Evidence retained	Action when outside planning window
Plate setup	Layout map and control locations	Pause before readout
Duplicate behavior	Paired raw values and CV note	Review pipetting and rerun only affected placement
Recovery	Synthetic spike reconciliation	Record method variance; do not infer sample behavior
Readout summary	Median planning target 6.8 ng/mL	Label as illustrative planning value
Reviewer handoff	Signed synthetic review note	Close only after controls and exception note agree

Conclusion

For this fictional run, the defensible conclusion is operational: the protocol exposes a reviewable chain from controls through duplicate behavior to a bounded synthetic summary. It does not make a clinical, biological, or product-performance claim.