

This is to certify that the equipment listed below has been calibrated in accordance with applicable quality procedures and metrology standards.

EQUIPMENT INFORMATION

Equipment Type : Digital Caliper
Equipment ID : **CAL-ALPHA-18**
Manufacturer : ForgeLine Metrology
Product : **Product Alpha**
Measurement Range : 0 - 150 mm
Resolution : 0.01 mm

Certificate Date

2026-08-11

Certificate No.

NCR-026

Tolerance

Tolerance: +/- 0.18 mm

CALIBRATION CONDITIONS

Ambient Temperature : 20.2 °C
Relative Humidity : 45 %
Calibration Standard : Gauge Block Set (Grade 1)
Traceability : Traceable to national standards

CALIBRATION RESULTS

Nominal Value (mm)	Indicated Value (mm)	Deviation (mm)	Maximum Allowable Deviation (mm)	Result
0.00	0.00	0.00	+/- 0.18	PASS
25.00	25.01	0.01	+/- 0.18	PASS
50.00	49.99	-0.01	+/- 0.18	PASS
75.00	75.02	0.02	+/- 0.18	PASS
100.00	100.01	0.01	+/- 0.18	PASS
125.00	124.99	-0.01	+/- 0.18	PASS
150.00	150.02	0.02	+/- 0.18	PASS

UNCERTAINTY OF MEASUREMENT

Expanded Uncertainty (k=2) : 0.05 mm

The measurement uncertainty has been determined in accordance with internal procedures and is valid at the time of calibration.

NOTES

1. This certificate applies only to the item identified above.
2. Recalibration is recommended at appropriate intervals.

