



191.0264
167.7078
154.5384
149.6492
145.2622
144.2751
143.9683
142.5726
141.3389
139.8237
132.2246
130.9732
130.1327
128.8594
127.5072
126.9526
125.7049
125.4990
119.9795
110.4184
108.9810
106.3721

61.6719
56.2001
56.0362

21.3320
14.1038

Current Data Parameters
NAME CLB
EXPNO 5
PROCNO 1

F2 - Acquisition Parameters
Date_ 20251219
Time_ 12.35 h
INSTRUM Avance
PROBHD z151574_0199 (
PULPROG zgpg30
TD 65536
SOLVENT CDCl3
NS 400
DS 4
SWH 30120.482 Hz
FIDRES 0.919204 Hz
AQ 1.0878977 sec
RG 101
DE 16.600 usec
TE 297.9 K
D1 2.00000000 sec
D11 0.03000000 sec
TD0 1
SFO1 125.7703643 MHz
NUC1 13C
P0 3.00 usec
P1 9.00 usec
PLW1 96.00000000 W
SFO2 500.1320005 MHz
NUC2 1H
CPDPRG2 A000
PCPD2 80.00 usec
PLW2 24.00000000 W
PLW12 0.23999999 W
PLW13 0.12072000 W

F2 - Processing parameters
SI 32768
SF 125.7577885 MHz
WDW EM
SSB 0
LB 1.00 Hz
GB 0
PC 1.40

