

Temperature Humidity Bias Test Specification

Project: Project Nimbus
Product: ThermaGuard M4
Status: RELEASED DEMO SPECIFICATION
Primary method: REL-006

1. Purpose and scope

This synthetic demonstration specification defines the validation intent, sample population, test sequence, operating conditions and acceptance criteria for ThermaGuard M4. It is provided only as seeded LabOS prototype content.

2. Test definition

Parameter	Requirement
Primary condition	85 C / 85%RH
Exposure / sequence	336 h exposure
Sample population	10 DUTs minimum; genealogy maintained by LabOS
Preconditioning	24 h at laboratory ambient unless otherwise specified
Data capture	Record test conditions, equipment, calibration, operator, DUT and evidence
Acceptance	No functional loss; leakage below released limit

3. Readiness requirements

Before execution: confirm the released method revision, correct DUT population, valid calibrated equipment, qualified operator, predecessor completion, sample availability and approved acceptance criteria. Any deviation shall be recorded as a LabOS issue and shall trigger replanning when it affects availability or duration.

4. Evidence and reporting

Raw data, result parameters, pass/fail evaluation, photographs where relevant, equipment and calibration traceability, operator identity and final report evidence shall be linked to the applicable test leg and validation requirements.