

Software HIL Regression Test Specification

Project: Project Nova

Product: SmartDAQ D9

Status: RELEASED DEMO SPECIFICATION

Primary method: SW-009

1. Purpose and scope

This synthetic demonstration specification defines the validation intent, sample population, test sequence, operating conditions and acceptance criteria for SmartDAQ D9. It is provided only as seeded LabOS prototype content.

2. Test definition

Parameter	Requirement
Primary condition	Nominal and boundary conditions
Exposure / sequence	Automated regression suite
Sample population	10 DUTs minimum; genealogy maintained by LabOS
Preconditioning	24 h at laboratory ambient unless otherwise specified
Data capture	Record test conditions, equipment, calibration, operator, DUT and evidence
Acceptance	100% critical test cases pass; no Severity-1 defect

3. Readiness requirements

Before execution: confirm the released method revision, correct DUT population, valid calibrated equipment, qualified operator, predecessor completion, sample availability and approved acceptance criteria. Any deviation shall be recorded as a LabOS issue and shall trigger replanning when it affects availability or duration.

4. Evidence and reporting

Raw data, result parameters, pass/fail evaluation, photographs where relevant, equipment and calibration traceability, operator identity and final report evidence shall be linked to the applicable test leg and validation requirements.

DEMO DOCUMENT - not a released engineering specification for production use.