

Acquisition Method Report

Acquisition Method Info

Method Name NH415_1-18-21_H.m
Method Path D:\MassHunter\methods\Thomas\NH415_1-18-21_H.m
Method Description Default Method

Device List

Sampler
 Binary Pump
 Column Oven
 QQQ

MS QQQ Mass Spectrometer

Ion Source AJS ESI **Tune File** D:\MassHunter\Tune\QQQ\G6470A
 \atunes.TUNE.XML
Stop Mode By StopTime **Stop Time (min)** 40
Time Filter On **Time Filter Width (min)** 0.1

Time Segments

Index	Start Time (min)	Scan Type	Ion Mode	Div Valve	Delta EMV (+)	Delta EMV (-)	Store
1	0	MRM	ESI+Agilent Jet Stream	To MS	300	300	Yes

Time Segment 1

Scan Segments

Cpd Group	Cpd Name	ISTD?	Prec Ion	MS1 Res	Prod Ion	MS2 Res	Dwell	Frag (V)	CE (V)	Cell Acc (V)	Polarity
	GMP	No	364	Unit/Enh (6490)	152	Unit/Enh (6490)	10	40	18	5	Positive
	IMP	No	349	Unit/Enh (6490)	137	Unit/Enh (6490)	10	40	19	5	Positive
	AMP	No	348	Unit/Enh (6490)	136	Unit/Enh (6490)	10	40	21	5	Positive
	UMP M+1	No	326	Unit/Enh (6490)	97	Unit/Enh (6490)	10	30	14	5	Positive
	UMP	No	325	Unit/Enh (6490)	97	Unit/Enh (6490)	10	30	14	5	Positive
	Guanosine	No	284	Unit/Enh (6490)	152	Unit/Enh (6490)	10	30	20	5	Positive
	Inosine	No	269	Unit/Enh (6490)	137	Unit/Enh (6490)	10	35	20	5	Positive
	deoxyguanosine	No	268.1	Unit/Enh (6490)	152	Unit/Enh (6490)	10	30	14	5	Positive
	Adenosine	No	268	Unit/Enh (6490)	136	Unit/Enh (6490)	10	40	18	5	Positive
	deoxyadenosine	No	252	Unit/Enh (6490)	136	Unit/Enh (6490)	10	35	20	5	Positive
	Uridine Standard	No	249.2	Unit/Enh (6490)	112.96	Unit/Enh (6490)	10	30	14	5	Positive
	Uridine M+1	No	246.2	Unit/Enh (6490)	115.96	Unit/Enh (6490)	10	30	14	5	Positive
	Cytidine M+1	No	245.22	Unit/Enh (6490)	111.92	Unit/Enh (6490)	10	28	10	5	Positive
	Uridine	No	245.2	Unit/Enh (6490)	112.96	Unit/Enh (6490)	10	30	14	5	Positive
	Cytidine	No	244.22	Unit/Enh (6490)	111.92	Unit/Enh (6490)	10	28	10	5	Positive
	deoxyuridine	No	229.1	Unit/Enh (6490)	113	Unit/Enh (6490)	10	28	20	5	Positive
	deoxycytidine	No	228.1	Unit/Enh (6490)	112	Unit/Enh (6490)	10	28	20	5	Positive
	Arginine Standard	No	185.2	Unit/Enh (6490)	64	Unit/Enh (6490)	10	40	14	5	Positive

Acquisition Method Report



Agilent

Trusted Answers

Cpd Group	Cpd Name	ISTD?	Prec Ion	MS1 Res	Prod Ion	MS2 Res	Dwell	Frag (V)	CE (V)	Cell Acc (V)	Polarity
	Citrulline	No	176.1	Unit/Enh (6490)	70	Unit/Enh (6490)	10	35	13	5	Positive
	Arginine	No	175.2	Unit/Enh (6490)	60	Unit/Enh (6490)	10	40	14	5	Positive
	Xanthene	No	153	Unit/Enh (6490)	110	Unit/Enh (6490)	10	30	20	5	Positive
	Hypoxanthine	No	137.1	Unit/Enh (6490)	109.9	Unit/Enh (6490)	10	44	20	5	Positive
	Aspartate	No	134	Unit/Enh (6490)	74	Unit/Enh (6490)	10	40	15	5	Positive
	Ornithine	No	133	Unit/Enh (6490)	70	Unit/Enh (6490)	10	35	12	5	Positive
	Leucine	No	132	Unit/Enh (6490)	86.1	Unit/Enh (6490)	10	40	15	5	Positive
	d2_ND	No	127	Unit/Enh (6490)	84	Unit/Enh (6490)	10	100	20	5	Positive
	Proline M+5	No	121	Unit/Enh (6490)	74.1	Unit/Enh (6490)	10	80	20	5	Positive
	Proline	No	116	Unit/Enh (6490)	70	Unit/Enh (6490)	10	80	20	5	Positive
	Serine Standard	No	109	Unit/Enh (6490)	63	Unit/Enh (6490)	10	40	13	5	Positive
	Serine	No	106	Unit/Enh (6490)	60	Unit/Enh (6490)	10	40	13	5	Positive
	Beta- Aminolsobuty rate M+3	No	105	Unit/Enh (6490)	39	Unit/Enh (6490)	10	40	30	5	Positive
	Beta- Aminolsobuty rate	No	104	Unit/Enh (6490)	39	Unit/Enh (6490)	10	40	30	5	Positive
	Alanine M+1	No	91	Unit/Enh (6490)	45	Unit/Enh (6490)	10	40	17	5	Positive
	Alanine	No	90	Unit/Enh (6490)	44	Unit/Enh (6490)	10	40	17	5	Positive
	Glycine	No	76	Unit/Enh (6490)	30	Unit/Enh (6490)	10	40	20	5	Positive
	ATP Standard	No	542.9	Unit/Enh (6490)	159	Unit/Enh (6490)	10	80	50	5	Negative
	dATP Standard	No	527	Unit/Enh (6490)	159	Unit/Enh (6490)	10	35	32	5	Negative
	GTP	No	522.2	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	32	5	Negative
	dGTP	No	506	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	32	5	Negative
	ATP	No	506	Unit/Enh (6490)	159	Unit/Enh (6490)	10	35	40	5	Negative
	dATP	No	490.1	Unit/Enh (6490)	159	Unit/Enh (6490)	10	35	32	5	Negative
	UTP M+1	No	484	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	CTP M+1	No	483.16	Unit/Enh (6490)	159	Unit/Enh (6490)	10	34	36	5	Negative
	UTP	No	483	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	CTP	No	482.16	Unit/Enh (6490)	159	Unit/Enh (6490)	10	34	36	5	Negative
	dTTP M+1	No	482	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	dTTP	No	481	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	46	5	Negative
	dUTP M+1	No	468	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	46	5	Negative
	dCTP M+1	No	467	Unit/Enh (6490)	159	Unit/Enh (6490)	10	34	35	5	Negative
	dUTP	No	467	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	dCTP	No	466	Unit/Enh (6490)	159	Unit/Enh (6490)	10	34	35	5	Negative

Acquisition Method Report



Agilent

Trusted Answers

Cpd Group	Cpd Name	ISTD?	Prec Ion	MS1 Res	Prod Ion	MS2 Res	Dwell	Frag (V)	CE (V)	Cell Acc (V)	Polarity
	GDP	No	442	Unit/Enh (6490)	150	Unit/Enh (6490)	10	35	24	5	Negative
	ADP	No	426	Unit/Enh (6490)	134	Unit/Enh (6490)	10	135	20	5	Negative
	UDP M+1	No	404.16	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	CDP M+1	No	403.18	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	UDP	No	403.16	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	CDP	No	402.18	Unit/Enh (6490)	159	Unit/Enh (6490)	10	30	26	5	Negative
	dTDP M+1	No	402	Unit/Enh (6490)	79	Unit/Enh (6490)	10	36	44	5	Negative
	dTDP	No	401	Unit/Enh (6490)	79	Unit/Enh (6490)	10	36	44	5	Negative
	AMP Neg	No	346	Unit/Enh (6490)	79	Unit/Enh (6490)	10	40	21	5	Negative
	UMP Neg M+1	No	324	Unit/Enh (6490)	79	Unit/Enh (6490)	10	40	30	5	Negative
	CMP M+1	No	323	Unit/Enh (6490)	79	Unit/Enh (6490)	10	35	30	5	Negative
	UMP Neg	No	323	Unit/Enh (6490)	79	Unit/Enh (6490)	10	40	30	5	Negative
	dTMP M+1	No	322.2	Unit/Enh (6490)	79	Unit/Enh (6490)	10	30	30	5	Negative
	CMP	No	322	Unit/Enh (6490)	79	Unit/Enh (6490)	10	35	30	5	Negative
	dTMP	No	321	Unit/Enh (6490)	79	Unit/Enh (6490)	10	30	30	5	Negative
	dUMP M+1	No	308	Unit/Enh (6490)	79	Unit/Enh (6490)	10	40	30	5	Negative
	dUMP	No	307	Unit/Enh (6490)	79	Unit/Enh (6490)	10	40	30	5	Negative
	Thymidine M+1	No	242	Unit/Enh (6490)	44	Unit/Enh (6490)	10	35	10	5	Negative
	Thymidine	No	241	Unit/Enh (6490)	42	Unit/Enh (6490)	10	35	10	5	Negative
	Dihydroorotate	No	157	Unit/Enh (6490)	113	Unit/Enh (6490)	10	35	12	5	Negative
	Carbamoyl-phosphate	No	140	Unit/Enh (6490)	79	Unit/Enh (6490)	10	70	20	5	Negative
	Fumaric acid d2 Std	No	117	Unit/Enh (6490)	72.8	Unit/Enh (6490)	10	50	20	5	Negative
	Uracil M+1	No	112	Unit/Enh (6490)	44	Unit/Enh (6490)	10	30	20	5	Negative
	Uracil	No	111	Unit/Enh (6490)	42	Unit/Enh (6490)	10	30	20	5	Negative
	Lactate	No	89	Unit/Enh (6490)	43	Unit/Enh (6490)	10	40	16	5	Negative

Source Parameters

Parameter	Value (+)	Value (-)
Gas Temp (°C)	350	350
Gas Flow (l/min)	9	9
Nebulizer (psi)	50	50
SheathGasHeater	375	375
SheathGasFlow	12	12
Capillary (V)	3500	3500
VCharging	1000	1000

Scan Parameters

Data Stg	Threshold
Centroid	0

Acquisition Method Report



Agilent

Trusted Answers

Chromatograms

Chrom Type	Label	Offset	Y-Range
TIC	TIC	0	10000000

Instrument Curves

Actual

Name: Sampler

Module: G7129B

Auxiliary

Draw Speed	100 µL/min
Eject Speed	400 µL/min
Wait Time After Draw	1.2 s
Needle Height Offset	0.0 mm

Injection

Injection Mode	Standard injection
Injection Volume	10.00 µL

High throughput

Sample Flush-Out Factor	5.0
Injection Valve to Bypass for Delay Volume Reduction	No

Overlapped Injection

Overlapped Injection Mode	Off
---------------------------	-----

Stop Time

Stoptime Mode	As pump/No limit
---------------	------------------

Post Time

Posttime Mode	Off
---------------	-----

Name: Binary Pump

Module: G7120A

Flow	0.350 mL/min
Use Solvent Types	Yes
Stroke Mode	Synchronized
Low Pressure Limit	0.00 bar
High Pressure Limit	700.00 bar
Max. Flow Ramp Up	100.000 mL/min ²
Max. Flow Ramp Down	100.000 mL/min ²
Expected Mixer	No check

Stroke A

Automatic Stroke Calculation A	Yes
--------------------------------	-----

Stop Time

Stoptime Mode	Time set
Stoptime	40.00 min

Post Time

Posttime Mode	Off
---------------	-----

Solvent Composition

	Channel	Ch. 1 Solv.	Name 1	Ch2 Solv.	Name 2	Selected	Used	Percent (%)
1	A	100.0 % Water V.02		100.0 % Water V.02		Ch. 1	Yes	15.00 %
2	B	100.0 % Acetonitrile V.02		100.0 % Acetonitrile V.02		Ch. 1	Yes	85.00 %

Timetable

	Time (min)	A (%)	B (%)	Flow (mL/min)
1	5.00 min	70.00 %	30.00 %	0.350 mL/min
2	30.00 min	98.00 %	2.00 %	0.350 mL/min
3	35.00 min	98.00 %	2.00 %	0.350 mL/min
4	40.00 min	15.00 %	85.00 %	0.350 mL/min



	Pressure (bar)
1	700.00 bar
2	700.00 bar
3	700.00 bar
4	700.00 bar

Acquisition Method Report



Agilent

Trusted Answers

	Time (min)	A (%)	B (%)	Flow (mL/min)
5	45.00 min	15.00 %	85.00 %	0.350 mL/min

Name: Column Oven

Module: G7130A

Temperature Control

Temperature Control Mode

Not Controlled

Temperature Not Ready Limit

Temperature Not Ready Limit On

Yes

Temperature Not Ready Limit Value

0.8 °C

Temperature Equilibration Time

0.0 min

Enforce column for run

Enforce column for run enabled

No

Stop Time

Stoptime Mode

As pump/injector

Post Time

Posttime Mode

Off

Timetable



	Pressure (bar)
5	700.00 bar