

Passive photodiode visibility — run report

RUN

214714_pd-visibility_pd-visibility

Status: failed | Started: 2026-09-21T20:47:14.003018+00:00

PURPOSE

Measure fringe contrast on the selected PD input under external drive. No RP output initialization or PAX connection.

REQUIRED OPTICAL SETUP

DC-coupled PD on the selected RP input; external drive must span complete fringes. Set its frequency and, when known, the blocked-light voltage offset. Existing Pyrpl monitor server required.

RECORDED BENCH CONTEXT (RECIPE)

In front of NPBS 1, C port

Red Pitaya (0-1 V command) -> Thorlabs MDT690 -> piezo

DATA

0 recorded samples

Source: data.csv; full acquisition settings: recipe.json; status: run.json

RUN ISSUE

RuntimeError: Cannot attach to existing RP monitor server; scope-only mode never starts a server or reloads the FPGA

Run log

RECORDED OUTPUT

WARNING:pyrpl.redpitaya_client:Socket error during connection attempt 0.

Run failed: RuntimeError: Cannot attach to existing RP monitor server; scope-only mode never starts a server or reloads the FPGA

Active parameters

RP commands: volts, bounded to 0–1. V_lambda: terminal volts for 2π. Gains: terminal V / RP V.

Parameter	Recorded value
label	"pd-visibility"
duration_s	12.0
report	"pdf"
pd_input	"in2"
visibility_frequency_hz	0.5
visibility_dark_voltage_v	null