



Info for small sample holder, update on angle file

From Anderson, Nikolaus J - (nja) <nanderson@optics.arizona.edu>

Date Thu 7/2/2026 3:27 PM

To Krell, Jacob Patrick - (jacobpkrell) <jacobpkrell@arizona.edu>

Cc Mavko, Matthew Edward - (mmavko) <mmavko@arizona.edu>

 1 attachment (2 MB)

PXL_20260702_221744981.jpg;

Jacob,

First off, to follow up on the smaller sample holder I think might work for a 3D print job to hold your 3x6" samples, it will need to include:

- 4x threaded $\frac{1}{4}$ -20 holes at 4.5" spacing apart (square pattern centered around center axis of sample)
 - I used these types of threaded inserts in other printed parts, they seem to work well: <https://www.mcmaster.com/94459A380/>
- An 0.750" height from the bottom of your new mounting base to the top surface of the sample under measurement.

I attached a photo of the mirror mount which is currently attached to the hub and baseplate we'll reuse for your mounting assembly. You can see the 4 screws I'm mentioning. Basically, we would remove the 4 screws holding this bare aluminum mirror assembly in place and install your new mount assembly to the black-coated baseplate.

Also, I went through and updated things to be ready for next week. I have your wavelengths programmed in (including 632 nm for this first set with Aeroglaze), and I went through the angle file piece by piece. I ended up modifying a couple of angles in addition to deleting the angles we agreed didn't work from the previous set. I realized looking back at our prior work that 12° (not 10°) is the minimum angle for the VNIR instrument away from the source to ensure no effect/shading. So that will look slightly different but otherwise we have a trimmed down and more useful overall angle set to run with next week. I'll get things going with Aeroglaze Monday morning and we can team up with Matt as we continue. If you can give us a heads up for when you'll be in, we can team up on a changeover or two before we hand things over once you're around.

Have a good long weekend!

Nik