REDLIGHT THERAPY CO.

Testing and Documentation

Every product in this range comes with supporting documents. The pages that follow are one-page summaries, prepared by Redlight Therapy Co. from third-party test reports, certificates and instrument measurements, with manufacturer details removed.

Product	Pages in this pack	Model named on the documents
300 LED Red Light Panel	EMC test summary, compliance certificate summary, light output measurement, LED measurement	RT1500S+
Red Light Therapy Hair Cap	EMC and electrical safety certificate summary	QZ022, QZ036
Red Light Therapy Knee Strap	EMC test summary, declaration of conformity summary	K8 (K6 and K9 also named)
Red Light Therapy Face Mask	EMC test summary, certificate of conformity summary	V-F2509 (V-F239 and V-F239Pro also listed)
Starter LED Face Mask	EMC test summary, certificate of compliance summary	SA-P02M01 (tested); SA-P02M01 to M09 covered

How to read these pages

- **EMC** (electromagnetic compatibility) tests check that a device does not disturb other electronics, and keeps working normally when exposed to interference.
- **Electrical safety** is covered by the low voltage certificate for the hair cap.
- **Light measurements** for the panel show the colours of light it gives out (660 nm red and near-infrared) and how steady the output is.
- **“Not applicable”** on a page means the lab recorded that test as not relevant to the product, for example mains tests on a battery-powered device.

Please note

- These pages are summaries, not certificates. Each is based on a single tested sample and does not assess every unit produced.
- The documents cover EMC, electrical safety and light measurements. They do not make or support any medical or health claim.
- Product images are for illustration.

EMC Test Summary

Light Therapy Mask | Summary of Test Report No. HX251230E001

Product	Light Therapy Mask	**Overall result**	**PASS**
Model tested	V-F2509	**Date of issue**	Dec. 24, 2025
Also listed	V-F239, V-F239Pro	**Test dates**	Dec. 22 - 24, 2025
Power supply	5V / 1A (DC)	**Operating mode**	Working, high speed
Test lab	Shenzhen Huaxin Information Technology Service Co., Ltd	**Lab accreditation**	A2LA Cert. No. 6855.01 (ISO/IEC 17025:2017)

Standards Tested Against

EN IEC 55014-1:2021 (emissions) • EN IEC 55014-2:2021 (immunity) • EN IEC 61000-3-2:2019+A1:2021 • EN 61000-3-3:2013+A1:2019+A2:2021

Emissions

Test	Standard	Result	Worst-case margin below limit
Radiated emission (30 MHz - 1 GHz)	EN IEC 55014-1	**Pass**	5.5 dB (horizontal, 208.8 MHz, QP)
Conducted emission, AC mains (150 kHz - 30 MHz)	EN IEC 55014-1	**Pass**	18.5 dB (average detector, 11.8 MHz)
Disturbance power	EN IEC 55014-1	Not applicable	-
Harmonic current emissions	EN IEC 61000-3-2	Not applicable	-
Voltage fluctuations and flicker	EN 61000-3-3	Not applicable	-

Immunity (EN IEC 55014-2:2021)

Test	Standard	Level applied	Criterion required / achieved	Result
Electrostatic discharge	EN 61000-4-2	Contact ±4 kV; Air ±8 kV	B / A	**Pass**
Radiated RF electromagnetic field	EN IEC 61000-4-3	3 V/m, 80-1000 MHz plus 1800, 2600, 3500, 5000 MHz; 80% AM	A / A	**Pass**
Electrical fast transient / burst	EN 61000-4-4	-	-	Not applicable
Surge	EN 61000-4-5	-	-	Not applicable
Conducted RF disturbances	EN 61000-4-6	-	-	Not applicable
Voltage dips and interruptions	EN IEC 61000-4-11	-	-	Not applicable

“Not applicable” means the lab recorded the test as not applicable for this product as tested (DC-powered, 5V/1A supply). Criterion A: normal operation continues during the test. Criterion B: normal operation continues after the test.

About this document

This page is a summary prepared by the seller from test report HX251230E001, issued by Shenzhen Huaxin Information Technology Service Co., Ltd. It is not the laboratory's report and is not a certificate. Test results apply only to the sample tested. The full original report is available on request.

Conformity Certificate Summary

Light Therapy Mask | Certificate of Conformity, EMC Directive 2014/30/EU

Document	Certificate of Conformity
Registration No.	HX251230E001C
Date of issue	Dec. 30, 2025
Issued by	Shenzhen Huaxin Information Technology Service Co., Ltd., Shenzhen, China
Product	Light Therapy Mask
Models	V-F2509, V-F239 / V-F239Pro
Directive	EMC Directive 2014/30/EU
Test report No.	HX251230E001

Standards Listed

EN IEC 55014-1:2021 • EN IEC 55014-2:2021 • EN IEC 61000-3-2:2019+A1:2021 • EN 61000-3-3:2013+A1:2019+A2:2021

What the certificate states

- The submitted products were tested against the listed standards and found in compliance with the EMC Directive.
- The applicant is authorised to use the certificate in connection with the EU declaration of conformity to the Directive; it applies only to the equipment described.
- The CE marking can be affixed after the necessary technical documentation is prepared.

About this document

This page is a summary prepared by the seller from a certificate of conformity issued by Shenzhen Huaxin Information Technology Service Co., Ltd. It is not a certificate and does not itself contain test results. The test report named above holds the measured results.