REDLIGHT THERAPY CO.

Testing and Documentation

Every product in this range comes with supporting documents. The pages that follow are one-page summaries, prepared by Redlight Therapy Co. from third-party test reports, certificates and instrument measurements, with manufacturer details removed.

Product	Pages in this pack	Model named on the documents
300 LED Red Light Panel	EMC test summary, compliance certificate summary, light output measurement, LED measurement	RT1500S+
Red Light Therapy Hair Cap	EMC and electrical safety certificate summary	QZ022, QZ036
Red Light Therapy Knee Strap	EMC test summary, declaration of conformity summary	K8 (K6 and K9 also named)
Red Light Therapy Face Mask	EMC test summary, certificate of conformity summary	V-F2509 (V-F239 and V-F239Pro also listed)
Starter LED Face Mask	EMC test summary, certificate of compliance summary	SA-P02M01 (tested); SA-P02M01 to M09 covered

How to read these pages

- **EMC** (electromagnetic compatibility) tests check that a device does not disturb other electronics, and keeps working normally when exposed to interference.
- **Electrical safety** is covered by the low voltage certificate for the hair cap.
- **Light measurements** for the panel show the colours of light it gives out (660 nm red and near-infrared) and how steady the output is.
- **“Not applicable”** on a page means the lab recorded that test as not relevant to the product, for example mains tests on a battery-powered device.

Please note

- These pages are summaries, not certificates. Each is based on a single tested sample and does not assess every unit produced.
- The documents cover EMC, electrical safety and light measurements. They do not make or support any medical or health claim.
- Product images are for illustration.

EMC Test Summary

Starter LED Face Mask | Summary of Test Report No. PRMS2607302-01ER

Product	Starter LED Face Mask; named “LED Facial Beauty Device” on the report	**Overall result**	**PASS**
Model tested	SA-P02M01	**Date of issue**	Aug. 7, 2026
Models covered	SA-P02M01 to SA-P02M09; report says models differ in name only	**Test dates**	Aug. 3 - 5, 2026
Power	DC 5V input; battery DC 3.7V, 3000mAh (7.4Wh)	**Tested in**	Charging mode (via AC 230V/50Hz adapter) and working mode
Test lab	Shenzhen Promise Test Technology Co., Ltd.	**Directive**	EMC Directive 2014/30/EU

Standards Tested Against

EN IEC 55014-1:2021 (emissions) • EN IEC 55014-2:2021 (immunity) • EN IEC 61000-3-2:2019+A1:2021+A2:2024 • EN 61000-3-3:2013+A1:2019+A2:2021

Emissions (EN IEC 55014-1)

Test	Range / standard	Result	Worst-case margin below limit
Conducted emission, AC mains (charging)	150 kHz - 30 MHz	**Pass**	8.1 dB (line, 0.154 MHz, QP)
Radiated emission (working, on battery)	30 MHz - 1 GHz, 3 m	**Pass**	8.4 dB (horizontal, 177.5 MHz, QP)
Disturbance power	30 - 300 MHz	Not applicable	-
Harmonic current emissions	EN IEC 61000-3-2	Not applicable	Rated power under 75 W
Voltage fluctuations and flicker	EN 61000-3-3	Not applicable	Not tested: input power about 5 W

Immunity (EN IEC 55014-2:2021)

Test	Standard	Level applied	Criterion	Result
Electrostatic discharge	EN 61000-4-2	Air ±2, ±4, ±8 kV; contact ±2, ±4 kV; coupling planes ±4 kV	B required; A observed	**Pass**
Radiated RF electromagnetic field*	EN IEC 61000-4-3	3 V/m, 80 - 1000 MHz, 80% AM, four sides, both polarisations	A required; A observed	**Pass**
Electrical fast transient / burst	EN 61000-4-4	AC line L, N, L+N; ±1 kV, 5 kHz	B required; A observed	**Pass**
Surge	EN 61000-4-5	L-N, ±1 kV	B required; A observed	**Pass**
Conducted RF disturbances	EN 61000-4-6	3 V, 0.15 - 230 MHz, 80% AM, AC power port	A required; A observed	**Pass**
Voltage dips and interruptions	EN IEC 61000-4-11	100% / 0.5 cycle; 60% / 10 cycles; 30% / 25 cycles	C required; B observed	**Pass**

**The report states this test was carried out at a different test site. “Not applicable” is as recorded by the lab; the report gives the reasons shown. Criterion A: normal operation continues during the test. Criterion B: normal operation continues after the test. Criterion C: temporary loss of function allowed if it recovers.*

About this document

This page is a summary prepared by the seller from test report PRMS2607302-01ER, issued by Shenzhen Promise Test Technology Co., Ltd. It is not the laboratory's report and is not a certificate. Test results apply only to the sample tested (model SA-P02M01).

Compliance Certificate Summary

Starter LED Face Mask | Certificate of Compliance, EMC Directive 2014/30/EU

Document	Certificate of Compliance
Certificate No.	PRMS2607302-01EC
Date of issue	Aug. 7, 2026
Issued by	Shenzhen Promise Test Technology Co., Ltd., Shenzhen, China
Product	Starter LED Face Mask; named “LED Facial Beauty Device” on the certificate
Models	SA-P02M01 to SA-P02M09
Directive	EMC Directive 2014/30/EU
Test report No.	PRMS2607302-01ER

Standards Listed

EN IEC 55014-1:2021 • EN IEC 55014-2:2021 • EN IEC 61000-3-2:2019+A1:2021+A2:2024 • EN 61000-3-3:2013+A1:2019+A2:2021

What the certificate states

- The product described was tested against the listed standards and found in compliance with the EMC Directive 2014/30/EU.
- Valid only in connection with test report PRMS2607302-01ER.
- The CE marking may be used if all relevant and effective European Commission Directives and Regulations are complied with.
- Based on a single evaluation of the submitted sample(s). It does not imply an assessment of the whole product, and other relevant directives have to be observed.

About this document

This page is a summary prepared by the seller from a certificate of compliance issued by Shenzhen Promise Test Technology Co., Ltd. It is not a certificate and does not itself contain test results. The test report named above holds the measured results.