

VALIDATION REPORT

OPERATIONAL QUALIFICATION

COMMUNITY SCRIPT SUITE (CORE + DOE)

JR Validated Environment — Community Script Suite

Field	Value
Document ID	JR-OQ-001
Document Version	5.0
Execution Date	2026-10-01
OQ Result	PASS — 204 / 204 tests passed, 0 failures, 0 deviations
Status	Final
Author	Joep Rous
Reviewer	[Name]
Approver	[Name]

Version History

Version	Date	Author	Description
1.0	2026-03-15	Joep Rous	Initial OQ execution report. All 116 test cases passed. 2 deviations identified and closed.
2.0	2026-03-29	Joep Rous	Phase 9: 25 new numeric correctness TCs added (all PASS). DoE and core infrastructure TCs added. Total 164 test cases.
3.0	2026-04-19	Joep Rous	Added jrc_verify_discrete (Clopper-Pearson discrete verification). 8 new OQ TCs (TC-VER-DISC-001..008). Total 177 test cases.
4.0	2026-06-07	Joep Rous	Regenerated from OQ evidence file. 191 test cases executed (191 passed, 0 failed). Results derived directly from evidence file.
5.0	2026-10-01	Joep Rous	jrscripsts v5.0.0, code review 2026-10: re-execution of the full suite against the revised validation plan. 204 / 204 tests passed.

Approval Signatures

Role	Name	Signature	Date
Author	Joep Rous		2026-10-01
Reviewer	[Name]		
Approver (QA)	[Name]		

1. Purpose

This Validation Report documents the execution and outcome of the Operational Qualification (OQ) test suite for the community script suite (Core module and DoE module) of the JR Validated Environment. It provides objective evidence that all 204 test cases defined in JR-VP-002 v3.0 (Validation Plan) were executed, confirming that each community script produces correct, reproducible outputs for its intended use-case under 21 CFR 820.70(i).

Results are derived directly from the OQ evidence file produced by admin/admin_oq, linking each test case outcome to the raw pytest output.

2. Scope

Same as JR-VP-002 v3.0 (Validation Plan). In scope:

- 23 community R scripts and 2 community Python scripts (Core module), including jrc_verify_discrete
- DoE scripts (jrc_doe_design.R, jrc_doe_analyse.R)
- Core infrastructure (jrc_R_hello.R, jrc_py_hello.py, integrity enforcement, bypass protection)
- Numeric correctness assertions for 9 high-risk scripts

Out of scope:

- Infrastructure scripts (bin/jrrun, admin/*) — covered by JR-IQ-001
- Non-normal distribution code paths for jrc_ss_attr and jrc_verify_attr (deferred)

3. Reference Documents

Document ID	Title	Location
JR-VP-002 v3.0	Validation Plan — OQ Community Script Suite	docs/statistics_validation_plan.pdf
JR-IQ-001	IQ Validation Evidence	docs/core_IQ_validation_plan.pdf
OQ Evidence	Community Script Suite OQ Execution Evidence (pytest output + header)	oq_execution_20261001T170347.txt

4. Test Environment

Item	Details
Execution date / time	2026-10-01T17:03:48
Host	Joeps-MacBook-Air.local
Operating system	26.6.2
R version	4.6.0
Python version	3.11.9
pytest version	pytest 8.3.4
OQ venv	~/venvs/<PROJECT_ID>_oq/ (dedicated OQ environment)
Test runner	admin/admin_oq
Test directory	oq/ (test_*.py files)
Test data directory	oq/data/ (synthetic datasets, fixed seeds)
Integrity hash	admin/project_integrity.sha256
OQ evidence file	oq_execution_20261001T170347.txt

5. Test Data

Synthetic datasets committed to oq/data/. No proprietary or clinical data is used. All files generated with fixed seeds for reproducibility.

File	Description	Used by
normal_n30_mean10_sd1_seed42.csv	30 values $N(10,1)$, seed 42	jrc_normality, jrc_outliers, jrc_capability (community), jrc_descriptive, jrc_verify_attr, jrc_ss_attr, jrc_ss_sigma, jrc_ss_attr_check, jrc_ss_attr_ci
skewed_n30_lognormal_seed42.csv	30 log-normal values (skewness ≈ 0.52), seed 42	jrc_verify_attr (Box-Cox path)
outlier_n30_seed42.csv	Normal(10,1), row 15 = 5-sigma outlier	jrc_outliers
bland_altman_method1_seed42.csv	25 values $N(10,2)$, seed 42	jrc_bland_altman
bland_altman_method2_seed42.csv	Method1 + $N(0,0.2)$ proportional bias, seed 99	jrc_bland_altman
method1_short.csv	First 10 rows of method1 (mismatch test)	jrc_bland_altman (TC-BA-003)
weibull_n20_seed42.csv	Weibull(2,1000), 15 failures / 5 censored, seed 42	jrc_weibull
all_censored.csv	20 rows, status=0 (error path)	jrc_weibull (TC-WEIB-002)
neg_times.csv	20 rows, cycles row 1 = -50 (error path)	jrc_weibull (TC-WEIB-003)
bad_status.csv	20 rows, status includes 2 (invalid, error path)	jrc_weibull (TC-WEIB-004)
convert_multicolumn.txt	Tab-delimited, 3 headers, 20 rows	jrc_convert_txt
convert_singlecolumn.txt	200 numeric values, one per line	jrc_convert_txt
verify_attr_known.csv	30 rows: 28×10.0 + rows 29/30 chosen so mean=10.0, sd=1.0, $K1(30,0.95,0.95)=2.220 \rightarrow TL=7.780$	jrc_verify_attr (TC-VER-009..011), jrc_ss_attr (TC-ATTR-008..009), jrc_ss_attr_check (TC-ATTRCK-004..006), jrc_ss_attr_ci (TC-ATTRCI-004..005)
doe_factors_*.csv / doe_results_*.csv	Multiple synthetic DoE factor and response files (full factorial, fractional, PB)	jrc_doe_design, jrc_doe_analyse (TC-DOE-DES-*, TC-DOE-ANA-*)

6. Test Execution Results

6.1 Summary

Metric	Value
Total test cases executed	204
Passed	204
Failed	0
Errors	0
OQ Result	PASS

6.2 Results by Test File

Test File	Scripts Covered	TCs	Result
test_ss_suite.py	jrc_ss_discrete, jrc_ss_discrete_ci, jrc_ss_attr, jrc_ss_attr_check, jrc_ss_attr_ci, jrc_ss_sigma, jrc_ss_fatigue, jrc_ss_paired, jrc_ss_equivalence, jrc_msa_grr_design	69	69 / 69 PASS
test_statistical_suite.py	jrc_bland_altman, jrc_weibull, jrc_verify_attr	37	37 / 37 PASS
test_diagnostic_suite.py	jrc_normality, jrc_outliers, jrc_capability (community), jrc_descriptive	19	19 / 19 PASS

test_gen_suite.py	jrc_gen_normal, jrc_gen_lognormal, jrc_gen_sqrt, jrc_gen_boxcox, jrc_gen_uniform	21	21 / 21 PASS
test_convert_suite.py	jrc_convert_csv.py, jrc_convert_txt.py	15	15 / 15 PASS
test_doe_suite.py	jrc_doe_design.R, jrc_doe_analyse.R	37	37 / 37 PASS
test_core.py	jrc_R_hello.R, jrc_py_hello.py (core infrastructure)	6	6 / 6 PASS
TOTAL		204	204 / 204 PASS

6.3 Numeric Correctness Test Results

The numeric correctness test cases assert computed values against independently derived expected results. Results below are read directly from the evidence file.

TC	Script	Expected value	Result
TC-DISC-006	jrc_ss_discrete	n(f=0) = 299	PASS
TC-DISC-007	jrc_ss_discrete	n(f=1) = 473	PASS
TC-DISC-008	jrc_ss_discrete	n(f=0) = 45	PASS
TC-DISC-009	jrc_ss_discrete	n(f=0) = 459	PASS
TC-DISCICI-005	jrc_ss_discrete_ci	p ≈ 0.9900 ± 0.0001	PASS
TC-DISCICI-006	jrc_ss_discrete_ci	p ≈ 0.8729 ± 0.0001	PASS
TC-ATTR-008	jrc_ss_attr	3 ≤ n ≤ 10	PASS
TC-ATTRCK-004	jrc_ss_attr_check	verdict = PASS	PASS
TC-ATTRCK-005	jrc_ss_attr_check	verdict = FAIL	PASS
TC-ATTRCI-004	jrc_ss_attr_ci	0.985 < p < 1.000	PASS
TC-ATTR-009	jrc_ss_attr	k = -2: no N reported	PASS
TC-ATTR-010	jrc_ss_attr	Box-Cox rejected (SW p ≤ 0.01): no N	PASS
TC-ATTRCK-006	jrc_ss_attr_check	k = -2: verdict = FAIL	PASS
TC-ATTRCI-005	jrc_ss_attr_ci	k = -2: no proportion reported	PASS
TC-SIGMA-005	jrc_ss_sigma	n = 5	PASS
TC-SIGMA-006	jrc_ss_sigma	n = 14	PASS
TC-SIGMA-007	jrc_ss_sigma	n = 3	PASS
TC-FAT-006	jrc_ss_fatigue	n(f=0) = 29	PASS
TC-FAT-007	jrc_ss_fatigue	n(f=0) = 8	PASS
TC-FAT-008	jrc_ss_fatigue	AF = 1: no β table	PASS
TC-PAIRED-006	jrc_ss_paired	n = 44 (2-sided)	PASS
TC-PAIRED-007	jrc_ss_paired	n = 36 (1-sided)	PASS
TC-EQUIV-005	jrc_ss_equivalence	n = 45	PASS
TC-EQUIV-006	jrc_ss_equivalence	n = 53	PASS
TC-VER-009	jrc_verify_attr	TL ≈ 7.780 ± 0.050, verdict 	PASS
TC-VER-010	jrc_verify_attr	TL ≈ 7.780, verdict 	PASS
TC-VER-011	jrc_verify_attr	LTL ≈ 7.445, UTL ≈ 12.555, verdict 	PASS

7. Deviations from OQ Plan

All previously identified deviations (DEV-001..005) are closed. No new deviations.

ID	Affected TC(s)	Description	Status
DEV-001	TC-DISC-001	Plan used log formula (n=299); script uses chi-squared method (n=300). Justification: chi-squared is mathematically equivalent and more conservative. Assertions updated.	CLOSED
DEV-002	TC-ATTR-001..003	Plan used ±1σ spec limits; script triggers Box-Cox for skewed data	CLOSED

		causing $n > \text{cap}$. Tests updated to use $\pm 3\sigma$ limits which exercise the same code path.	
DEV-003	TC-VER-009..011	Plan had incorrect K1 derivation (1.778 vs correct 2.220). TC assertions updated before execution.	CLOSED
DEV-004	TC-PAIRED-006/007, TC-EQUIV-005/006	Plan applied ceiling() incorrectly. Correct values: 44, 36, 36, 45. Assertions updated before execution.	CLOSED
DEV-005	TC-ATTRCI-004	Plan expected proportion > 0.990 based on incorrect $K1=1.778$. With $K1=2.220$, achieved proportion = 0.9886. Assertion updated to range (0.985, 1.000).	CLOSED

8. Requirements Traceability Matrix

All URs from JR-VP-002 v2.0 mapped to executed test cases.

UR	Requirement (summary)	Test Cases	Result
UR-001	Sample size — discrete (chi-squared exact)	TC-DISC-001..009	PASS
UR-002	Proportion CI — exact Clopper-Pearson	TC-DISCICI-001..006	PASS
UR-003	Sample size — attribute tolerance interval	TC-ATTR-001..010	PASS
UR-004	Attribute sample size check (pass/fail boundary)	TC-ATTRCK-001..006	PASS
UR-005	Achieved proportion for attribute tolerance interval	TC-ATTRCI-001..005	PASS
UR-006	Sample size — sigma (normal approximation)	TC-SIGMA-001..007	PASS
UR-007	Sample size — fatigue (Weibull-adjusted)	TC-FAT-001..008	PASS
UR-008	Sample size — paired t-test	TC-PAIRED-001..007	PASS
UR-009	Sample size — equivalence (TOST)	TC-EQUIV-001..006	PASS
UR-010	Sample size — Gauge R&R	TC-GRR-001..005	PASS
UR-011	Normality testing	TC-NORM-001..006	PASS
UR-012	Outlier detection (Grubbs)	TC-OUT-001..005	PASS
UR-013	Process capability (Cp, Cpk, Pp, Ppk) — community script	TC-CAP-001..004	PASS
UR-014	Descriptive statistics	TC-DESC-001..004	PASS
UR-015	Bland-Altman method comparison	TC-BA-001..006	PASS
UR-016	Weibull reliability analysis	TC-WEIB-001..006	PASS
UR-017	Tolerance interval verification ( /  verdict)	TC-VER-001..011	PASS
UR-018	Synthetic data generation (5 distributions)	TC-GEN-N/LN/SQ/BC/U (21 TCs)	PASS
UR-019	CSV file conversion (multi-column → id/value)	TC-CCSV-001..008	PASS
UR-020	Text file conversion (plain-text → id/value CSV)	TC-CTXT-001..007	PASS
UR-021	Input validation and usage messages (all scripts)	Error-path TCs across all suites	PASS
UR-022	RENV_PATHS_ROOT enforcement (R bypass protection)	TC-CORE-OQ-004	PASS
UR-023	DoE — design generation and analysis	TC-DOE-DES-001..019, TC-DOE-ANA-001..018	PASS
UR-024	Core infrastructure (hello scripts, integrity, bypass)	TC-CORE-IQ-001, TC-CORE-OQ-001..005	PASS
UR-025	Discrete (pass/fail) verification — Clopper-Pearson exact bound	TC-VER-DISC-001..008	PASS
Total		All 204 community OQ tests	PASS

9. OQ Conclusion

The Operational Qualification of the JR Validated Environment Community Script Suite (Core + DoE modules) is complete. All 204 test cases were executed on 2026-10-01 and passed with no failures, no errors, and no deviations.

All 25 user requirements (UR-001 through UR-025) are covered by at least one passing test case. The OQ acceptance criteria stated in JR-VP-002 Section 10 are fully met.

The Community Script Suite (Core and DoE modules) is hereby declared OPERATIONALLY QUALIFIED and released for use in design verification activities in accordance with the JR Validated Environment quality system.

OQ Evidence file:

/Users/joeprous/.jrscrip/MyProject/validation/oq_execution_20261001T170347.txt