Cyber-Physical Quality Assurance of Analytical Laboratories: Joint Permutation-Corruption Inference with Low-Cost Optical Witnesses and Conformal Calibration Bounds

Sudeep Sharma[1], *Student Member, IEEE*, Armaan Srivastava[2], *Student Member, IEEE*, Udit Shokeen[3], *Student Member, IEEE*, and Dr. Priya Saminathan[4], *Senior Member, IEEE, Faculty Guide*

School of Computer Science and Engineering, Vellore Institute of Technology (VIT), Vellore 632014, Tamil Nadu, India

[1]Reg. No. 23BDS0315 (sudeep.sharma2023@vitstudent.ac.in), [2]Reg. No. 23BCE2318 (armaan.srivastava2023@vitstudent.ac.in), [3]Reg. No. 23BCB0001 (udit.shokeen2023@vitstudent.ac.in),

[4]Faculty Guide, Department of Computational Intelligence, SCSE, VIT Vellore (priya.s@vit.ac.in).

Manuscript received October 2, 2026; revised October 26, 2026. Corresponding author: Sudeep Sharma.

***Abstract*—High-throughput analytical testing laboratories in agronomy, environmental chemistry, and clinical pathology govern safety-critical decisions, including regional fertilizer allocations, soil carbon sequestration underwriting, and remediation protocols. However, analytical reports are susceptible to silent mechanical and human failure modes: specimen transpositions (sample swaps) and clerical transcription errors (decimal point slips, unit confusion, digit inversions). Classical univariate autoverification and longitudinal delta checks fail in environmental batch testing because samples lack historical time-series baselines and swapped specimens satisfy marginal domain boundaries. We present EdaphNet, a cyber-physical quality assurance framework that couples an independent, non-contact optical observation (a physical witness) captured before chemical testing with a joint batch-level probabilistic graphical model. The engine computes Bayesian Maximum A Posteriori (MAP) scores over structured tray-permutation subgroups $S_{\text{tray}} \subset S_n$ (exact within candidate set S_{tray}) and a formal library of typed numerical corruption operators C, producing calibrated suspicion scores, root-cause attributions, and repaired values.**

The framework is evaluated on 19,746 real soil samples from the Kellogg Soil Survey Laboratory (KSSL) and 3,473 samples from the ICRAF-ISRIC repository (23,219 samples total) across a conceptual four-rung sensor ladder (W_0–W_3), with three sensor tiers (W_0, W_1, W_3) empirically benchmarked on KSSL under pre-registered evaluation protocols. In confirmatory testing, the primary pre-registered hypothesis A1—that a commodity smartphone camera proxy would outperform the best existing baseline in overall error recall across an equal mix of errors—was refuted (+0.014 [−0.008, +0.038] with good phone optics $\Delta E = 2$). In contrast, with continuous laboratory Vis-NIR spectroscopy (W_3), EdaphNet achieves a statistically significant confirmatory gain of +9.2 percentage points (+0.092 [+0.069, +0.113], $p < 0.001$) across all error categories. An exploratory post-hoc sub-analysis establishes that commodity phone optics provide a decisive +10.3 percentage point advantage (95% CI: [+0.064, +0.139]) specifically in isolating physical specimen swaps, offset by a −8.2 percentage point deficit on isolated numerical value slips; the method crosses into an overall advantage once swaps comprise ≥ 60–70% of batch failures. For root-cause fault attribution, EdaphNet achieves 84–92% accuracy in simulation, 89–98% on ICRAF color data, and 97–100% on KSSL spectrometer data (against 24–29% on KSSL and 32–51% on ICRAF/simulation for rule heuristics), with 87–99% exact swap-partner identification accuracy under high-information witnesses. Furthermore, we demonstrate that conformal false discovery rate (FDR) control collapses under calibration sample re-use (pseudo-replication, inflating realized FDR to 57–92%), whereas restricting calibration to one score per distinct physical sample reliably restores empirical FDR control (0.000–0.090 in pre-registered 50-batch pools, and 0.03–0.11 across pool sizes 1 to 50 batches, nominal $q = 0.10$).

***Index Terms*—**Analytical laboratory quality control, cyber-physical systems, sample permutation, conformal inference, false discovery rate (FDR), diffuse reflectance spectroscopy, fault attribution, label noise.

I. INTRODUCTION

High-throughput analytical chemistry laboratories constitute the primary quantitative bedrock for modern agriculture, environmental compliance, food safety, and clinical healthcare. In agriculture, wet-chemistry reports quantifying pH, organic carbon (SOC), total nitrogen, cation exchange capacity (CEC), and primary nutrients directly govern multi-ton fertilizer allocations, soil amendment investments, and verifiable soil carbon sequestration credits. Despite rigid instrument calibration, sample batch handling remains susceptible to silent failure modes. Recent empirical studies across commercial analytical service laboratories revealed that sample-processing differences (sieving, grinding, drying, and analytical quantification) induce extensive inter-laboratory variability in reported soil organic carbon [1].

Analytical batch discrepancies stem from two fundamentally distinct error mechanisms:

1) Specimen Swaps (Permutations): Misplacement of test tubes or robotic rack offsets cause Specimen j to be reported under Specimen k's identifier. The report delivered to farmer j is physically and biochemically valid—it is simply measured on the wrong sample.

2) Numerical Cell Corruptions: Data entry errors or automated parser glitches inject typed distortions, such as orders-of-magnitude decimal slips (×10, ÷10), transposed digits (0.16 ↔ 0.61), or column shifts.

Traditional quality assurance (QA) mechanisms are structurally incapable of detecting this failure spectrum. In clinical diagnostics, automated "delta checks" compare fresh patient samples against longitudinal historical archives [2], [3]. In environmental and agricultural batch testing, historical baselines do not exist for incoming specimens. Furthermore, rule-based range filters (B_0) and agronomic consistency rules (B_1) are blind to specimen swaps because swapped samples satisfy marginal bounds.

To overcome this bottleneck, we introduce a cyber-physical testing paradigm: capturing an independent, non-contact optical measurement of each sample—termed a *physical witness* w_i—prior to destructive laboratory processing. The witness can range from a low-cost smartphone photograph under ambient illumination to a laboratory Vis-NIR diffuse reflectance spectrometer. Because the witness is acquired directly from physical matter, it is impervious to downstream laboratory clerical errors.




Fig. 1. System architecture of the EdaphNet cyber-physical laboratory verification engine. An optical sensor captures a physical witness w_i of each specimen. A multi-property latent Gaussian process estimates the joint property distribution $p(z \mid w_i)$. The joint inference engine enumerates tray-permutation hypotheses $\pi \in S_{\text{tray}}$ against corruption operators $T_c \in C$ across the entire analytical batch R, producing calibrated suspicion ranks, fault attributions, and repair values.

In this paper, we propose **EdaphNet**, a batch-level probabilistic verification framework. Rather than evaluating specimens in isolation, EdaphNet performs Bayesian joint inference over the combinatorial space of structured sample permutations (S_{tray}) and typed value-corruption operators (C). Within the structured candidate set S_{tray}, the posterior evaluation is computed exactly. We evaluate the framework on 23,219 real soil samples from global repositories under pre-registered evaluation protocols.

II. RELATED WORK & PRIOR ART TAXONOMY

A targeted literature survey across IEEE Xplore, Scopus, Web of Science, and PubMed examined prior art across four relevant methodological fields:

A. Optical Soil Sensing: Visible and Near-Infrared (Vis-NIR, 350–2500 nm) diffuse reflectance spectroscopy is an established method for non-destructive soil characterization [5], [6]. Deployed mobile applications (e.g., LandPKS [7]) record soil color and estimate physical soil characteristics using mobile devices. However, existing optical sensing methods are deployed exclusively as *predictive estimators*; EdaphNet inverts this paradigm, using optical sensing as an independent *verification witness* to audit high-precision laboratory reports.

B. Autoverification and Delta Checks: Checking measured values against predictive models is established in process industries (US Patent 8,352,394 [8]) and clinical laboratories. Multi-analyte machine learning delta checks detect "wrong-blood-in-tube" (WBIT) errors by tracking patient longitudinal trajectories [2], [3]. EdaphNet extends this to cross-sectional batches lacking historic baselines.

C. Permutation Matching in Genomics & Soil Predictability: In high-throughput genomics, sample mix-ups are detected and repaired by matching genotype arrays against RNA sequencing (Broman et al. [9]). While genomics leverages near-deterministic signatures ($R^2 \approx 1.0$), agricultural soil witnesses are noisy, low-dimensional, and weak: with good smartphone camera color (W_1, $\Delta E = 2$), optical witnesses explain at most $R^2 \leq 0.15$–0.16 for mineral, cation, and texture properties on ICRAF (clay 0.07, sand ≤ 0, CEC 0.05, Mg 0.06, K ≤ 0, Ca 0.13, pH 0.15), rising to $R^2 = 0.35$ for ICRAF SOC and $R^2 = 0.70$ (70%) for KSSL SOC. Under perfect color (W_1, $\Delta E = 0$), ICRAF SOC reaches $R^2 = 0.37$, pH reaches $R^2 = 0.27$, and Ca reaches $R^2 = 0.23$, while texture and other cations remain ≤ 0.16. Under laboratory Vis-NIR spectroscopy (W_3), predictability reaches $R^2 = 0.60$–0.66 across ICRAF properties and $R^2 = 0.90$ for KSSL SOC and C_{tot}. Selective uncertainty rejection strategies in diffuse reflectance spectroscopy (Schmidinger et al. [10]) allow models to abstain on out-of-distribution spectra, but do not model clerical laboratory failures; EdaphNet complements spectral uncertainty with a structured permutation and transcription error model.

D. Conformal Outlier Testing: Distribution-free outlier detection using conformal p-values with Benjamini-Hochberg (BH) false discovery rate control has been formalized by Bates et al. [11] and Jin & Candès [12]; foundational theory is established by Angelopoulos & Bates [13] and Vovk et al. [14]. Cleanlab (Zhou et al. [4]) detects label noise in continuous numerical and regression tasks. EdaphNet incorporates conformal inference but reveals a critical failure mode: pseudo-replication under calibration sample re-use.

TABLE I: COMPARISON OF EDAPHNET AGAINST PRIOR ART QA METHODOLOGIES

Methodology	Joint Batch Inference	Physical Witness	Permutation Model (π)	Typed Error Library (C)	Calibrated FDR Control	Root-Cause Attribution
Range Filters (B_0, B_1) [This work]	No	No	No	Heuristic bounds	None	24–29% (B_6, KSSL)
Clinical Delta Checks [3]	No	Longitudinal	Pairwise	No	Empirical	No
Genomic Matching [9]	Yes	DNA (Exact)	Pairwise	No	None	Swap Only
Cleanlab / Label Noise [4]	No	Covariates	No	Continuous Residual	Heuristic	No
Conformal Outliers [11], [12]	No	Covariates	No	No	Distribution-Free	No
EdaphNet (Proposed)	**Yes**	**RGB / Vis-NIR**	**Structured S_{tray}**	**Formal Library C**	**Distribution-Free**	**97–100% (KSSL W_3)**

III. MATHEMATICAL PROBLEM FORMULATION

Consider an automated laboratory processing rack with n physical positions ($n = 24$ in standard multi-well trays). For each position $i \in \{1, ..., n\}$, an optical sensor records witness $w_i \in \mathbb{R}^d$. The laboratory outputs reported batch $R = (r_1, ..., r_n)$, with property vector $r_j \in \mathbb{R}^P$.

A. Structured Permutation Subgroup S_{tray}

Specimen swaps represent a bijection π: $\{1, ..., n\} \to \{1, ..., n\}$. While the full symmetric group has size $n!$, physical robotic and technician mistakes follow structured mechanics. We define S_{tray} as the union of: (i) identity $\pi_0(j) = j$ (1 permutation); (ii) all pairwise transpositions (j, k) across the tray ($C(24, 2) = 276$ permutations, which directly subsume the 23 adjacent swaps $\pi_{\text{adj}}(j, j+1)$); and (iii) cyclic block tray shifts ±1 and ±2 positions (4 permutations). In total, $|S_{\text{tray}}| = 1 + 276 + 4 = 281$ unique candidate permutations for a standard 24-well rack.

B. Typed Corruption Library C

Clerical and analytical errors are modeled as operator transformations $r_j = T_c(r_j^*)$: $C = \{T_{\times 10}, T_{\div 10}, T_{\text{unit}}, T_{\text{digit-swap}}, T_{\text{col-swap}}, T_{\text{gross}}\}$ where $T_{\text{gross}}(x) = \gamma x$ with scaling factor $\gamma \in [0.2, 5.0]$.

C. Permutation Prior Formulation

The prior distribution $P(\pi)$ over S_{tray} assigns probability mass $1 - \rho$ to the identity π_0 (no swap, $\rho = 0.10$). The remaining mass ρ is partitioned across three mechanical classes: adjacent transpositions

(class mass 0.50), cyclic block rotations (class mass 0.30), and non-adjacent pair transpositions (class mass 0.20), distributed uniformly within each class.

D. Multi-Property Latent Gaussian Process

Raw properties are converted to latent coordinates $z = g(r) \in \mathbb{R}^P$ via property-specific monotonic scalers. Conditional on witness w_i, latent vector z_i follows a multivariate Gaussian distribution:

$$z_i \mid w_i \sim N(\mu(w_i), \Sigma) \tag{1}$$

where $\mu(w_i)$ is an out-of-fold gradient-boosted regression estimator, and $\Sigma \in \mathbb{R}^{P \times P}$ is the full residual covariance matrix. The covariance structure allows a corrupted value to be detected through inconsistency with uncorrupted properties, even when the optical witness is weakly informative for that property.

IV. THE JOINT INFERENCE ENGINE

The inference workflow is formalized in Algorithm 1. For candidate pairing between report r_j and physical sample i, the marginal likelihood integrates across the corruption library:

$$L(j, i) = \sum_{c \in C} P(c) \cdot \varphi(g(T_c^{-1}(r_j)); \mu(w_i), \Sigma) \cdot |J_{g \circ T_c^{-1}}(r_j)| \tag{2}$$

where $\varphi(\cdot; \mu, \Sigma)$ denotes the multivariate Gaussian density and $|J|$ is the Jacobian determinant.

Algorithm 1: Batch Permutation-Attribution Inference (EdaphNet-JBC)

Input: Reports $R=(r_1..r_n)$, witnesses $W=(w_1..w_n)$, predictor $\mu(\cdot)$, covariance Σ, permutation space S_{tray}, corruption library C.

Output: Anomaly score $s \in [0, 1]^n$, fault attribution A, repaired reports $\hat{R}$.

Compute predicted latent means: $\hat{z}_i \leftarrow \mu(w_i), \forall i \in \{1, ..., n\}$.

for $j = 1$ to n **do**

  for $i = 1$ to n **do**

    Compute cell log-likelihood $\ell(j, i) \leftarrow \log L(j, i)$ via Eq. (2).

for $\pi \in S_{\text{tray}}$ **do**

  $\log P(\pi \mid R, W) \leftarrow \log P(\pi) + \sum_{j:\, \pi(j) \neq j} [\ell(j, \pi(j)) - \ell(j, j)]$.

Normalize posterior $P(\pi \mid R, W)$ over S_{tray}.

for $j = 1$ to n **do**

  $P(\text{moved}_j) \leftarrow \sum_{\pi:\, \pi(j) \neq j} P(\pi \mid R, W)$.

  $P(\text{clean}_j) \leftarrow (1 - P(\text{moved}_j)) \cdot P(c = \text{none} \mid r_j, w_j)$.

  Suspicion score $s_j \leftarrow 1 - P(\text{clean}_j)$.

  Assign MAP attribution A_j and repaired report $\hat{r_j}$.

return s, A, $\hat{R}$

The posterior distribution of tray permutation π given batch tuple (R, W) is calculated exactly within candidate subgroup S_{tray}:

$$P(\pi \mid R, W) \propto P(\pi) \cdot \prod_{j=1}^{n} [L(j, \pi(j)) / L(j, j)] \tag{3}$$

Because $|S_{\text{tray}}| = 281$ for a 24-well rack, full normalization executes in 18 ms (17.9 ± 1.2 ms) on an AMD Ryzen 5 7520U laptop-class mobile processor (4 cores, 8 threads, 2.80 GHz base clock). The 18 ms latency encompasses feature prediction $\mu(w_i)$, cell likelihood matrix evaluation $\ell(j, i)$, hypothesis summation over all 281 permutations, and exact normalization.

V. CONFORMAL CALIBRATION & PSEUDO-REPLICATION

To control false alarms, scores s_j are converted into conformal p-values using clean calibration data $D_{\text{cal}} = \{(r_k^{\text{cal}}, w_k^{\text{cal}})\}_{k=1}^{N}$:

$$p_j = [1 / (N + 1)] \cdot [1 + \sum_{k=1}^{N} \mathbf{1}\{s_k^{cal} \geq s_j\}] \tag{4}$$

Applying Benjamini-Hochberg (BH) [15] at nominal level $q = 0.10$ guarantees distribution-free False Discovery Rate (FDR) control under exchangeability [11], [14]. Under covariate shift, conformal bounds extend via weighted conformality [16].




Fig. 2. Empirical verification of conformal calibration across pool sizes. In the initial real-data experiment, reusing calibration soils across simulated batches resulted in severe FDR inflation (57–92%). Enforcing non-replicated calibration (one score per distinct physical sample) restores empirical FDR control (0.03–0.11 across pool sizes 1 to 50 batches for EdaphNet, and 0.000–0.090 in pre-registered 50-batch pools, nominal $q = 0.10$). However, statistical power at fixed nominal q depends strongly on evaluation pool size relative to calibration denominator N.

Methodological Finding 1 (Calibration Pseudo-Replication): In the initial real-data experiment, empirical FDR collapsed for all baselines (realized FDR 57–92%). Diagnostic investigation revealed that sampling the same finite set of calibration specimens repeatedly across batches violates exchangeability (identified in `scripts/diag_calibration.py:80-96`). Sample re-use creates discrete duplicate spikes in the empirical distribution of s^{cal}, artificially driving conformal p-values down and triggering false alarm surges. Enforcing *one score per distinct physical sample* ($N \approx$ 2,800 on KSSL) restored empirical FDR below the target 0.10 threshold (0.000–0.090 in pre-registered pools, and 0.03–0.11 across pool sizes 1 to 50 batches).

VI. EXPERIMENTAL BENCHMARK SETUP

A. Datasets: Evaluations span two major soil spectral libraries: (1) *KSSL Vis-NIR*: 19,746 samples from 6,999 US locations (OC, N, C_{tot}); protocol registered in `docs/18_preregistered_amendment_v2.md` on 2026-10-01 (commit `095deaa`). (2) *ICRAF-ISRIC*: 3,473 samples from 553 global sites (pH, OC, Clay, Sand, CEC, K, Ca, Mg). Geographic profile-disjoint splits: 50% training, 15% calibration, 35% test.

B. Sensor Ladder: Conceptual four-rung ladder: W_0 (uninformative), W_1 (RGB colorimetry, noise $\Delta E \in \{0, 2, 4, 8\}$), W_2 (10-band multispectral, piloted on ICRAF; omitted from KSSL due to lack of camera data), and W_3 (Vis-NIR spectrometer, 350–2500 nm).




Fig. 3. Cross-validated coefficient of determination (R^2) across witness sensor tiers on unseen geographic locations. Under good smartphone camera optics (W_1, ΔE = 2), predictability reaches R^2 = 0.70 for KSSL SOC and R^2 = 0.35 for ICRAF SOC (R^2 = 0.15 for pH), but drops to $R^2 \leq 0.07$ for texture and potassium. Under perfect color (W_1, ΔE = 0), ICRAF SOC reaches R^2 = 0.37 and pH reaches R^2 = 0.27, while minerals and texture remain $R^2 \leq 0.16$. Laboratory spectroscopy (W_3) achieves R^2 = 0.60–0.66 across ICRAF properties and R^2 = 0.90 for KSSL SOC.

VII. EMPIRICAL RESULTS AND ANALYSIS

A. Primary Pre-Registered KSSL Benchmark

The primary evaluation criterion was specified in a pre-registered protocol prior to unblinding test evaluations: *Recall at a 5% re-test budget across 50-batch pools over 5 random seeds*. Table II and Fig. 4 present the pre-registered results on 19,746 KSSL soils.

TABLE II: PRE-REGISTERED RECALL AT 5% BUDGET ON KSSL (19,746 SOILS)

Sensor Witness	EdaphNet	Best Baseline	Difference	95% CI
W_0 (None)	0.238	0.271 (B_5)	-0.033	[-0.049, -0.016]
W_1 (Color, ΔE = 0)	0.416	0.415 (B_5)	+0.001	[-0.022, +0.023]
W_1 (Phone, ΔE = 2)	0.410	0.396 (B_5)	+0.014	[-0.008, +0.038]
W_1 (Noisy, ΔE = 4)	0.382	0.354 (B_5)	+0.028	[+0.010, +0.049]
W_3 (Spectrometer)	**0.604**	**0.512 (B_5)**	**+0.092**	**[+0.069, +0.113]**




Fig. 4. Pre-registered KSSL benchmark recall at 5% re-test budget. Whiskers indicate 95% bootstrap confidence intervals. With commodity phone optics (ΔE = 2), overall recall matches the best supervised baseline (+0.014 [−0.008, +0.038], refuting Hypothesis A1); with a laboratory spectrometer (W_3), EdaphNet achieves a statistically significant confirmatory gain of +9.2 percentage points ($p < 0.001$).

Under pre-registered Hypothesis A1, the headline hypothesis that commodity RGB color would outperform all existing methods by ≥ +5 points across an equal mix of errors was **refuted**. However, with continuous spectroscopy (W_3), EdaphNet achieved a statistically significant confirmatory gain of +9.2 points (+0.092 [+0.069, +0.113], $p < 0.001$) over the best supervised baseline.

B. Decomposing the Trade-Off: Swaps vs. Numerical Corruptions

Exploratory decomposition (Fig. 5) reveals the underlying trade-off:

1) Specimen Swap Dominance: On sample swaps, EdaphNet decisively outperformed B_5 by **+10.3 percentage points** (95% CI: [+0.064, +0.139]) with phone color, and by **+23.1 points** ([+0.205, +0.258]) with spectroscopy. Because swapped samples represent plausible soils, methods lacking optical witnesses cannot identify them.

2) Numeric Corruption Deficit: On isolated numerical corruptions, EdaphNet lagged B_5 by −8.2 points ([−0.116, −0.050]). Diagnostic runs revealed that B_5 heavily exploited the hard domain inequality SOC ≤ C_{tot}. While EdaphNet's Gaussian process captured continuous correlation, it did not strictly enforce the hard ceiling inequality.




Fig. 5. Breakdown of detection recall for specimen swaps versus numerical value corruptions. EdaphNet holds a decisive +10.3 point exploratory advantage on specimen swaps under phone optics, while supervised baselines excel at isolated gross errors on constrained properties.

C. Root-Cause Fault Attribution

While existing baselines merely generate uninterpretable anomaly scores, EdaphNet accurately identifies the underlying error mechanism. As shown in Table III, EdaphNet achieved 84–92% accuracy in simulation, 89–98% on ICRAF color data, and 97–100% on KSSL spectrometer data (against 24–29% on KSSL and 32–51% on ICRAF/simulation for rule heuristics). Exact swap-partner identification reached 87–99% across high-information witness settings (W_3 and simulation $\lambda \geq 2$), and 95–98% on KSSL W_3.

TABLE III: ROOT-CAUSE FAULT ATTRIBUTION AND SWAP-PARTNER IDENTIFICATION ACCURACY

Dataset & Setting	Fault Type (EdaphNet)	Swap Partner	Rule Heuristic (B_6)
Simulation (T_0)	**84–92%**	**86–98%**	32–50%
ICRAF (Color W_1 / W_3)	**89–98%**	**75–99%**	37–51%
KSSL (Spectro W_3)	**97–100%**	**95–98%**	24–29%

D. Error-Mix Sensitivity Analysis

In commercial laboratories, the ratio of specimen swaps to typographical slips varies by automation level. We evaluated performance across varying swap shares ω ∈ [0.1, 0.9] (Fig. 6). With commodity phone optics (W_1), EdaphNet's performance advantage expands monotonically, dominating supervised baselines once swaps represent ≥ 60–70% of batch failures.




Fig. 6. Sensitivity to batch error composition. As the share of specimen swaps among batch errors varies from 10% to 90%, EdaphNet's advantage expands. With phone color, EdaphNet surpasses all baselines once swaps exceed 60–70%; with spectroscopy, it dominates above 30%.

VIII. CYBER-PHYSICAL DEPLOYMENT

A. Hybrid Combined Policy: We implemented and smoke-tested a two-stage hybrid verification policy on KSSL (W_1, $\Delta E = 2$): EdaphNet serves as the primary permutation sieve, while deterministic boundary rules audit single-cell bounds. The combined policy achieved balanced recall: 0.20–0.25 on swaps and 0.222 on value errors, eliminating the single-checker deficit.

B. Computational Throughput: Full joint inference over 281 unique permutation hypotheses for a 24-well sample tray executes in **18 milliseconds** (17.9 ± 1.2 ms) on an AMD Ryzen 5 7520U laptop-class mobile processor (4 cores, 8 threads, 2.80 GHz base clock). The 18 ms latency encompasses feature prediction, cell likelihood evaluation, hypothesis summation across all 281 candidates, and exact posterior normalization, enabling inline verification before chemical digestion.

IX. CONCLUSION

This paper presented EdaphNet, a cyber-physical laboratory verification framework that combines joint permutation-corruption inference with optical physical witnesses. Through pre-registered benchmarks on 23,219 soil samples, we delineated the operational boundaries of optical verification: while the primary pre-registered claim of an overall advantage for phone color was refuted (+0.014 [−0.008, +0.038]), confirmatory results established a decisive +9.2 percentage point advantage under laboratory spectroscopy (+0.092 [+0.069, +0.113], $p < 0.001$). Exploratory sub-analyses revealed that commodity RGB phone color provides a potent +10.3 percentage point boost specifically in catching specimen swaps, offset by a deficit on isolated numerical slips. In root-cause fault attribution, EdaphNet achieved 84–92% accuracy in simulation, 89–98% on ICRAF color data, and 97–100% on KSSL spectrometer data (against 24–29% on KSSL for rule-based heuristics), with 87–99% exact swap-partner identification. Furthermore, we showed that conformal FDR control collapses under calibration pseudo-replication and demonstrated that non-replicated calibration reliably restores distribution-free validity. Future work includes field validation with in-situ Indian agricultural soil trays and robotic integration.

ACKNOWLEDGMENT

The authors express their deepest gratitude to their faculty guide, **Dr. Priya Saminathan**, School of Computer Science and Engineering, Vellore Institute of Technology, Vellore, for her dedicated mentorship, invaluable domain insights, and critical review throughout this project. The authors also thank Vellore Institute of Technology, Vellore, for providing computational infrastructure, and acknowledge the USDA-NRCS Kellogg Soil Survey Laboratory (KSSL) and the World Agroforestry Centre (ICRAF-ISRIC) for their open-access soil spectral datasets.

REPRODUCIBILITY STATEMENT

All code, model weights, experiment configs, and run records are open-source and version-controlled at: https://github.com/Sudeep-Sharma1/edaphnet-v2. Every metric and figure in this paper traces directly to automated experiment logs.

REFERENCES

1. R. J. Even, M. B. Machmuller, J. M. Lavallee, T. J. Zelikova, and M. F. Cotrufo, "Large errors in soil carbon measurements attributed to inconsistent sample processing," *SOIL*, vol. 11, no. 1, pp. 17–34, Jan. 2025, doi: 10.5194/soil-11-17-2025.
2. J. O. Westgard and S. A. Westgard, *Basic QC Practices: Training in Statistical Quality Control for Healthcare Laboratories*, 4th ed., Madison, WI, USA: Westgard Quality Corporation, 2020.
3. M. W. Rosenbaum and J. M. Baron, "Using machine learning-based multianalyte delta checks to detect wrong blood in tube errors," *Am. J. Clin. Pathol.*, vol. 150, no. 6, pp. 555–566, Dec. 2018, doi: 10.1093/ajcp/aqy085.
4. H. Zhou, J. Mueller, M. Kumar, J.-L. Wang, and J. Lei, "Detecting errors in a numerical response via any regression model," *J. Data-centric Mach. Res.*, vol. 1, pp. 1–25, 2024, arXiv:2305.16583.
5. R. A. Viscarra Rossel *et al.*, "A global spectral library to characterize the world's soil," *Earth-Sci. Rev.*, vol. 155, pp. 198–230, Apr. 2016, doi: 10.1016/j.earscirev.2016.01.012.
6. J. L. Safanelli *et al.*, "Open Soil Spectral Library (OSSL): Building reproducible soil calibration models through open development and community engagement," *PLOS ONE*, vol. 20, no. 1, p. e0296545, Jan. 2025, doi: 10.1371/journal.pone.0296545.
7. J. E. Herrick *et al.*, "The global Land-Potential Knowledge System (LandPKS): Supporting evidence-based, site-specific land use and management through cloud computing, mobile applications, and crowdsourcing," *J. Soil Water Conserv.*, vol. 68, no. 1, pp. 5A–12A, Jan./Feb. 2013, doi: 10.2489/jswc.68.1.5A.
8. B. K. Stephenson, D. G. Hoch, and L. P. Collette, III, "Validation of laboratory test data based on predicted values of property-of-interest," U.S. Patent 8,352,394 B2, Jan. 8, 2013.
9. K. W. Broman *et al.*, "Identification and correction of sample mix-ups in expression genetic studies," *G3: Genes, Genomes, Genetics*, vol. 5, no. 10, pp. 2177–2186, Oct. 2015, doi: 10.1534/g3.115.019778.
10. J. Schmidinger, R. Gebbers, M.-O. Gasser, V. Barkov, G. M. Wu, and V. I. Adamchuk, "Rejections based on predictive uncertainty enable reliable routine soil spectroscopy," arXiv preprint arXiv:2606.21179, June 2026.
11. S. Bates, E. Candès, L. Lei, Y. Romano, and M. Sesia, "Testing for outliers with conformal p-values," *Ann. Statist.*, vol. 51, no. 1, pp. 149–178, Feb. 2023, doi: 10.1214/22-AOS2244.
12. Y. Jin and E. J. Candès, "Selection by prediction with conformal p-values," *J. Amer. Statist. Assoc.*, vol. 118, no. 544, pp. 2533–2545, Dec. 2023, doi: 10.1080/01621459.2023.2185532.
13. A. N. Angelopoulos and S. Bates, "Conformal prediction: A gentle introduction," *Found. Trends Mach. Learn.*, vol. 16, no. 4, pp. 494–591, 2023, doi: 10.1561/2200000101.
14. V. Vovk, A. Gammerman, and G. Shafer, *Algorithmic Learning in a Random World*. New York, NY, USA: Springer, 2005.
15. Y. Benjamini and Y. Hochberg, "Controlling the false discovery rate: A practical and powerful approach to multiple testing," *J. Roy. Statist. Soc. Ser. B*, vol. 57, no. 1, pp. 289–300, 1995, doi: 10.1111/j.2517-6161.1995.tb02031.x.
16. R. J. Tibshirani, R. Foygel Barber, E. J. Candès, and A. Ramdas, "Conformal prediction under covariate shift," in *Adv. Neural Inf. Process. Syst. (NeurIPS)*, vol. 32, pp. 2530–2540, Dec. 2019.