Reported result

c_Cd = (1002.7 ± 1.7) mg/L, where the expanded uncertainty is the combined standard uncertainty multiplied by the coverage factor k = 2.00, corresponding to a coverage probability of approximately 95 %. U is rounded to 2 significant digits: ordinary rounding unless it lowers U by more than 5 %, then upwards (EA-4/02 M:2022, 6.3).

GUM uncertainty framework (JCGM 100)

Estimate	1002.70 mg/L
Combined standard uncertainty	0.864 mg/L
Effective degrees of freedom	∞
Coverage factor	2.00
Coverage probability	95.45 %
Expanded uncertainty	1.73 mg/L
Kragten u_c	0.863 mg/L

Uncertainty budget

Quantity	Estimate	Standard uncertainty	Degrees of freedom	Sensitivity coefficient	Contribution	Share of variance	Kragte
P	0.999900	5.77×10^{-5}	∞	1003	0.0579	0.449 %	0.057
m	100.280	0.0500	∞	10.0	0.500	33.5 %	0.500
V	100.000	0.0700	∞	-10.0	-0.702	66.0 %	-0.701

Contributions to the combined variance

