
(a) Findings by source document
c1
JCGM GUM-5:2026 3 13
Eurachem QUAM 11
UKAS M3003 3 7
EMPIR EMUE 2 7
EURAMET cg-18 6
Nordtest TR 537 6
Eurachem UfS 5
EA-4/02 1 4
ISO 5725 4
Eurachem MUPR 4
JCGM 100 (GUM) 4
NIST TN 1900 1 3
Bettencourt da Silva 2026 3
Nordtest TR 569 3
ISO 13528 3
ISO 6143 1 2
ISO 21748 1 2
14 further documents 21
0 5 10 15 20 25
Findings (a finding with two classes counts in both)
c1: changes a final printed result
other c: c2, m, e, open
a: intermediate rounding
b: misprint without effect
n: no discrepancy
(b) How far the 12 confirmed c1 findings move the printed result
one expanded uncertainty
F28 NIST TN 1900 E4
F78 JCGM GUM-5 Example 1
F82 JCGM GUM-5 Example 4
F57 UKAS M3003 K.3
F21 EA-4/02 S13
F90 EMPIR EMUE E3.3
F42 ISO 6143:2025 Example 3
F87 EMPIR EMUE E3.4
F56 UKAS M3003 K.5
F77 JCGM GUM-5 Example 15
F39 ISO 21748 C.3
F58 UKAS M3003 K.6
estimate y: |Δy|/U
uncertainty: |ΔU|/U
1 %
10 %
1
10
Change of the printed result, relative to its U