

Medical Robotics

Marilena Vendittelli

Robots and Robotic Technologies for Interventional Radiology



SAPIENZA
UNIVERSITÀ DI ROMA



- “manually executed” interventional radiology (IR) procedures
- value added by robotic systems
- challenges in IR robotic systems design
- a (nonexhaustive) catalogue of existing systems
- open problems
- toward teleoperated needle insertion



IR procedures



- minimally invasive procedures: needles, probes or catheters, are used to reach the surgical target through or to deliver some kinds of therapy
 - the patient skin → percutaneous procedures
 - e.g., core biopsy of tumors, embolization, tumor ablation
 - anatomical structures → vascular procedures
 - e.g., angioplasty and stenting
- performed under the guidance of imaging systems:
 - ultrasound (US) there is a probe that admits ultrasounds signal which is reflected depending on the impedance of the tissue and it is reflected back to the probe and based on the delay, the image is reconstructed
 - mainly soft tissues and superficial anatomical structures
 - fluoroscopy (XA) x-ray imaging with a contrast medium
 - computed tomography (CT) 3D reconstruction of the part of human body that is being image through the acquisition of several slices (NO REAL-TIMEit requires elaboration of images)
 - magnetic resonance (MR) usually it is not real-time but today most modern system provide a quasi-real-time image acquisition

real-time image acquisition



percutaneous procedures

criticalities of traditional procedure




the needle is not inserted in one step but gradually



identify the target and plan the procedure

manually perform the procedure

assess results and complications



reach the target manually



Phase risks

poorly planned access, inappropriate probe route, “no-touch” structures, real-time imaging visualization limitations

Phase risks

patient movements, operator fatigue, in-procedure complications, need for repositioning (iterative process)

Phase risks

subjective assessment of procedure outcome, lack of quantitative post-procedure validation

this process is not automatical



percutaneous procedures

how robots can help

this reduces the risk and the errors related to the operator. The problem is that there are some system (ex ROBIO) to which miss the flexibility and sensitivity (i.e. intelligence)



the setup of these system is more expensive in terms of time, organization and economic


we want to combine the flexibility and intelligence of the human operator with the aspect of accuracy, reproducibility and also the possibility to add some features to help the operator

today imaging guided interventions

today robotics-based alternatives

tomorrow imaging guided interventions

CT scan



manipulator at the patient side which performs the needle insertion in the patient. It is not completely automatic because there is an operator in the operating room that guiding the needle under the guidance of images and also using an haptic interface to feel the forces that are exchanged between the needle or probe and the tissue



performed by human operators

- + operator experience factor
- + high flexibility/adaptability
- + high sensibility/discernibility
- operator training and costs
- subjective result evaluation
- mental stress
- exposure to ionizing radiation

mainly guidance systems

- + reduced operator-related costs and risks
- + high reproducibility
- + reduced stress-related errors
- less flexibility
- non-Intelligent Platforms
- high development costs

automated or hybrid human-robot procedures

- + reliability, reproducibility, procedure monitoring and optimization
- + post-operative data analysis
- + reduced training and personnel costs and risks
- + improved comfort for operators
- + remote control when trained operator not available
- + more procedures per session
- + less complications



introduction

technical challenges



need of

- robot **kinematic architectures**
 - fitting the workspace constrained by the imaging device
 - complying with the fixed point of access to the patient body
- sensors and actuators
 - compatible with the imaging modality
- modeling and control methodologies for
 - tool motion planning and steering
 - compensation of physiological motions and tissue deformation
 - constrained planning and motion

probe and needle deformation
can due to a deviation from the
plan



introduction

technical challenges (cont'd)



- pre- and intra-operative image fusion
- force sensing and rendering can compromise sometimes transparency but can also improve other kind of properties in particular stability. Stability and transparency are in contrast each other
- user interfaces providing
 - haptic and visual feedback which kind of feedback
 - augmented information



percutaneous procedures

navigation systems



- commercial devices


SimplyCT (Neorad)

how is obtain this direction that the laser follows?

this navigation system helps the operator to align the needle with the projected laser ray that points to the internal target



laser guide





percutaneous procedures

navigation systems



SimplyCT

planning phase (relies on previous laser/CT scan calibration)

the operator plans the needle trajectory to reach this target and since this point, visible in the CT scan, and the guided grid is on the patient, this point is directly marked on the patient skin with a special marker like a pen

guiding grid
this is visible in the CT scan





percutaneous procedures

navigation systems




from one reference system to another. If we know where this point is in the reference frame of the imaging system and by using the calibration between the imaging system and the laser, we can express the coordinates of this point in the reference frame of the laser. The laser can point directly to the internal target

SimplyCT

alignment, insertion, verification → sampling



the operator aligns the needle with this laser beam



red spot indicates that the needle is aligned with the laser beam



further verification by taking again another CT scan



percutaneous procedures


patient-mounted systems



- commercial devices mainly passive → not robot

SeeStar (Apriomed)

four are attached to the patient skin



rotating arches of half spheres for orbital
needle positioning

arches along which the needle can be moved so as to oriented to
work in the right direction pointing at the internal target. Once in
position the needle is inserted manually

note: both for SimplyCT and Seestar the needle is oriented and inserted manually






percutaneous procedures

patient-mounted systems



- research prototypes

constructed to go in the CT scan. It doesn't produce artifact in the images



Robopsy

- orbital positioning
- 2 stepper motors for insertion of the needle
- manual alignment

but the insertion is automatic through stepper motors

CT-bot

- 5-dof, parallel architecture
- ultrasonic piezo motors
- fiducials for registration
- 2-dof steering device for needle spinning and insertion

Light Puncture Robot (LPR)

- MR-compatible
- 2-dof, pneumatic actuators
- frame fixed to the scanner table
- fiducials for registration
- rapid needle insertion (9cm/s) and release

for insertion and orientation of the needle




percutaneous procedures

gantry/floor-mounted systems



- commercial devices



higher degree of accuracy



it provides a guidance for the needle



INNOMOTION

- one of the first MR-compatible, CE cleared
- 6-dof, pneumatic motors 6 motors
- fornt-end accomodates coaxial probes (cannulae, laser probes, endoscopes)
- no longer commercialized

Soteria

- MR-guided prostate biopsy
- 5-dof hybrid parallel-serial architecture
- automatic planning and motion execution

iSYS1

- CT and XA guide
- 2-dof orienting the needle guide
- passive positioning arm
- planning software
- manual insertion

ROBIO

- CT/PET-CT guide
- 5-dof
- floor-mounted
- image-based registration
- manual insertion




percutaneous procedures

gantry/floor-mounted systems



- a very recently commercialized system

no forces sensing



the insertion is automatic. The operator plans the trajectory of the needle based on preoperative and real-time images.

step by step operation that can be corrected by the operator at a discrete time

<https://www.youtube.com/watch?v=3eJCBAUcIMQ>





percutaneous procedures

gantry/floor-mounted systems



• research prototypes

the main material challenge is the compatibility with high magnetic field




Siemens

- XA guide
- 2-dof RCM + a needle guide
- controlled by joystick or through visual servoing



B-Rob

precursor of iSYS



AcuBot

- manual positioning to the entry point relying on the CT laser
- automatic orientation through the RCM module
- registration by alignment with the CT laser



MrBot

- MR-guided
- 6-dof, pneumatic stepper motors
- transperineal access to the prostate for biopsy, thermal ablation, brachytherapy
- registration markers





percutaneous procedures

gantry/floor-mounted systems



- an open-access CT-compatible prototype



- 7-dof
- belt/cable driven
- user interface for remote control
- V-REP simulator



percutaneous procedures

summing up: limitations of current systems



semi-autonomous
positioning devices


advanced guidance
technologies

advanced image
processing




Limitations

rigid, non-intelligent
positioning devices, real-
time control of insertion
not available, bulky and
difficult to position



Limitations

require image
registration or patient
sensor, no real-time
control, patient
movement



Limitations

limited to post-
procedure phase, not
routinely used to modify
procedure outcome



percutaneous procedures

summing up: challenges of tomorrow's interventions



fully autonomous
positioning devices



Challenge

advanced robotic arm
technology and control
methodologies for safe
physical interaction with
patients and staff

remote, real-time
haptic control



Challenge

integrated haptic
controller with extra-fine
force feedback for robot
teleoperation by medical
personnel

real-time imaging +
prediction models

advanced procedure
monitoring



Challenge

high-speed
communication between
scanner and control
software, real-time 3D
procedure adaptation




vascular procedures

the flexible probe is guided inside the body through haptic interface. The couple is between the haptic handle and the tip of the probe

the Sensei X2 system (similar to Magellan)



system to introducing a flexible catheter inside the human body to reach the heart starting from the femoral arteria (for ex. used for treatment of aritmia). The procedure is performed under the guidance of a c arm used for fluoroscopy.





vascular procedures

commercial systems



Magellan (Hansens)



CorPath

- similar to Magellan and SenseiX
- used for coronary and peripheral vascular interventions



Amigo

- closer to traditional steering methods (knobs and buttons)
- no possibility of haptic feedback



Niobe (Stereotaxis Inc.)

- XA guide
- electrophysiology procedures
- remote magnetic catheter control: uniform magnetic field generated by arms motion for omnidirectional steering of a catheter tip




vascular procedures


commercial systems



accuracy




stability



softness

Genesis (Stereotaxis Inc.)

- 70-80% faster than Niobe
- reduced size
- increased workspace





open problems

endovascular procedures

it is easier to design a teleoperated system for endovascular procedure.



although the illustrated systems have changed the standard of care in both percutaneous and intravascular procedures, challenging aspects remain, like e.g.,

- smaller, lighter, intrinsically safe robots
- control techniques for safe human-robot interaction
- prediction models
- real-time (re)planning
- force sensing at the needle tip and haptic rendering
- shape and force sensing on the whole catheter body
- compensation of cardiac heart motion
- transmission of sufficient force to execute the treatment