



S2225 Pediatric HAL® Final Inspection Form

1. Documentation (Provided by Production)

Serial # : _____

Customer: _____

Options: (Special Orders):

Voltage: _____

New ☐ Repair ☐ Upgrade ☐ Show ☐ Loaner ☐

Color: Light ☐ Medium ☐ Dark ☐

2. Cosmetic

	Pass	Fail		Pass	Fail
1 Skin overall condition	_____	_____	6 ECG & Defibrillation Snaps OK	_____	_____
2 Male Genitalia placement	_____	_____	7 Tracheostomy insert	_____	_____
3 Female Genitalia placement	_____	_____	8 Bilateral IV arms(Air Test)	_____	_____
4 Urinary Catheterization (5 ml min)	_____	_____	9 Right I/O access (correct position)	_____	_____
5 Urinary Control	_____	_____	10 Glucose Finger	_____	_____

Notes: _____

3. Hardware

	Pass	Fail		Pass	Fail
1 Battery Level without charger	_____	_____	3 Battery Level with Charger	_____	_____
2 Face Movement & Tracking Test	_____	_____		_____	_____

Notes: _____

4. Neurological Functions

	Pass	Fail		Pass	Fail
1 Mouth movement	_____	_____	7 Skin appearance red (redness)	_____	_____
a Mouth Open	_____	_____	8 Skin appearance yellow (jaundice)	_____	_____
b Mouth Closed	_____	_____	9 Skin appearance Cyanosis(central)	_____	_____
c Teeth Chattering	_____	_____	10 Programmable Tears Left	_____	_____
5 Skin appearance Calibration	_____	_____	11 Programmable Tears Right	_____	_____
6 Skin appearance white (paleness)	_____	_____	12 Seizures	_____	_____

Notes: _____



5. Circulatory Functions

	Pass	Fail		Pass	Fail
1 Carotid Right Pulse	_____	_____	9 Pedal Right Pulse	_____	_____
2 Carotid Left Pulse	_____	_____	10 Pedal Left Pulse	_____	_____
3 Brachial Right Pulse	_____	_____	11 Tongue Edema	_____	_____
4 Brachial Left Pulse	_____	_____	12 Laryngospasms	_____	_____
5 Radial Right Pulse	_____	_____	13 Manual CRT 5 sec (Right Knee)	_____	_____
6 Radial Left Pulse	_____	_____	14 Osat Left Index Finger (Calib)	_____	_____
7 Femoral Right Pulse	_____	_____			
8 Femoral Left Pulse	_____	_____			

18 _____ Pulse Strength Check (Verify ✓)

	Carotid	Femoral	Radial	Pedal
BP = 49/40	_____	_____	_____	_____
BP = 54/40	_____	_____	_____	_____
BP = 59/40	_____	_____	_____	_____
BP = 100/70	_____	_____	_____	_____

19 _____ Bp = 120/80 (Using Stethoscope/Blood Pressure Cuff verify ✓)

20 _____ NIBP left arm (calibration required using calibration kit)

Korotkoff (Sound)
(Systolic) 115-120 (Diastolic) 75-80
Left Arm _____

180/110 (+/- 10) _____
160/90 (+/- 10) _____
120/80 (+/- 10) _____
100/60 (+/- 10) _____

Notes: _____

6. Sounds ✓

	Sound OK		Sound OK
1 Lung Right Front	_____	10 Bowel Sound Lower Right	_____
2 Lung Left Front	_____	11 Crying	_____
3 Lung Right Upper Back	_____	12 Aortic Heart Sound	_____
4 Lung Left Upper Back	_____	13 Pulmonic Heart Sound	_____
5 Lung Right Lower Back	_____	14 Mitral Heart Sound	_____
6 Lung Left Lower Back	_____	15 Throat Sound	_____
7 Bowel Sound Upper Left	_____	16 Speech	_____
8 Bowel Sound Upper Right	_____	17 Streaming Voice Spkr	_____
9 Bowel Sound Lower Left	_____	18 Streaming Voice M-phone	_____

Notes: _____



7. Breathing Functions

	Pass	Fail		Pass	Fail
1 Lung Compliance	_____	_____	9 Right Lung Disable	_____	_____
2 Lung Resistance	_____	_____	10 Left Lung Disable	_____	_____
3 Esophageal intubation detection	_____	_____	11 ET Tube Position 3 (Too Deep)	_____	_____
4 CO ₂	_____	_____	12 Position 3 Ventilation Right Lung	_____	_____
5 Chest Rise	_____	_____	13 Right Mainstem Intubation	_____	_____
6 Respiration Rate (14/20/30/ b min	_____	_____	14 ET Tube Position 1 (Normal)	_____	_____
7 Needle decompression	_____	_____	15 Position 1 Ventilation Both Lungs	_____	_____
8 Chest tube detection	_____	_____			

Notes: _____

8. Airway, Cardiac, Respiratory, and Connectivity Functions

	Pass	Fail		Pass	Fail
1 Artificial Ventilation with Mask	_____	_____	12 Rhythms and HR's OK	_____	_____
2 Simultaneous compressions/ventilat	_____	_____	13 Respiratory Graph	_____	_____
3 Auto Lung Sound	_____	_____	14 Change HAL 5 Year to Channel 4	_____	_____
4 Chest Compression	_____	_____	15 Bluetooth Connection testing	_____	_____
5 Auto Carotid Pulse	_____	_____	16 Labels	_____	_____
6 AP Pacing	_____	_____	a USB "Connect to Tablet USB Port"	_____	_____
7 ECG defibrillation snaps	_____	_____	b Tablet, Desktop "This is a dedica..."	_____	_____
8 ECG Rhythm at ECG site	_____	_____	c "Pediatric Hal Five Years" Charger Label	_____	_____
9 ECG at Shock Sites	_____	_____	d "CE" Label	_____	_____
10 Pacing Detection	_____	_____	e "Patented; other..." Label	_____	_____
11 Pacing Capture	_____	_____	17 Check S/N back	_____	_____

Notes: _____

Inspected by: _____ Date: _____