



VICTORIA™ S2200 Final Inspection Form

1. Documentation (Provided by Production)

Serial # simulator : _____ Voltage: _____
Serial number for computer used for testing and is going to be shipped with simulator _____
Customer: _____ New ☐ Repair ☐ Upgrade ☐ Show ☐ Loaner ☐
Options: (Special Orders): _____ Color: Light ☐ Medium ☐ Dark ☐

2. Cosmetic

	Pass	Fail		Pass	Fail
1 Skin overall condition	_____	_____	3 ECG & Defibrillation snaps OK	_____	_____
2 Power / Communication Silicone Cover	_____	_____			

Notes: _____

3. Hardware

	Pass	Fail		Pass	Fail
1 Battery Level when charging test	_____	_____	4 Power Supply	_____	_____
2 Compressor Pressure up to 180 mmHg	_____	_____	5 Epidural (too deep)	_____	_____
3 Battery Level without charger	_____	_____	6 Tracking Eyes Test	_____	_____

Notes: _____

4. Functions

	Pass	Fail		Pass	Fail
1 Left Eye Dilation "Light"	_____	_____	14 Seizures (Mild)	_____	_____
2 Right Eye Dilation "Light"	_____	_____	15 Seizures (Severe)	_____	_____
3 Eyes Blinking	_____	_____	16 Laryngospasm	_____	_____
4 Carotid Right Pulse	_____	_____	17 Tongue Edema	_____	_____
5 Carotid Left Pulse	_____	_____	18 Pharyngeal swelling	_____	_____
6 Brachial Right Pulse	_____	_____	19 O ₂ SAT Index Left (70, 80, 95)	_____	_____
7 Brachial Left Pulse	_____	_____	20 Left IV Arm test	_____	_____
8 Radial Right Pulse	_____	_____	21 Drug Recognition – Right arm		
9 Radial Left Pulse	_____	_____	a Identify tag	_____	_____
10 Right Radial Pulse Disabled	_____	_____	b Programming tab	_____	_____
11 Left Radial Pulse Disabled	_____	_____	c Detection 10 ml drug	_____	_____
12 Chest Rise	_____	_____	d Delete drugs	_____	_____
13 Respiration Rate (30/20/13/ bp/min)	_____	_____	22 CO ₂ System Capability (port in right side, power connection box)	_____	_____
			23 CO ₂ (Optional)	_____	_____



(Systolic) 115-120
Left Arm

200/120 (+/- 10) _____
 160/90 (+/- 10) _____
 120/80 (+/- 10) _____
 100/60 (+/- 10) _____

Notes:

Sound OK

- 1 Lung Right Upper Front
- 2 Lung Left Upper Front
- 3 Heart Sound

Sound OK

- 4 Throat Sound
- 5 Speech
- 6 Streaming Voice Spkr
- 7 Streaming Voice M-phone

Notes:

Pass	Fail
1. The system shall be able to handle up to 1000 concurrent users. 2. The system shall be able to process transactions within 2 seconds. 3. The system shall be able to store data for at least 5 years. 4. The system shall be able to generate reports in PDF format. 5. The system shall be able to integrate with existing systems.	1. The system shall be able to handle up to 1000 concurrent users. 2. The system shall be able to process transactions within 2 seconds. 3. The system shall be able to store data for at least 5 years. 4. The system shall be able to generate reports in PDF format. 5. The system shall be able to integrate with existing systems.

- 1 ET Tube Position 3 (Too Deep)
- 2 ET Tube Position 1 (Normal)
- 3 Nose Intubation
- 4 Artificial Ventilation with Mask
- 5 Auto Lung Sound
- 6 Chest Compression
- 7 Auto Carotid Pulse
- 8 Simultaneous compressions/ventilator
- 9 ECG at ECG snaps
- 10 Rhythms and HR's OK
- 11 ECG at defib snaps sites

Pass	Fail
1. The system shall be able to handle up to 1000 concurrent users. 2. The system shall be able to process transactions within 2 seconds. 3. The system shall be able to store up to 100,000 records. 4. The system shall be able to generate reports within 5 minutes. 5. The system shall be able to handle data from multiple sources. 6. The system shall be able to handle data from multiple formats. 7. The system shall be able to handle data from multiple languages. 8. The system shall be able to handle data from multiple time zones. 9. The system shall be able to handle data from multiple currencies. 10. The system shall be able to handle data from multiple countries.	1. The system shall be able to handle up to 1000 concurrent users. 2. The system shall be able to process transactions within 2 seconds. 3. The system shall be able to store up to 100,000 records. 4. The system shall be able to generate reports within 5 minutes. 5. The system shall be able to handle data from multiple sources. 6. The system shall be able to handle data from multiple formats. 7. The system shall be able to handle data from multiple languages. 8. The system shall be able to handle data from multiple time zones. 9. The system shall be able to handle data from multiple currencies. 10. The system shall be able to handle data from multiple countries.

- 12 Pacing detection
- 13 Pacing capture
- 14 Labels on Simulator
 - a Simulator S/N Label
 - b URINE 200 ml max
 - c BLOOD 200ml max
 - d LUBE 10 ml max
 - e DRAIN after delivery
- 15 Right leg rotation sensor
- 16 Left leg rotation sensor

Notes: _____

Pass	Fail
-------------	-------------

- 1 Rotation Motor Encoder
- 2 Translation Motor
- 3 Lubrication
- 4 Urinary Catheterization
- 5 Check birth canal snaps and hoses
- 6 Check Victoria cervixes
- 7 Check Drain
- 8 Hemorrhage

Pass	Fail
1. The system is able to handle the load without crashing or significant performance degradation. 2. The system is able to handle the load without crashing or significant performance degradation. 3. The system is able to handle the load without crashing or significant performance degradation. 4. The system is able to handle the load without crashing or significant performance degradation. 5. The system is able to handle the load without crashing or significant performance degradation. 6. The system is able to handle the load without crashing or significant performance degradation. 7. The system is able to handle the load without crashing or significant performance degradation. 8. The system is able to handle the load without crashing or significant performance degradation. 9. The system is able to handle the load without crashing or significant performance degradation. 10. The system is able to handle the load without crashing or significant performance degradation.	1. The system is unable to handle the load, crashing or experiencing significant performance degradation. 2. The system is unable to handle the load, crashing or experiencing significant performance degradation. 3. The system is unable to handle the load, crashing or experiencing significant performance degradation. 4. The system is unable to handle the load, crashing or experiencing significant performance degradation. 5. The system is unable to handle the load, crashing or experiencing significant performance degradation. 6. The system is unable to handle the load, crashing or experiencing significant performance degradation. 7. The system is unable to handle the load, crashing or experiencing significant performance degradation. 8. The system is unable to handle the load, crashing or experiencing significant performance degradation. 9. The system is unable to handle the load, crashing or experiencing significant performance degradation. 10. The system is unable to handle the load, crashing or experiencing significant performance degradation.

- 9 Test baby connects to motor
VDB S/N _____
- 10 Perform one full delivery
- 11 Visual inspection for lower torso
leaking
- 12 Change Channel
- 13 Wire Communication

Inspected by: _____ **Date:** _____



8. Abdominal Covers

Contraction Abdominal Cover

	Pass	Fail
1 Check contraction intensity w/monitor	_____	_____
2 Upper Fetal TOCO Sound	_____	_____
3 Lower Fetal TOCO Sound	_____	_____
4 Bowel Sound Left/Right	_____	_____
5 Suprapubic Pressure	_____	_____
6 Overall condition, fits and latch to torso	_____	_____

PPH Abdominal Cover

	Pass	Fail
1 Urinary Catheterization	_____	_____
2 Uterine Hemorrhage	_____	_____
3 Vaginal Hemorrhage	_____	_____
4 Suppository	_____	_____
5 Switch position	_____	_____
6 Overall condition and fits in torso	_____	_____
7 Uterine pressure	_____	_____
8 Check power connection between Victoria and the PPH cover	_____	_____

Palpation Abdominal Cover

	Pass	Fail
1 Overall condition and fits in torso	_____	_____
	_____	_____

C-section Abdominal Cover

	Pass	Fail
1 Overall condition, fits and latch to torso	_____	_____
2 C-Section baby head/breech assembly fits to belly	_____	_____
3 C-Section insert snaps inspection	_____	_____

Episiotomy Repair Insert

	Pass	Fail
1 Snaps position and orientation	_____	_____
2 Overall condition	_____	_____

PPH Abdominal Cover Inserts (regular and 4th degree)

	Pass	Fail
1 Urinary Catheterization	_____	_____
2 Uterine Hemorrhage	_____	_____
3 Vaginal Hemorrhage	_____	_____
4 Snaps position and orientation	_____	_____

Miscellaneous:

- Check snaps in lower legs: Pass _____ Fail _____ N/A _____
- Assembly lower legs to upper legs and make sure they bent properly: Pass _____ Fail _____ N/A _____
- Attach pictures of the four C-sections attached to lower torso (digital) Yes _____ No _____ N/A _____ Date _____

Notes: _____

Inspected by: _____ Date: _____

Victoria Tracking Eyes Inspection Chart

Blinking Function	Pass/ Fail	Pupil Reaction	Pass/ Fail
10 blinks per min (60 secs)	/	Consensual reaction to light ON/OFF	/
15 blinks per min (60 secs)	/	Right eye reaction ON/OFF	/
20 blinks per min (60 secs)	/	Left eye reaction ON/ OFF	/
Eye Strabismus	Pass/Fail	Pupil Dilation	Pass/Fail
Crossed eyes	/	Left pupil dilation	/
Left esotropia	/	Right pupil dilation	/
Left exotropia	/	Mydriasis right 9/left 9 pupils	/
Right esotropia	/	Anisocoria right 7/left 2 pupils	/
Right exotropia	/		
Eye Auto Tracking	Pass/Fail	Eye/Idle Movement	Pass/Fail
Both eyes	/	Idle-Normal	/
Right eye only	/	Idle-Worried	/
Left eye only	/	Idle-Anxious	/
Horizontal nystagmus	/	Left	/
Right nystagmus	/	Right	/
Left nystagmus	/	Up	/
Vertical nystagmus	/	Down	/
Blepharospasm Severity	Pass/Fail	Joystick Control	Pass/Fail
Severity 2	/	Horizontal/Vertical	/
Severity 3	/	Horizontal	/
Severity 4	/	Vertical	/
Severity 5	/		
Ptosis Severity	Pass/Fail	Nystagmus Option (Joystick)	Pass/Fail
Severity 2	/	Horizontal	/
Severity 3	/	Right beat	/
Severity 4	/	Left beat	/
Severity 5	/	Vertical	/
		Downbeat	/
		Upbeat	/
		Pendular	/

Comments: