



S3005 Pediatric HAL® 5 Year Old Final Inspection Form

1. Documentation (Provided by Production)

Serial # : _____

Customer: _____

Options: (Special Orders):

Voltage: _____

New ☐ Repair ☐ Upgrade ☐

Color: Light ☐ Medium ☐ Dark ☐

2. Cosmetic

	Pass	Fail		Pass	Fail
1 Skin overall condition	_____	_____	3 ECG & Defibrillation Patches OK	_____	_____
2 Both arms with Injection Cover	_____	_____	4 Lower left ECG patch cover	_____	_____

Notes: _____

3. Hardware

	Pass	Fail		Pass	Fail
1 Battery Level holding test	_____	_____	2 External Charger Connected	_____	_____

Notes: _____

4. Functions

	Pass	Fail		Pass	Fail
1 Right Eye Blinking	_____	_____	13 Right Radial Pulse Disabled	_____	_____
2 Right Eye Dilatation "Light"	_____	_____	14 Left Radial Pulse Disabled	_____	_____
3 Left Eye Blinking	_____	_____	15 Chest Rise	_____	_____
4 Left Eye Dilatation "Light"	_____	_____	16 Right Lung Disable	_____	_____
5 Carotid Right Pulse	_____	_____	17 Left Lung Disable	_____	_____
6 Carotid Left Pulse	_____	_____	18 Respiration Rate (14/20/30/ b min)	_____	_____
7 Brachial Right Pulse	_____	_____	19 Tongue Edema	_____	_____
8 Brachial Left Pulse	_____	_____	20 Seizures (Mild)	_____	_____
9 Radial Right Pulse	_____	_____	21 Seizures (Severe)	_____	_____
10 Radial Left Pulse	_____	_____	22 Nose Intubation	_____	_____
11 Femoral Right Pulse	_____	_____	23 Urinary Catheterization	_____	_____
12 Femoral Left Pulse	_____	_____	24 Check Urine Bag	_____	_____
			25 Physiological Model	_____	_____



_____ Pulse Strength Check (Verify ✓)

BP = 50/40
BP = 55/40
BP = 60/40
BP = 100/70

Carotid Femoral Radial

_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

_____ Bp = 100/70 (Using Stethoscope/Blood Pressure Cuff verify ✓)

Korotkoff (Sound)

(Systolic) 100

(Diastolic) 70

Right Arm _____

Left Arm _____

Notes: _____

5. Sounds ✓

Sound OK

Sound OK

- 1 Lung Right Front _____
- 2 Lung Left Front _____
- 3 Lung Right Back _____
- 4 Lung Left Back _____
- 5 Bowel Sound Upper Left _____
- 6 Bowel Sound Upper Right _____

- 7 Bowel Sound Lower Left _____
- 8 Bowel Sound Lower Right _____
- 9 Heart Sound _____
- 10 Throat Sound _____
- 11 Speech _____
- 12 Streaming Voice Spkr (Optional) _____
- 13 Streaming Voice M-phone (Optional) _____

Notes: _____

6. Other Functions

Pass Fail

Pass Fail

- 1 Cyanosis Level – Central _____
- 2 ET Tube Position 3 (Too Deep) _____
- a Position 3 –Right Lung Up w/Air Bag _____
- 3 ET Tube Position 1 (Normal) _____
- a Position 1 –Both Lungs Up w/Air Bag _____
- 4 Cyanosis Level - Healthy _____
- 5 Artificial Ventilation with Mask _____
- 6 Auto Lung Sound _____
- 7 Belly Bag inflate/deflate _____
- 8 Chest Compression _____
- 9 Auto Carotid Pulse _____

- 10 ECG at ECG _____
- 11 Rhythms and HR's OK _____
- 12 ECG at Shock Sites _____
- 13 Pacing Detection _____
- 14 Pacing Capture _____
- 15 Change HAL 5 Year to Channel 4 _____
- 16 Labels _____
- a USB "Connect to Tablet USB Port" _____
- b Tablet, Desktop "This is a dedica..." _____
- c "Pediatric Hal Five Years" Charger Label _____
- d "CE" Label _____
- 17 Check S/N Left Upper Leg _____

Notes: _____

Inspected by: _____ Date: _____