



## Super Tory® S2220 Final Inspection Form

### 1. Documentation (Provided by Production)

Serial # simulator: \_\_\_\_\_

Voltage: \_\_\_\_\_

Customer: \_\_\_\_\_

New ☐ Repair ☐ Upgrade ☐ Show ☐ Loaner ☐

Options: (Special Orders): \_\_\_\_\_

Color: Light ☐ Medium ☐ Dark ☐

### 2. Cosmetic

|                                        | Pass  | Fail  |                                       | Pass  | Fail  |
|----------------------------------------|-------|-------|---------------------------------------|-------|-------|
| 1 Skin overall condition               | _____ | _____ | 7 ECG & Defibrillation snaps OK       | _____ | _____ |
| 2 Charger port silicone cover          | _____ | _____ | 8 Male Genitalia                      | _____ | _____ |
| 3 Left I/O access (correct position)   | _____ | _____ | 9 Bilateral IV arms                   | _____ | _____ |
| 4 Right I/O access (correct position)  | _____ | _____ | 10 Umbilical cord (alcohol test)      | _____ | _____ |
| 5 Chest drain left (correct position)  | _____ | _____ | 11 Scalp IV test (air test)           | _____ | _____ |
| 6 Chest drain right (correct position) | _____ | _____ | 12 Urinary Catheterization (4 ml min) | _____ | _____ |

Notes: \_\_\_\_\_

### 3. Hardware

|                                    | Pass  | Fail  |                                       | Pass  | Fail  |
|------------------------------------|-------|-------|---------------------------------------|-------|-------|
| 1 Battery Level when charging test | _____ | _____ | 2 Battery Level without charger (10V) | _____ | _____ |

Notes: \_\_\_\_\_

### 4. Neurological/Circulatory Functions

|                                                      | Pass  | Fail  |                                      | Pass  | Fail  |
|------------------------------------------------------|-------|-------|--------------------------------------|-------|-------|
| 1 Eyes open                                          | _____ | _____ | 13 Fontanelle (sunken)               | _____ | _____ |
| 2 Eyes close                                         | _____ | _____ | 14 Fontanelle (bulging)              | _____ | _____ |
| 3 Eyes Blinking (10 b/m)                             | _____ | _____ | 15 Fontanelle (normal)               | _____ | _____ |
| 4 Mouth movement (active, smacking, closed, relaxed) | _____ | _____ | 16 Right brachial pulse              | _____ | _____ |
| 5 Muscle Tone (Active)                               | _____ | _____ | 17 Left brachial pulse               | _____ | _____ |
| 6 Muscle Tone (Jittery)                              | _____ | _____ | 18 Umbilical pulse                   | _____ | _____ |
| 7 Muscle Tone (Limp)                                 | _____ | _____ | 19 Fontanel pulse                    | _____ | _____ |
| 8 Skin appearance white (paleness)                   | _____ | _____ | 20 Femoral left pulse                | _____ | _____ |
| 9 Skin appearance red (redness)                      | _____ | _____ | 21 Femoral right pulse               | _____ | _____ |
| 10 Skin appearance yellow (jaundice)                 | _____ | _____ | 22 Manual CRT 5 sec (left foot)      | _____ | _____ |
| 11 Skin appearance Cyanosis (central)                | _____ | _____ | 23 Osat right foot (Calib. required) | _____ | _____ |
| 12 Skin appearance Cyanosis (peripher)               | _____ | _____ | 24 Osat right hand (Calib. Required) | _____ | _____ |

|                                                                                                                                                                                                                                             |                                                                                                                                                        |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------|
| BP = 60/40 (Using Stethoscope/Blood Pressure Cuff verify <input checked="" type="checkbox"/> )<br><div style="text-align: center;">Korotkoff (Sound)</div> (Systolic) 60+/-5    (Diastolic) 40+/- 5<br>Right Arm                      _____ | NIBP right arm (No Calibration required)<br>120/80    (+/- 10) _____<br>100/60    (+/- 10) _____<br>60/40    (+/- 10) _____<br>50/30    (+/- 10) _____ |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------|

\_\_\_\_\_ Pulse Strength Check (Verify ☒)  
 Fontanel Brachial Femoral Umbilical

BP = 34/25    \_\_\_\_\_  
 BP = 60/40    \_\_\_\_\_

### 5. Sounds

|                          | Sound OK |                        | Sound OK |
|--------------------------|----------|------------------------|----------|
| 1 Lung Right Upper Front | _____    | 4 Throat Sound         | _____    |
| 2 Lung Left Upper Front  | _____    | 5 Crying               | _____    |
| 3 Heart Sound            | _____    | 6 Bowel sounds         | _____    |
|                          |          | 7 Abdominal Distension | _____    |

Notes: \_\_\_\_\_

### 6. Airway, Cardiac, Respiratory, and Connectivity Functions

|                                        | Pass  | Fail  |                                       | Pass  | Fail  |
|----------------------------------------|-------|-------|---------------------------------------|-------|-------|
| 1 Chest Rise                           | _____ | _____ | 17 Chest compression                  | _____ | _____ |
| 2 Respiration Rate (60/20/40/ bp/min)  | _____ | _____ | 18 Auto pulse detected                | _____ | _____ |
| 3 Left lung disable                    | _____ | _____ | 19 Simultaneous compressions/ventilat | _____ | _____ |
| 4 Right lung disable                   | _____ | _____ | 20 Lung Compliance(2-normal, 1-low)   | _____ | _____ |
| 5 Head Position (Hyper Flexed 1)       | _____ | _____ | 21 Ventilator Options (Check resp.    | _____ | _____ |
| 6 Head Position (Hyper Extended 2)     | _____ | _____ | Sound in the head changes)            | _____ | _____ |
| (Mild retractions activated with both) | _____ | _____ | 22 CO2 Capability ( left port )       | _____ | _____ |
| 7 Head Position (Normal 0)             | _____ | _____ | 23 ECG Rhythms at ECG monitor         | _____ | _____ |
| 8 Retractions- (Severe w/ hunger)      | _____ | _____ | 24 HR at ECG snaps                    | _____ | _____ |
| 9 Retractions- (Hernia)                | _____ | _____ | 25 Respiratory curve at ECG monitor   | _____ | _____ |
| 10 Nose Intubation                     | _____ | _____ | 26 ECG at defib. Snaps sites          | _____ | _____ |
| 11 ET Tube Position 3 (Too Deep)       | _____ | _____ | 27 HR at defib. snaps                 | _____ | _____ |
| 12 ET Tube Position 1 (Normal)         | _____ | _____ | 28 Pacing detected ( 80-83 mA )       | _____ | _____ |
| 13 Hypoxia model with cyanosis         | _____ | _____ | 29 Change RF Channel                  | _____ | _____ |
| 14 Artificial Ventilation with Mask    | _____ | _____ | 30 Bluetooth connection ready         | _____ | _____ |
| 15 Auto Lung Sound                     | _____ | _____ | 31 labels                             | _____ | _____ |
| 16 Manual Sealing of mouth and nose    | _____ | _____ | a Serial number ( left lower arm)     | _____ | _____ |
| (check airway obstruction)             | _____ | _____ | b Charger box label                   | _____ | _____ |

Inspected by: \_\_\_\_\_ Date: \_\_\_\_\_