

	Title: Form: Recertification Wired Probe Final QC Checklist	Rev 00	Page 1 of 2
	Number: FO-00156		

Wired Probe Serial # _____ Probe Size _____ mm

1. Verify that the Serial Number is present on handle	☐ Pass ☐ Fail
2. Verify Probe surfaces are of acceptable finish (no major dents, major scratches)	☐ Pass ☐ Fail
3. Verify that the Probe ODU connector is not physically damaged (no missing pins, no bent pins)	☐ Pass ☐ Fail
4. Verify that the Probe cable connector slides in and out smoothly	☐ Pass ☐ Fail
5. Verify that the Probe cable connector locks and unlocks properly	☐ Pass ☐ Fail
6. Verify that the Probe Window has a smooth finish, with no visible cracks	☐ Pass ☐ Fail
7. Plug in the Unit Under Test into the Reference Control Unit. Verify that the Reference Control Unit displays (0) when probe is away from the source.	☐ Pass ☐ Fail
Quick Check	
8. Plug in Unit Under Test into the Reference Control Unit and align with a Navigator Probe Fixture (PR-9000-20). Perform the following 10 second count test. After the double beep, divide the number of counts by 10, and record the results (round off the number). Set the Calibration switch to the “-” position _____ CPS Set the Calibration switch to the “0” position _____ CPS Set the Calibration switch to the “+” position _____ CPS Verify that the “0” position has the highest number of counts.	☐ Pass ☐ Fail

Note: If the Unit under Test does not meet specification, record the observation below and contact Dilon Technologies, Inc. to request an RMA number. Unit Under Test must be re-inspected after correction.

	Title: Form: Recertification Wired Probe Final QC Checklist Number: FO-00156	Rev 00	Page 2 of 2
---	--	---------------	------------------------

Reject Reason: _____ _____	
Corrected by: _____	Date: _____

Pass ☐ Fail ☐

Inspected by: _____ Date: _____