

Standard Operating Procedure

Implant and Medical Device Management

«Hospital Name» — Supports MOM.10.c, MOM.10.d, MOM.10.e (Achievement)

SOP No.	[Hospital to define]	Version	1.0
Effective Date	_____	Review Frequency	Annual, or on change
Prepared By	Quality Coordinator Name: _____ Signature: _____		
Approved By	Medical Superintendent Name: _____ Signature: _____		

1. Purpose

This SOP describes how «Hospital Name» counsels patients, tracks batch/serial numbers, and manages recalls for implants and medical devices.

2. Scope

Applies to every implant or medical device used in patient care.

3. Responsibility

Treating Doctor / Nurse: counsels the patient and ensures batch/serial traceability at every required location.

4. Procedure

1. Before the procedure, counsel the patient and family on precautions specific to the implant, and document this within the informed consent.
2. Confirm the implant/device carries applicable regulatory approval before use.
3. Record the batch/serial number in all three required locations for every implant: the patient's medical record, the master logbook, and the discharge summary.
4. Where no pre-labelled sticker exists, transcribe the batch/serial number manually into all three locations and have a second person verify the transcription.
5. On learning of a recall — whether from internal feedback or an external notice — identify every patient who received that batch, using the master logbook, and notify the regulatory authority and manufacturer for internally identified recalls.

5. Records

Supported by Records 3, 4, and 5 of the MOM.10 Required Records pack.

6. Related Documents

- HCO_MOM_10_v2_REWRITE_DRAFT — MOM.10 policy

7. Review

Reviewed annually, and immediately after any implant recall.

8. References

- Ministry of Health and Family Welfare, Government of India (CDSCO). (2017). Medical Devices Rules, 2017 (G.S.R. 78(E)). https://cdsco.gov.in/opencms/resources/UploadCDSCOWeb/2022/m_device/Medical%20Devices%20Rules,%202017.pdf