

ABBOTT
SPECTRUM[®]

II
SERIES

OPERATION

The Operation Manual is an instructional and reference manual for general operation of the ABBOTT SPECTRUM® SERIES II™ System. It contains the following:

- Discussion of System components and associated hardware
- Discussion of the theory of operation
- Installation and relocation specifications
- Operational procedures
- Discussion of touch screens
- Discussion of hazards and general safety issues

Do not operate the ABBOTT SPECTRUM SERIES II System before becoming thoroughly familiar with the information in this manual.

In addition to the Operation Manual, the ABBOTT SPECTRUM SERIES II Manual Set includes a Maintenance & Troubleshooting Manual and a Reagent Manual. Additional copies of the manual sets may be ordered.

If additional information or assistance is needed, Abbott Laboratories provides telephone technical assistance.

Abbott Instrument Warranty

Abbott Laboratories warrants instruments sold by the Abbott Diagnostic Division to be free from defects in workmanship and materials during normal use by the original purchaser. This warranty shall continue for a period of one year, commencing twenty-one (21) days from the date of shipment to the original purchaser, or until title is transferred from the original purchaser, whichever occurs first (the "Warranty Period").

If any defects occur during the Warranty Period, contact the Abbott Customer Support Center immediately, and be prepared to furnish pertinent details concerning the defect, the model number, and the serial number.

Warranty service is provided 8:30 a.m. through 5:00 p.m., Monday through Friday, except on Abbott-observed holidays. Any service performed at other times, and all service required to correct defects or malfunctions not covered by this Warranty, will be billed at Abbott's labor rates then in effect.

This Warranty does not cover defects or malfunctions which: (1) are not reported to Abbott during the Warranty Period and within one week of occurrence; (2) result from chemical decomposition or corrosion; (3) are described in the applicable Abbott Operation Guide; or (4) result from maintenance, repair, or modification performed without Abbott's prior written authorization.

Abbott's liability for all matters arising from the supply, installation, use, repair, and maintenance of the instrument, whether arising under this Warranty or otherwise, shall be limited solely to the repair or (at Abbott's sole discretion) replacement of the instrument or of components thereof. In no event shall Abbott be liable for injuries sustained by third parties, incidental or consequential damages, or lost profits. Replaced parts shall become the property of Abbott Laboratories.

THE FOREGOING IS THE SOLE WARRANTY MADE BY ABBOTT LABORATORIES REGARDING THE INSTRUMENT, AND ABBOTT SPECIFICALLY DISCLAIMS ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING THE WARRANTIES OF MERCHANTABILITY AND OF FITNESS FOR A PARTICULAR PURPOSE.

Revision Status

The ABBOTT SPECTRUM® SERIES II™ System is manufactured by Abbott Laboratories, Diagnostics Division, 1921 Hurd, P.O. Box 152020, Irving, Texas 75015, U.S.A. Please direct all inquiries concerning information in this manual to the foregoing address.

The Revision Status of the manual is indicated below. Be sure that the manual contains the latest revision number of all pages.

Note: Direct all inquiries regarding equipment problems to the Customer Support Center (CSC).

Document Control Number	Revision Date	Pages Revised and Added
Originally Issued — 18472-101	11/90	Not applicable
18472-102	1/91	All pages
18472-103	6/92	iii, 1-6, 1-41, 2-55, 2-61, 2-65, 2-71, and 3-112
18472-104	9/92	i through iv, and 3-64
18472-105	10/93	iii through xvi, 1-7, 1-8, 1-39 through 1-44, 2-55, 2-56, 2-65, 2-67 through 2-69, and 3-56
18472-105M	2/95	All pages (Document was converted to another publishing software. The field was not updated with converted pages.) Pages deleted: 3-23 through 3-27
18472-106	12/95	All pages

This page is blank.

Revision Log

This log provides a record of incorporation of revised pages.

- 1. Remove and replace revised pages.
- 2. Record the following information in the appropriate column:
 - The document control number*
 - The revision date*
 - The initials of person incorporating revised pages
 - The date revised pages are incorporated

Document Control Number	Revision Date	Revision Incorporated by	Date Incorporated

* Refer to the Revision Status page for this information.

This page is blank.

Review Status

This page provides a record of persons who review this manual on a periodic basis.

Signature

Date _____

[illegible]

This page is blank.