

0.2 Micron filter, part #, 3-18  
24V Distribution Board (Lower, Main) (RR-14.4/14.7), 4-223  
24V Distribution Board (Upper) (RR-15.10), 4-244  
24V Power Supply  
    adjustment (VP-1), 5-8  
    isolation procedure (IP-1), 2-27  
    part # (PL-14), 3-28  
    removal/replacement (RR-14.1), 4-217  
    verification procedure (VP-1), 5-8  
    wiring diagram, A-3  
24V Power Supply Adjust. (VP-1), 5-8  
24V Power Supply (gnd/ungnd), part #, 3-28  
24V Power Supply (IP-1), 2-27  
24V Power Supply (RR-14.1), 4-217  
24V Power Supply Wiring Diagram, A-3  
25-Pin Wire Tap Breakout Plug, part #, 3-32  
512K Dynamic RAM (DRAM) Board, jumpers, B-8

## **A**

AC Power Distribution, overview block diagram, A-7  
Active Current Heat Sink (RR-15.3/15.4/15.8/15.9), 4-236

AD Converter Board  
    jumpers, B-36  
    part # (PL-7), 3-14  
    removal/replacement (RR-7.11), 4-102  
AD Converter Bd, part #, 3-14  
AD Converter Board (RR-7.11), 4-102  
AD Offset Check (VP-22), 5-54  
AD OFFSET IS ZERO. CHANNEL NUMBER: XX (1-16)  
    (EC 15) (IP-2), 2-28  
AD Read Check (VP-23), 5-56  
adjustments  
    24V Power Supply Adjust. (VP-1), 5-8  
    Calibration Wheel Adjust. (VP-2), 5-10  
    Cuvette Carrier Height Adjust. (VP-4), 5-15  
    Fluid Sensitivity Adjust. (Reagent Arm Robotics) (VP-38), 5-107  
    Incubator Level Sense Adjust. (VP-5), 5-17  
    Incubator Servo Adjust. (VP-6), 5-19  
    Lamp Adjust. (VP-7), 5-21  
    Mixer Stroke Adjust. (Mix Arm Robotics) (VP-36), 5-87  
    Multi-Output Power Supply Adjust. (VP-11), 5-30  
    overview, 5-7  
    Sample Arm Fluid Sense Adjust.  
        Top LED (Sample Arm Robotics) (VP-37), 5-95  
        Side LED (Sample Arm Robotics) (VP-37), 5-97

## adjustments (cont)

- Sample Diluent Level Adjust. (VP-16), 5-40
- Sample Diluent Pump Adjust. (VP-17), 5-42
- Top Deck Height Adjust. (VP-20), 5-47
- Water Quality Station Adjust. (VP-21), 5-50

## alignments

- Cuvette Carrier Centering (VP-3), 5-12
- Light Beam Align. (VP-8), 5-24
- Mirror 1 Align. (VP-9), 5-26
- Mixer Drive Assembly Align. (VP-10), 5-28
- overview, 5-7
- Reagent Bar Code Reader Align. (VP-12), 5-32
- Sample Arm Align. (VP-13), 5-33
- Sample Carousel Align. (VP-14), 5-36
- Sample Carousel ID Reader Board Align. (VP-15), 5-38
- Sample Wash Cup Align. (VP-18), 5-43
- Syringe Drive Motor Align. (VP-19), 5-46

## Analyzer Decontamination (VP-39), 5-115

## analyzer introduction

- operation overview, 1-21
- Power Off Sequence (VP-47), 5-131
- Power On Sequence (VP-48), 5-134
- specifications, 1-11
- surface decontamination, 5-115
- system overview, 1-17

## analyzer specifications, 1-11

## arm

- Mix, *see* Mix Arm
- Reagent, *see* Reagent Arm
- Sample, *see* Sample Arm

## audience for this manual, 1-3

**B**

## Ball Allen™ Wrench Set, part #, 3-32

## banner messages

- see* alphabetical list in chapter 2
- see* specific message

bar code reader, *see* Reagent Bar Code Reader

## Bar Code Reader Bd Interface Bd, part #, 3-14

## Bar Code Reader Board (RR-7.6), 4-90

## Benzalkonium Chloride, part #, 3-32

## bezel (CRT)

- cleaning (VP-40), 5-117
- see also* CRT Bezel

## biosafety, 1-8

## block diagrams

- 24V Power Supply Wiring Diagram, A-3
- AC Power Distribution Overview, A-7
- Card Cages Main Block Diagram, A-9
- CRT and Touchscreen, A-23
- Incubator Optics Interface Board, A-18
- I/O Peripherals, A-24
- Lamp Drive, A-17

## block diagrams

- Master Card Cage, A-10
- Mixer Drive, A-19
- Multi-Output Power Supply Wiring Diagram, A-5
- Optics, A-16
- Reagent Arm, A-12
- Reagent Bar Code Reader Assembly Interconnect, A-21
- Reagent Cooler, A-20
- Pumps and Valves, A-25
- Sample Arm, A-13
- Sample Carousel Drive, A-15
- Slave Card Cage, A-11
- Syringes (Reagent and Sample), A-14
- Water System, A-26
- see also* wiring diagrams

## boards

- block diagrams, Appendix A
- jumpers, Appendix B
- see also* specific board name

Bubble Memory Set, jumpers, B-9

## Bus I/O Board

- jumpers, B-28
- part # (PL-7), 3-14
- removal/replacement (RR-7.2), 4-82

Bus I/O Bd (gnd/ungnd), part #, 3-14

Bus I/O Board (RR-7.2), 4-82

**C**

Calcium/AST Run (VP-51), 5-140

calibration overview, 1-26

## calibration wheel

- adjustment (VP-2), 5-10
- part # (PL-10), 3-20
- removal/replacement (RR-10.7), 4-155
- robotics errors (IP-3), 2-30

Calibration Wheel Adjust. (VP-2), 5-10

Calibration Wheel Assembly (RR-10.7), 4-155

Calibration Wheel Assy, part #, 3-20

Calibration Wheel Dual Optics Sensor

- part #, 3-20
- removal/replacement (RR-10.2), 4-145

Calibration Wheel Dual Optics Sensor (RR-10.2), 4-145

Calibration Wheel robotics errors (EC 34, 36, 185) (IP-3), 2-30

## calibrations

ISE Calibration (VP-29), 5-71

Temperature Calibration (VP-34), 5-79

calibrators, handling, 1-9

Card Cage Fan, 115 VAC, part #, 3-14

## card cage (Master)

- block diagram, A-10
- jumpers, B-2
- part # (PL-7), 3-14
- removal/replacement (RR-7.1), 4-79

- card cage (Slave)
  - block diagram, A-11
  - jumpers, B-30
  - part # (PL-7), 3-14
  - removal/replacement, 4-108
- Card Cage (Slave) (RR-7.14), 4-108
- card cages
  - main block diagram, A-9
  - see also* Master Card Cage
  - see also* Slave Card Cage
- carousel, *see* Sample Carousel
- cautions and warnings
  - conventions used in this manual, 1-5
  - decontamination, 1-8
  - disposal, 1-10
  - electrical safety, 1-7
  - general biosafety, 1-8
  - handling controls, calibrators, reagents, and samples, 1-9
  - personal protective equipment, 1-7
  - physical safety, 1-7
  - spills, 1-10
- centering
  - Cuvette Carrier Centering (VP-3), 5-12
  - see also* adjustments
  - see also* alignments
  - see also* positioning
- chapters in the manual, descriptions of, 1-4
- charts
  - Example Probe Positioning Summary Chart, 5-85
  - Procedure Order Flowchart, 5-84
  - see also* tables
- checks
  - AD Offset Check (VP-22), 5-54
  - AD Read Check (VP-23), 5-56
  - Dual Optics Sensor Check (VP-26), 5-62
  - Ground Modification Check (VP-27), 5-65
  - Static Temperature Check (VP-33), 5-77
- cleaning procedures
  - CRT Bezel/Touchscreen Array (VP-40), 5-117
  - Diluent System (VP-41), 5-119
  - Fan Screens (VP-43), 5-122
  - Incubator (VP-44), 5-124
  - Relay Lens (VP-50), 5-138
  - see also* Diluent System Purge (VP-42)
- CODE RAM 19 32781 (banner msg) (IP-4), 2-32
- cooler, *see* Reagent Cooler
- communication, host (VP-28), 5-67
- Compound 111 (Dow Corning®), part #, 3-32
- configuration
  - Line Voltage Selection Plug (VP-45), 5-126
  - Oneac® Line Conditioner (VP-46), 5-128
- Connector (Inlet Water) (female/male), part #, 3-18
- controls, handling, 1-9
- conventions used in this manual, 1-5
- covers, part # (PL-1), 3-2

CRT assembly  
  block diagram, A-23  
  part # (PL-13), 3-26  
  removal/replacement (RR-13.5), 4-211  
CRT Assy with Touchscreen, part #, 3-26  
CRT Assembly (with Touchscreen) (RR-13.5), 4-211  
CRT Bezel, Insert, part #, 3-26  
CRT Bezel/Touchscreen Array Cleaning (VP-40), 5-117  
CRT Controller Bd (gnd/ungnd), part #, 3-14  
CRT Controller Board  
  jumpers, B-26  
  part # (PL-7), 3-14  
  removal/replacement (RR-7.3), 4-84  
CRT Controller Board (RR-7.3), 4-84  
CRT display is blank (observ. prob.) (IP-5), 2-33  
CRT Junction Bd (gnd/ungnd), part #, 3-14  
CRT Junction Board (RR-7.16), 4-113  
CRT/Keyboard Bezel, part #, 3-26  
CRT/Keyboard Bezel (RR-13.1), 4-207  
Customer Tool Kit, part #, 3-32  
Cuvette Carrier Assy, part #, 3-22  
Cuvette Carrier Assembly (RR-11.1), 4-162  
Cuvette Carrier Bearing (RR-11.8), 4-177  
Cuvette Carrier Bearings, part #, 3-22  
Cuvette Carrier Belt Tensioning Tool, part #, 3-32  
Cuvette Carrier Centering (VP-3), 5-12  
Cuvette Carrier Cover, part #, 3-22  
Cuvette Carrier Dual Optics Sensor, part #, 3-22  
Cuvette Carrier Dual Optics Sensor (RR-11.5), 4-171

Cuvette Carrier Height Adjust. (VP-4), 5-15  
Cuvette Carrier Highest Bottom Determination (VP-36B), 5-90  
  (Mix Arm Robotics)  
Cuvette Carrier Motor, part #, 3-20  
Cuvette Carrier Motor (RR-10.4), 4-149  
Cuvette Carrier Positioning (Reagent Arm Robotics) (VP-38B), 5-105  
Cuvette Carrier robotics errors [EC 20, 104, 175 (Motor 6), 184]  
  (IP-6), 2-34  
Cuvette Cell Positioning (Sample Arm Robotics) (VP-37A), 5-94  
Cuvette Change Panel, part #, 3-20  
Cuvette Change Panel (RR-10.8), 4-158  
Cuvette Clamps, part #, 3-22  
Cuvette Clamp(s) (RR-11.2), 4-164

## **D**

Dark Current Test (VP-24), 5-58  
Dark Cuvette Set, part #, 3-32  
Dark Cuvette Test (VP-25), 5-60  
decontamination  
  analyzer surface (VP-39), 5-115  
  overview, 1-8  
Depth Micrometer, part #, 3-32  
Diluent System, part # (PL-8) (PL-9), 3-16, 3-18  
Diluent System Cleaning (VP-41), 5-119

Diluent System Purge (VP-42), 5-121  
disposal, 1-10  
documents related to this manual, 2-3  
Dow Corning® Compound 111, part #, 3-32  
Drain Assembly (RR-9.1), 4-124  
Drain Assy, part #, 3-18  
Drain Connector (female/male), part #, 3-18  
drainage requirements for analyzer, 1-14  
DRAM (512K Dynamic RAM) Board, jumpers, B-8  
Dual Optics Sensor Check (VP-26), 5-62  
Dual Port RAM Bd, part #, 3-14  
Dual Port RAM Board  
    jumpers, B-3  
    part # (PL-7), 3-14  
    removal/replacement (RR-7.10), 4-100  
Dual Port RAM Board (RR-7.10), 4-100  
DUAL PORT RAM TIMEOUT OR FAIL (banner msg) (IP-7), 2-36

## E

electrical requirements for analyzer, 1-12  
electrical safety, 1-7  
error codes, *see also* numerical list in chapter 2  
External Wash/Waste Tubing Kit, part #, 3-18

## F

Fan Filter, part #, 3-20  
Fan Screen, part #, 3-24  
Fan Screens Cleaning (VP-43), 5-122  
flowcharts, symbols used in, 2-6  
Fluid Sense Adjust. (Sample Arm)  
    top LED (VP-37C), 5-97  
    side LED (VP-37C), 5-99  
Fluid Sensitivity Adjust. (Reagent Arm Robotics) (VP-38E), 5-110  
Front End Bd (High Gain), part #, 3-20  
Front End Bd (Low Gain), part #, 3-20  
Front End Board (RR-10.3), 4-147  
Front Panel, part #, 3-2  
    part # (PL-1), 3-2  
    removal/replacement (RR-1.4), 4-11  
    *see also* Panel, Front

## G

general biosafety, 1-8  
Ground Modification Check (VP-27), 5-65  
Ground Strap, part #, 3-6

**H**

Heat Glass, part #, 3-22  
 Heat Glass (RR-11.10), 4-184  
 Heat Sink, Active Current (RR-15.3/15.4/15.8/15.9), 4-236  
 Heat Sink, Incubator Drive (RR-15.7), 4-242  
 Heat Sink, Lamp Drive (RR-15.6), 4-240  
 Heat Sink, Lower Resistive (RR-15.5), 4-238  
 Heat Sink, Resistive Current (RR-15.2/15.11), 4-234  
 heat sinks, part # (PL-15), 3-30  
     *see also* specific heat sink name  
 height  
     Cuvette Carrier Height Adjust. (VP-4), 5-15  
     Top Deck Height Adjust. (VP-20), 5-47  
 Highest Bottom Determination of Cuvette Carrier  
     (Mix Arm Robotics) (VP-36B), 5-90  
 Home Station Interface Board (old style)  
     part # (PL-2), 3-4  
     removal/replacement (RR-2.3), 4-25  
     *see also* Pump & Valve/Home Station Interface Board  
 Home Station Interface Bd (old style), part #, 3-4  
 Home Station Interface Bd (RR-2.3), 4-25  
 host communication  
     isolation procedures (IP-8), 2-37  
     verification procedure (VP-28), 5-67  
 Host Communication (VP-28), 5-67  
 Host communication errors (EC 18) (IP-8), 2-37

**I**

incubator  
     cleaning (VP-44), 5-124  
     level sense adjust. (VP-5), 5-17  
     servo adjust. (VP-6), 5-19  
 Incubator Cleaning (VP-44), 5-124  
 Incubator Drive Heat Sink, part #, 3-30  
 Incubator Drive Heat Sink (RR-15.7), 4-242  
 Incubator Lenses, part #, 3-22  
 Incubator Lenses (RR-11.13), 4-191  
 Incubator Level Sense Adjust. (VP-5), 5-17  
 Incubator Level Sense Probe, part #, 3-22  
 Incubator Level Sense Probe (RR-11.3), 4-166  
 Incubator Optics  
     block diagram, A-16  
     part # (PL-10) (PL-11), 3-20, 3-22  
 Incubator Optics Assembly (RR-10.1), 4-141  
 Incubator Optics Assy, part #, 3-20  
 Incubator Optics Fan, part #, 3-20  
 Incubator Optics Fan (RR-10.9), 4-160  
 Incubator Optics Interface Bd, part #, 3-22  
 Incubator Optics Interface Board  
     block diagram, A-18  
     part # (PL-11), 3-22  
     removal/replacement (RR-11.7), 4-175  
 Incubator Optics Interface Board (RR-11.7), 4-175

- Incubator Servo Adjust. (VP-6), 5-19  
Incubator Servo Board  
    jumpers, B-22  
    removal/replacement (RR-7.4), 4-86  
Incubator Servo Board (RR-7.4), 4-86  
Incubator Servo Calibration/Adjustment Tool, 30K Plug, part #, 3-32  
Incubator Thermistor, part #, 3-22  
Incubator Thermistor (RR-11.6., 4-173  
INCUBATOR WILL NOT FILL (EC 191) (IP-9), 2-38  
INCUBATOR WILL NOT STABILIZE (EC 231) (IP-10), 2-39  
Initialization (Reinitialization), SRAM Board (VP-53), 5-144  
Inlet Water Connector (female/male), part #, 3-18  
Insert, CRT Bezel, part #, 3-26  
Internal Wash/Waste Tubing Kit, part #, 3-18  
Interrupt Cable, part #, 3-14  
Interrupt Cable (RR-7.17), 4-116  
I/O peripherals, block diagram, A-24  
IP-1: 24V Power Supply, 2-27  
IP-2: AD OFFSET IS ZERO. CHANNEL NUMBER: XX (1-16)  
    (EC 15), 2-28  
IP-3: Calibration Wheel robotics errors (EC 34, 36, 185), 2-30  
IP-4: CODE RAM 19 32781 (banner msg), 2-32  
IP-5: CRT display is blank (observ. prob.), 2-33  
IP-6: Cuvette Carrier robotics errors [EC 20, 104, 175 (Motor 6), 184],  
    2-34  
IP-7: DUAL PORT RAM TIMEOUT OR FAIL (banner msg), 2-36  
IP-8: Host communication errors (EC 18), 2-37  
IP-9: INCUBATOR WILL NOT FILL (EC 191), 2-38  
IP-10: INCUBATOR WILL NOT STABILIZE (EC 231), 2-39  
IP-11: LAMP DID NOT TURN ON (EC 218), 2-41  
IP-12: Lamp Drive problems (EC 177), 2-42  
IP-12: Lamp Drive problems (Lockup on AD Offset), 2-42  
IP-12: Lamp Drive problems (Photometer Check Fail), 2-42  
IP-13: MIX ARM NOT MIXING (observ. prob.), 2-43  
IP-14: Mix Arm robotics errors [EC 78, 79, 80, 81, 82, 84, 85,  
    175 (Motor 10), 175 (Motor 11)], 2-44  
IP-15: Multi-Output Power Supply, 2-45  
IP-16: NMI ATTEMPT TO WRITE PROTECTED RAM (banner msg),  
    2-47  
IP-17: NMI ATTEMPT TO WRITE ROM (banner msg), 2-48  
IP-18: NMI POWER FAIL (banner msg), 2-49  
IP-19: NO MASTER/SLAVE CLOCK INTERRUPT (banner msg),  
    2-50  
IP-20: PACKET LOAD ADDRESS ERROR (banner msg), 2-51  
IP-21: POWER ON CHECK FAIL (banner msg), 2-52  
IP-22: Pumps & Valves errors [EC 175 (Motor 18), 175 (Motor 19)],  
    2-53  
IP-23: Reagent Arm horizontal robotics errors [EC 50, 52, 66, 67, 68,  
    75, 76, 175 (Motor 1), 175 (Motor 2), 226, 227], 2-54  
IP-24: Reagent Arm vertical robotics errors [EC 32, 105, 175  
    (Motor 0)], 2-55  
IP-25: REAGENT ARM WILL NOT DETECT FLUID (EC 8, 21, 39, 40,  
    179, 324, 325), 2-57  
IP-26: REAGENT BAR CODE READER NOT FUNCTIONING  
    (EC 176), 2-58  
IP-27: Reagent Cooler not cooling (observ. prob.), 2-60  
IP-28: REAGENT REFRIGERATOR TEMPERATURE NOT STABLE  
    (EC 112), 2-62  
IP-29: Reagent Syringe errors [EC 28, 30, 53, 175 (Motor 7)], 2-63  
IP-30: Results-related problems, 2-64



IP-31: Sample Arm horizontal robotics errors [EC 57, 58, 59, 61, 63, 64, 175 (Motor 4)], 2-66

IP-32: Sample Arm vertical robotics errors [EC 60, 65, 175 (Motor 3)], 2-68

IP-33: SAMPLE ARM WILL NOT DETECT FLUID (EC 21, 39, 178, 224, 236), 2-70

IP-34: Sample Carousel robotics errors [EC 19, 100, 175 (Motor 5), 182, 183], 2-71

IP-35: SAMPLE DILUENT LEVEL LOW (EC 190), 2-73

IP-36: SAMPLE DILUENT PRESSURE HIGH (EC 187), 2-74

IP-37: SAMPLE DILUENT PRESSURE LOW (EC 186), 2-75

IP-38: Sample ID Reader not functioning (observ. prob.), 2-77

IP-39: Sample Syringe robotics errors [EC 24, 53, 175 (Motor 8)], 2-78

IP-40: WAITING ON TEMP (banner msg), 2-79

IP-41: Wash Cups overflowing (EC 39), 2-80

IP-42: WASH WATER PRESSURE HIGH/LOW (EC 188, 189), 2-82

IP-43: Water Quality Station Light not functioning (observ. prob.), 2-83

ISE

- calibration (VP-29), 5-71
- part # (PL-2), 3-4
- probe positioning (VP-35), 5-87
- shroud removal/replacement (RR-2.2), 4-23
- (for additional information, refer to ISE Service Manual)

ISE Assy (for Spectrum Series II), part #, 3-4

ISE Calibration (VP-29), 5-71

ISE Probe Positioning (VP-35), 5-87

ISE Shroud, part #, 3-4

ISE Shroud (RR-2.2), 4-23

ISE Starter Kit, part #, 3-32

Isolated AC Distribution Bd, part #, 3-28

Isolated AC Distribution Board (RR-14.8), 4-230

isolation procedures (IPs)

- alphabetical list of IPs, 2-9
- conventions used in flowcharts, 2-6
- error code list (numerical order), 2-11
- generic troubleshooting flowchart, 2-7
- symbols used in flowcharts, 2-6

Isolation Transformer, part #, 3-28

Isolation Transformer (RR-14.3), 4-221

## J

### jumpers

- AD Converter Board, B-36
- Bubble Memory Set (3.3 only), B-9
- Bus I/O Board, B-28
- CRT Controller Board, B-26
- DRAM (512K Dynamic RAM) Board (3.3, 3.7 only), B-8
- Dual Port RAM Board, B-3
- Incubator Servo Board, B-22
- Lamp Servo Board, B-20
- Master Card Cage, B-2

jumpers (cont)

- Master CPU Board, B-5

Motor Controller Board, B-31  
Power Monitor Board, B-24  
Pump & Valve/Home Station Interface Board (new style), B-38  
Quad UART Board, B-16  
Reagent Bar Code Reader Board, B-18  
Real Time Processor Board (Slave CPU), B-34  
ROM Set, B-13  
Slave Card Cage, B-30  
SRAM Board (3.7, 4.8), B-10

**K**

keyboard assembly  
    part # (PL-13), 3-26  
    removal/replacement (RR-13.3), 4-209  
Keyboard Assembly (RR-13.3), 4-209  
Keyboard Assy, part #, 3-26  
Keyboard Interface Bd, part #, 3-26  
kinetic noise test, 5-74  
Krytox® (DuPont Krytox® 240AC Fluorinated Grease), part #, 3-32

**L**

lamp  
    adjustment (VP-7), 5-21  
    removal/replacement (RR-11.12), 4-188  
    *see also* Source Lamp  
Lamp Adjust. (VP-7), 5-21  
LAMP DID NOT TURN ON (EC 218) (IP-11), 2-41  
Lamp Drive  
    block diagram, A-17  
    problems, *see* Lamp Drive problems  
Lamp Drive Heat Sink, part #, 3-30  
Lamp Drive Heat Sink (RR-15.6), 4-240  
Lamp Drive problems  
    EC 177 (IP-12), 2-42  
    Lockup on AD Offset (IP-12), 2-42  
    Photometer Check Fail (IP-12), 2-42  
Lamp Housing Assembly (RR-11.9), 4-181  
Lamp Housing Assy, part #, 3-22  
Lamp Servo Bd (gnd/ungnd), part #, 3-14  
Lamp Servo Board  
    jumpers, B-20  
    part # (PL-7), 3-14  
    removal/replacement (RR-7.5), 4-88  
Lamp Servo Board (RR-7.5), 4-88  
Lamp Socket Assembly (RR-11.11), 4-186  
Lamp Socket Assy, part #, 3-22  
Left Side Door, part #, 3-2

## lens

- Incubator Lenses (RR-11.13), 4-191
- Relay Lens Cleaning (VP-50), 5-138

## level

- Incubator Level Sense Adjust. (VP-5), 5-17
- Sample Diluent Level Adjust. (VP-16), 5-40

## Light Beam Align. (VP-8), 5-24

## Line Conditioner (Oneac®), configuration (VP-46), 5-128

## Line Voltage Selection Bd, part #, 3-28

## Line Voltage Selection Board (RR-14.5), 4-226

## Line Voltage Selection Plug Configuration (VP-45), 5-126

## Line Voltage Selection Plug, part #, 3-28

## Lithium 12-Hydroxy Grease, part #, 3-32

## LO-17 Oil, part #, 3-32

## Lockup on AD Offset (Lamp Drive problem) (IP-12), 2-42

## Loopback Test (VP-30), 5-72

## Low Kinetic Noise Test (VP-31), 5-74

## Lower 24V Distribution Bd, part #, 3-28

## Lower Center Back Panel, part #, 3-2

## Lower Left Back Panel, part #, 3-2

## Lower Resistive Heat Sink, part #, 3-30

## Lower Resistive Heat Sink (RR-15.5), 4-238

## Lubrication, Syringe (VP-54), 5-145

**M**

## Main 24V Distribution Bd, part #, 3-28

## Main AC Distribution Bd, part #, 3-28

## Main AC Distribution Board (RR-14.6), 4-228

## Main Power Switch

- see Power Off Sequence

- see Power On Sequence

## Maintenance Power Switch

- see Power Off Sequence

- see Power On Sequence

## maintenance

- Pre-Warming Chamber (VP-49), 5-136

- Water Quality Station (VP-55), 5-149

## manual

- conventions used in manual, 1-5

- descriptions of chapters, 1-4

- intended audience, 1-3

- overview, 1-3

- related documents, 1-3

## Master Card Cage

- block diagram, A-10

- jumpers, B-2

- part # (PL-7), 3-14

- removal/replacement (RR-7.1), 4-79

## Master Card Cage Fan (RR-7.15), 4-111

Master Card Cage (gnd/ungnd), part #, 3-14  
Master Card Cage (RR-7.1), 4-79  
Master CPU Bd, part #, 3-14  
Master CPU Board  
    jumpers, B-5  
    part # (PL-7), 3-14  
    removal/replacement (RR-7.9), 4-98  
Master CPU Board (RR-7.9), 4-98  
Mirror 1 Align. (VP-9), 5-26  
Mix Arm Assembly (RR-10.6), 4-153  
Mix Arm Assy, part #, 3-20  
MIX ARM NOT MIXING (observ. prob.) (IP-13), 2-43  
Mix Arm robotics errors ([EC 78, 79, 80, 81, 82, 84, 85, 175  
    (Motor 10), 175 (Motor 11)] (IP-14), 2-44  
Mix Arm Robotics Training (VP-36)  
    Highest Bottom Determination of Cuvette Carrier (VP-36B), 5-90  
    Mixer Stroke Adjust. (VP-36A), 5-89  
    overview, 5-88  
    Wash Cup Positioning (VP-36C), 5-92  
Mix Wash Cup, part #, 3-22  
Mix Wash Cup (RR-11.4), 4-168  
Mixer Drive Assembly  
    alignment (VP-10), 5-28  
    block diagram, A-19  
    part # (PL-10), 3-20  
    removal/replacement (RR-10.5), 4-151  
Mixer Drive Assy, part #, 3-20  
Mixer Drive Assembly (RR-10.5), 4-151  
Mixer Drive Assembly Align. (VP-10), 5-28  
Mixer Stroke Adjust. (Mix Arm Robotics) (VP-36A), 5-89

modes of operation, overview, 1-24  
motor, Syringe Drive Motor Align. (VP-19), 5-46  
Motor Controller Bd, part #, 3-14  
Motor Controller Board  
    jumpers, B-31  
    part # (PL-7), 3-14  
    removal/replacement (RR-7.13), 4-106  
Motor Controller (MC) Board (RR-7.13), 4-106  
Multi-Output Power Supply  
    adjustment (VP-11), 5-30  
    isolation procedure (IP-15), 2-45  
    part # (PL-14), 3-28  
    removal/replacement (RR-14.2), 4-219  
    wiring diagram, A-5  
Multi-Output Power Supply Adjust. (VP-11), 5-30  
Multi-Output Power Supply (gnd/ungnd), part #, 3-28  
Multi-Output Power Supply (IP-15), 2-45  
Multi-Output Power Supply (RR-14.2), 4-219

## **N**

NMI ATTEMPT TO WRITE PROTECTED RAM (banner msg)  
    (IP-16), 2-47  
NMI ATTEMPT TO WRITE ROM (banner msg) (IP-17), 2-48  
NMI POWER FAIL (banner msg) (IP-18), 2-49  
NO MASTER/SLAVE CLOCK INTERRUPT (banner msg)  
    (IP-19), 2-50  
noise, Low Kinetic Noise Test (VP-31), 5-74

## normal operations

Calcium/AST Run (VP-51), 5-140

overview, 1-21

Reproducibility Run (VP-52), 5-142

## overview (cont)

robotics training, 5-84

Spectrum system, 1-17

tests, 5-53

**O**observed problems, *see* alphabetical list in chapter 2

offset, AD Offset Check (VP-22), 5-54

Oneac® Line Conditioner (110 V), part #, 3-32

Oneac® Line Conditioner (220 V), part #, 3-32

Oneac® Line Conditioner Configuration (VP-46), 5-128

operation of analyzer, overview, 1-21

## optics

block diagram, A-16

drift test (VP-32), 5-75

*see also* Incubator Optics

Optics Drift Test (VP-32), 5-75

## overview

analyzer operation, 1-21

calibration, 1-26, 5-53

checks, 5-53

how to use this manual, 1-3

quality control, 1-40

process, 1-21

**P**

PACKET LOAD ADDRESS ERROR (banner msg) (IP-20), 2-51

Panel, Front, part #, 3-2

Panel, Front (RR-1.4), 4-11

Panel, Lower Center Back (RR-1.9), 4-13

Panel, Lower Left Back (RR-1.13), 4-21

Panel, Right Back (RR-1.12), 4-19

Panel, Upper Center Back (RR-1.11), 4-17

Panel, Upper Left Back (RR-1.10), 4-15

part numbers, *see* specific part by name

personal protective equipment, 1-7

Photometer Check Fail (Lamp Drive problem) (IP-12), 2-42

physical requirements for analyzer, 1-11

Pipe Check Solution, part #, 3-32

PL-1: Covers, 3-2

PL-2: ISE, 3-4

PL-3: Sample Carousel Drive Assembly, 3-6

PL-4: Sample Arm, 3-8

PL-5: Reagent Arm, 3-10

PL-6: Sample and Reagent Syringes, 3-12  
PL-7: Card Cages (Master and Slave), 3-14  
PL-8: Water and Diluent Systems (Part 1), 3-16  
PL-9: Water and Diluent Systems (Part 2), 3-18  
PL-10: Incubator Optics (Part 1), 3-20  
PL-11: Incubator Optics (Part 2), 3-22  
PL-12: Reagent Cooler, 3-24  
PL-13: CRT and Keyboard Assemblies, 3-26  
PL-14: System Power, 3-28  
PL-15: Heat Sinks, 3-30  
PL-16: Supplemental Tools and Supplies, 3-32  
PM Kit, part #, 3-32  
positioning  
    Cuvette Carrier (VP-38B), 5-105  
    Cuvette Carrier Centering (VP-3), 5-12  
    Cuvette Cell (VP-37A), 5-94  
    Reagent Tray (VP-38C), 5-107  
    Reagent Tray Perimeter (VP-38D), 5-109  
    Sample Carousel (VP-37B), 5-95  
    Wash Cup (Mix Arm) (VP-36C), 5-92  
    Wash Cup (Reagent Arm) (VP-38A), 5-103  
    Wash Cup (Sample Arm) (VP-37D), 5-101  
    *see also* adjustments  
    *see also* alignments  
    *see also* centering  
    *see also* Mix Arm Robotics Training  
    *see also* Reagent Arm Robotics Training  
    *see also* Sample Arm Robotics Training  
power distribution (AC), block diagram, A-7

Power Monitor Board, jumpers, B-24  
Power Off Sequence (VP-47), 5-131  
POWER ON CHECK FAIL (banner msg) (IP-21), 2-52  
Power On Sequence (VP-48), 5-134  
power supply  
    24V, *see* 24V Power Supply  
    Multi-Output, *see* Multi-Output Power Supply  
power (system), part # (PL-14), 3-28  
Pre-Warming Chamber Maintenance (VP-49), 5-136  
Pressure Bypass Plug, part #, 3-32  
preventive maintenance, 6-2  
Printer [Okidata® 184 (refurb)], part #, 3-32  
Printer Ribbon, part #, 3-32  
printouts, examples, *see* specific procedure  
process overview, 1-21  
probe positioning  
    chart (summary, example), 5-85  
    ISE probe (VP-35), 5-87  
pump  
    Sample Diluent Pump Adjust. (VP-17), 5-42  
    *see also* Sample Diluent Pump  
Pump & Valve Board (old style)  
    part # (PL-2), 3-4  
    removal/replacement (RR-2.4), 4-27  
    *see also* Pump & Valve/Home Station Interface Board  
Pumps & Valves  
    block diagram, A-25  
    isolation procedure (IP-22), 2-53  
Pumps & Valves errors [EC 175 (Motor 18), 175 (Motor 19)]  
    (IP-22), 2-53

Pump & Valve/Home Station Interface Board (new style)  
jumpers, B-38  
part # (PL-2), 3-4  
removal/replacement (RR-2.5), 4-29  
Pump & Valve/Home Station Interface Board (new style) (RR-2.5),  
4-29  
Purge, Diluent System (VP-42), 5-121

## **Q**

Quad Rings  
Reagent Syringe (RR-6.3), 4-56  
Sample Syringe (RR-6.5), 4-60  
Quad UART Board, jumpers, B-16  
quality control overview, 1-40  
Quick Disconnect (female/male), part #, 3-16

## **R**

read, AD Read Check (VP-23), 5-56  
reader  
Reagent Bar Code Reader  
Sample Carousel ID Reader

Reagent Arm  
block diagram, A-12  
part # (PL-5), 3-10  
removal/replacement (RR-5.1), 4-47  
robotics errors  
horizontal (IP-23), 2-54  
vertical (IP-24), 2-55  
Reagent Arm (RR-5.1), 4-47  
Reagent Arm Active Heat Sink, part #, 3-30  
Reagent Arm Fluid Sense Bd  
part # (PL-5), 3-10  
removal/replacement (RR-5.4), 4-49  
Reagent Arm Fluid Sense Board (RR-5.4), 4-49  
Reagent Arm horizontal robotics errors [EC 50, 52, 66, 67, 68, 75, 76,  
175 (Motor 1), 175 (Motor 2), 226, 227] (IP-23), 2-54  
Reagent Arm Left Cover, part #, 3-10  
Reagent Arm Middle Bd, part #, 3-10  
Reagent Arm Resistive Heat Sink, part #, 3-30  
Reagent Arm Right Cover, part #, 3-10  
Reagent Arm Robotics Training (VP-38)  
Cuvette Carrier Positioning (VP-38B), 5-105  
Fluid Sensitivity Adjustment (VP-38E), 5-110  
overview, 5-102  
Reagent Tray Perimeter Positioning (VP-38D), 5-109  
Reagent Tray Positioning (VP-38C), 5-107  
Wash Cup Positioning (VP-38A), 5-103

Reagent Arm Upper Bd, part #, 3-10  
Reagent Arm vertical robotics errors [EC 32, 105, 175 (Motor 0)]  
    (IP-24), 2-55  
REAGENT ARM WILL NOT DETECT FLUID (EC 8, 21, 39, 40, 179,  
    324, 325) (IP-25), 2-57  
Reagent Bar Code Dual Optics Sensor, part #, 3-24  
Reagent Bar Code Dual Optics Sensor (RR-12.1B), 4-198  
Reagent Bar Code Reader  
    alignment (VP-12), 5-32  
    isolation procedures (IP-25), 2-57  
Reagent Bar Code Reader Align. (VP-12), 5-32  
Reagent Bar Code Reader Assembly Interconnect, block diagram,  
    A-21  
Reagent Bar Code Reader Board, jumpers, B-18  
REAGENT BAR CODE READER NOT FUNCTIONING (EC 176)  
    (IP-26), 2-58  
Reagent Cooler  
    block diagram, A-20  
    fan (RR-12.4), 4-204  
    not cooling (isolation procedure) (IP-27), 2-60  
    part # (PL-12), 3-24  
Reagent Cooler Assembly (RR-12.1A), 4-195  
Reagent Cooler Assy  
    CCx™ Series II™, part #, 3-24  
    Spectrum® Series II™, part #, 3-24  
Reagent Cooler Controller, part #, 3-24  
Reagent Cooler Controller (RR-12.3), 4-202  
Reagent Cooler Cover, part #, 3-24  
Reagent Cooler Fan, 115 VAC, part #, 3-24  
Reagent Cooler Fan (RR-12.4), 4-204

Reagent Cooler not cooling (observ. prob.) (IP-27), 2-60  
Reagent Cooler Thermistor, part #, 3-24  
Reagent Cooler Thermistor (RR-12.2), 4-200  
Reagent Probe, part #, 3-10  
Reagent Probe (RR-5.6), 4-51  
Reagent Probe Adapter, part #, 3-10  
Reagent Probe Retainer, part #, 3-10  
REAGENT REFRIGERATOR TEMPERATURE NOT STABLE  
    (EC 112) (IP-28), 2-62  
Reagent Syringe  
    block diagram, A-14  
    isolation procedures (IP-29), 2-63  
    part # (PL-6), 3-12  
    removal/replacement (RR-6.4), 4-58  
Reagent Syringe (RR-6.4), 4-58  
Reagent Syringe Active Heat Sink, part #, 3-30  
Reagent Syringe Drive Assembly (RR-6.1), 4-53  
Reagent Syringe Drive Assy, part #, 3-12  
Reagent Syringe errors [EC 28, 30, 53, 175 (Motor 7)] (IP-29), 2-63  
Reagent Syringe Limit Bd, part #, 3-12  
Reagent Syringe Limit Board (RR-6.11), 4-75  
Reagent Syringe Quad Rings, part #, 3-12  
Reagent Syringe Quad Rings (RR-6.3), 4-56  
Reagent Syringe Tubing, part #, 3-12  
Reagent Syringe Tubing (RR-6.9A), 4-68  
Reagent Tray Perimeter Positioning (Reagent Arm) (VP-38D), 5-109  
Reagent Tray Positioning (Reagent Arm Robotics) (VP-38C), 5-107  
Reagent Wash Cup, part #, 3-22  
Reagent Wash Cup (RR-11.4), 4-168  
Reagents, handling, 1-9



- Real Time Processor Board (Slave CPU)
  - jumper, B-34
  - part # (PL-7), 3-14
  - removal/replacement (RR-7.12), 4-104
- Real Time Processor Board (Slave CPU) (RR-7.12), 4-104
- Rear Power Panel (120 VAC), part #, 3-30
- Rear Power Panel (220 VAC), part #, 3-30
- Rear Power Panel (RR-15.1), 4-232
- reinitialization (initialization), SRAM board (VP-53), 5-144
- Relay Lens Cleaning (VP-50), 5-138
- removal/replacement procedures
  - alphabetical list, 4-5
  - see also specific procedure by name
  - see also specific procedure by RR number
- Reproducibility Run (VP-52), 5-142
- Resistive Current Heat Sink (RR-15.2/15.11), 4-234
- Results-related problems (IP-30), 2-64
- Right Back Panel, part #, 3-2
- Right Side Door, part #, 3-2
- Robotics training
  - Example Probe Positioning Summary Chart, 5-85
  - Example Step Tables, 5-86
  - Mix Arm (VP-36), 5-88
  - procedure order, 5-84
  - overview, 5-84
  - Procedure Order Flowchart, 5-84
  - Reagent Arm (VP-28), 5-102
  - Sample Arm (VP-37), 5-93
- ROM Decode Board, jumpers, B-13
- ROM Memory Board, jumpers, B-13
- ROM set (ROM Decode Bd and ROM Memory Bd)
  - jumper, B-13
  - part # (PL-7), 3-14
  - removal/replacement (RR-7.7), 4-93
- ROM Set (RR-7.7), 4-93
- RR-1.3: Top Deck, 4-9
- RR-1.4: Panel, Front, 4-11
- RR-1.9: Panel, Lower Center Back, 4-13
- RR-1.10: Panel, Upper Left Back, 4-15
- RR-1.11: Panel, Upper Center Back, 4-17
- RR-1.12: Panel, Right Back, 4-19
- RR-1.13: Panel, Lower Left Back, 4-21
- RR-2.2: ISE Shroud, 4-23
- RR-2.3: Home Station Interface Board (old style), 4-25
- RR-2.4: Pump & Valve Board (old style), 4-27
- RR-2.5: Pump & Valve/Home Station Interface Board (new style), 4-29
- RR-3.1: Sample Carousel Drive Assembly, 4-32
- RR-3.2: Sample Carousel Dual Optics Sensor, 4-34
- RR-3.5: Sample Carousel ID Reader Board, 4-36
- RR-4.1: Sample Arm, 4-39
- RR-4.4: Sample Probe, 4-41
- RR-4.7A: Sample Arm Fluid Sense Board (new style), 4-43
- RR-4.7B: Sample Arm Fluid Sense Board (old style), 4-45
- RR-5.1: Reagent Arm, 4-47
- RR-5.4: Reagent Arm Fluid Sense Board, 4-49
- RR-5.6: Reagent Probe, 4-51
- RR-6.1/6.2: Syringe Drive Assembly (6.1 Reagent, 6.2 Sample), 4-53

- RR-6.3: Reagent Syringe Quad Rings, 4-56  
RR-6.4: Reagent Syringe, 4-58  
RR-6.5: Sample Syringe Quad Rings, 4-60  
RR-6.6: Sample Syringe, 4-62  
RR-6.7: Sample Diluent Valve, 4-64  
RR-6.8: Sample Diluent 35-Micron Filter, 4-66  
RR-6.9A: Reagent Syringe Tubing, 4-68  
RR-6.9B: Sample Syringe Tubing, 4-70  
RR-6.10/6.11: Syringe Limit Board (6.11 Reagent, 6.10 Sample), 4-75  
RR-7.1: Master Card Cage, 4-79  
RR-7.2: Bus I/O Board, 4-82  
RR-7.3: CRT Controller Board, 4-84  
RR-7.4: Incubator Servo Board, 4-86  
RR-7.5: Lamp Servo Board, 4-88  
RR-7.6: Bar Code Reader Board, 4-90  
RR-7.7: ROM Set, 4-93  
RR-7.8: SRAM Board, 4-96  
RR-7.9: Master CPU Board, 4-98  
RR-7.10: Dual Port RAM Board, 4-100  
RR-7.11: AD Converter Board, 4-102  
RR-7.12: Real Time Processor Board (Slave CPU), 4-104  
RR-7.13: Motor Controller (MC) Board, 4-106  
RR-7.14: Slave Card Cage, 4-108  
RR-7.15: Master Card Cage Fan, 4-111  
RR-7.16: CRT Junction Board, 4-113  
RR-7.17: Interrupt Cable, 4-116  
RR-8.3: Sample Diluent 70-Micron Filter, 4-118  
RR-8.5: Sample Diluent Tray Level Switch, 4-120  
RR-8.6: Sample Diluent Tray, 4-122  
RR-9.1: Drain Assembly, 4-124  
RR-9.2: Waste Pump, 4-127  
RR-9.4: Sample Diluent Pump, 4-130  
RR-9.5: Wash Manifold Assembly, 4-132  
RR-9.6: Sample Diluent Manifold Assembly, 4-135  
RR-9.7: Water Quality Station, 4-137  
RR-10.1: Incubator Optics Assembly, 4-141  
RR-10.2: Calibration Wheel Dual Optics Sensor, 4-145  
RR-10.3: Front End Board, 4-147  
RR-10.4: Cuvette Carrier Motor, 4-149  
RR-10.5: Mixer Drive Assembly, 4-151  
RR-10.6: Mix Arm Assembly, 4-153  
RR-10.7: Calibration Wheel Assembly, 4-155  
RR-10.8: Cuvette Change Panel, 4-158  
RR-10.9: Incubator Optics Fan, 4-160  
RR-11.1: Cuvette Carrier Assembly, 4-162  
RR-11.2: Cuvette Clamp(s), 4-164  
RR-11.3: Incubator Level Sense Probe, 4-166  
RR-11.4: Wash Cups (Mix, Reagent, Sample), 4-168  
RR-11.5: Cuvette Carrier Dual Optics Sensor, 4-171  
RR-11.6: Incubator Thermistor, 4-173  
RR-11.7: Incubator Optics Interface Board, 4-175  
RR-11.8: Cuvette Carrier Bearing, 4-177  
RR-11.9: Lamp Housing Assembly, 4-181  
RR-11.10: Heat Glass, 4-184  
RR-11.11: Lamp Socket Assembly, 4-186  
RR-11.12: Source Lamp, 4-188  
RR-11.13: Incubator Lenses, 4-191  
RR-11.14: Shutter Assembly, 4-193  
RR-12.1A: Reagent Cooler Assembly, 4-195  
RR-12.1B: Reagent Bar Code Dual Optics Sensor, 4-198

RR-12.2: Reagent Cooler Thermistor, 4-200  
RR-12.3: Reagent Cooler Controller, 4-202  
RR-12.4: Reagent Cooler Fan, 4-204  
RR-13.1: CRT/Keyboard Bezel, 4-207  
RR-13.3: Keyboard Assembly, 4-209  
RR-13.5: CRT Assembly (with Touchscreen), 4-211  
RR-13.6: Touchscreen Assembly, 4-213  
RR-13.7: Side Power Panel Assembly, 4-215  
RR-14.1: 24V Power Supply, 4-217  
RR-14.2: Multi-Output Power Supply, 4-219  
RR-14.3: Isolation Transformer, 4-221  
RR-14.4/14.7: 24V Distribution Board (Lower, Main), 4-223  
RR-14.5: Line Voltage Selection Board, 4-226  
RR-14.6: Main AC Distribution Board, 4-228  
RR-14.8: Isolated AC Distribution Board, 4-230  
RR-15.1: Rear Power Panel, 4-232  
RR-15.2/15.11: Heat Sink, Resistive Current, 4-234  
RR-15.3/15.4/15.8/15.9: Heat Sink, Active Current, 4-236  
RR-15.5: Heat Sink, Lower Resistive, 4-238  
RR-15.6: Heat Sink, Lamp Drive, 4-240  
RR-15.7: Heat Sink, Incubator Drive, 4-242  
RR-15.10: 24V Distribution Board, Upper, 4-244  
Run: Calcium/AST (VP-51), 5-140  
Run: Reproducibility (VP-52), 5-142  
Run: Sample, overview, 1-37

## S

Sample Arm  
    block diagram, A-13  
    part # (PL-4), 3-8  
    removal/replacement (RR-4.1), 4-39  
    robotics errors  
        horizontal (IP-31), 2-66  
        vertical (IP-32), 2-68  
Sample Arm (RR-4.1), 4-39  
Sample Arm Active Heat Sink, part #, 3-30  
Sample Arm Align. (VP-13), 5-33  
Sample Arm Fluid Sense Adjust. (Sample Arm Robotics)  
    top LED (VP-37C), 5-97  
    side LED (VP-37C), 5-99  
Sample Arm Fluid Sense Bd, part #, 3-8  
Sample Arm Fluid Sense Board (new style) (RR-4.7A), 4-43  
Sample Arm Fluid Sense Board (old style) (RR-4.7B), 4-45  
Sample Arm horizontal robotics errors [EC 57, 58, 59, 61, 63, 64, 175 (Motor 4)] (IP-31), 2-66  
Sample Arm Left Cover, part #, 3-8  
Sample Arm Middle Bd, part #, 3-8  
Sample Arm Resistive Heat Sink, part #, 3-30  
Sample Arm Right Cover, part #, 3-8  
Sample Arm Robotics Training (VP-37), 5-94  
Sample Arm Upper Bd, part #, 3-8

Sample Arm vertical robotics errors [EC 60, 65, 175 (Motor 3)]  
    (IP-32), 2-68  
SAMPLE ARM WILL NOT DETECT FLUID (EC 21, 39, 178, 224, 236)  
    (IP-33), 2-70  
Sample Carousel Align. (VP-14), 5-36  
Sample Carousel Cover, part #, 3-6  
Sample Carousel Drive Assembly  
    block diagram, A-15  
    part # (PL-3), 3-6  
    removal/replacement (RR-3.1), 4-32  
Sample Carousel Drive Assembly (RR-3.1), 4-32  
Sample Carousel Dual Optics Sensor, part #, 3-6  
Sample Carousel Dual Optics Sensor (RR-3.2), 4-34  
Sample Carousel ID Kit, part #, 3-6  
Sample Carousel ID Reader Board  
    alignment (VP-15), 5-38  
    part # (PL-3), 3-6  
    removal/replacement (RR-3.5), 4-36  
Sample Carousel ID Reader Board (RR-3.5), 4-36  
Sample Carousel ID Reader Board Align. (VP-15), 5-38  
Sample Carousel Motor, part #, 3-6  
Sample Carousel Positioning (Sample Arm Robot.) (VP-37B), 5-95  
Sample Carousel robotics errors [EC 19, 100, 175 (Motor 5), 182, 183]  
    (IP-34), 2-71  
Sample Carousel Tray, part #, 3-6  
Sample Diluent 35-Micron Filter, part #, 3-12  
Sample Diluent 35-Micron Filter (RR-6.8), 4-66  
Sample Diluent 70-Micron Filter, part #, 3-16  
Sample Diluent 70-Micron Filter (RR-8.3), 4-118  
Sample Diluent Bottle, part #, 3-16

Sample Diluent Level Adjust. (VP-16), 5-40  
SAMPLE DILUENT LEVEL LOW (EC 190) (IP-35), 2-73  
Sample Diluent Manifold, part #, 3-18  
Sample Diluent Manifold Assembly (RR-9.6), 4-135  
SAMPLE DILUENT PRESSURE HIGH (EC 187) (IP-36), 2-74  
SAMPLE DILUENT PRESSURE LOW (EC 186) (IP-37), 2-75  
Sample Diluent Pump  
    adjustment (VP-17), 5-42  
    part # (PL-8), 3-18  
    removal/replacement (RR-9.4), 4-130  
Sample Diluent Pump (RR-9.4), 4-130  
Sample Diluent Pump Adjust. (VP-17), 5-42  
Sample Diluent Reservoir Access Door, part #, 3-2  
Sample Diluent Tray, part #, 3-16  
Sample Diluent Tray (RR-8.6), 4-122  
Sample Diluent Tray Level Switch, part #, 3-16  
Sample Diluent Tray Level Switch (RR-8.5), 4-120  
Sample Diluent Valve, part #, 3-12  
Sample Diluent Valve (RR-6.7), 4-64  
Sample ID Reader not functioning (observ. prob.) (IP-38), 2-77  
Sample Probe (RR-4.4), 4-41  
sample run, overview, 1-37  
Sample Syringe  
    block diagram, A-14  
    part # (PL-6), 3-12  
    removal/replacement (RR-6.6), 4-62  
    robotics errors (IP-38), 2-78  
Sample Syringe (RR-6.6), 4-62  
Sample Syringe Active Heat Sink, part #, 3-30  
Sample Syringe Drive Assembly (RR-6.2), 4-53

- Sample Syringe Drive Assy, part #, 3-12
- Sample Syringe Limit Bd, part #, 3-12
- Sample Syringe Limit Board (RR-6.10), 4-75
- Sample Syringe Quad Rings, part #, 3-12
- Sample Syringe Quad Rings (RR-6.5), 4-60
- Sample Syringe robotics errors [EC 24, 53, 175 (Motor 8)]  
(IP-39), 2-78
- Sample Syringe Set Screw, part #, 3-12
- Sample Syringe Tubing, part #, 3-12
- Sample Syringe Tubing (RR-6.9B), 4-70
- Sample Wash Cup
  - alignment (VP-18), 5-43
  - removal/replacement (RR-11.4), 4-168
- Sample Wash Cup, part #, 3-22
- Sample Wash Cup (RR-11.4), 4-168
- Sample Wash Cup Align. (VP-18), 5-44
- samples, handling, 1-9
- screen displays, examples, *see specific procedure*
- screens, fan, cleaning (VP-43), 5-122
- sensors
  - Calibration Wheel Dual Optics Sensor (RR-10.2), 4-145
  - Dual Optics Sensor Check (VP-26), 5-62
- service call
  - beginning the call, 2-5
  - preventive maintenance, 2-5
  - total service call, 6-2
- servo, incubator, *see Incubator Servo*
- Shutter Assembly, part #, 3-22
- Shutter Assembly (RR-11.14), 4-193
- Side Power Panel Assembly (RR-13.7), 4-215
- Side Power Panel, part #, 3-26
- Slave Card Cage
  - block diagram, A-11
  - jumper, B-30
  - part # (PL-7), 3-14
  - removal/replacement (RR-7.14), 4-108
- Slave Card Cage (RR-7.14), 4-108
- Slave CPU, *see Real Time Processor Board*
- Source Lamp, part #, 3-22
- Source Lamp (RR-11.12), 4-188
- space requirements for analyzer, 1-11
- specifications
  - analyzer, 1-15
  - drainage requirements, 1-14
  - electrical requirements, 1-12
  - physical requirements, 1-11
  - water quality requirements, 1-13
- Spectrum analyzer
  - normal operations overview, 1-21
  - specifications, 1-11
  - system overview, 1-17
  - total service call, 6-2
- spills, 1-10
- SRAM Battery, part #, 3-14

## SRAM board

initialization (VP-53), 5-144

jumpers, B-10

part # (PL-7), 3-14

removal/replacement (RR-7.8), 4-96

SRAM Board (RR-7.8), 4-96

SRAM Board Initialization (VP-53), 5-144

Static Mat, part #, 3-32

Static Temperature Check (VP-33), 5-77

step tables, example, 5-86

supplies, part # (PL-16), 3-32

supplemental tools, part # (PL-16), 3-32

Support Stand Key, part #, 3-32

Syringe Bracket Thumb Screw, part #, 3-12

Syringe Drive Assembly (6.1 Reagent, 6.2 Sample) (RR-6.1/6.2), 4-53

Syringe Drive Motor, part #, 3-12

Syringe Drive Motor Align. (VP-19), 5-46

Syringe Limit Board (Reagent, Sample) (RR-6.10/6.11), 4-75

Syringe Lubrication (VP-54), 5-145

Syringe Motor Alignment Tool, part #, 3-32

## syringes

block diagram, A-14

lubrication, 5-145

part # (PL-6), 3-12

## system power

part # (PL-14), 3-28

specifications, 1-12

Syringe Tubing (Reagent), part #, 3-12

Syringe Tubing (Sample), part #, 3-12

**T**

## table

step tables (probe positioning example), 5-86

*see also* charts

Teflon® Tubing Installation Kit, part #, 3-32

Temperature Calibration (VP-34), 5-79

temperature check (static), 5-77

## tests

Dark Current Test (VP-24), 5-58

Dark Cuvette Test (VP-25), 5-60

Loopback Test (VP-30), 5-72

Low Kinetic Noise Test (VP-31), 5-74

Optics Drift Test (VP-32), 5-75

overview, 5-53

tools (supplemental), part # (PL-16), 3-32

## Top Deck

adjustment (VP-20), 5-47

part # (PL-1), 3-2

removal/replacement (RR-1.3), 4-9

Top Deck (RR-1.3), 4-9

Top Deck Height Adjust. (VP-20), 5-47

total service call, 6-2

touchscreen, block diagram, A-23

touchscreen array, cleaning (VP-40), 5-117

Touchscreen Assembly (RR-13.6), 4-213

Touchscreen Assy, part #, 3-26

transistors (used in heat sinks), part #, 3-30

## troubleshooting

## error codes

see specific code

see *also* cross-reference list (alpha order) in chapter 2

## isolation procedures

see chapter 2 in this manual

for ISE-related procedures, refer to ISE Service Manual

## overview, 2-5

Tubing (Reagent Syringe), part #, 3-12

Tubing (Sample Syringe), part #, 3-12

U

UART, *see* Quad UART

Upper 24V Distribution Bd, part #, 3-30

Upper Center Back Panel, part #, 3-2

Upper Center Door, part #, 3-2

Upper Left Back Panel, part #, 3-2

Upper Left Door, part #, 3-2

Upper Right Door, part #, 3-2

V

valves, *see also* Pumps & Valves

## verification procedures

alphabetical list, 5-3

calibrations, overview, 5-53

checks, overview, 5-53

Example Probe Positioning Summary Chart, 5-85

Example Step Tables, 5-86

Procedure Order Flowchart, 5-84

robotics training overview, 5-84

tests, overview, 5-53

*see also* specific procedure by name

*see also* specific procedure by VP number

## voltage

configuring the Line Voltage Selection Plug (VP-45), 5-126

specifications, 1-12

VP-1: 24V Power Supply Adjustment, 5-8

VP-2: Calibration Wheel Adjustment, 5-10

VP-3: Cuvette Carrier Centering, 5-12

VP-4: Cuvette Carrier Height Adjustment, 5-15

VP-5: Incubator Level Sense Adjustment, 5-17

VP-6: Incubator Servo Adjustment, 5-19

VP-7: Lamp Adjustment, 5-21

VP-8: Light Beam Alignment, 5-24

VP-9: Mirror 1 Alignment, 5-26

- VP-10: Mixer Drive Assembly Alignment, 5-28
- VP-11: Multi-output Power Supply Adjustment, 5-30
- VP-12: Reagent Bar Code Reader Alignment, 5-32
- VP-13: Sample Arm Alignment, 5-33
- VP-14: Sample Carousel Alignment, 5-36
- VP-15: Sample Carousel ID Reader Board Alignment, 5-38
- VP-16: Sample Diluent Level Adjustment, 5-40
- VP-17: Sample Diluent Pump Adjustment, 5-42
- VP-18: Sample Wash Cup Alignment, 5-43
- VP-19: Syringe Drive Motor Alignment, 5-46
- VP-20: Top Deck Height Adjustment, 5-47
- VP-21: Water Quality Station Adjustment, 5-50
- VP-22: AD Offset Check, 5-54
- VP-23: AD Read Check, 5-56
- VP-24: Dark Current Test, 5-58
- VP-25: Dark Cuvette Test, 5-60
- VP-26: Dual Optics Sensor Check, 5-62
- VP-27: Ground Modification Check, 5-65
- VP-28: Host Communication, 5-67
- VP-29: ISE Calibration, 5-71
- VP-30: Loopback Test, 5-72
- VP-31: Low Kinetic Noise Test, 5-74
- VP-32: Optics Drift Test, 5-75
- VP-33: Static Temperature Check, 5-77
- VP-34: Temperature Calibration, 5-79
- VP-35: ISE Probe Positioning, 5-87
- VP-36: Mix Arm Robotics Training, 5-88
  - A. Mixer Stroke Adjustment, 5-89
  - B. Highest Bottom Determination of Cuvette Carrier, 5-90
  - C. Wash Cup Positioning, 5-92
- VP-37: Sample Arm Robotics Training
  - A. Cuvette Cell Positioning, 5-94
  - B. Sample Carousel Positioning, 5-95
  - C-1. Sample Arm Fluid Sense Adjustment (top LED), 5-97
  - C-2. Sample Arm Fluid Sense Adjustment (side LED), 5-99
  - D. Wash Cup Positioning, 5-101
- VP-38: Reagent Arm Robotics Training, 5-102
  - A. Wash Cup Positioning, 5-103
  - B. Cuvette Carrier Positioning, 5-105
  - C. Reagent Tray Positioning, 5-107
  - D. Reagent Tray Perimeter Positioning, 5-109
  - E. Fluid Sensitivity Adjustment, 5-110
- VP-39: Analyzer Decontamination, 5-115
- VP-40: CRT Bezel/Touchscreen Array Cleaning, 5-117
- VP-41: Diluent System Cleaning, 5-119
- VP-42: Diluent System Purge, 5-121
- VP-43: Fan Screens Cleaning, 5-122
- VP-44: Incubator Cleaning, 5-124
- VP-45: Line Voltage Selection Plug Configuration, 5-126
- VP-46: Oneac® Line Conditioner Configuration, 5-128
- VP-47: Power Off Sequence, 5-131
- VP-48: Power On Sequence, 5-134
- VP-49: Pre-Warming Chamber Maintenance, 5-136
- VP-50: Relay Lens Cleaning, 5-138
- VP-51: Run: Calcium/AST, 5-140
- VP-52: Run: Reproducibility, 5-142
- VP-53: SRAM Board Initialization, 5-144
- VP-54: Syringe Lubrication, 5-145
- VP-55: Water Quality Station Maintenance, 5-149



**W, X, Y, Z**

WAITING ON TEMP (banner msg) (IP-40), 2-79

Wash Cup Cover (Reagent Cup, Mix Cup, Sample Cup), part #, 3-22

Wash Cup Positioning

Mix Arm Robotics (VP-36C), 5-92

Reagent Arm Robotics (VP-38A), 5-103

Sample Arm Robotics (VP-37D), 5-101

wash cups

alignment, Sample Wash Cup (VP-18), 5-43

isolation procedures (IP-41), 2-80

positioning, *see* Wash Cup Positioning

removal/replacement (RR-11.4), 4-168

Wash Cups (Mix, Reagent, Sample) (RR-11.4), 4-168

Wash Cups overflowing (EC 39) (IP-41), 2-80

Wash Manifold, part #, 3-18

Wash Manifold Assembly (RR-9.5), 4-132

WASH WATER PRESSURE HIGH/LOW (EC 188, 189) (IP-42), 2-82

Waste Manifold, part #, 3-18

Waste Pump, part #, 3-18

Waste Pump (RR-9.2), 4-127

water quality requirements for analyzer, 1-13

Water Quality Station

adjustment (VP-21), 5-50

maintenance (VP-55), 5-149

part # (PL-9), 3-18

removal/replacement (RR-9.7), 4-137

Water Quality Station (RR-9.7), 4-137

Water Quality Station Adjust. (VP-21), 5-50

Water Quality Station Light

amber, part #, 3-18

amber & green, part #, 3-18

Water Quality Station Light not functioning (observ. prob.)

(IP-43), 2-83

Water Quality Station Maintenance (VP-55), 5-149

Water System

block diagram, A-26

part # (PL-8) (PL-9), 3-16, 3-18

wiring diagrams

24V Power Supply, A-3

AC Power Distribution, A-7

Multi-Output Power Supply, A-5

