



## VP-1: POWER SUPPLY CHECK (pg 1 of 1)

Time: 5 minutes

Purpose: To verify correct Main Power Supply outputs

1. Remove Enclosure ([RR-1.3](#)).
2. Measure voltage at test points on top of power supply.

<u>Specifications</u>			
	+24	+21.6 to 26.4	VDC
	+15	+14.7 to 15.3	VDC
	+12	+11.4 to 12.6	VDC
	N/A		
	+5	+4.75 to 5.25	VDC
	-12	-11.4 to -12.6	VDC
	-15	-14.7 to -15.3	VDC
	REF		

VP1.ds4