
10 s of contiguous frames
Face and skin region
Mean R, G, B
Window norm.
Projection to pulse signal
Spectrum 0.7 to 3 Hz
Peak and SNR
Control: the same clips with time-shuffled frames (pooled, at least 10 shuffles per clip)
Pass, or report: pulse not recoverable

schematic