

QUOTATION NO. KEBS/QUOTE/034/2017/2018 SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF OIL COMPRESSER SYSTEM (REQUOTE)

Kenya Bureau of Standards Head office P. O. Box 54974 – 00200

Tel: 020 6948333/518 Fax: +254 20 609660

NAIROBI

AUGUST 2017

- 1. You are invited to submit a quotation on the item listed.
- 2. This quotation should be deposited in the quotation box provided in the main reception on or before **10.00AM**. on **Tuesday 31st October 2017**.
- 3. Your quotation should include all government taxes

1. QUOTATION PRICE

Kenya shillings (in words)
In figures

2. VALIDITY PERIOD

➤ The bid shall remain valid for a period of 120 (One Hundred and twenty) days from the date of submission

3. TECHNICAL SPECIFICATIONS

Specifications for oil free compressor and accessories for the Acoustics and Vibration laboratory.

Technical Evaluation Criteria

Item	Weight
	(%)
Oil Free Compression	20
Minimum air output at 5 bar (70 psi) 1.9 litres/sec, 6.9 m ³ /h,	10
4 cfm	10
Minimum 24 litre capacity tank, with anti-oxidant treatment	20
Condensate absorption dryer with filters	20
Anti-vibration damping and low noise operation (65 dBA or	10
less)	10
Power supplier: 220/240 V, 50 Hz	10
Pressure regulator with gauge	5
Safety valve/thermal cut-out	5
Total	100

Accessories

- Pressure stainless steel tube 6 mm internal diameter with a total length of 50 meters from the compressor to the point of application of pressure.
- 2 inch diameter PVC ducting pipe including bends from the compressor position to the room of application
- Power supply to the compressor and including ducting to source
- A secured cage next to the building for compressor as per the diagrams KEBS 87 & 88 attached

4. MANDATORY REQUIREMENTS

All bidders are required to attach copies of the following mandatory documents:

- 1. Valid Business Registration Certificate
- 2. Valid KRA Tax Compliance Certificate
- 3. Site visit to be undertaken before quoting- All bidders are required to attend Mandatory site visit on **Monday 30th** October 2017.



