

Bay 2 Reagent Spill Report

1. Identifiers

Site: Chemistry Analyzer Bay 2

Incident Date: 2026-07-03

Severity: Minor

2. Description

A reagent spill on the Bay 2 analyzer halted processing over a one hour window. See Figure 2 for the tray layout.

[FIGURE: reagent_seal_cross_section.png]

3. Timeline

Reagent-level warning at 08:12; spill detected 08:20; resumed after decontamination at 11:20.

[FIGURE: figureA.png]

4. Contributing Factors

Control-software 7.2.0 was running on the affected analyzer. A secondary catch tray was absent.

5. Corrective Actions

Due Date	Action
----------	--------

6. Approvals

Reviewer:

Signoff Date: