Chapter 7

Raw Spectral Quality Control and Instrument Diagnostics

Lucas VHH TRAN[1]

Abstract

Technical artifacts and experimental confounding in mass spectrometry invalidate downstream statistical modeling if left unaddressed [12, 13]. Samples exhibiting spray collapse, progressive retention time shifts, mass calibration drift, or carryover from high-concentration standards must be flagged during acquisition [4, 9]. Experimental designs where biological cohorts are collinear with acquisition batches collapse model degrees of freedom, rendering treatment effects inseparable from technical variance [14]. In R 4.6.1 and Bioconductor 3.20, quality control auditing operates at two distinct levels: low-level spectral trace extraction using `Spectra` *[15, 16] and standardized metric calculation via the* `MsQuality` *framework [1] conforming to HUPO-PSI* `mzQC` *controlled vocabularies [2, 3]. This chapter details the physical failure modes of liquid chromatography--mass spectrometry (LC-MS) instruments, mathematical formulations for mass accuracy drift and autosampler carryover ratios [11, 10], degrees-of-freedom auditing in design matrices, and empirical coefficient of variation (CV) distributions in pooled quality control samples that set the lower bound for biological effect detection [5, 6].*

Keywords: Quality control, mzQC, Total Ion Chromatogram, Mass accuracy drift, Carryover, Confounding, Pooled QC, Coefficient of variation.

Chapter Overview and Architectural Roadmap

Table 7.1 outlines the diagnostic checks required to audit mass spectrometry data before preprocessing and statistical analysis.

Methodological Note

Tutor Warning: The QC Gate Principle

Quality control is an irreversible analytical gate, not an exploratory afterthought. Applying complex batch correction or imputation to datasets with severe spray collapse or confounded study designs

[1]CK-CARE Davos, Christine Kühne--Center for Allergy Research and Education, Davos, Switzerland. Email: lucas.tran@ck-care.ch

Table 7.1: Diagnostic roadmap for mass spectrometry raw data quality control.

Diagnostic Tier	Physical Target	Analytical Metric	Rejection Threshold
1. Design Matrix	Batch-group confounding [12]	Cross-tabulation rank, Pearson r	$\text{rank}(X) < p$, $r = 1.0$
2. Chromatography	Column stability and gradient	TIC/BPC overlay, apex RT drift [4]	$\Delta t_R > 30$ seconds
3. Mass Accuracy	Analyzer calibration drift [18]	Parts-per-million error vs lock mass	$\lvert\text{ppm}\rvert > 5.0$ ppm [17]
4. Carryover	Autosampler wash efficiency [9]	Solvent blank area post-ULOQ	$> 20\%$ of LLOQ area [11]
5. Technical Noise	Pooled QC repeatability [5]	Feature intensity CV across batch	Feature CV $> 30\%$ [8]

generates clean plots of artificial signal. If an instrument run fails basic chromatographic or mass accuracy tolerances, no statistical test can recover the lost biological measurement. Flag and exclude corrupted acquisitions before feature detection.

7.1 Experimental Design Auditing and Batch Confounding

The most severe quality failure in mass spectrometry is experimental confounding [12, 13]. Confounding occurs when a biological condition is completely collinear with a technical batch factor [14].

7.1.1 Mathematical Formulation of Design Rank Deficiency

Let $\mathbf{X} \in \mathbb{R}^{N \times p}$ denote the experimental design matrix for N samples across p predictor variables:

$$\mathbf{y} = \mathbf{X}\beta + \varepsilon, \quad \varepsilon \sim \mathcal{N}(\mathbf{0}, \sigma^2 \mathbf{I}) \tag{7.1}$$

In a confounded study where all Control samples are analyzed in Batch 1 and all Disease samples are analyzed in Batch 2, the indicator column for Disease ($\mathbf{x}_{\text{Disease}}$) is identical to the indicator column for Batch 2 ($\mathbf{x}_{\text{Batch2}}$):

$$\mathbf{x}_{\text{Disease}} \equiv \mathbf{x}_{\text{Batch2}} \implies \text{Corr}(\mathbf{x}_{\text{Disease}}, \mathbf{x}_{\text{Batch2}}) = 1.000 \tag{7.2}$$

The matrix $\mathbf{X}^T\mathbf{X}$ is rank-deficient:

$$\text{rank}(\mathbf{X}) < p \implies \det(\mathbf{X}^T\mathbf{X}) = 0 \tag{7.3}$$

The ordinary least squares normal equations $(\mathbf{X}^T\mathbf{X})\hat{\beta} = \mathbf{X}^T\mathbf{y}$ possess no unique solution. The biological treatment effect and technical batch difference are algebraically inseparable. A valid study design balances conditions across batches and randomizes injection sequence:

$$\sum_{i \in \text{Batch}_k} \mathbb{I}(\text{Group}_i = \text{Control}) = \sum_{i \in \text{Batch}_k} \mathbb{I}(\text{Group}_i = \text{Disease}), \quad \forall k \tag{7.4}$$

7.2 Chromatographic Stability: TIC, BPC, and Retention Drift

The total ion current (TIC) represents the summed intensity across all measured m/z channels at retention time t_i [4, 7]:

$$\mathrm{TIC}(t_i) = \sum_{j=1}^{P_i} I_{ij} \tag{7.5}$$

where P_i is the number of peaks detected in scan i. The base peak chromatogram (BPC) records the maximum observed peak intensity at time t_i:

$$\mathrm{BPC}(t_i) = \max_{j \in \{1,\ldots,P_i\}} I_{ij} \tag{7.6}$$

7.2.1 Physical Failure Modes Detected by Chromatograms

- **Electrospray Instability (Spray Collapse):** Characterized by high-frequency signal drops exceeding 10-fold between adjacent scans (`msSignal10xChange.fall`). Triggered by nanobore emitter clogs, air bubbles, or electrical arcing.
- **Retention Time Shift:** Caused by column stationary phase degradation, mobile phase composition drift, or column temperature fluctuations. Retention time drift exceeding 30 seconds between replicate QC injections requires chromatographic alignment before feature grouping.
- **Global Signal Suppression:** A uniform 2-fold drop in Area Under the TIC (AUC) between consecutive runs indicates autosampler injection failure, sample under-loading, or severe ion suppression.

7.3 Mass Accuracy Drift and Lock Mass Monitoring

High-resolution mass analyzers (Orbitrap, Time-of-Flight) maintain mass accuracies within narrow tolerances [18, 17]. Mass measurement error is expressed in parts per million (ppm):

$$\delta_{\mathrm{ppm}} = \frac{m_{\mathrm{obs}} - m_{\mathrm{theor}}}{m_{\mathrm{theor}}} \times 10^6 \tag{7.7}$$

Laboratory temperature fluctuations cause mechanical expansion of Orbitrap electrodes or flight tube drift in TOF analyzers, shifting calibration by 0.5 to 1.5 ppm per hour. Internal lock masses (e.g., ambient polydimethylcyclosiloxane ions at $m/z = 445.120025$) or regular QC calibrator injections enable real-time detection and recalibration. Acquisitions where $|\delta_{\mathrm{ppm}}| > 5.0$ ppm fail standard quality criteria.

7.4 Solvent Blank Contamination and Carryover Dynamics

Autosampler needles, rotor valves, and column trap cartridges retain hydrophobic or basic analytes. When an injection of a high-abundance biological sample or Upper Limit of Quantitation (ULOQ) standard is followed by a solvent blank, retained analyte elutes into the blank.

7.4.1 Carryover Assessment Formulations

The carryover percentage is defined relative to the preceding calibrator:

$$\mathrm{Carryover}_{\mathrm{ULOQ}}\% = \frac{\mathrm{Area}_{\mathrm{blank,\ post}}}{\mathrm{Area}_{\mathrm{ULOQ}}} \times 100\% \tag{7.8}$$

Regulatory bioanalytical guidelines (FDA, EMA) enforce a strict criterion evaluated against the Lower Limit of Quantitation (LLOQ) [9, 10, 11]:

$$\text{Carryover}_{\text{LLOQ}}\% = \frac{\text{Area}_{\text{blank, post}}}{\text{Area}_{\text{LLOQ}}} \times 100\% \leq 20.0\% \tag{7.9}$$

If carryover in the solvent blank exceeds 20% of the LLOQ peak area, subsequent low-concentration clinical samples are contaminated, producing false-positive detections.

7.5 Pooled QC Precision and Technical Noise Limits

A pooled quality control (QC) sample combines equal aliquots of every biological sample in the cohort [5, 6]. Injected periodically throughout the analytical sequence (typically every 5--10 injections), pooled QCs provide an empirical measure of technical variation across the entire pipeline [4, 8].

For each feature j, its technical coefficient of variation (CV_j) across K pooled QC injections is:

$$\text{CV}_j = \frac{\sigma_j}{\mu_j} \times 100\% = \frac{\sqrt{\frac{1}{K-1}\sum_{k=1}^{K}(y_{jk} - \bar{y}_j)^2}}{\frac{1}{K}\sum_{k=1}^{K} y_{jk}} \times 100\% \tag{7.10}$$

Theorem 7.1 (Minimum Resolvable Biological Effect). *Let a feature have technical measurement error CV in pooled QC samples. In a two-group biological comparison with equal sample sizes, the minimum statistically detectable fold-change ($FC_{\min}$) distinguishing biological variation from instrument noise with 95% confidence is bounded by:*

$$FC_{\min} \geq 1 + 2 \times \frac{CV}{100} \tag{7.11}$$

Proof. Under the null hypothesis of no biological difference, the distribution of observed fold changes is driven exclusively by technical noise, forming a 95% confidence interval $[\bar{y} - 2\sigma, \bar{y} + 2\sigma]$. A measured biological fold change cannot be distinguished from noise unless its magnitude exceeds this two-sigma technical boundary. □

Features with CV $> 30\%$ in pooled QC samples must be filtered prior to differential abundance testing.

7.6 Laboratory Workflow and Executable Implementations

The laboratory exercises demonstrate experimental design auditing, raw TIC/BPC extraction from Orbitrap Fusion data, mass accuracy drift regression, solvent carryover assessment, and pooled QC precision modeling in R 4.6.1.

7.6.1 Listing 7.1: Experimental Design Confounding Audit

R Lab Listing 1: Listing 7.1: Detecting Collinearity in Experimental Design Matrices [R 4.4 / Bioconductor]

```
library(Spectra)
library(msdata)
library(tibble)
library(dplyr)
library(tidyr)

# Confounded design: Control run in Batch 1; Disease run in Batch 2
confounded_meta <- tibble(
```

```
  sample_id = paste0("Sample_", sprintf("%02d", 1:12)),
  condition = rep(c("Control", "Disease"), each = 6),
  batch     = c(rep("Batch_1", 6), rep("Batch_2", 6))
)

tbl_conf <- table(confounded_meta$batch, confounded_meta$condition)
print(tbl_conf)

# Check for rank deficiency
is_confounded <- any(rowSums(tbl_conf > 0) == 1)
cat(sprintf("Confounding Flag: %s\n", if (is_confounded) "CRITICAL_FAIL" else "PASS"))
```

```
Control Disease Batch_1 6 0 Batch_2 0 6 Confounding Flag: CRITICAL_FAIL
```

7.6.2 Listing 7.2: Raw File Chromatographic Profiling via Spectra

R Lab Listing 2: Listing 7.2: Auditing Survey Scan TIC and Base Peak Intensities [R 4.4 / Bioconductor]

```
f_raw <- system.file("proteomics", "MS3TMT11.mzML", package = "msdata")
sps <- Spectra(f_raw, backend = MsBackendMzR())

sps_ms1 <- filterMsLevel(sps, 1L)
rts  <- rtime(sps_ms1)
tics <- tic(sps_ms1)
bpis <- sps_ms1$basePeakIntensity

cat(sprintf("Survey Scans: %d | RT Range: [%.2f, %.2f] s\n",
            length(sps_ms1), min(rts), max(rts)))
cat(sprintf("Area Under TIC (AUC): %.2e counts\n",
            sum(tics * c(diff(rts), median(diff(rts))))))
cat(sprintf("TIC CV across survey scans: %.2f%%\n", sd(tics) / mean(tics) * 100))
cat(sprintf("Median Base Peak Intensity: %.2e counts\n", median(bpis)))
```

```
Survey Scans: 30 | RT Range: [2727.07, 2824.49] s Area Under TIC (AUC): 2.84e+12 counts TIC CV across survey scans:
32.96% Median Base Peak Intensity: 4.74e+09 counts
```

7.6.3 Listing 7.3: Mass Accuracy Drift Regression

R Lab Listing 3: Listing 7.3: Longitudinal Calibration Error Tracking [R 4.4 / Bioconductor]

```
theor_mz <- 508.1234
inj_seq  <- 1:10

# Simulated progressive drift across 10 sequential injections
obs_mz <- c(508.1227, 508.1233, 508.1233, 508.1238, 508.1240,
            508.1246, 508.1248, 508.1252, 508.1255, 508.1259)

mass_df <- tibble(
  Injection = inj_seq,
  Observed  = obs_mz,
```

```
  Error_PPM = round((obs_mz - theor_mz) / theor_mz * 1e6, 2)
)

fit_drift <- lm(Error_PPM ~ Injection, data = mass_df)
cat(sprintf("Mass Drift Rate: %.3f ppm per injection (p = %.2e)\n",
            coef(fit_drift)[2], summary(fit_drift)$coefficients[2, 4]))
cat(sprintf("Pass Tolerance (|ppm| <= 5.0): %s\n", all(abs(mass_df$Error_PPM) <= 5.0)))
```

```
Mass Drift Rate: 0.679 ppm per injection (p = 1.41e-09) Pass Tolerance (|ppm| <= 5.0): TRUE
```

7.6.4 Listing 7.4: Carryover Auditing Post-High Standard

R Lab Listing 4: Listing 7.4: Autosampler Residual Contamination Audit [R 4.4 / Bioconductor]

```
carryover_df <- tibble(
  Analyte         = c("Tryptophan", "Kynurenine", "Phenylalanine", "Tyrosine"),
  High_Std_Area   = c(4500000, 2800000, 6200000, 3100000),
  Post_Blank_Area = c(4200, 18500, 3100, 89000),
  Carryover_Pct   = round(c(4200, 18500, 3100, 89000) / c(4500000, 2800000, 6200000, 3100000) * 100, 3),
  Pass_Threshold  = Carryover_Pct <= 1.0
)

print(as.data.frame(carryover_df))
```

```
Analyte High_Std_Area Post_Blank_Area Carryover_Pct Pass_Threshold 1 Tryptophan 4500000 4200 0.093 TRUE 2 Kynurenine
2800000 18500 0.661 TRUE 3 Phenylalanine 6200000 3100 0.050 TRUE 4 Tyrosine 3100000 89000 2.871 FALSE
```

7.7 Five Deadly Sins of Mass Spectrometry Quality Control

1. **Confounding Batches with Cohorts:** Processing disease samples in one batch and controls in another makes technical variation indistinguishable from disease biology, rendering the study invalid.

2. **Skipping Blank Carryover Checks:** Failing to run solvent blanks after high calibrators contaminates subsequent low-abundance clinical samples, producing false-positive hits.

3. **Ignoring Retention Time Drift:** Overlooking a 40-second drift in column retention causes feature grouping algorithms to misclassify true peaks as distinct metabolites.

4. **Trusting Low-Abundance Peaks without QC CV Auditing:** Claiming a 1.3-fold biological difference on a feature that exhibits 35% CV in pooled QC injections reports pure instrument noise as biology.

5. **Post-Hoc Data Manipulation:** Applying mathematical normalization to salvage datasets with electrospray collapse or column failure produces clean plots of meaningless data.

7.8 Pre-Flight Diagnostic Checklist

Table 7.2 provides an audit checklist to execute on all incoming files prior to preprocessing.

Table 7.2: Diagnostic checklist for raw mass spectrometry quality control.

Category	Diagnostic Check	Acceptance Criteria	Status
Design	Biological condition balanced across batches.	Zero cross-tabulation cells with zero frequency.	[ ]
Spray Stability	Adjacent-scan TIC drop frequency tracked.	Less than 5 drops $>$ 10-fold per acquisition run.	[ ]
RT Drift	EIC apex time tracked across sequence.	Peak shift $\leq$ 15 seconds between consecutive QCs.	[ ]
Mass Error	PPM calibration verified against standard.	Error $\lvert\delta_{\text{ppm}}\rvert \leq 5.0$ ppm across all runs.	[ ]
Carryover	Blank peak area post-ULOQ injection audited.	Peak area $\leq 20.0\%$ of LLOQ standard peak area.	[ ]
Precision	Pooled QC feature reproducibility binned.	$\geq 70\%$ of features exhibit CV $\leq 20\%$.	[ ]

7.9 Chapter Exercises

Exercise 7.1 (Confounded Experimental Design Degrees of Freedom). A clinical study analyzes plasma samples from 20 participants (10 healthy controls, 10 rheumatoid arthritis patients) across two instrument batches. Batch 1 contains all 10 healthy controls; Batch 2 contains all 10 patients.

1. Construct the experimental design matrix $\mathbf{X}$ for the model $\mathbf{y} = \beta_0 + \beta_1 \mathbf{x}_{\text{Disease}} + \beta_2 \mathbf{x}_{\text{Batch2}} + \varepsilon$.
2. Calculate the rank of the design matrix and prove that β_1 and β_2 cannot be estimated independently.

Exercise 7.2 (Mass Accuracy Drift Analysis). A bovine serum albumin digest peptide has theoretical monoisotopic mass $m/z = 1,296.6853$. In a 20-run sequence without lock-mass correction, injection 1 yields observed $m/z = 1,296.6885$, while injection 20 yields observed $m/z = 1,296.6742$.

1. Calculate the mass error (in ppm) for injections 1 and 20.
2. Compute the total sequence drift and determine whether injection 20 satisfies the standard ± 5.0 ppm high-resolution tolerance.

Exercise 7.3 (Autosampler Carryover Evaluation). A clinical targeted assay establishes an Upper Limit of Quantitation (ULOQ) with peak area $8,500,000$ counts and a Lower Limit of Quantitation (LLOQ) with peak area $25,000$ counts. A solvent blank analyzed immediately following the ULOQ calibrator yields an analyte peak area of $34,000$ counts.

1. Compute the carryover percentage relative to the ULOQ calibrator.
2. Compute the carryover percentage relative to the LLOQ calibrator and evaluate against FDA bioanalytical criteria ($\leq 20\%$ of LLOQ).

Exercise 7.4 (Pooled QC Feature CV and Biological Limits). In an untargeted metabolomics experiment, a candidate metabolite exhibits a mean peak area of $450,000$ counts with a standard deviation of $99,000$ counts across 6 pooled QC injections.

1. Calculate the coefficient of variation (CV%) for this metabolite.
2. Using the 2σ technical noise boundary, compute the minimum biological fold-change required to distinguish a true concentration change from instrument noise.
3. A biological comparison between patient groups yields an apparent fold-change of $1.25\times$ for this metabolite. State whether this finding is statistically reliable.

7.10 Worked Solutions to Chapter Exercises

7.10.1 Solution to Exercise 7.1: Confounded Experimental Design Degrees of Freedom

Part 1: Design Matrix Construction The sample vectors are:

$$\mathbf{x}_{\text{Disease}} = [0,0,0,0,0,0,0,0,0,0,\ 1,1,1,1,1,1,1,1,1,1]^T \tag{7.12}$$

$$\mathbf{x}_{\text{Batch2}} = [0,0,0,0,0,0,0,0,0,0,\ 1,1,1,1,1,1,1,1,1,1]^T \tag{7.13}$$

Part 2: Rank Deficiency and Inseparability Because $\mathbf{x}_{\text{Disease}} \equiv \mathbf{x}_{\text{Batch2}}$, their Pearson correlation is $r = 1.000$. The design matrix $\mathbf{X} = [\mathbf{1}, \mathbf{x}_{\text{Disease}}, \mathbf{x}_{\text{Batch2}}]$ has 3 columns but rank 2:

$$\text{rank}(\mathbf{X}) = 2 < 3 \tag{7.14}$$

The matrix $\mathbf{X}^T\mathbf{X}$ is singular and non-invertible. The estimated model can only identify the combined sum $\gamma = \beta_1 + \beta_2$. The disease effect is completely confounded with the technical batch effect.

7.10.2 Solution to Exercise 7.2: Mass Accuracy Drift Analysis

Part 1: Mass Error in PPM

- Injection 1:

$$\delta_{\text{ppm, 1}} = \frac{1296.6885 - 1296.6853}{1296.6853} \times 10^6 = +2.47 \text{ ppm} \tag{7.15}$$

- Injection 20:

$$\delta_{\text{ppm, 20}} = \frac{1296.6742 - 1296.6853}{1296.6853} \times 10^6 = -8.56 \text{ ppm} \tag{7.16}$$

Part 2: Sequence Drift and Tolerance Verdict The total drift across 20 runs is:

$$\Delta\delta_{\text{ppm}} = -8.56 - (+2.47) = -11.03 \text{ ppm} \tag{7.17}$$

Injection 1 meets the ± 5.0 ppm specification ($|+2.47| \leq 5.0$). Injection 20 violates the specification ($|-8.56| > 5.0$). Peak alignment without recalibration will drop peptide identifications.

7.10.3 Solution to Exercise 7.3: Autosampler Carryover Evaluation

Part 1: Carryover Relative to ULOQ

$$\text{Carryover}_{\text{ULOQ}}\% = \frac{34,000}{8,500,000} \times 100\% = 0.40\% \tag{7.18}$$

Part 2: Carryover Relative to LLOQ and Regulatory Verdict

$$\text{Carryover}_{\text{LLOQ}}\% = \frac{34,000}{25,000} \times 100\% = 136.0\% \tag{7.19}$$

FDA bioanalytical validation criteria mandate that solvent blank carryover must not exceed 20.0% of the LLOQ peak area. A carryover ratio of 136.0% fails regulatory acceptance, indicating that residual carryover alone generates signal greater than a true low-concentration sample.

7.10.4 Solution to Exercise 7.4: Pooled QC Feature CV and Biological Limits

Part 1: Feature CV Calculation

$$\mathrm{CV} = \frac{99,000}{450,000} \times 100\% = 22.0\% \tag{7.20}$$

Part 2: Minimum Resolvable Biological Fold-Change Using the 2σ technical error limit:

$$\mathrm{FC}_{\mathrm{min,up}} = 1 + 2 \times 0.22 = 1.44\times \quad (+44.0\%) \tag{7.21}$$

$$\mathrm{FC}_{\mathrm{min,down}} = \frac{1}{1+2\times 0.22} \approx 0.69\times \quad (-31.0\%) \tag{7.22}$$

Part 3: Evaluation of Apparent 1.25x Biological Fold-Change An apparent fold-change of $1.25\times$ ($+25\%$) falls inside the technical noise interval $[0.69\times, 1.44\times]$. The observed difference cannot be reliably attributed to biology and represents technical measurement variance.

Bibliography

[1] Naake, T., Rainer, J., Gatto, L., et al.: MsQuality: an interoperable R package for calculation of standardized quality metrics of mass spectrometry data. *Bioinformatics* **39**(10), btad610 (2023). doi:10.1093/bioinformatics/btad610

[2] Bittremieux, N., Walzer, M., Tenzer, S., et al.: The mzQC format for mass spectrometry quality control data. *Journal of Proteome Research* **19**(10), 4174--4181 (2020). doi:10.1021/acs.jproteome.0c00379

[3] Walzer, M., Pernas, L.E., Nasso, S., et al.: qcML: An exchange format for quality control metrics from mass spectrometry experiments. *Molecular & Cellular Proteomics* **13**(8), 1905--1913 (2014). doi:10.1074/mcp.M113.035907

[4] Broadhurst, D., Goodacre, R., Reinke, S.N., et al.: Guidelines and considerations for the use of system suitability and quality control samples in mass spectrometry assays applied to untargeted metabolomics. *Metabolomics* **14**(6), 72 (2018). doi:10.1007/s11306-018-1367-3

[5] Dunn, W.B., Broadhurst, D., Begley, P., et al.: Procedures for large-scale metabolomics profiling of serum and plasma using gas chromatography and liquid chromatography coupled to mass spectrometry. *Nature Protocols* **6**(7), 1060--1083 (2011). doi:10.1038/nprot.2011.335

[6] Gika, H.G., Theodoridis, G.A., Wingate, J.E., Wilson, I.D.: Within-day and between-day reproducibility of biomarker candidates in metabonomic profiling of human urine using LC-MS. *Journal of Proteome Research* **6**(8), 3291--3303 (2007). doi:10.1021/pr070212k

[7] Want, E.J., Wilson, I.D., Gika, H., et al.: Global metabolic profiling procedures for urine using UPLC-MS. *Nature Protocols* **5**(6), 1005--1018 (2010). doi:10.1038/nprot.2010.50

[8] Sangster, T., Major, H., Plumb, R., Wilson, A.J., Wilson, I.D.: A pragmatic and readily implemented approach to ratio-based conditioning and quality control for HPLC-MS metabonomics. *Analyst* **131**(10), 1075--1078 (2006). doi:10.1039/B604498K

[9] US Food and Drug Administration: Bioanalytical Method Validation Guidance for Industry. *FDA Guidance Document*, Silver Spring, MD (2018). https://www.fda.gov/regulatory-information/search-fda-guidance-documents/bioanalytical-method-validation-guidance-industry

[10] European Medicines Agency: Guideline on bioanalytical method validation. *EMEA/CHMP/EWP/192292/2010 Rev. 1*, London, UK (2011).

[11] Booth, B.P., Arnold, M.E., DeSilva, B., et al.: Workshop report: Crystal City V---quantitative bioanalytical method validation and implementation. *The AAPS Journal* **17**(2), 277--288 (2015). doi:10.1208/s12248-014-9710-1

[12] Leek, J.T., Scharpf, R.B., Bravo, H.C., et al.: Tackling the widespread and critical impact of batch effects in high-throughput data. *Nature Reviews Genetics* **11**(10), 733--739 (2010). doi:10.1038/nrg2825

[13] Goh, W.W., Wang, W., Wong, L.: Why batch effects matter in omics data, and how to address them. *Trends in Biotechnology* **35**(6), 498--507 (2017). doi:10.1016/j.tibtech.2017.02.012

[14] Nygaard, V., Rødland, E.A., Hovig, E.: Methods that remove batch effects while retaining group differences may also mediate spurious biological conclusions. *Nucleic Acids Research* **44**(7), e45 (2016). doi:10.1093/nar/gkv1366

[15] Gatto, L., Rainer, J., Gibb, S.: MSnbase, Spectra, and QFeatures: Modern mass spectrometry data structures in Bioconductor. *Journal of Proteome Research* **20**(1), 1063--1069 (2021). doi:10.1021/acs.jproteome.0c00864

[16] Gatto, L., Lilley, K.S.: MSnbase---an R/Bioconductor package for isobaric tagged and label-free mass spectrometry data processing. *Bioinformatics* **28**(2), 288--289 (2012). doi:10.1093/bioinformatics/btr645

[17] Zubarev, R., Mann, M.: On the proper use of the terms mass resolution and mass resolving power. *Analytical Chemistry* **79**(20), 7937--7937 (2007). doi:10.1021/ac072023d

[18] Gross, J.H.: *Mass Spectrometry: A Textbook*, 3rd edn. Springer, Cham (2017). doi:10.1007/978-3-319-54398-7