
Deployed system
f_0 (source: FDG)
segmenter + agent
$\mathcal{A}$: adapt
propose updates
$\Lambda = \{\lambda_1, \ldots, \lambda_m\}$
$\mathcal{B}$: certify (LTT)
$H_\lambda : R(\lambda) > \delta$
FWER $\leq \alpha$
certified
deploy $\hat{\lambda}$
none
roll back
to f_0
$\mathcal{C}$: locked
test, read
once
New domain (PSMA)
377 patients, split
$\mathcal{A}|\mathcal{B}|\mathcal{C}$ = 126|126|125
risk $R(\lambda) = \Pr$[patient harmed by λ vs. f_0],
harmed = a lesion lost, or Dice drop > 0.01
locked $\mathcal{C}$: Dice $0.11 \rightarrow 0.55$, FP $734 \rightarrow 14.5$ mL, harm 3/125