

=====  
Analysis BegunLogged In Analyst: Administrator  
Spectrometer: FIMS-100, S/N B050-9550Technique: AA FIMS-MHS  
Autosampler: S10Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb  
=====

## Method Loaded

Method Name: Hg\_Fish+Tissue\_FAS

Method Last Saved: 11/4/2025 8:58:49 PM

Method Description: Hg Method for Fish Tissue Sample\_FAS 10162025  
=====

Sequence No.: 1

Autosampler Location: 1

Sample ID: Calib Blank 1

Date Collected: 12/16/2025 11:57:24 AM

Analyst:

Data Type: Original  
-----

Replicate Data: Calib Blank 1

Analyte: Hg 253.7

| Repl | SampleConc | StndConc | BlkCorr | Peak    | Peak    | Time        | Peak   |
|------|------------|----------|---------|---------|---------|-------------|--------|
| #    | ug/g       | ug/L     | Signal  | Area    | Height  |             | Stored |
| 1    |            | [0.00]   | -0.0046 | -0.0046 | -0.0000 | 11:58:14 AM | Yes    |

User canceled analysis.  
=====

## Analysis Begun

Logged In Analyst: Administrator  
Spectrometer: FIMS-100, S/N B050-9550Technique: AA FIMS-MHS  
Autosampler: S10Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb  
=====

Sequence No.: 1

Autosampler Location: 1

Sample ID: Calib Blank 1

Date Collected: 12/16/2025 12:00:36 PM

Analyst:

Data Type: Original  
-----

Replicate Data: Calib Blank 1

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | BlkCorr | Peak    | Peak    | Time        | Peak   |
|-------|------------|----------|---------|---------|---------|-------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area    | Height  |             | Stored |
| 1     |            | [0.00]   | -0.0024 | -0.0024 | -0.0000 | 12:01:25 PM | Yes    |
| 2     |            | [0.00]   | -0.0014 | -0.0014 | 0.0000  | 12:01:56 PM | Yes    |
| 3     |            | [0.00]   | -0.0019 | -0.0019 | -0.0000 | 12:02:26 PM | Yes    |
| Mean: |            | [0.00]   | -0.0019 |         |         |             |        |
| SD:   |            | 0.0000   | 0.0005  |         |         |             |        |
| %RSD: |            | 0.00%    | 25.56   |         |         |             |        |

Auto-zero performed.  
=====

Sequence No.: 2

Autosampler Location: 2

Sample ID: Calib Std 1

Date Collected: 12/16/2025 12:03:07 PM

Analyst:

Data Type: Original  
-----

Replicate Data: Calib Std 1

Analyte: Hg 253.7

| Repl | SampleConc | StndConc | BlkCorr | Peak   | Peak   | Time        | Peak   |
|------|------------|----------|---------|--------|--------|-------------|--------|
| #    | ug/g       | ug/L     | Signal  | Area   | Height |             | Stored |
| 1    |            | [0.1]    | 0.0071  | 0.0052 | 0.0008 | 12:03:57 PM | Yes    |

|                                                               |       |        |        |        |             |     |
|---------------------------------------------------------------|-------|--------|--------|--------|-------------|-----|
| 2                                                             | [0.1] | 0.0067 | 0.0048 | 0.0007 | 12:04:27 PM | Yes |
| 3                                                             | [0.1] | 0.0075 | 0.0056 | 0.0008 | 12:04:57 PM | Yes |
| Mean:                                                         | [0.1] | 0.0071 |        |        |             |     |
| SD:                                                           | 0.000 | 0.0004 |        |        |             |     |
| %RSD:                                                         | 0.00% | 5.75   |        |        |             |     |
| Standard number 1 applied. [0.1]                              |       |        |        |        |             |     |
| Correlation Coef.: 1.000000 Slope: 0.07110 Intercept: 0.00000 |       |        |        |        |             |     |

Sequence No.: 3 Autosampler Location: 3  
Sample ID: Calib Std 2 Date Collected: 12/16/2025 12:05:39 PM  
Analyst: Data Type: Original

Replicate Data: Calib Std 2 Analyte: Hg 253.7

| Repl #                                                        | SampleConc ug/g | StdConc ug/L | BlkCorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|---------------------------------------------------------------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1                                                             |                 | [0.5]        | 0.0350         | 0.0332    | 0.0045      | 12:06:29 PM | Yes         |
| 2                                                             |                 | [0.5]        | 0.0344         | 0.0325    | 0.0044      | 12:06:59 PM | Yes         |
| 3                                                             |                 | [0.5]        | 0.0360         | 0.0341    | 0.0046      | 12:07:29 PM | Yes         |
| Mean:                                                         |                 | [0.5]        | 0.0352         |           |             |             |             |
| SD:                                                           |                 | 0.000        | 0.0008         |           |             |             |             |
| %RSD:                                                         |                 | 0.00%        | 2.27           |           |             |             |             |
| Standard number 2 applied. [0.5]                              |                 |              |                |           |             |             |             |
| Correlation Coef.: 1.000000 Slope: 0.07015 Intercept: 0.00010 |                 |              |                |           |             |             |             |

Sequence No.: 4 Autosampler Location: 4  
Sample ID: Calib Std 3 Date Collected: 12/16/2025 12:08:11 PM  
Analyst: Data Type: Original

Replicate Data: Calib Std 3 Analyte: Hg 253.7

| Repl #                                                        | SampleConc ug/g | StdConc ug/L | BlkCorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|---------------------------------------------------------------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1                                                             |                 | [5]          | 0.3438         | 0.3419    | 0.0454      | 12:09:01 PM | Yes         |
| 2                                                             |                 | [5]          | 0.3375         | 0.3356    | 0.0449      | 12:09:32 PM | Yes         |
| 3                                                             |                 | [5]          | 0.3375         | 0.3356    | 0.0449      | 12:10:02 PM | Yes         |
| Mean:                                                         |                 | [5]          | 0.3396         |           |             |             |             |
| SD:                                                           |                 | 0.00         | 0.0037         |           |             |             |             |
| %RSD:                                                         |                 | 0.00%        | 1.08           |           |             |             |             |
| Standard number 3 applied. [5]                                |                 |              |                |           |             |             |             |
| Correlation Coef.: 0.999866 Slope: 0.06875 Intercept: 0.00026 |                 |              |                |           |             |             |             |

Sequence No.: 5 Autosampler Location: 5  
Sample ID: Calib Std 4 Date Collected: 12/16/2025 12:10:44 PM  
Analyst: Data Type: Original

Replicate Data: Calib Std 4 Analyte: Hg 253.7

| Repl #                                                        | SampleConc ug/g | StdConc ug/L | BlkCorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|---------------------------------------------------------------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1                                                             |                 | [10]         | 0.6874         | 0.6855    | 0.0910      | 12:11:34 PM | Yes         |
| 2                                                             |                 | [10]         | 0.6761         | 0.6742    | 0.0901      | 12:12:05 PM | Yes         |
| 3                                                             |                 | [10]         | 0.6772         | 0.6753    | 0.0900      | 12:12:35 PM | Yes         |
| Mean:                                                         |                 | [10]         | 0.6802         |           |             |             |             |
| SD:                                                           |                 | 0.00         | 0.0062         |           |             |             |             |
| %RSD:                                                         |                 | 0.00%        | 0.91           |           |             |             |             |
| Standard number 4 applied. [10]                               |                 |              |                |           |             |             |             |
| Correlation Coef.: 0.999902 Slope: 0.06846 Intercept: 0.00029 |                 |              |                |           |             |             |             |

## Calibration data for Hg 253.7

Equation: Weighted Linear

| ID            | Mean Signal (Abs) | Entered Conc. ug/L | Calculated Conc. ug/L | Standard Deviation | %RSD  |
|---------------|-------------------|--------------------|-----------------------|--------------------|-------|
| Calib Blank 1 | 0.0000            | 0                  | -0.004                | 0.00               | 25.56 |

|             |        |      |       |      |      |
|-------------|--------|------|-------|------|------|
| Calib Std 1 | 0.0071 | 0.1  | 0.100 | 0.00 | 5.75 |
| Calib Std 2 | 0.0352 | 0.5  | 0.510 | 0.00 | 2.27 |
| Calib Std 3 | 0.3396 | 5.0  | 4.956 | 0.00 | 1.08 |
| Calib Std 4 | 0.6802 | 10.0 | 9.933 | 0.01 | 0.91 |

Correlation Coef.: 0.999902    Slope: 0.06846    Intercept: 0.00029

=====  
Analysis Begun

Logged In Analyst: Administrator  
Spectrometer: FIMS-100, S/N B050-9550

Technique: AA FIMS-MHS  
Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 1

Sample ID: QAQC

Analyst:

User canceled analysis.

Autosampler Location: 6

Date Collected: 12/16/2025 12:19:23 PM

Data Type: Original

=====  
Analysis Begun

Logged In Analyst: Administrator  
Spectrometer: FIMS-100, S/N B050-9550

Technique: AA FIMS-MHS  
Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 36

Sample ID: QAQC

Analyst:

Autosampler Location: 6

Date Collected: 12/16/2025 12:20:34 PM

Data Type: Original

-----  
Replicate Data: QAQC

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdConc ug/L | Blncorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|--------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1      |                 | 1.053        | 0.0724         | 0.0705    | 0.0093      | 12:21:25 PM | Yes         |
| 2      |                 | 0.992        | 0.0682         | 0.0663    | 0.0089      | 12:21:55 PM | Yes         |
| 3      |                 | 0.998        | 0.0686         | 0.0667    | 0.0090      | 12:22:26 PM | Yes         |
| Mean:  |                 | 1.015        | 0.0697         |           |             |             |             |
| SD:    |                 | 0.0338       | 0.0023         |           |             |             |             |
| %RSD:  |                 | 3.33%        | 3.31           |           |             |             |             |

QC value within limits for Hg 253.7    Recovery = 101.46%  
All analyte(s) passed QC.

Sequence No.: 37

Sample ID: Standard blank

Analyst:

Autosampler Location: 1

Date Collected: 12/16/2025 12:23:09 PM

Data Type: Original

-----  
Replicate Data: Standard blank

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdConc ug/L | Blncorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|--------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1      |                 | 0.030        | 0.0023         | 0.0004    | 0.0001      | 12:23:57 PM | Yes         |
| 2      |                 | 0.040        | 0.0031         | 0.0012    | 0.0001      | 12:24:27 PM | Yes         |
| 3      |                 | 0.014        | 0.0012         | -0.0007   | 0.0000      | 12:24:58 PM | Yes         |
| Mean:  |                 | 0.028        | 0.0022         |           |             |             |             |
| SD:    |                 | 0.0133       | 0.0009         |           |             |             |             |
| %RSD:  |                 | 47.56%       | 41.32          |           |             |             |             |

QC value less than the lower limit for Hg 253.7 Recovery = 2.79%  
QC Failed. Continue with analysis.

Sequence No.: 38  
Sample ID: 21\_gas  
Analyst:

Autosampler Location: 38  
Date Collected: 12/16/2025 12:25:39 PM  
Data Type: Original

## Replicate Data: 21\_gas

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdConc ug/L | Blncorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|--------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1      | 0.476           | 0.478        | 0.0330         | 0.0311    | 0.0049      | 12:26:30 PM | Yes         |
| 2      | 0.425           | 0.427        | 0.0296         | 0.0277    | 0.0038      | 12:27:00 PM | Yes         |
| 3      | 37.03           | 37.22        | 2.5481         | 2.5462    | 1.2048      | 12:27:30 PM | Yes         |

Sample concentration is greater than that of the highest standard.

Mean: 12.64 12.71 0.8702  
SD: 21.120 21.226 1.4531  
%RSD: 167.03% 167.03% 166.98

Sample concentration is greater than that of the highest standard.

Sequence No.: 39  
Sample ID: 50\_gas  
Analyst:

Autosampler Location: 39  
Date Collected: 12/16/2025 12:28:14 PM  
Data Type: Original

## Replicate Data: 50\_gas

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdConc ug/L | Blncorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|--------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1      | 26.16           | 26.95        | 1.8450         | 1.8431    | 2.6665      | 12:29:08 PM | Yes         |
| 2      | -16.31          | -16.80       | -1.1496        | -1.1514   | 0.0716      | 12:29:40 PM | Yes         |
| 3      | 0.023           | 0.023        | 0.0019         | 0.0000    | 0.0000      | 12:30:11 PM | Yes         |

Sample concentration is greater than that of the highest standard.

Mean: 3.293 3.391 0.2325  
SD: 21.4227 22.0653 1.5105  
%RSD: 650.62% 650.62% 649.81

Sequence No.: 40  
Sample ID: 11\_lung  
Analyst:

Autosampler Location: 40  
Date Collected: 12/16/2025 12:30:54 PM  
Data Type: Original

## Replicate Data: 11\_lung

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdConc ug/L | Blncorr Signal | Peak Area | Peak Height | Time        | Peak Stored |
|--------|-----------------|--------------|----------------|-----------|-------------|-------------|-------------|
| 1      | 1.038           | 1.023        | 0.0703         | 0.0684    | 0.0084      | 12:31:47 PM | Yes         |

User canceled analysis.

## Analysis Begun

Logged In Analyst: Administrator  
Spectrometer: FIMS-100, S/N B050-9550

Technique: AA FIMS-MHS  
Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 36  
Sample ID: QAQC  
Analyst:

Autosampler Location: 6  
Date Collected: 12/16/2025 1:00:26 PM  
Data Type: Original

Replicate Data: QAQC

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StndConc ug/L | BlnkCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|---------------|-----------------|-----------|-------------|------------|-------------|
| 1      |                 | 0.907         | 0.0624          | 0.0605    | 0.0083      | 1:01:16 PM | Yes         |
| 2      |                 | 0.851         | 0.0586          | 0.0567    | 0.0077      | 1:01:46 PM | Yes         |
| 3      |                 | 1.008         | 0.0693          | 0.0674    | 0.0098      | 1:02:16 PM | Yes         |
| Mean:  |                 | 0.922         | 0.0634          |           |             |            |             |
| SD:    |                 | 0.0795        | 0.0054          |           |             |            |             |
| %RSD:  |                 | 8.62%         | 8.59            |           |             |            |             |

QC value within limits for Hg 253.7 Recovery = 92.20%  
All analyte(s) passed QC.

Sequence No.: 37  
Sample ID: Standard blank  
Analyst:

Autosampler Location: 1  
Date Collected: 12/16/2025 1:02:59 PM  
Data Type: Original

Replicate Data: Standard blank

| Repl # | SampleConc ug/g | StndConc ug/L | BlnkCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|---------------|-----------------|-----------|-------------|------------|-------------|
| 1      |                 | -0.034        | -0.0021         | -0.0040   | -0.0000     | 1:03:48 PM | Yes         |
| 2      |                 | -0.041        | -0.0025         | -0.0044   | -0.0000     | 1:04:19 PM | Yes         |
| 3      |                 | -0.031        | -0.0018         | -0.0037   | -0.0000     | 1:04:49 PM | Yes         |
| Mean:  |                 | -0.035        | -0.0021         |           |             |            |             |
| SD:    |                 | 0.0051        | 0.0003          |           |             |            |             |
| %RSD:  |                 | 14.48%        | 16.45           |           |             |            |             |

QC value less than the lower limit for Hg 253.7 Recovery = -3.52%  
QC Failed. Continue with analysis.

Sequence No.: 38  
Sample ID: 21\_gas  
Analyst:

Autosampler Location: 38  
Date Collected: 12/16/2025 1:05:30 PM  
Data Type: Original

Replicate Data: 21\_gas

| Repl # | SampleConc ug/g | StndConc ug/L | BlnkCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|---------------|-----------------|-----------|-------------|------------|-------------|
| 1      | 0.446           | 0.448         | 0.0310          | 0.0291    | 0.0042      | 1:06:21 PM | Yes         |
| 2      | 0.452           | 0.455         | 0.0314          | 0.0295    | 0.0042      | 1:06:51 PM | Yes         |
| 3      | 0.442           | 0.444         | 0.0307          | 0.0288    | 0.0041      | 1:07:21 PM | Yes         |
| Mean:  | 0.447           | 0.449         | 0.0310          |           |             |            |             |
| SD:    | 0.0053          | 0.0053        | 0.0004          |           |             |            |             |
| %RSD:  | 1.18%           | 1.18%         | 1.17            |           |             |            |             |

Sequence No.: 39  
Sample ID: 50\_gas  
Analyst:

Autosampler Location: 39  
Date Collected: 12/16/2025 1:08:04 PM  
Data Type: Original

Replicate Data: 50\_gas

| Repl # | SampleConc ug/g | StndConc ug/L | BlnkCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|---------------|-----------------|-----------|-------------|------------|-------------|
| 1      | 0.965           | 0.994         | 0.0683          | 0.0664    | 0.0096      | 1:08:55 PM | Yes         |
| 2      | 0.907           | 0.934         | 0.0642          | 0.0623    | 0.0093      | 1:09:25 PM | Yes         |
| 3      | 0.923           | 0.950         | 0.0654          | 0.0635    | 0.0093      | 1:09:55 PM | Yes         |
| Mean:  | 0.931           | 0.959         | 0.0660          |           |             |            |             |
| SD:    | 0.0302          | 0.0311        | 0.0021          |           |             |            |             |
| %RSD:  | 3.24%           | 3.24%         | 3.23            |           |             |            |             |

Sequence No.: 40  
Sample ID: 11\_lung  
Analyst:

Autosampler Location: 40  
Date Collected: 12/16/2025 1:10:38 PM  
Data Type: Original

Replicate Data: 11\_lung

| Repl | SampleConc | StndConc | BlnkCorr | Peak | Peak   | Time | Peak   |
|------|------------|----------|----------|------|--------|------|--------|
|      |            |          |          | Area | Height |      | Stored |

| #     | ug/g   | ug/L   | Signal | Area   | Height |            | Stored |
|-------|--------|--------|--------|--------|--------|------------|--------|
| 1     | 0.884  | 0.871  | 0.0599 | 0.0580 | 0.0080 | 1:11:28 PM | Yes    |
| 2     | 0.815  | 0.803  | 0.0552 | 0.0534 | 0.0078 | 1:11:58 PM | Yes    |
| 3     | 0.833  | 0.820  | 0.0564 | 0.0545 | 0.0079 | 1:12:29 PM | Yes    |
| Mean: | 0.844  | 0.831  | 0.0572 |        |        |            |        |
| SD:   | 0.0359 | 0.0353 | 0.0024 |        |        |            |        |
| %RSD: | 4.25%  | 4.25%  | 4.23   |        |        |            |        |

Sequence No.: 41

Autosampler Location: 41

Sample ID: 17\_lung

Date Collected: 12/16/2025 1:13:11 PM

Analyst:

Data Type: Original

Replicate Data: 17\_lung

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 1.165      | 1.183   | 0.0812  | 0.0793 | 0.0111 | 1:14:02 PM | Yes    |
| 2     | 1.090      | 1.106   | 0.0760  | 0.0741 | 0.0108 | 1:14:32 PM | Yes    |
| 3     | 1.165      | 1.182   | 0.0812  | 0.0793 | 0.0112 | 1:15:03 PM | Yes    |
| Mean: | 1.140      | 1.157   | 0.0795  |        |        |            |        |
| SD:   | 0.0435     | 0.0441  | 0.0030  |        |        |            |        |
| %RSD: | 3.81%      | 3.81%   | 3.80    |        |        |            |        |

Sequence No.: 42

Autosampler Location: 42

Sample ID: 53\_brain

Date Collected: 12/16/2025 1:15:45 PM

Analyst:

Data Type: Original

Replicate Data: 53\_brain

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 1.041      | 1.046   | 0.0719  | 0.0700 | 0.0097 | 1:16:36 PM | Yes    |
| 2     | 0.956      | 0.960   | 0.0660  | 0.0641 | 0.0094 | 1:17:07 PM | Yes    |
| 3     | 0.989      | 0.994   | 0.0683  | 0.0664 | 0.0094 | 1:17:37 PM | Yes    |
| Mean: | 0.995      | 1.000   | 0.0687  |        |        |            |        |
| SD:   | 0.0429     | 0.0431  | 0.0030  |        |        |            |        |
| %RSD: | 4.31%      | 4.31%   | 4.29    |        |        |            |        |

Sequence No.: 43

Autosampler Location: 43

Sample ID: 14\_brain

Date Collected: 12/16/2025 1:18:20 PM

Analyst:

Data Type: Original

Replicate Data: 14\_brain

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 0.518      | 0.523   | 0.0361  | 0.0342 | 0.0047 | 1:19:11 PM | Yes    |
| 2     | 0.495      | 0.500   | 0.0345  | 0.0326 | 0.0046 | 1:19:41 PM | Yes    |
| 3     | 0.486      | 0.491   | 0.0339  | 0.0320 | 0.0046 | 1:20:12 PM | Yes    |
| Mean: | 0.500      | 0.505   | 0.0348  |        |        |            |        |
| SD:   | 0.0167     | 0.0169  | 0.0012  |        |        |            |        |
| %RSD: | 3.35%      | 3.35%   | 3.33    |        |        |            |        |

Sequence No.: 44

Autosampler Location: 44

Sample ID: 55\_pec

Date Collected: 12/16/2025 1:20:55 PM

Analyst:

Data Type: Original

Replicate Data: 55\_pec

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 0.651      | 0.638   | 0.0440  | 0.0421 | 0.0059 | 1:21:46 PM | Yes    |
| 2     | 0.621      | 0.609   | 0.0419  | 0.0401 | 0.0059 | 1:22:16 PM | Yes    |
| 3     | 0.637      | 0.624   | 0.0430  | 0.0411 | 0.0060 | 1:22:47 PM | Yes    |
| Mean: | 0.636      | 0.624   | 0.0430  |        |        |            |        |

SD: 0.0151 0.0148 0.0010  
%RSD: 2.37% 2.37% 2.35

Sequence No.: 45  
Sample ID: 14\_pec  
Analyst:

Autosampler Location: 45  
Date Collected: 12/16/2025 1:23:30 PM  
Data Type: Original

## Replicate Data: 14\_pec

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.562      | 0.567    | 0.0391  | 0.0372 | 0.0053 | 1:24:21 PM | Yes    |
| 2     | 0.520      | 0.525    | 0.0363  | 0.0344 | 0.0051 | 1:24:51 PM | Yes    |
| 3     | 0.542      | 0.547    | 0.0377  | 0.0358 | 0.0052 | 1:25:21 PM | Yes    |
| Mean: | 0.541      | 0.547    | 0.0377  |        |        |            |        |
| SD:   | 0.0207     | 0.0209   | 0.0014  |        |        |            |        |
| %RSD: | 3.82%      | 3.82%    | 3.80    |        |        |            |        |

Sequence No.: 46  
Sample ID: 50\_gi  
Analyst:

Autosampler Location: 46  
Date Collected: 12/16/2025 1:26:04 PM  
Data Type: Original

## Replicate Data: 50\_gi

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.954      | 0.978    | 0.0672  | 0.0653 | 0.0093 | 1:26:55 PM | Yes    |
| 2     | 0.902      | 0.925    | 0.0636  | 0.0617 | 0.0090 | 1:27:25 PM | Yes    |
| 3     | 0.948      | 0.971    | 0.0668  | 0.0649 | 0.0093 | 1:27:56 PM | Yes    |
| Mean: | 0.935      | 0.958    | 0.0659  |        |        |            |        |
| SD:   | 0.0282     | 0.0289   | 0.0020  |        |        |            |        |
| %RSD: | 3.02%      | 3.02%    | 3.00    |        |        |            |        |

Sequence No.: 47  
Sample ID: 46\_gi  
Analyst:

Autosampler Location: 47  
Date Collected: 12/16/2025 1:28:39 PM  
Data Type: Original

## Replicate Data: 46\_gi

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.925      | 0.952    | 0.0655  | 0.0636 | 0.0090 | 1:29:30 PM | Yes    |
| 2     | 0.857      | 0.883    | 0.0607  | 0.0588 | 0.0087 | 1:30:00 PM | Yes    |
| 3     | 0.870      | 0.896    | 0.0616  | 0.0597 | 0.0087 | 1:30:30 PM | Yes    |
| Mean: | 0.884      | 0.910    | 0.0626  |        |        |            |        |
| SD:   | 0.0359     | 0.0370   | 0.0025  |        |        |            |        |
| %RSD: | 4.06%      | 4.06%    | 4.04    |        |        |            |        |

Sequence No.: 48  
Sample ID: QAQC  
Analyst:

Autosampler Location: 6  
Date Collected: 12/16/2025 1:31:13 PM  
Data Type: Original

## Replicate Data: QAQC

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     |            | 1.128    | 0.0775  | 0.0756 | 0.0105 | 1:32:04 PM | Yes    |
| 2     |            | 1.080    | 0.0743  | 0.0724 | 0.0103 | 1:32:34 PM | Yes    |
| 3     |            | 1.096    | 0.0753  | 0.0735 | 0.0104 | 1:33:04 PM | Yes    |
| Mean: |            | 1.102    | 0.0757  |        |        |            |        |
| SD:   |            | 0.0240   | 0.0016  |        |        |            |        |
| %RSD: |            | 2.18%    | 2.17    |        |        |            |        |

QC value greater than the upper limit for Hg 253.7 Recovery = 110.15%  
QC Failed. Continue with analysis.

Sequence No.: 49

Sample ID: Standard blank

Analyst:

Autosampler Location: 1

Date Collected: 12/16/2025 1:33:47 PM

Data Type: Original

Replicate Data: Standard blank

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdndConc ug/L | BlndCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|----------------|-----------------|-----------|-------------|------------|-------------|
| 1      |                 | 0.027          | 0.0021          | 0.0002    | 0.0001      | 1:34:36 PM | Yes         |
| 2      |                 | -0.006         | -0.0001         | -0.0020   | 0.0000      | 1:35:06 PM | Yes         |
| 3      |                 | 0.028          | 0.0022          | 0.0003    | 0.0001      | 1:35:37 PM | Yes         |
| Mean:  |                 | 0.016          | 0.0014          |           |             |            |             |
| SD:    |                 | 0.0192         | 0.0013          |           |             |            |             |
| %RSD:  |                 | 117.07%        | 93.12           |           |             |            |             |

QC value less than the lower limit for Hg 253.7 Recovery = 1.64%  
QC Failed. Continue with analysis.

Sequence No.: 50

Sample ID: 16\_liver

Analyst:

Autosampler Location: 48

Date Collected: 12/16/2025 1:36:18 PM

Data Type: Original

Replicate Data: 16\_liver

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdndConc ug/L | BlndCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|----------------|-----------------|-----------|-------------|------------|-------------|
| 1      | 0.628           | 0.635          | 0.0437          | 0.0418    | 0.0059      | 1:37:09 PM | Yes         |
| 2      | 0.564           | 0.570          | 0.0393          | 0.0374    | 0.0057      | 1:37:39 PM | Yes         |
| 3      | 0.613           | 0.619          | 0.0427          | 0.0408    | 0.0060      | 1:38:09 PM | Yes         |
| Mean:  | 0.602           | 0.608          | 0.0419          |           |             |            |             |
| SD:    | 0.0336          | 0.0339         | 0.0023          |           |             |            |             |
| %RSD:  | 5.58%           | 5.58%          | 5.55            |           |             |            |             |

Sequence No.: 51

Sample ID: 40\_liver

Analyst:

Autosampler Location: 49

Date Collected: 12/16/2025 1:38:52 PM

Data Type: Original

Replicate Data: 40\_liver

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdndConc ug/L | BlndCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|----------------|-----------------|-----------|-------------|------------|-------------|
| 1      | 0.645           | 0.667          | 0.0460          | 0.0441    | 0.0062      | 1:39:43 PM | Yes         |
| 2      | 0.599           | 0.620          | 0.0428          | 0.0409    | 0.0061      | 1:40:13 PM | Yes         |
| 3      | 0.632           | 0.654          | 0.0451          | 0.0432    | 0.0062      | 1:40:44 PM | Yes         |
| Mean:  | 0.625           | 0.647          | 0.0446          |           |             |            |             |
| SD:    | 0.0234          | 0.0242         | 0.0017          |           |             |            |             |
| %RSD:  | 3.73%           | 3.73%          | 3.71            |           |             |            |             |

Sequence No.: 52

Sample ID: 26\_heart

Analyst:

Autosampler Location: 50

Date Collected: 12/16/2025 1:41:27 PM

Data Type: Original

Replicate Data: 26\_heart

Analyte: Hg 253.7

| Repl # | SampleConc ug/g | StdndConc ug/L | BlndCorr Signal | Peak Area | Peak Height | Time       | Peak Stored |
|--------|-----------------|----------------|-----------------|-----------|-------------|------------|-------------|
| 1      | 0.401           | 0.403          | 0.0279          | 0.0260    | 0.0036      | 1:42:18 PM | Yes         |
| 2      | 0.346           | 0.348          | 0.0241          | 0.0222    | 0.0033      | 1:42:49 PM | Yes         |
| 3      | 0.364           | 0.366          | 0.0253          | 0.0234    | 0.0034      | 1:43:19 PM | Yes         |
| Mean:  | 0.371           | 0.372          | 0.0258          |           |             |            |             |
| SD:    | 0.0279          | 0.0281         | 0.0019          |           |             |            |             |
| %RSD:  | 7.54%           | 7.54%          | 7.46            |           |             |            |             |

Sequence No.: 53

Sample ID: 33\_heart

Autosampler Location: 51

Date Collected: 12/16/2025 1:44:02 PM



Analyst:

Data Type: Original

## Replicate Data: 33\_heart

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.510      | 0.518    | 0.0357  | 0.0339 | 0.0047 | 1:44:53 PM | Yes    |
| 2     | 0.494      | 0.501    | 0.0346  | 0.0327 | 0.0047 | 1:45:24 PM | Yes    |
| 3     | 0.481      | 0.488    | 0.0337  | 0.0318 | 0.0047 | 1:45:54 PM | Yes    |
| Mean: | 0.495      | 0.502    | 0.0347  |        |        |            |        |
| SD:   | 0.0148     | 0.0150   | 0.0010  |        |        |            |        |
| %RSD: | 2.99%      | 2.99%    | 2.96    |        |        |            |        |

Sequence No.: 54

Autosampler Location: 52

Sample ID: 44\_gi

Date Collected: 12/16/2025 1:46:37 PM

Analyst:

Data Type: Original

## Replicate Data: 44\_gi

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.561      | 0.572    | 0.0394  | 0.0375 | 0.0052 | 1:47:28 PM | Yes    |
| 2     | 0.524      | 0.535    | 0.0369  | 0.0350 | 0.0051 | 1:47:59 PM | Yes    |
| 3     | 0.542      | 0.553    | 0.0381  | 0.0362 | 0.0052 | 1:48:29 PM | Yes    |
| Mean: | 0.542      | 0.553    | 0.0382  |        |        |            |        |
| SD:   | 0.0181     | 0.0185   | 0.0013  |        |        |            |        |
| %RSD: | 3.34%      | 3.34%    | 3.31    |        |        |            |        |

Sequence No.: 55

Autosampler Location: 53

Sample ID: 12\_pec

Date Collected: 12/16/2025 1:49:12 PM

Analyst:

Data Type: Original

## Replicate Data: 12\_pec

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.467      | 0.486    | 0.0336  | 0.0317 | 0.0046 | 1:50:03 PM | Yes    |
| 2     | 0.444      | 0.462    | 0.0319  | 0.0300 | 0.0045 | 1:50:34 PM | Yes    |
| 3     | 0.457      | 0.475    | 0.0328  | 0.0309 | 0.0045 | 1:51:04 PM | Yes    |
| Mean: | 0.456      | 0.474    | 0.0328  |        |        |            |        |
| SD:   | 0.0117     | 0.0122   | 0.0008  |        |        |            |        |
| %RSD: | 2.57%      | 2.57%    | 2.55    |        |        |            |        |

Sequence No.: 56

Autosampler Location: 54

Sample ID: 37\_pec

Date Collected: 12/16/2025 1:51:47 PM

Analyst:

Data Type: Original

## Replicate Data: 37\_pec

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.315      | 0.305    | 0.0212  | 0.0193 | 0.0028 | 1:52:38 PM | Yes    |
| 2     | 0.275      | 0.266    | 0.0185  | 0.0166 | 0.0026 | 1:53:08 PM | Yes    |
| 3     | 0.302      | 0.293    | 0.0204  | 0.0185 | 0.0027 | 1:53:39 PM | Yes    |
| Mean: | 0.297      | 0.288    | 0.0200  |        |        |            |        |
| SD:   | 0.0204     | 0.0198   | 0.0014  |        |        |            |        |
| %RSD: | 6.88%      | 6.88%    | 6.78    |        |        |            |        |

Sequence No.: 57

Autosampler Location: 55

Sample ID: 12\_liver

Date Collected: 12/16/2025 1:54:22 PM

Analyst:

Data Type: Original

## Replicate Data: 12\_liver

Analyte: Hg 253.7

| Repl | SampleConc | StndConc | Blncorr | Peak | Peak | Time | Peak |
|------|------------|----------|---------|------|------|------|------|
|------|------------|----------|---------|------|------|------|------|

| #     | ug/g   | ug/L   | Signal | Area   | Height |            | Stored |
|-------|--------|--------|--------|--------|--------|------------|--------|
| 1     | 0.906  | 0.928  | 0.0638 | 0.0620 | 0.0087 | 1:55:13 PM | Yes    |
| 2     | 0.838  | 0.859  | 0.0591 | 0.0572 | 0.0084 | 1:55:43 PM | Yes    |
| 3     | 0.860  | 0.882  | 0.0607 | 0.0588 | 0.0085 | 1:56:14 PM | Yes    |
| Mean: | 0.868  | 0.890  | 0.0612 |        |        |            |        |
| SD:   | 0.0344 | 0.0353 | 0.0024 |        |        |            |        |
| %RSD: | 3.96%  | 3.96%  | 3.95   |        |        |            |        |

Sequence No.: 58

Autosampler Location: 56

Sample ID: 51\_brain

Date Collected: 12/16/2025 1:56:57 PM

Analyst:

Data Type: Original

Replicate Data: 51\_brain

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | BlkCorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 0.467      | 0.471   | 0.0325  | 0.0306 | 0.0044 | 1:57:48 PM | Yes    |
| 2     | 0.437      | 0.441   | 0.0305  | 0.0286 | 0.0043 | 1:58:18 PM | Yes    |
| 3     | 0.446      | 0.451   | 0.0312  | 0.0293 | 0.0044 | 1:58:49 PM | Yes    |
| Mean: | 0.450      | 0.454   | 0.0314  |        |        |            |        |
| SD:   | 0.0151     | 0.0152  | 0.0010  |        |        |            |        |
| %RSD: | 3.35%      | 3.35%   | 3.32    |        |        |            |        |

Sequence No.: 59

Autosampler Location: 57

Sample ID: 47\_brain

Date Collected: 12/16/2025 1:59:33 PM

Analyst:

Data Type: Original

Replicate Data: 47\_brain

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | BlkCorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 0.766      | 0.778   | 0.0535  | 0.0516 | 0.0073 | 2:00:24 PM | Yes    |
| 2     | 0.691      | 0.701   | 0.0483  | 0.0464 | 0.0071 | 2:00:55 PM | Yes    |
| 3     | 0.747      | 0.759   | 0.0522  | 0.0503 | 0.0073 | 2:01:25 PM | Yes    |
| Mean: | 0.735      | 0.746   | 0.0514  |        |        |            |        |
| SD:   | 0.0394     | 0.0399  | 0.0027  |        |        |            |        |
| %RSD: | 5.36%      | 5.36%   | 5.33    |        |        |            |        |

Sequence No.: 60

Autosampler Location: 6

Sample ID: QAQC

Date Collected: 12/16/2025 2:02:09 PM

Analyst:

Data Type: Original

Replicate Data: QAQC

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | BlkCorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     |            | 1.126   | 0.0774  | 0.0755 | 0.0104 | 2:03:00 PM | Yes    |
| 2     |            | 0.906   | 0.0623  | 0.0604 | 0.0081 | 2:03:30 PM | Yes    |
| 3     |            | 0.959   | 0.0659  | 0.0640 | 0.0084 | 2:04:00 PM | Yes    |
| Mean: |            | 0.997   | 0.0685  |        |        |            |        |
| SD:   |            | 0.1146  | 0.0078  |        |        |            |        |
| %RSD: |            | 11.49%  | 11.45   |        |        |            |        |

QC value within limits for Hg 253.7 Recovery = 99.69%

All analyte(s) passed QC.

Sequence No.: 61

Autosampler Location: 1

Sample ID: Standard blank

Date Collected: 12/16/2025 2:04:44 PM

Analyst:

Data Type: Original

Replicate Data: Standard blank

Analyte: Hg 253.7

| Repl | SampleConc | StdConc | BlkCorr | Peak    | Peak   | Time       | Peak   |
|------|------------|---------|---------|---------|--------|------------|--------|
| #    | ug/g       | ug/L    | Signal  | Area    | Height |            | Stored |
| 1    |            | -0.011  | -0.0004 | -0.0023 | 0.0000 | 2:05:33 PM | Yes    |
| 2    |            | 0.011   | 0.0010  | -0.0009 | 0.0000 | 2:06:03 PM | Yes    |

3 0.038 0.0029 0.0010 0.0002 2:06:33 PM Yes  
Mean: 0.013 0.0012  
SD: 0.0245 0.0017  
%RSD: 192.62% 144.64

QC value less than the lower limit for Hg 253.7 Recovery = 1.27%  
QC Failed. Continue with analysis.

Sequence No.: 62  
Sample ID: 20\_lung  
Analyst:

Autosampler Location: 58  
Date Collected: 12/16/2025 2:07:15 PM  
Data Type: Original

## Replicate Data: 20\_lung

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.811      | 0.799    | 0.0550  | 0.0531 | 0.0072 | 2:08:07 PM | Yes    |
| 2     | 0.769      | 0.758    | 0.0522  | 0.0503 | 0.0071 | 2:08:37 PM | Yes    |
| 3     | 0.821      | 0.809    | 0.0556  | 0.0538 | 0.0072 | 2:09:08 PM | Yes    |
| Mean: | 0.800      | 0.788    | 0.0543  |        |        |            |        |
| SD:   | 0.0274     | 0.0270   | 0.0019  |        |        |            |        |
| %RSD: | 3.43%      | 3.43%    | 3.41    |        |        |            |        |

Sequence No.: 63  
Sample ID: 16\_lung  
Analyst:

Autosampler Location: 59  
Date Collected: 12/16/2025 2:09:52 PM  
Data Type: Original

## Replicate Data: 16\_lung

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.397      | 0.397    | 0.0275  | 0.0256 | 0.0034 | 2:10:43 PM | Yes    |
| 2     | 0.365      | 0.365    | 0.0253  | 0.0234 | 0.0032 | 2:11:14 PM | Yes    |
| 3     | 0.342      | 0.342    | 0.0237  | 0.0218 | 0.0032 | 2:11:44 PM | Yes    |
| Mean: | 0.368      | 0.368    | 0.0255  |        |        |            |        |
| SD:   | 0.0275     | 0.0275   | 0.0019  |        |        |            |        |
| %RSD: | 7.46%      | 7.46%    | 7.38    |        |        |            |        |

Sequence No.: 64  
Sample ID: 40\_gas  
Analyst:

Autosampler Location: 60  
Date Collected: 12/16/2025 2:12:28 PM  
Data Type: Original

## Replicate Data: 40\_gas

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.320      | 0.316    | 0.0219  | 0.0200 | 0.0027 | 2:13:20 PM | Yes    |
| 2     | 0.327      | 0.323    | 0.0224  | 0.0205 | 0.0028 | 2:13:50 PM | Yes    |
| 3     | 0.319      | 0.314    | 0.0218  | 0.0199 | 0.0028 | 2:14:20 PM | Yes    |
| Mean: | 0.322      | 0.317    | 0.0220  |        |        |            |        |
| SD:   | 0.0047     | 0.0046   | 0.0003  |        |        |            |        |
| %RSD: | 1.46%      | 1.46%    | 1.45    |        |        |            |        |

Sequence No.: 65  
Sample ID: 24\_gas  
Analyst:

Autosampler Location: 61  
Date Collected: 12/16/2025 2:15:04 PM  
Data Type: Original

## Replicate Data: 24\_gas

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.456      | 0.451    | 0.0312  | 0.0293 | 0.0038 | 2:15:56 PM | Yes    |
| 2     | 0.414      | 0.410    | 0.0284  | 0.0265 | 0.0037 | 2:16:26 PM | Yes    |
| 3     | 0.435      | 0.430    | 0.0297  | 0.0278 | 0.0038 | 2:16:56 PM | Yes    |
| Mean: | 0.435      | 0.431    | 0.0298  |        |        |            |        |
| SD:   | 0.0205     | 0.0203   | 0.0014  |        |        |            |        |

%RSD: 4.72% 4.72% 4.68

Sequence No.: 66  
Sample ID: Tuna\_3  
Analyst:

Autosampler Location: 62  
Date Collected: 12/16/2025 2:17:41 PM  
Data Type: Original

## Replicate Data: Tuna\_3

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     | 3.908      | 4.944   | 0.3387  | 0.3368 | 0.0451 | 2:18:32 PM | Yes    |
| 2     | 3.849      | 4.869   | 0.3336  | 0.3317 | 0.0464 | 2:19:02 PM | Yes    |
| 3     | 3.995      | 5.053   | 0.3462  | 0.3443 | 0.0456 | 2:19:33 PM | Yes    |
| Mean: | 3.917      | 4.955   | 0.3395  |        |        |            |        |
| SD:   | 0.0732     | 0.0926  | 0.0063  |        |        |            |        |
| %RSD: | 1.87%      | 1.87%   | 1.87    |        |        |            |        |

Sequence No.: 67  
Sample ID: QAQC  
Analyst:

Autosampler Location: 6  
Date Collected: 12/16/2025 2:20:17 PM  
Data Type: Original

## Replicate Data: QAQC

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|---------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area   | Height |            | Stored |
| 1     |            | 1.033   | 0.0710  | 0.0691 | 0.0088 | 2:21:08 PM | Yes    |
| 2     |            | 0.935   | 0.0643  | 0.0624 | 0.0084 | 2:21:39 PM | Yes    |
| 3     |            | 0.931   | 0.0640  | 0.0621 | 0.0081 | 2:22:09 PM | Yes    |
| Mean: |            | 0.966   | 0.0664  |        |        |            |        |
| SD:   |            | 0.0577  | 0.0039  |        |        |            |        |
| %RSD: |            | 5.97%   | 5.94    |        |        |            |        |

QC value within limits for Hg 253.7 Recovery = 96.62%  
All analyte(s) passed QC.

Sequence No.: 68  
Sample ID: Standard blank  
Analyst:

Autosampler Location: 1  
Date Collected: 12/16/2025 2:22:52 PM  
Data Type: Original

## Replicate Data: Standard blank

Analyte: Hg 253.7

| Repl  | SampleConc | StdConc | Blncorr | Peak    | Peak   | Time       | Peak   |
|-------|------------|---------|---------|---------|--------|------------|--------|
| #     | ug/g       | ug/L    | Signal  | Area    | Height |            | Stored |
| 1     |            | 0.028   | 0.0022  | 0.0003  | 0.0001 | 2:23:41 PM | Yes    |
| 2     |            | 0.018   | 0.0015  | -0.0004 | 0.0000 | 2:24:11 PM | Yes    |
| 3     |            | 0.028   | 0.0022  | 0.0003  | 0.0001 | 2:24:41 PM | Yes    |
| Mean: |            | 0.024   | 0.0020  |         |        |            |        |
| SD:   |            | 0.0057  | 0.0004  |         |        |            |        |
| %RSD: |            | 23.52%  | 20.06   |         |        |            |        |

QC value less than the lower limit for Hg 253.7 Recovery = 2.44%  
QC Failed. Continue with analysis.  
User canceled analysis.

## Analysis Begun

Logged In Analyst: Administrator  
Spectrometer: FIMS-100, S/N B050-9550

Technique: AA FIMS-MHS  
Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 1

Autosampler Location: 6

Sample ID: QAQC

Date Collected: 12/16/2025 2:30:14 PM

Analyst:

Data Type: Original

User canceled analysis.

=====  
Analysis Begun

Logged In Analyst: Administrator

Technique: AA FIMS-MHS

Spectrometer: FIMS-100, S/N B050-9550

Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 67

Autosampler Location: 6

Sample ID: QAQC

Date Collected: 12/16/2025 2:30:53 PM

Analyst:

Data Type: Original

-----  
Replicate Data: QAQC

Analyte: Hg 253.7

| Repl<br># | SampleConc<br>ug/g | StdConc<br>ug/L | Blncorr<br>Signal | Peak<br>Area | Peak<br>Height | Time       | Peak<br>Stored |
|-----------|--------------------|-----------------|-------------------|--------------|----------------|------------|----------------|
| 1         |                    | 0.937           | 0.0644            | 0.0625       | 0.0082         | 2:31:44 PM | Yes            |
| 2         |                    | 0.960           | 0.0660            | 0.0641       | 0.0084         | 2:32:14 PM | Yes            |
| 3         |                    | 0.948           | 0.0652            | 0.0633       | 0.0084         | 2:32:44 PM | Yes            |
| Mean:     |                    | 0.949           | 0.0652            |              |                |            |                |
| SD:       |                    | 0.0117          | 0.0008            |              |                |            |                |
| %RSD:     |                    | 1.24%           | 1.23              |              |                |            |                |

QC value within limits for Hg 253.7 Recovery = 94.86%

All analyte(s) passed QC.

User canceled analysis.

=====  
Analysis Begun

Logged In Analyst: Administrator

Technique: AA FIMS-MHS

Spectrometer: FIMS-100, S/N B050-9550

Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 62

Autosampler Location: 58

Sample ID: 22\_lung

Date Collected: 12/16/2025 2:33:18 PM

Analyst:

Data Type: Original

-----  
Replicate Data: 22\_lung

Analyte: Hg 253.7

| Repl<br># | SampleConc<br>ug/g | StdConc<br>ug/L | Blncorr<br>Signal | Peak<br>Area | Peak<br>Height | Time       | Peak<br>Stored |
|-----------|--------------------|-----------------|-------------------|--------------|----------------|------------|----------------|
| 1         | 0.622              | 0.632           | 0.0435            | 0.0416       | 0.0056         | 2:34:09 PM | Yes            |
| 2         | 0.569              | 0.577           | 0.0398            | 0.0379       | 0.0052         | 2:34:40 PM | Yes            |
| 3         | 0.568              | 0.577           | 0.0398            | 0.0379       | 0.0050         | 2:35:10 PM | Yes            |
| Mean:     | 0.586              | 0.595           | 0.0410            |              |                |            |                |
| SD:       | 0.0309             | 0.0314          | 0.0022            |              |                |            |                |
| %RSD:     | 5.28%              | 5.28%           | 5.24              |              |                |            |                |

Sequence No.: 63

Autosampler Location: 59

Sample ID: 48\_gi

Date Collected: 12/16/2025 2:35:54 PM

Analyst:

Data Type: Original

## Replicate Data: 48\_gi

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.667      | 0.680    | 0.0469  | 0.0450 | 0.0060 | 2:36:46 PM | Yes    |
| 2     | 0.612      | 0.625    | 0.0430  | 0.0411 | 0.0057 | 2:37:16 PM | Yes    |
| 3     | 0.650      | 0.663    | 0.0457  | 0.0438 | 0.0060 | 2:37:46 PM | Yes    |
| Mean: | 0.643      | 0.656    | 0.0452  |        |        |            |        |
| SD:   | 0.0281     | 0.0286   | 0.0020  |        |        |            |        |
| %RSD: | 4.36%      | 4.36%    | 4.34    |        |        |            |        |

Sequence No.: 64

Sample ID: 32\_heart

Analyst:

Autosampler Location: 60

Date Collected: 12/16/2025 2:38:31 PM

Data Type: Original

## Replicate Data: 32\_heart

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.727      | 0.741    | 0.0510  | 0.0491 | 0.0065 | 2:39:22 PM | Yes    |
| 2     | 0.671      | 0.684    | 0.0471  | 0.0452 | 0.0062 | 2:39:52 PM | Yes    |
| 3     | 0.695      | 0.709    | 0.0488  | 0.0469 | 0.0063 | 2:40:22 PM | Yes    |
| Mean: | 0.698      | 0.711    | 0.0490  |        |        |            |        |
| SD:   | 0.0279     | 0.0284   | 0.0019  |        |        |            |        |
| %RSD: | 3.99%      | 3.99%    | 3.97    |        |        |            |        |

Sequence No.: 65

Sample ID: 44\_liver

Analyst:

Autosampler Location: 61

Date Collected: 12/16/2025 2:41:07 PM

Data Type: Original

## Replicate Data: 44\_liver

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.703      | 0.735    | 0.0506  | 0.0487 | 0.0065 | 2:41:58 PM | Yes    |
| 2     | 0.629      | 0.657    | 0.0453  | 0.0434 | 0.0062 | 2:42:29 PM | Yes    |
| 3     | 0.672      | 0.702    | 0.0483  | 0.0464 | 0.0064 | 2:42:59 PM | Yes    |
| Mean: | 0.668      | 0.698    | 0.0481  |        |        |            |        |
| SD:   | 0.0374     | 0.0391   | 0.0027  |        |        |            |        |
| %RSD: | 5.60%      | 5.60%    | 5.56    |        |        |            |        |

Sequence No.: 66

Sample ID: 6\_liver

Analyst:

Autosampler Location: 62

Date Collected: 12/16/2025 2:43:43 PM

Data Type: Original

## Replicate Data: 6\_liver

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | Blncorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.699      | 0.710    | 0.0489  | 0.0470 | 0.0053 | 2:44:35 PM | Yes    |
| 2     | 0.408      | 0.414    | 0.0286  | 0.0267 | 0.0039 | 2:45:05 PM | Yes    |
| 3     | 0.424      | 0.430    | 0.0297  | 0.0278 | 0.0039 | 2:45:35 PM | Yes    |
| Mean: | 0.510      | 0.518    | 0.0357  |        |        |            |        |
| SD:   | 0.1636     | 0.1661   | 0.0114  |        |        |            |        |
| %RSD: | 32.07%     | 32.07%   | 31.81   |        |        |            |        |

User canceled analysis.

## Analysis Begun

Logged In Analyst: Administrator

Spectrometer: FIMS-100, S/N B050-9550

Technique: AA FIMS-MHS

Autosampler: S10

Sample Information File: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Sample Information\  
15dec2025\_pumaorgan\_er.sifx

Batch ID: FAS AOT

Results Data Set: results\_15dec2025\_puma\_er

Results Library: C:\Users\Public\PerkinElmer Syngistix\AA\Data\Results\Results\_15dec2025\_puma\_er.mdb

Sequence No.: 66

Sample ID: 6\_liver

Analyst:

Autosampler Location: 62

Date Collected: 12/16/2025 2:46:13 PM

Data Type: Original

Replicate Data: 6\_liver

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | BlkCorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     | 0.452      | 0.459    | 0.0317  | 0.0298 | 0.0041 | 2:47:04 PM | Yes    |
| 2     | 0.444      | 0.450    | 0.0311  | 0.0292 | 0.0041 | 2:47:35 PM | Yes    |
| 3     | 0.428      | 0.435    | 0.0301  | 0.0282 | 0.0039 | 2:48:05 PM | Yes    |
| Mean: | 0.441      | 0.448    | 0.0310  |        |        |            |        |
| SD:   | 0.0119     | 0.0121   | 0.0008  |        |        |            |        |
| %RSD: | 2.69%      | 2.69%    | 2.67    |        |        |            |        |

Sequence No.: 67

Sample ID: QAQC

Analyst:

Autosampler Location: 6

Date Collected: 12/16/2025 2:48:49 PM

Data Type: Original

Replicate Data: QAQC

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | BlkCorr | Peak   | Peak   | Time       | Peak   |
|-------|------------|----------|---------|--------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area   | Height |            | Stored |
| 1     |            | 0.938    | 0.0645  | 0.0626 | 0.0082 | 2:49:40 PM | Yes    |
| 2     |            | 0.903    | 0.0621  | 0.0602 | 0.0081 | 2:50:10 PM | Yes    |
| 3     |            | 0.934    | 0.0642  | 0.0623 | 0.0082 | 2:50:41 PM | Yes    |
| Mean: |            | 0.925    | 0.0636  |        |        |            |        |
| SD:   |            | 0.0189   | 0.0013  |        |        |            |        |
| %RSD: |            | 2.04%    | 2.03    |        |        |            |        |

QC value within limits for Hg 253.7 Recovery = 92.51%  
All analyte(s) passed QC.

Sequence No.: 68

Sample ID: Standard blank

Analyst:

Autosampler Location: 1

Date Collected: 12/16/2025 2:51:24 PM

Data Type: Original

Replicate Data: Standard blank

Analyte: Hg 253.7

| Repl  | SampleConc | StndConc | BlkCorr | Peak    | Peak   | Time       | Peak   |
|-------|------------|----------|---------|---------|--------|------------|--------|
| #     | ug/g       | ug/L     | Signal  | Area    | Height |            | Stored |
| 1     |            | 0.034    | 0.0026  | 0.0007  | 0.0001 | 2:52:13 PM | Yes    |
| 2     |            | -0.015   | -0.0008 | -0.0027 | 0.0000 | 2:52:43 PM | Yes    |
| 3     |            | 0.028    | 0.0022  | 0.0003  | 0.0002 | 2:53:13 PM | Yes    |
| Mean: |            | 0.016    | 0.0014  |         |        |            |        |
| SD:   |            | 0.0271   | 0.0019  |         |        |            |        |
| %RSD: |            | 173.38%  | 136.50  |         |        |            |        |

QC value less than the lower limit for Hg 253.7 Recovery = 1.56%  
QC Failed. Continue with analysis.