

DEPARTMENTAL UPDATE REPORT

SUBMITTED TO : DR. RAZEEB HASSAN (MEDICAL DIRECTOR)

SERVICES : BIOMEDICAL SERVICE

SUBJECT : UPDATE REPORT OF MARCH' 2026

Dear Sir,

As referred to the above subject kindly find below the summary of departmental activities –

A. Man Power:

SL	Name	Designation	Date of join
1	Suman Kumar Das	Assistant Manager- Biomedical	14 March,2014
2	Md Zahangir Alam	Assistant Biomedical Engineer	01 December,2016
3	Md. Rakib Hossain	Assistant Biomedical Engineer	01 December, 2021

B. New Equipment Installation Update:

SL	Equipment Name	Brand	Quantity
1	Syringe Pump	Mindray	05
2	Patient Monitor	Mindray	04
3	Voltage Stabilizer 10Kva	SAKO	02
4	Air Pump	Rocket	01

C. In House Breakdown & PPM Service (As per PMS):

S/L	Request No	Dept.	Request	Job Done	Done By
1	sf-bms-0292-59198	OPD	Calibration on Audiometry machine	Yes	BMES TEAM
2	sf-bms-8585-21572	OT	Check and solve the Laparoscopy display problem	Yes	BMES TEAM
3	sf-bms-8528-27333	CSSD	Training for OT and CSSD staff on how to clean medical instruments by air machine.	Yes	BMES TEAM
4	sf-bms-8445-90983	OT	C-arm machine and found image storage full. We empty image storage.	Yes	BMES TEAM
5	sf-bms-0171-26341	IPD	Nebulizer machine and found a motor problem need to be repair this motor	In progress	BMES TEAM

S/L	Request No	Department	Request	Job Done	Done By
6	sf-bms-6277-54960	OT	OT light adjustment and anesthesia machine calibration	Yes	BMES TEAM
7	sf-bms-6202-79238	OT	OT light adjustment	Yes	BMES TEAM
8	sf-bms-5414-72185	LAB	Distilled water plant filter replace		BMES TEAM
9	sf-bms-8864-89109	Endoscopy	Spot remove from colonoscopy video lens.	Yes	BMES TEAM
10	sf-bms-7692-68023	Dialysis	Dialyzer re-processor leakage repair.	Yes	BMES TEAM
11	sf-bms-7618-83924	Dialysis	Dialysis machine air leakage repair.	Yes	BMES TEAM
12	sf-bms-7501-13473	Dialysis	RO machine filter replace.	Yes	BMES TEAM
13	sf-bms-7440-43389	Dialysis	Dialysis machine outside water leakage repair and PPM .	Yes	BMES TEAM
14	sf-bms-7377-10054	Dialysis	Electric patient bed power supply repair	Yes	BMES TEAM
15	sf-bms-6554-61233	ICU	Electric patient bed power supply repair	Yes	BMES TEAM
16	sf-bms-6507-55306	ICU	03 Electric patient bed power supply repair	Yes	BMES TEAM
17	sf-bms-2010-68105	OPD	USG machine printer setting	Yes	BMES TEAM
18	sf-bms-2082-51552	A&E	Ventilator Machine Calibration	Yes	BMES TEAM
19	sf-bms-0544-96035	ICU	Patient monitor power supply burn out for high voltage in ICU.	In progress	BMES TEAM
20	sf-bms-0471-37185	ICU	New two Patient monitor (mindray) power supply burn out for high voltage in ICCU.	In progress	BMES TEAM
21	sf-bms-0402-85035	ICU	New one Ventilator (Aeonmed) Machine power supply burn out for high voltage in ICU.	In progress	BMES TEAM
22	sf-bms-0275-75107	ICU	3 unit Ventilator (Hamilton) Machine power supply burn out for high voltage in ICU	In progress	BMES TEAM
23	sf-bms-0219-45953	ICU	3 unit Electric patient power supply burn we repair	Yes	BMES TEAM
24	sf-bms-8056-70330	Dialysis	Dialyzer re-processor leakage repair.	Yes	BMES TEAM

S/L	Request No	Department	Request	Job Done	Done By
25	sf-bms-0219-48053	CCU	Oxygen flow meter (Cylinder type) connector broken we repair.	Yes	BMES TEAM
26	sf-bms-7237-19367	ICU	Ventilator (Hamilton) battery repair	Yes	BMES TEAM
27	sf-bms-7172-84776	ICU	Cardiac monitor power supply need to repair	In progress	
28	sf-bms-5549-41810	NICU	Baby warmer and CPAP machine PPM	Yes	BMES TEAM
29	sf-bms-5481-38615	NICU	Ventilator(Aeonmed) FIO2 Module Calibration	Yes	BMES TEAM
30	sf-bms-5285-38966	NICU	Ventilator(Velavista) calibration Oxygen sensor calibration And PPM	Yes	BMES TEAM
31	sf-bms-5253-23920	NICU	Ventilator(Drager) calibration	Yes	BMES TEAM
32	sf-bms-3525-52255	CCU	Patient bed power repair	Yes	BMES TEAM
33	sf-bms-9286-85971	Dialysis	R O Water Back Wash	Yes	BMES TEAM
34	sf-bms-5141-61808	A&E	Syringe Pump(Mindray) installation and testing & commissioning	Yes	BMES TEAM
35	sf-bms-0792-31699	Dialysis	B-Braun Dialysis Machine motor repair.	Yes	BMES TEAM
36	sf-bms-3693-93797	R&I	MRI UPS PPM	Yes	BMES TEAM
37	sf-bms-8853-55262	SPD	ETT machine software re-installation	Yes	BMES TEAM
38	sf-bms-8816-29414	SPD	ECG machine repair	Yes	BMES TEAM
39	sf-bms-9370-56695	IPD	Patient bed repair in room 803.	Yes	BMES TEAM
40	sf-bms-3489-14308	L:AB	RO Water plant Filter replaced	Yes	BMES TEAM
41	sf-bms-4464-92364	Dialysis	Ro Water Back Wash.	Yes	BMES TEAM
42	sf-bms-3272-41073	SPD	Portable ECHO machine software unlock	Yes	BMES TEAM
43	sf-bms-5806-28166	Maternity suit	Neonatal vital sing motor NIBP cup leakage repair.	Yes	BMES TEAM
44	sf-bms-7465-46824	Endoscopy	Endoscopic LED monitor Installation.	Yes	BMES TEAM

S/L	Request No	Department	Request	Job Done	Done By
45	sf-bms-7387-10649	Endoscopy	EEG & NCV equipment's transferred from Endoscopy to 8th Floor.	Yes	BMES TEAM
46	sf-bms-6494-51227	NICU	ventilator (Drager)calibration	Yes	BMES TEAM
47	sf-bms-6347-60185	NICU	High Flow nasal cannula humidifier repair.	Yes	BMES TEAM
48	sf-bms-6300-23476	NICU	Baby warmer check and PPM	Yes	BMES TEAM
49	sf-bms-0303-98982	LAB	Microbiology lab Biosafety cabinet air filter need to replace.	In progress	BMES TEAM
50	sf-bms-7980-88908	OT	EYE OT Equipment check and PPM	Yes	BMES TEAM
51	sf-bms-7911-93526	CSSD	RO water filter replace in CSSD	Yes	BMES TEAM
52	sf-bms-7739-50415	OT	Anesthesia machine calibration	Yes	BMES TEAM
53	sf-bms-7585-45029	IPD	Cylinder type oxygen cylinder leakage repair.	Yes	BMES TEAM
54	sf-bms-5496-17700	Dialysis	Re Processor Machine air leakage repair	Yes	BMES TEAM
55	sf-bms-4339-54397	A&E	Medication Refrigerator temperature controller repair.	Yes	BMES TEAM
56	sf-bms-6591-52528	CSSD	Autoclave machine water filter replace.	Yes	BMES TEAM
57	sf-bms-6537-73276	OT	GA machine Calibration in OT-1	Yes	BMES TEAM
58	sf-bms-3751-38383	Maternity suit	Colposcopy Printer repair	Yes	BMES TEAM
59	sf-bms-3707-74761	Maternity suit	Colposcopy Reporting Software installation.	Yes	BMES TEAM
60	sf-bms-7078-72032	OT	Anesthesia machine Oxygen sensor calibration in OT-2	Yes	BMES TEAM
61	sf-bms-3522-31167	A&E	USG machine probe repair.	Yes	BMES TEAM
62	sf-bms-6335-51937	IPD	Patient bed problem repair in room no 640	Yes	BMES TEAM
63	sf-bms-4102-39881	A&E	Syringe pump battery repair	In progress	BMES TEAM
64	sf-bms-8539-77041	IPD	Patient bed problem repair in room no 635	Yes	BMES TEAM
65	sf-bms-5355-69293	Maternity suit	colposcopy reporting software re-installation	Yes	BMES TEAM
66	sf-bms-2975-86163	IPD	7th floor CD block dressing room spot light power supply need to replace.	In progress	BMES TEAM

S/L	Request No	Department	Request	Job Done	Done By
67	sf-bms-1560-83696	OT	Exercise Chair power supply need to replace.	In progress	BMES TEAM
68	sf-bms-9533-68926	Maternity suit	Doctor name add to Colposcopy software reporting software	Yes	BMES TEAM
69	sf-bms-8717-22147	Endoscopy	Endoscopy Monitor exchanged.	Yes	BMES TEAM
70	sf-bms-5187-10672	Maternity suit	Colposcopy machine reporting software installation.	Yes	BMES TEAM
71	sf-bms-3703-61336	IPD	6th Floor AB Block Pulse Oximeter SPO2 Probe Replace.	Yes	BMES TEAM
72	sf-bms-2385-22372	OT	UPS mother board and charging card repair.	Yes	BMES TEAM
73	sf-bms-5987-85527	IPD	Patient bed repair in room no 801 and 803	Yes	BMES TEAM
74	sf-bms-2284-61393	NICU	HI flow machine humidifier repair.	Yes	BMES TEAM
75	sf-bms-8354-11228	Dialysis	Ro Water Back Wash.	Yes	BMES TEAM
76	sf-bms-8312-47192	Dialysis	Ro machine PPM.	Yes	BMES TEAM
77	sf-bms-9265-19917	Physiotherapy	IRR repair	Yes	BMES TEAM
78	sf-bms-5391-37716	Endoscopy	Laryngoscope lighting replace.	Yes	BMES TEAM
79	sf-bms-3639-97636	ICU Surgery	5th floor S.ICU one cardiac monitor (advance) SPO2 module repair.	Yes	BMES TEAM
80	sf-bms-9217-54399	OPD	OAE machine calibration	Yes	BMES TEAM
81	sf-bms-4934-37715	A&E	Cylinder type flow meter repair.	Yes	BMES TEAM
82	sf-bms-4852-30270	A&E	laryngoscope repair	Yes	BMES TEAM
83	sf-bms-2358-73639	A&E	Medication refrigerator controller repair	Yes	BMES TEAM
84	sf-bms-9635-20207	OT	Drager patient monitor calibration.	Yes	BMES TEAM
85	sf-bms-9590-84380	OT	Anesthesia machine calibration.	Yes	BMES TEAM
86	sf-bms-1280-64669	IPD	Baby weight scale power supply repair.	Yes	BMES TEAM

S/L	Request No	Department	Request	Job Done	Done By
87	sf-bms-1251-50926	IPD	Portable Sport light lamp replace.	Yes	BMES TEAM
88	sf-bms-1155-21340	IPD	New Oxygen flow meter(wall type) installation	Yes	BMES TEAM
89	sf-bms-1094-68546	IPD	New Syringe pump installation and user training	Yes	BMES TEAM
90	sf-bms-8461-70097	ICU M	patient bed repair	Yes	BMES TEAM
91	sf-bms-5503-52828	SPD	ETT machine software installation	Yes	BMES TEAM
92	sf-bms-8460-39973	SPD	Machine invasive pressure dome repair.	Yes	BMES TEAM
93	sf-bms-8401-15610	SPD	ETT room printer repair.	Yes	BMES TEAM
94	sf-bms-1158-21342	IPD	New Oxygen flow meter(wall type) installation	Yes	BMES TEAM
95	sf-bms-2556-81033	Dialysis	Dialysis machine water leakage repair.	Yes	BMES TEAM
96	sf-bms-9907-49005	R&I	Fluoroscopy machine Shifting	Yes	BMES TEAM
97	sf-bms-0007-73115	R&I	BMD machine shifting and reinstallation	Yes	BMES TEAM
98	sf-bms-0060-29373	R&I	Portable X-ray power supply problem	Yes	BMES TEAM

D. Vendor Breakdown Service:

SL	Job No	Dept.	Request	Price(BDT)	Job Done	Remarks
1	507	Dialysis	Nipro BTB assembly part	16,500	Yes	
2	508	CSSD	RO membrane & switch	13,643	Yes	
3	509	Dialysis	Gear pump B. B	30,000	Yes	
4	510	R&I	MRI PSU	20,00,000	No	
5	511	R&I	BMD Reinstallation	50,000	Yes	
Total				21,10,143		

E. Demonstration Update:

- ☐ Endoscopy and Colposcopy system
- ☐ Body warmer

F. Medical Gas Update:
Liquid Oxygen filled in March: 5,616 Liters

Date	Refill Gas in liters	Remarks
2-Mar-26	598	
5-Mar-26	630	
7-Mar-26	504	
12-Mar-26	1106	
16-Mar-26	630	
19-Mar-26	378	
24-Mar-26	945	
29-Mar-26	825	
Total	5,616 Liters	

☐ **Portable cylinder filled in March:**

S/N	Cylinder Name	Quantity	Remarks
01	Oxygen Bullnose Type	53	
02	Oxygen Pin Type	0	
03	Oxygen Manifold	0	
04	Nitrous Oxide Pin Type	0	
05	Nitrous Oxide Manifold	0	
06	Liquid Nitrous Oxide (20L)	01	
07	Carbon Dioxide	03	
08	Medical Air	0	
Total		57	

Above report will give a summarize idea of our departmental activities for the month of March 2026.

Sincerely yours

Suman Kumar Das

Assistant Manager- Biomedical

KPJ Specialized Hospital