

DEPARTMENTAL UPDATE REPORT

SUBMITTED TO : DR. RAZEEB HASSAN (MEDICAL DIRECTOR)

SERVICES : BIOMEDICAL SERVICE

SUBJECT : UPDATE REPORT OF APRIL' 2026

Dear Sir,

As referred to the above subject kindly find below the summary of departmental activities –

A. Man Power:

SL	Name	Designation	Date of join
1	Suman Kumar Das	Assistant Manager- Biomedical	14 March,2014
2	Md Zahangir Alam	Assistant Biomedical Engineer	01 December,2016
3	Md. Rakib Hossain	Assistant Biomedical Engineer	01 December, 2021

B. New Equipment Installation Update:

SL	Equipment Name	Brand	Quantity
1	Endoscopic System	Olympus	01
2	Body warmer	Lifotronic	01
3	OT Light	Weyuan	02
4	OT Table	Triup	01
5	OT Table	OET	01
5	Eight Function Electric patient bed	EB-8X	02
6	Five Function Electric patient bed	EB-3E	05
6	Three Function Electric patient bed	EB-3	30
7	Medication Trolley	YMHM02	04

C. In House Breakdown & PPM Service (As per PMS):

S/L	Request No	Dept.	Request	Job Done	Done By
1	sf-bms-2310-99521	Dialysis	Dialysis bed power supply repair	Yes	BMES TEAM
2	sf-bms-0357-87573	OPD	Camera Unit Halogen Light Source repair	Yes	BMES TEAM
3	sf-bms-7621-26597	Dialysis	RO machine pipe connector Repair	Yes	BMES TEAM
4	sf-bms-8057-36201	OPD	Dental unit chair repair and PPM	Yes	BMES TEAM
5	sf-bms-9171-62918	Maternity	Nebulizer machine repair.	Yes	BMES TEAM
6	sf-bms-6646-92617	OT	Covidien Diathermy mother board repair	In progress	BMES TEAM
7	sf-bms-1632-21927	ICU	ICU patient bed power supply repair	Yes	BMES TEAM
8	sf-bms-5316-65295	OT	C-arm machine repair and PPM		BMES TEAM
9	sf-bms-5283-86622	R&I	CT SCAN Error code remove	Yes	BMES TEAM
10	sf-bms-5248-28896	R&I	MRI CHILER repair	Yes	BMES TEAM
11	sf-bms-1469-19992	Endoscopy	Endoscopy machine repair	Yes	BMES TEAM
12	sf-bms-4871-92248	LAB	Lab Biosafety cabin Air Damage need to replace.	In progress	BMES TEAM
13	sf-bms-2156-37547	OT	N20 connector replaced	Yes	BMES TEAM
14	sf-bms-2123-60100	CSSD	Autoclave controller repair.	Yes	BMES TEAM
15	sf-bms-2952-13515	A&E	Syringe pump head drive lock repair	Yes	BMES TEAM
16	sf-bms-2502-20190	NICU	Vela vista Ventilator Calibration and PPM	Yes	BMES TEAM
17	sf-bms-6238-52353	ICU	Ventilator Oxygen connector replaced	Yes	BMES TEAM
18	sf-bms-6191-11816	ICU	Electric patient bed power supply repair.	Yes	BMES TEAM
19	sf-bms-6865-41089	Dialysis	RO Water Backwash.	Yes	BMES TEAM
20	sf-bms-0467-75647	NICU	Ventilator (Drager) flow sensor calibration and PPM	Yes	BMES TEAM
21	sf-bms-7455-61245	CCU	Digital Height & Weight Repair	Yes	BMES TEAM
22	sf-bms-7373-20425	CCU	Patient Bed power supply repair.	In progress	BMES TEAM
23	sf-bms-5156-78501	OT	UPS repair	Yes	BMES TEAM
24	sf-bms-1566-62396	OT	Anesthesia machine (Drager) Calibration	Yes	BMES TEAM
25	sf-bms-1262-68551	NICU	Ventilator (Drager) flow sensor calibration.	Yes	BMES TEAM
26	sf-bms-0780-67184	LAB	RO water filter replace.	Yes	BMES TEAM
24	sf-bms-1566-62396	OT	Anesthesia machine (Drager) Calibration	Yes	BMES TEAM

S/L	Request No	Department	Request	Job Done	Done By
24	sf-bms-1566-62396	OT	Anesthesia machine (Drager) Calibration	Yes	BMES TEAM
25	sf-bms-1262-68551	NICU	Ventilator (Drager) flow sensor calibration.	Yes	BMES TEAM
26	sf-bms-0780-67184	LAB	RO water filter replace.	Yes	BMES TEAM
27	sf-bms-8090-69689	A&E	laryngoscope repair	Yes	
28	sf-bms-8044-93861	A&E	BP machine repair	Yes	BMES TEAM
29	sf-bms-8014-95962	A&E	Nebulizer Machine Repair	Yes	BMES TEAM
30	sf-bms-8712-80175	OT	N02 line leakage check and repair.	Yes	BMES TEAM
31	sf-bms-1226-64307	LAB	Embedding center repair.	Yes	BMES TEAM
32	sf-bms-6734-73538	A&E	Vacuum Regulator Adjustment	Yes	BMES TEAM
33	sf-bms-6731-25237	A&E	Ventilator machine calibration	Yes	BMES TEAM
34	sf-bms-4309-23159	A&E	Laryngoscope blade Check	Yes	BMES TEAM
35	sf-bms-4362-84490	OT	Laparoscopy machine monitor repair	Yes	BMES TEAM
36	sf-bms-5809-57151	Physiotherapy	Treadmill UPS repair	Yes	BMES TEAM
37	sf-bms-5849-74837	SDP	ETT software reinstallation and report format setting	Yes	BMES TEAM
38	sf-bms-5793-54707	SPD	ECG Cable replace	Yes	BMES TEAM
39	sf-bms-8883-51088	ICU	Cylinder Oxygen flow meter repair	Yes	BMES TEAM
34	sf-bms-4309-23159	A&E	Laryngoscope blade Check	Yes	BMES TEAM
35	sf-bms-4362-84490	OT	Laparoscopy machine monitor repair	Yes	BMES TEAM
36	sf-bms-5809-57151	Physiotherapy	Treadmill UPS repair	Yes	BMES TEAM
37	sf-bms-5849-74837	SDP	ETT software reinstallation and report format setting	Yes	BMES TEAM
38	sf-bms-5793-54707	SPD	ECG Cable replace	Yes	BMES TEAM
39	sf-bms-8883-51088	ICU	Cylinder Oxygen flow meter repair	Yes	BMES TEAM
40	sf-bms-8739-31141	ICU	Central Oxygen regulator repair	Yes	BMES TEAM
41	sf-bms-0060-29373	R&I	Portable x-ray stabilizer installation.	Yes	BMES TEAM
42	sf-bms-0007-73115	R&I	BMD re installation needed	Yes	BMES TEAM
43	sf-bms-9907-49005	R&I	Fluoroscopy x-ray dismantling needed	Yes	BMES TEAM
44	sf-bms-0292-59198	NICU	Audiometry machine Calibration	Yes	BMES TEAM
45	sf-bms-8585-21572	OT	Laparoscopy monitor repair	Yes	BMES TEAM
46	sf-bms-8528-27333	CSSD	Instruments air machine operational Training	Yes	BMES TEAM
47	sf-bms-8445-90983	OT	C-arm machinery repair	Yes	BMES TEAM

D. Vendor Breakdown Service:

SL	Job No	Dept.	Request	Price(BDT)	Job Done	Remarks
1	512	R&I	MRI ramp up/ down	11,00,000	Yes	
2	513	Dialysis	Nipro BtB part	16,500	Yes	
3	514	ICU	Ventilator -2 (H)	4,77,000	No	
4	515	R&I	Stabilizer 10Kva	2,600	Yes	
5	516	OT	UPS 3KVA	3,400	No	
6	517	R&I	Portable X ray light	48,000	No	
7	518	Lab	Refrigerator (panasonic)	3,000	Yes	
8	519	OT	Refrigerator (Haier)	13,000	Yes	
Total				16,63,500		

E. Demonstration Update:

- ☐ Body warmer
- ☐ Syringe Pump
- ☐ Infusion Pump
- ☐ Multi parameter Patient Monitor
- ☐ Video Laryngoscope
- ☐ DVT Pump
- ☐ Cardiac Monitor

F. Medical Gas Update:

Liquid Oxygen filled in April: 5,435 Liters

Date	Refill Gas in liters	Remarks
1-Apr-26	473	
6-Apr-26	910	
9-Apr-26	630	
13-Apr-26	599	
16-Apr-26	441	
20-Apr-26	693	
23-Apr-26	518	
28-Apr-26	1171	
Total	5,435 Liters	

☐ **Portable cylinder filled in August**

S/N	Cylinder Name	Quantity	Remarks
01	Oxygen Bullnose Type	65	
02	Oxygen Pin Type	0	
03	Oxygen Manifold	0	
04	Nitrous Oxide Pin Type	1	
05	Nitrous Oxide Manifold	0	
06	Liquid Nitrous Oxide (20L)	0	
07	Carbon Dioxide	06	
08	Medical Air	0	
	Total	72	

Above report will give a summarize idea of our departmental activities for the month of March 2026.

Sincerely yours
Suman Kumar Das
Assistant Manager- Biomedical
KPJ Specialized Hospital