

31st May 2018

- Wiping of all incubator and laminar hood surfaces with 70% isopropanol.
 - ↳ This incubator does not have an automatic heat sterilization cycle → only 1st wipe-down of all surfaces
- The tray has been filled with autoclaved distilled water.

→ The CO_2 valve has been turned on much later due to a negligent error - left overnight to equilibrate.