

Certificate of Analysis

Product name:

Zero Oxygen Solution Set (2 components)

Product code:

HI7040L

Lot number:

50021/18

Best use before:

or one month after mixing

Date of analysis:

March 2023 2018/03/14

Certified value, % Oz: 0.0 ± 0.1 at 25 °C

Method of standardization:

This quality product is standardized by lodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-0 B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty $\ensuremath{\textbf{U}}$ with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Components racing in the contract of the contra			
Component code: Lot number:	Lot number:	Expiration Date:	
HI7040-1	2684	March 2023	
HI7040-2	2681	March 2023	

Reference number: 13C81

QA manager:

Andrea Coman

Hanna Instruments Inc. 584 Park East Drive Woonsocket, RI 02895 www.hannainst.com