

Journal of Medical Engineering & Technology




ISSN: (Print) (Online) Journal homepage: https://www.tandfonline.com/loi/ijmt20

Role of machine learning in gait analysis: a review

Preeti Khera & Neelesh Kumar

To cite this article: Preeti Khera & Neelesh Kumar (2020): Role of machine learning in gait analysis: a review, Journal of Medical Engineering & Technology, DOI: 10.1080/03091902.2020.1822940

To link to this article: https://doi.org/10.1080/03091902.2020.1822940



Taylor & Francis Taylor & Francis Group

RESEARCH ARTICLE



Role of machine learning in gait analysis: a review

Preeti Khera^{a,b} and Neelesh Kumar^{a,b}

^aAcademy of Scientific and Innovative Research, Ghaziabad, India: ^bBiomedical Instrumentation Unit, CSIR-CSIO, Chandigarh, India

ABSTRACT

Human biomechanics and gait form an integral part of life. The gait analysis involves a large number of interdependent parameters that were difficult to interpret due to a vast amount of data and their inter-relations. To simplify evaluation, the integration of machine learning (ML) with biomechanics is a promising solution. The purpose of this review is to familiarise the readers with key directions of implementation of ML techniques for gait analysis and gait rehabilitation. An extensive literature survey was based on research articles from nine databases published from 1980 to 2019. With over 943 studies identified, finally, 43 studies met the inclusion criteria. The outcome reported illustrates that supervised ML techniques showed accuracies above 90% in the identified gait analysis domain. The statistical results revealed support vector machine (SVM) as the best classifier (mean-score = 0.87 ± 0.07) with remarkable generalisation capability even on small to medium datasets. It has also been analysed that the control strategies for gait rehabilitation are benefitted from reinforcement learning and (deep) neural-networks due to their ability to capture participants' variability. This review paper shows the success of ML techniques in detecting disorders, predicting rehabilitation length, and control of rehabilitation devices which make them suitable for clinical diagnosis.

ARTICLE HISTORY

Received 26 June 2020 Revised 8 September 2020 Accepted 9 September 2020

KEYWORDS

Artificial intelligence (AI); machine learning (ML); gait analysis; gait rehabilitation; pathology detection

1. Introduction

Human mobility is significant to perceive quality life and degradation to it can lead to abnormalities. The study of human biomechanics termed as gait analysis aimed at quantifying factors governing the functionality of lower limbs. This is crucial for the detection of gait abnormalities, recognition of postural instabilities, and assessment of clinical interventions and rehabilitation programs. The clinicians anticipate gait analysis for diagnosis and treatment selection but have been challenged by underlined complexities such as volume and non-unique correlations between an individual's gait. Therefore, it is evident that gait analysis would be benefitted from artificial intelligence (AI) techniques. These techniques are capable enough in handling high dimensional, temporal and complex data [1,2]. They build models that learn automatically from available databases, make accurate predictions, and behave intelligently [3,4]. The machine learning (ML) is being extensively used in multiple fields such as medical diagnosis [5-7], pattern recognition [8,9], image processing [10,11], classification [12–14], prediction analysis [15–17], monitoring [12,18], therefore, making suitable for gait studies. Although,

techniques have been used for numerous gait applications, viz. to diagnose gait disorders [14,19,20], to predict early intervention for fall-related risks due to a disability or aging [3,6,21], to determine motor recovery tasks [22,23], or in planning a rehabilitation or therapeutic interventions [24–26].

Al techniques are used to analyse the data, images, and recognise patterns by creating algorithms that help physicians to diagnose a specific disease/disorder timely and accurately. Moreover, such algorithms learn continuously, thus improving the diagnostic results. However, sometimes physicians need to simultaneously consider symptoms of the patient, treatment plans, potential side effects, another disease with similar signs, previous medical history, and many other aspects. The Al-technology, in turn, provides a solution to assist physicians by analysing a vast amount of data and ensure a holistic understanding of patient health records. Even AI techniques can predict the onset of disease by processing the large volume of data gathered longitudinally by monitoring the health status of an individual. The surgical robots to handle many complex operations are the advent of AI techniques for disease management. Thus, the ML models built using complex relationships can reduce diagnosis

time, enhance patient monitoring, and aids the clinicians in the selection of treatment but requires computing facilities at the implementation site. Nevertheless, early diagnosis can prevent loss of mobility [1] and can reduce healthcare costs which are the rising concern of developing countries [27].

However, the success rate of deployment of such techniques in its initial phase of evolvement was less due to a lack of technological advent in data processing. The data collection and storage capabilities were unmet. The gait analysis was limited to laboratories using video-based systems. Thereafter, with the technological advent vision and sensor capture methods for gait analysis were developed which led to the accumulation of a large number of interdependent parameters. Thus, a need to process such temporal and highly complex data comes in. Although statistical tools are used widely in gait analysis, they lack predictive power to generalise on unseen data. Hence, the integration of ML and biomedical gait analysis can help in providing the solution.

A challenge is to determine how well a machine performs the tasks on the unknown dataset. The variety of supervised and unsupervised learning techniques are explored for classification and prediction tasks using approaches like vision-based [21,24] and sensor-based [20,28,29]. The dynamic learning, i.e., reinforcement learning (RL) can be used to achieve better tracking control performance for rehabilitation strategies [30]. Another aspect of ML that cannot be left intact is an algorithm devised for feature extraction and selection phase [14,23,31] that can be used to enhance computational efficiency and reduces the burden of classification or learning by reducing computing intervals.

The main objectives of this review are highlighted below:

- 1. To identify the studies that have used AI or ML techniques in gait analysis and rehabilitation, and conduct a selection criterion for their inclusion.
- 2. To classify the gait studies based on recent applications areas in this domain which are categorised as gait activity detection (GAD), gait event detection (GED), gait disorder detection (a) due to pathology (GDDP) (b) due to aging (GDDA), gait asymmetry detection (GAsD) and neurological effects on gait (GNE), to assess the performance capabilities and their limitations. Future directions for the researchers are also highlighted.
- 3. To study the qualitative analysis of ML techniques in gait rehabilitation for the intent detection to

- generate the triggering and feedback for trajectory control.
- To determine which AI or ML techniques are most widely used in gait analysis and rehabilitation based on performance measures.

The review structure is as follows: Section 2 describes the basic definition of ML models, types and applicability in gait-related studies, which are used in this review. Section 3 presents the steps of data collection. Section 4 describes a summary of studied literature, highlighting the outcomes, limitations and active areas of research in the gait domain using ML techniques. Section 5 mentions the advantages of the usage of ML in gait and future expectations. This review is expected to provide a significant impact in the field of human biomechanics to promote the clinical application of ML techniques to aid the diagnosis and assessment of lower limb pathologies.

2. Machine learning in gait studies

The growth of ML in gait analysis is modelling a biomechanical system T(x) by determination of the relationship between input data f(x) and outputs y(x), although the input data are corrupted by noise n(t)which demands the need for preprocessing of the input data. The input data are raw multidimensional array $\{u_i, v_i\}$, where, u_i corresponds to a number of subjects or their trails, and v_i depicts the data features such as kinematics, kinetics or neuromuscular signals. The model output is representative of the classification of gait events, activities or disorders. Evaluation of biomechanical system T(x) using ML techniques uses an iterative process where the input dataset is divided into a training set, testing set and validation set. After the selection of a particular ML technique, the model is trained using a training set and validated to determine the level of the fitting. Finally, the performance is evaluated on an unseen test dataset. When suitable accuracy is reached the process is terminated otherwise the model parameters need retuning and retraining until the desired accuracy is achieved. Too many parameters lead to system complexity; therefore, feature selection can be done. The most commonly used techniques include supervised, unsupervised and RL.

2.1. Supervised learning

In this type of learning, a feature vector comprises of labelled data with the aim to determine a function that best maps the relation between input feature vectors and the corresponding label. Various algorithms that were explored in gait studies consist of support vector machine (SVM), neural networks (NNs), random forest (RF), hidden Markov model (HMM), ensemble learning, k-nearest neighbour (kNN) and decision trees (DTs). The popularity of using SVM for gait analysis was that it has good generalisation capability even for datasets that are not too large. It has kernels to deal with linear and nonlinear problems. The classification performance was not only limited to binary class but can be extended to multiclass [27,28] which is quite beneficial in gait studies.

The prolific technique in gait analysis is the use of NNs consisting of a single or multi-layer perceptron. The NN uses feed-forward and backward propagation algorithms and is free from the need for hand-crafted features but many times acts as a black box. NN was used widely in gait studies to deal with the problems related to prediction and pattern recognition. DT, a sub-class of RF were used for the highly nonlinear and complex relationships between variables. Besides being interpretable, lacks to provide the optimal solution. Though, being an ensemble of randomised DTs, RF chooses the prediction with the maximum number of votes. The kNN classifier is based on distance metric and was widely used in real-time applications [8,22] as it is free from underlying assumptions about the distribution of the dataset. Gait asymmetry studies correspond to the use of varying membership grades to account for linguistic information which cannot be expressed numerically uses fuzzy techniques [12,32]. Such techniques are less explored in gait studies due to problems in defining a number of linguistic variables and optimal selection of membership functions.

2.2. Unsupervised learning

In unsupervised learning, no labels were given to the learning algorithm. The algorithm itself devises the relationship between various inputs to determine an output. Most clustering techniques, determine clusters based on the distance between all feature vectors. These techniques were less explored in gait studies because accurately defining the learning objectives and to deal with a large number of feature vectors was a tedious task. However, such techniques can be used when the relationship between different observations is unknown. To deal with the large datasets, it is desirable to combine the classifier with some dimensionality reduction approaches. These unsupervised techniques are capable of learning distinct patterns for particular disorders. An explanatory study can ensure the appropriate selection of distance metrics for the given problem. Apart from the distance metric, the latent profile analysis of clustering can also be used for subgroup classification.

2.3. Reinforcement learning

In rehabilitation devices such as exoskeletons and walking assistive devices, RL is required for the systems to interact with a dynamic environment. The various control strategies were developed for gait rehabilitation. The RL and (deep) neural networks (DNNs) were widely used with rehabilitation devices due to their ability to better capture the participants' variability and thus, resulting in automation according to subject-specific needs. The summary of presently deployed techniques in gait studies with their merits, demerits, complexity and interpretability is presented in Table 1.

2.4. Feature extraction and selection in machine learning

To improve the processing capabilities and to reduce the complexities, dimensionality reduction techniques [23,31] were employed. These techniques were categorised as feature selection that selects a subset of features from the original feature set without changing their originality, and feature extraction leading to the mathematical transformation of original features to derive new features. Some classifiers like artificial neural network (ANN), DNN and convolutional neural network (CNN) automate the feature extraction and selection process. Techniques most frequently explored in gait analysis were principal component analysis (PCA) [20,27] which possibly constructs principal components by converting correlated variables into a set of linearly uncorrelated variables through orthogonal transformation. The Hill climbing approach [33] being a heuristic search tries to reduce the feature vector by the selection of the best possible attributes. In forward selection, initially, one feature vector is selected to determine the accuracy, and then the features are added until the optimal solution is generated. In the backward selection, the initialisation is done by using all features and then removing one feature at a time to search for an optimal solution. The linear classifiers such as linear discriminate analysis (LDA) [22,34] can be used to project the features from higher dimension space to lower dimensions. If the dataset cannot be separated linearly than non-discriminate classifiers can be used such as quadratic

Table 1. Summary of learning model, their merits and demerits of presently deployed ML techniques.

Learning model	Gait applications	Complexity	Interpretability	Merits	Demerits
				9	VM
Supervised	Activity detection Event detection Disorder detection Asymmetry detection Neurological studies	Moderate	Moderate	Good generalisation capability Stability	Feature scaling Selection of kernel function
					NN
				Removes the need for handcrafting features complex relationship detection	Network function lacks explainability
				<u></u>	DT
				Probabilistic approach	Lacks predictability of continuous attributes Instability RF
				Combination of multiple decision trees with probabilistic approach Fast training	For large datasets, the memory requirement is high Tuning of hyper-parameters
					:NN
				Instance-based learning	Feature scaling Sensitive to outliers
					uzzy
				Rule-based with less complexity	Selection of membership function A large number of linguistic features
Unsupervised	Clustering of disorder Asymmetry detection Determination of gait activities	Moderate	Difficult	Ease of implementation	Time complexity
Reinforcement	Assistive devices Rehabilitation Automation	Moderate	Moderate	Continuous learning No model training	Limited learning

discriminate analysis (QDA), flexible discriminate analysis (FDA) and regularized discriminate analysis (RDA). They had to be tested first to avoid over-fitting of the model and were implemented offline to reduce the computational burden. A challenge is to devise the relationship between optimal attributes for observed gait patterns and to use some nonlinear dimensionality reduction techniques for gait studies. The technique explored for gait disorder detection is kernel PCA [14]. These methods usually suffer from slow convergence, local minima and high computational complexities. Gait related studies can also be benefitted in near future by incorporating advanced feature selection methods based on removing highly correlated features, ranking features based on imparting high accuracies, and wrapper approach for selection and elimination of feature vectors.

3. Methods

This section describes the search approach used for the selection of studies. The inclusion and exclusion criteria are discussed together with the information extraction to carry out the review.

3.1. Search approach

The review was accomplished according to the preferred reporting items for systematic reviews and meta-analyses (PRISMA) checklist as illustrated by Moher et al. [35]. It includes four phases viz. identification, screening, eligibility and finally inclusion of the searched articles. The consideration was focussed on search terms viz. artificial intelligence, gait rehabilitation, lower limb and machine learning for searched articles (i.e., journal and magazine articles, proceedings of international and national conferences, and extended articles published from January 1980 to May 2019) in the recommended databases including PubMed, IEEE Explore, Science Direct, Wiley online library, Mendeley, Springer Link, Web of Science, Elsevier (Gait & Posture, Artificial intelligence in Medicine, Engineering Applications of Al, Journal of Knowledge-based Biomechanics, Systems and Neurocomputing), and Cochrane.




Figure 1. PRISMA flowchart of the selection process for review flow.

3.2. Selection criteria

The title and abstract screening were carried out by the authors (PK and NK) and independent reviewer(s) (VI) independently. The articles were screened based on the following questionnaires. (1) Has the article targeted the lower limb? (2) The gait analysis and gait rehabilitation he the focussed area? (3) Implementation of AI or ML for gait analysis and rehabilitation.

Based on the screening process, outcomes were reported. Throughout the study, the basic criterion was to traverse whether it is related to the lower limb or not, the inclusion of gait analysis or rehabilitation discussed in the articles or not. Whether or not the ML or Al approaches were used for gait analysis and rehabilitation. By considering all these factors, the selection of the papers was established based on keywords and outlined in Figure 1.

3.3. Inclusion and exclusion criteria

Machine learning and AI were the only keywords for the search. However, other possibilities may also exist and can be used with different names. This review includes only the lower limb whereas the upper limb is not targeted. The study focussing on sports biomechanics, gait for biometric, and security were not part of the inclusion criteria although they used ML. The review articles, book chapters and short communications were excluded. Related studies published in the English language were considered.

3.4. Data extraction

The authors (PK and NK) studied the full-length articles and information extracted is the objective of the study, measurement techniques, number of participants, pathology, parameters, methods, outcomes of the study, limitations, and its future scope (Table 2). The information about the dataset, as presented in Table 2 showing the size of data used for research which varies from small to large population size.

4. Results and discussion

Unlike, previous studies or methods which mainly focussed on gait analysis through observation by naked eyes lacks scientific and objective measures for its determination. The results were entirely based on the experience of the clinical expert or therapist. Nowa-days, with the advent of technological advances in

			,							
S. no.	Source o. of study	Objective of the study	Measurement techniques	Participants	Pathology	Parameters/ activity performed	Method	Outcomes	Limitations	Future scope
Gait c	Gait activity detection 1. Lau et al., 2007 [7]	Development of accelerometer and gyroscope based sensor unit. Classification of five walking conditions.	Sensor unit at shank and foot (accelerometer and gyroscope)	1 3H (3M)	1	Stair ascending (SA) Stair descending (SD) Level ground (LG) Upslope (SU) Downslope (SD)	SVM RBF-ANN Bayesian Belief Network	Using SVM maximum accuracy for all classification tasks was achieved. 100% accuracy to classify SA from other activities, SD from other activities, and SA, SD, and other walking conditions. 85% for all 5 class classification. Accuracy increases from 78.82% to 84.71% by the addition of foot features together with the		To determine the effect of adding more input features from other sensors on accuracy. To perform real-time applications to aid in the decision making process a classifier is required.
~	Pham et al., 2015 [36]	Use of muscle activities for recognition of human walking movement.	Motion capture system (kinematic data – joints (hip, knee and ankle) heel, toe) EMG data (iliopsoas, gluteus maximus, biceps femoris, rectus femoris, vastus lateralis, vastus medialis, tibialis anterior, gastrocnemius)	3H (3M)		Two principal components from EMG data in maximum voluntary contraction state in relation with stance and swing phase for Walking Running	Segmentation Feature extraction using LLE Classification using HMM	shank readures. In walking, stance phase occurs for 60% of gait cycle at 4km/h and 52%, and 93% in stance phase and 80% and 82% in swing phase, respectively. In running, stance phase is 40% of gait cycle at 7 km/h and 9 km/h, with accuracy 85% and 96% in stance phase and 89% and 96% in stance phase and 89% and 96% or more depending on different speeds for walking and strunning movements.	Reliability of analysis needs to be explored using more subject data.	Enhancement of framework for more speed changes. Controlling of wearable TTI-exo for supporting gait movements.
m ⁱ	Lonini et al., 2016 [37]	To differentiate between personal (patients group) model, global – patient models, global – healthy models for activity	Triaxial accelerometer	11H (6M, 5F), 10P (6M, 4F)	Knee-ankle-foot- orthosis (KAFO) for walking	131 time and frequency domain features for activity classification Sitting Standing Walking Stair down Stair up	Classification of five activities using RF 100 trees for personal models 50 trees for global models	_ e, u	Need to improve generalisation of models. Accuracy is restricted for stair climbing due to tri-axial accelerometer on waist.	Use of different feature selection methods needs to be included for patient population. To explore Transfer learning methods.
										(continued)

7		
	1	
	Ξ	
	_	
9	Ξ	
	_	
4	_	
9	=	
	_	
٠.	_	
`	٠	
•	N	
•		
	1	
	*	
7	=	
4	0	
- (Ť	
	۰	

Future scope		Use of statistical methods to handle noisy or missing data. Inclusion of different other kinds of movements for sport gesture optimisation and rehabilitation, rehabilitation.	Usi	To study the nature of knee stiffness and drop foot. To explore the possibilities of hierarchicalweighted classifier.
Limitations		Static abnormality detection is not possible like FoG.	Able to detect subtle differences between semi tandem and tandem using proposed feature set.	Limited generalisation capability
Outcomes	respectively. Personal shuffled model increases the accuracy from 84.2% to 96.7%. Sitting is easy to classify while stair climbing has lowest classification accuracy. For patients population condition-specific activity recognition algorithms are sufficient.	LL and RL have better detection results compared to FoG.	To differentiate good mobility and poor mobility RF produced a maximum average recognition rate of 87.10%. GRBM producing lowermost average result of 73.08%. Determination of rehabilitation through the recognition of frames based on experimentally selected threshold.	The position of sensor selection on body is dependent on ARC from being 5 to 3 sensors. User dependent model allows more reduction in number of sensors as compared to user-independent models.
Method		Hierarchical Clustering algorithm PCA	SVM RF ANN GRBM AdaBoost LPBoost Total Boost Total Boost Bagging	Hierarchical weighted classifier
Parameters/ activity performed		Walking Stair climbing	Balance (eyes open) Chair rising Semi tandem Tandem balance Walking (4 m)	Sit Stand Transitions (sit to stand and stand to sit) Side steps Stairs and motion
Pathology		Knee injury (knee camot be bent) Foot dragged (using crutch) Left leg lead (LL) Right leg lead (RL) Freeze of gait (FG)		Stroke
Participants		12P (SPHERE dataset) DAI dataset 5H (frontal gait dataset)	10H	d11
Measurement techniques		3D data using depth sensor Kinect	Microsoft Kinect (MoCap 10H skeleton) Joint angles Both arms Both legs Torso	Kinematics data using wireless IMMU system at thorax, peivis, right/left upper leg, right/left lower leg, right/left foot
Objective of the study	recognition in individuals with lower limb injuries.	Abnormal gait detection. Re-identification through gait recognition.	Recognition of motions. Evaluation of human mobility impairments.	Four activity recognition chain (ARC). Reducing energy consumption. Identification of position at body to locate the sensor.
Source . of study		Devanne and Berretti, 2016 [20]	Leightley et al., 2017 [11]	Recher et al., 2018 [38]
S. no.		4.	νi	ø

Table	Fable 2. Continued									
	Source	Objective of	Measurement			Parameters/				
S. no.	of study	the study	techniques	Participants	Pathology	activity performed	Method	Outcomes	Limitations	Future

cope	metric based	on. model for ns.	sensor terrains terrains lope and stair of gait Auring Atries.
Future scope	To embed a wearable metric with clip-based	activity recognition. To test the model for variety of patient populations.	To evaluate sensor systems at different terains such as slope walking and stair climbing. Monitoring of gait patterns during daily activities.
Limitations	More misclassifications between sit and	wheeling or between walk and stair climbing, due to difficulty to distinguish the clips as accelerometer produces the same result. Challenging to interpret the criteria for decision making.	1
Outcomes	Preferable location for sensors being upper legs due to less variability and positioning the sensor on non-affected leg for user-independent model. Decision Fusion ARCs are more robust in performance than PARCs even during sensor failure. Classification of six activities, i.e., lying, sit, stand, walk.	wheeling and stair climb. Intra subject 20-fold cross-validation using hybrid RF static HMM classification accuracy of 88.9% i.e., 2.6% higher compared to static RF classifier with high precision and recall values. Subject wise cross-validation using 12 subjects data as training and remaining one as testing provides an accuracy of 64.3% using SVM-HMM higher by 1.2% using only SVM.	DF differs from normal gait as it has more peaks and troughs of high amplitude in stance and swing phase. Significant changes appear in gait patterns for different walking conditions. High performance in TPs appears at the time of phase transitions between stance and
Method	Static classification: SVM, NB, regularised LR, k- NN, DT, RF	HMM with probability distributions (state transition and emission probabilities)	Threshold detection method (based on turning points (TPs))
Parameters/ activity performed	Parameters Mean X, Y and Z Absolute value of mean	Moments: SD, skew, kurtosis RMS Extremes: min, max, abs min, abs max Cross product mean: xy, xz, yz Absolute mean of cross products Overall mean acceleration Activities Lying Sit Stand Walk Wheeling Stair climb	Classification of turning points in stance and swing phase
Pathology	Incomplete spinal cord injury		Hemi paretic patients with dropped foot (DF)
Participants	13P (9M, 4F)		3H (3M), 10P (2F, 8M)
Measurement techniques	Accelerometer data		Sensor units at thigh, shank, foot (accelerometer and gyroscope) FSR at 2nd middle phalanx, 1st metatarsal head, 5th metatarsal head, heel
Objective of the study	Recognition of activities using waist worn	accelerometer for ISCI subjects.	Identification of gait events and evaluate the reliability of system based on accelerometer and gyroscope data.
Source of study	Sok et al., 2018 [39]		Gait event detection 8. Lau and Tong, 2008 [40]
S. no.			Gait ever

~	
>	•
ď	Ĺ
=	3
-	
_	-
- =	
7	
_	
c	
Conti	1
•	•
_	1
r	١
9	ı
_	Ė
•	3
-	
٠,٠	٤
_	

2		;								
S. no.	Source of study	Objective of the study	Measurement techniques	Participants	Pathology	Parameters/ activity performed	Method	Outcomes	Limitations	Future scope
o's	Jung et al., 2012 [41]	Intent detection method consisting of a gait phase recogniser (to determine the intent to start moving unsound leg to swing) and gait pattern generator by estimating the unsound leg stance pattern using sound leg movement and echo control (echoing the swing pattern of	Ground reaction force sensor Joint angles	ı	Stroke	Two joint (hip and knee) NN angles Foot status value of one leg	¥	A single foot sensor data provides good performance in detection of TPs but in combination with other sensors, it can combine two directions for reliable operation. Classification of one stride into seven subphases using binary representation. Effectiveness of gait pattern generator is evaluated by comparing similarities between output and actual values in shape and time to evaluate unsound leg active joints.	1	Implementation of the developed hybrid method to real exoskeleton robot on walking rehabilitation.
10.	Yoo et al., 2013 [16]	previous gait). To predict movement related to pain and radiographic severity associated with ascending stairs. Determination of broso OA	previous gait). To predict movement Kinematics data using related to pain 3D motion and radiographic analysis system severity associated with ascending stairs. Determination of the previous	20H, 18P	Knee osteoarthritis (OA)	Time of stair ascent Maximal anterior pelvis tilting Knee flexion at initial foot contact Ankle dorsiflexion at initial foot contact	NVS	Detection of knee OA with accuracy of 97.4%. Prediction of pain determination with accuracy of 83.3% and radiographic 52.00.00.00.00.00.00.00.00.00.00.00.00.00	No optimal method to determine penalty parameter and scaling factor of Gaussian kernel function in SVM. Sample size	To develop prediction model for progressive knee OA. Use of other machine learning methods for realization
Ë	Jiang et al., 2018 [42]	inital stage. Wearable shoe pressure sensor graphene porous network system for human	Plantar pressure of walking	10Н	ı	Normal pattern Toe-in Toe out Lame feet Heel feet	Ensemble leaming algorithm – SVM + RF + LR	Normal pattern – 94.3% – 10 – 94.3% – 10 e out – 94.2% Lame feet – 93.6% Heel feet – 96.8%	000	- Lephragae
12.	Nazmi et al., 2018 [43]	gail detection. Detection of gait events, i.e., stance and swing phase using sEMG.	sEmg signal from lower leg muscles – tibialis anterior (TA) and medial gastrocnemius (mGAS) Two FSR's at hallux and	8H (8M)	1	Time domain features RMS, SD, MAV, WL and IEMG	ANN using Levenberg Marquardt (LM) and scaled conjugate gradient (SCG)	During stance and swing hase TA and mGAS muscle activate, respectively. ANN accuracy statistics 87.5% using	High difference in detection of phases using foot switch signal and ANN model being 29% and 10%	Incorporation of other TD features. Variations in hidden layer neurons.
								ובמווובת מוומ		.,

)	(

Table 2. Continued.

	activity performed	nts Pathology	techniques Participants
PNN 9999 LVQ DNN KNN DN Naïve Bayes Random forest SyM MLP	Oblique angle of pelvis PNN Pelvis angle of rotation LVC Left/right lip flexion/ DNI adduction angles kNN angles Raion Naï angles Ran Left/right ankle SVN flexion angles MLI		1
reported by LVQ. SVM, MLP, NB, RF resulted in performance in the range of 97%. arning LF-MAS depicted good Rec performance of Apptive unaffected leg. The subjects affected leg	Hip joint Rein Knee joint T Smart shoes with Poli plantar sensors		Hemiplegia patients (healthy subjects
programming together with each hair exoskeleton leg will gait cycle. lead to half gait cycle delay. Higher tracking error for unaffected leg during gait transitions. High knee joint error in comparison to hip joint due to large motion variation of knee as compared to hip during normal walking reported			simulating hemiplegic condition)
C45 decision tree For larger training set — MLP C4.5 outperforms NARX MLP and NARX with 100% accuracy as compared to 94,79% for MLP and 98,76% for MLP and 98,76% for NARX in gait	Knee IMU C4.5 Hip IMU MILI Toe FSR NAI Heel FSR		1
Logistic model decision GPR model provided an The error rate of accuracy of 90.60% 10% not suitable Transition sequence using LMT and it for assistive verification and increases to 98.61% control devices correction (TSVC) using TSVC. using TSVC is applied to improve the	Gait phases Loading response Tran Push off Swing Valking surfaces Level Down-slope		1

	=
	\sim
	2
	≓
≣	:=
_	7
⊒`	₹
₹	ŭ
ש	$\overline{}$

S. no.	Source of study	Objective of the study	Measurement techniques	Participants	Pathology	Parameters/ activity performed	Method	Outcomes	Limitations	Future scope
a a	Paulo et al., 2019 [3]	To automatically detect shifts in gait pattem.	Lower limb parameters such as thigh, knee, shin using IMU	SH (4M, 1F)	1	Up-slope Right cross-slope Left cross-slope Spatiotemporal features Similarity rate Walking patterns – natural walk, left and right knee joint bending restrictions, respectively	One class SVM	The more distinct input and output results in low value of similarity rates. Can be used as an automatic progress detection tool for healthcare professionals.	overall performance. Less variability of gait pattern acquired.	Improvement based on medically selected gait parameter shift. To determine a quantifiable shift in gait patterns.
₹ ₹	out alsotter generation 18. Muniz et al., 2006 [31]	To determine the usefulness of PCA in GRF in differentiating gait pattems of normal and impaired subjects.	GRF using force platforms with an instrumented treadmill gaitway	25H (13M, 12F), 6P (5M, 1F)	Lower limb fracture	Vertical component of GRF	~	In PCA, the principal components determine the variance of GRF curve useful for analysis. Scatter diagram of PCC (first two coefficients) allows objective assessment of each patient and also able to determine the treatment effects.	Lacks the ability to access the non-linearity of PCA in classifying groups.	Use of more PCC with nonlinear separators.
<i>a</i>	Lai et al., 2007 [33]	Classification of PFPS patients from healthy group with reduced features using hill-climbing feature selection algorithm.	Rear-foot kinematics Foot GRF	14H, 13P (14F, 13F)	Patellofemoral pain syndrome (PFPS) on the right knee	14 features from GRF 16 kinematics features	W/S	14 GRF feature set results in 85.19% accuracy using polynomial kernel. 16 kinematics features resulted in an accuracy of 44.07%. Using optimal feature set from both GRF and kinematic data achieved accuracy of 88.88%. The optimal set of features are time to max rear foot abduction and adduction, time to max eak eversion and second vertical peak GRFs.		To investigate the direct relationship between rear-foot kinematics and GRFs.
a	Nair et al., 2010 [34]	To identify muscles required to distinguish arthritis and healthy group.	EMG patterns of each leg	7H, 18P (8 RA, 10 OA)	Rheumatoid arthritis (RA) and osteoarthritis (OA)	RA + OA groups: Soleus Gastrocnemius medialis Tibialis anterior Vastus lateralis Biceps femoris	LSK (developed algo) SOM LVQ MLP LDA	Classification accuracy of 91.48 ± 7.57% for RA patients and 96.67 ± 3.77% for OA patients using CGF LSK kernel algorithm.	Current rehabilitative To compare neural measures are nets, kemel needed to methods and understand regression muscular analysis to map nonlinearity in	To compare neural nets, kernel methods and regression analysis to map nonlinearity in
l										(bountinued)

Table 2. Continued.

S. no.	Source of study	Objective of the study	Measurement techniques	Participants	Pathology	Parameters/ activity performed	Method	Outcomes	Limitations	Future scope
		algorithm for the treatment of biological data.				Peroneus brevis (RA only) Gluteus medialis (OA only)		The most dominant muscle for OA is gluteus medialis and for RA are soleus and biceps femoris.	differences during gait.	EMG and kinematics approach.
21.	Mu et al., 2010 [14]	Classification of patients with different foot lesions.	Foot kinematics	27P	Pathological plantar hyperkeratosis (PPH)	Leg Heel Midfoot First metatarsal Hallux	KPCA FLDA	Reduced feature space of d = 46 and accuracy 94.1% using KPCA, DCT and FLDA. Foot kinematics contains information relevant for pathology classification and KPCA produces lower dimensionality.	In consideration of foot kinematics, the motion of proximal segments affecting plantar surfaces needs to be taken into account.	Inclusion of kinematics dataset using proximal segments, i.e., thigh and trunk.
52.	Alaqtash et al., 2011 [19]	Classification of gaits of healthy, CP and multiple sclerosis participants using 3D GRFs components.	Classification of gaits Three GRF components of healthy, CP and multiple sclerosis participants using 3D GRFs components.	and 4MS	Cerebral palsy (CP) Multiple sclerosis (MS)	F_{p} F_{p	Nearest neighbour classifier (NNC) Artificial neural network (ANN)	GRF parameters using NNC has higher accuracy rate than DWT. The healthy and impaired gait patterns can be distinguished using one feature, i.e., M-shaped value of F _z . Optimal feature set, i.e., F _z , F _{x1} , T _{x3} , F _{x3} , T _{y3} and T _{y2} . 85% accuracy without feature selection and 95% for selection and 95% for selected set of six features.	ı	
23.	Gestel et al., 2011 [13]	A classification approach that characterises CP gait by membership degrees.	3D gait analysis of ankle and knee	hemiplegia, 80 diplegic and 1 triplegic)	Cerebral palsy (CP)	Ankle and knee joint movement patterns	Bayesian network	A Bayesian network is standard tool for 3DGA with probabilistic, non-linearity handling approach and enhanced clinical interpretation. Able to identify mixed joint movement patterns of CP. Accuracy is high, i.e., 82–91% for distinct movements	Reliability of outcome depends upon expert input.	Strengthening of clinical relevance by inclusion of other joints and plane.
24.	Lee et al., 2015 [44]	A wearable platform to measure knee kinematics. To develop detection algorithm to	Electro goniometer angular sensor to measure knee angles 3-axis accelerometer on thigh	17H (scripted data), 4H (unscripted data)	ı	Mean and SD of time series Cross covariance Signal entropy Range of amplitude	One class SVM (LibSVM) Binary classifier RF with 100 trees Multi-class RF with 100 trees	LibSVM separates walking data from other movements of daily activities. Same feature set for	More precise decision boundaries are required to recognise fast	Extension to other activities using wearable technologies in human context

	Č	Ĺ
	Č	
•	2	
(_	
(
	9	
	2	

ed.	with	of the state of th		
Future scope	for patients with knee OA.	To refine classification by adding modelling with GRF, joint moments and powers as well as EMG.		
Limitations	stairs down climbing, slow walking from stairs up climbing, and slow walking from transitional activities.	Applicable only to Traparticipants that walk without physical assistance.		
Outcomes	and walking at slow speed as well as for stairs down climbing and walking at fast speed. Hybrid dassifier used to distinguish occurrence of walking from stairs using one class classifier and to distinguish different walking speeds from stairs up and down climbing using multiclass classifiers. Precision and recall vary inversely with decision boundaries. Multi-class classifier (RF) showed the best performance, 90% precision and	Classification into 6 classes, i.e., spastic hemiparesis, nonspastic hemiparesis, unilateral ataxia/ dyspraxia, spastic bilateral paresis, nonspastic bilateral paresis, in ataxia/dyspraxia. Unaffected limb classification accuracy results in 82.35%. Affected limb classification accuracy results in 75.47% footiles in	96. 939.	
Method	and RF	Multiclass SVM	Classification using ELM, SVM and NN	
Parameters/ activity performed	Dominant frequency Correlation coefficients	Kinematic signals from pelvis and lower limb (hip, knee and ankle)	Time domain features RMS, MAV, WL, ZC,VAR	
Pathology		TBI related gait disorders	Foot drop	
Participants		102P	5H, 5P	
Measurement techniques		Joint angles (pelvis, hip, knee and ankle)	sEMG signal from both the legs using muscles Gastrocnemius, tibialis anterior GRF	
Objective of the study	occurrences of walking using accelerometer.	To develop prediction system for TBI gait disorders.	Classification of foot drop drop (crouch gait) patients from healthy subjects using sEMG signal and to explore the effect of GRF on minor, moderate and serious spastic	
Source of study		Williams et al., 2015 [28]	Adil et al., 2016 [45]	
S. no.		25.	26.	

₹	3
đ	j
-	3
7	_
.≥	
+	
2	
-	٦
٠,	1
_	•
`	•
_	•
C	
٠,	,
0	
٠,	,
٠,	
٠,	2
٠,	

S. no.	Source of study	Objective of the study	Measurement techniques	Participants	Pathology	Parameters/ activity performed	Method	Outcomes	Limitations	Future scope
27.	Cho et al., 2018 [46]	diplegic cerebral palsy patients. Automatic recognition of scoliosis and control group. Classification of mild, moderate scoliosis and normal group.	IMU-based system (seven sensors on both lower limbs)	24P, 18H	Scoliosis (right tilting of trunk)	72 features (Kinematic aspects)	Relieff for feature selection SVM	spastic patients have maximum GRF. 20 features showed maximum accuracy of 95,2% (Relieff + SVM) and 85,7% using only SVM. 90.5% accuracy for distinguishing healthy and scoliosis group and SI% between mild, moderate and normal group.	Limited space size. According to location of apex gait patterns are not analysed.	Inclusion of more factors related to study for patients group classification.
28.	Guo et al., 2017 [27]	Classification of normal and pathological gait using foot pressure data in young children.	Foot-pressure data (GAITRite)	95H, 6P	Equinus gait pattern (common gait pattern for cerebral palsy)	Spatiotemporal foot pressure data	PCA Discriminative mapping SVM RF	94.36% and 97.50% accuracy achieved using SVM and RF, respectively, based on age information together with other spatiotemporal features. Possibility to detect minimal variations of gait development during early states.	Less sample size for pathological group	Development of classifier to determine multiple degrees of abnormality in children with spastic gait.
29.	Cui et al., 2018 [47]	Development of an automatic gait analysis system for recognition and assessment of the gait abnormality.	and EMG and EMG	21P, 21H (16M, 5F)	Post stroke hemi paretic patients	28 markers 2 force plates 8 pair of electrodes	Five classification methods – NB, SVM, NN, KNN, RF Two rule-based fusion methods – AR and MR	Recognition performance Single modal – RF, accuracy 92.26% based on GRF data. Two modal – SVM fusion, accuracy 95.83% based on MT and GRF data. Three modal – SVM fusion, accuracy 98.21% based on MT, GRF and EMG. Assessment WAMS has potential in clinical application for given pathology class and subjects walking ability is expressed as a real number.	Time normalisation weakens subject velocities Lack of standard evaluation metrics Architecture limited to only abnormal gait after stroke Lacks assessment about therapeutic treatment	Different modal gait data can increase both recognition and performance.
30.	Zeng et al, 2019 [18]	Classification of gait patterns between ACL-D and contralateral ACL-I knee in impaired individuals.	Kinematic signals	43P	Anterior cruciate ligament (ACL) (full tear)	Flexion/extension of knee, hip, ankle Ankle internal/ external rotation	PSR (phase space reconstruction) ED (Euclidean distance) NN	The classification accuracy of 95.4%, 93.3% and 94.8% to distinguish ACL-D and ACL-I under normal, fast and both	Detection deficient due to the presence of other injuries like posterior cruciate ligament having	To assist in diagnosis by inclusion of other dataset having same gait patterns that result in knee injury.

palsy Classification groups iplegia) True equinus Jump gait Apparent equinus Crouch gait	Cerebral palsy (spastic diplegia)
24 features were extracted from minimum foot clearance (MFC) during walking occurring at midswing phase of gait cycle	28 elderly)
Temporal-spatial Foot GRF Lower limb joint angular data	12Y, 12O (6M, 6F) –

Table 2. Continued.

	in led	5)5
,	U	;
	Alde	2

	spoo p	ā p 6		
Future scope	other feature selection methods like backward elimination techniques and genetic algorithms.	Inclusion of more data from other activities based on rehabilitation protocol. To consider anthropometric data for evaluation of recovery.		Need to obtain complete gait features.
Limitations		Semi-automated process based on expert knowledge and intelligent famework.	Complex applications depending upon many factors lead to an exponential increase in rule-set.	Less gait dataset results in instabilities in new recovery trajectory.
Outcomes	(i.e., knee range of motion during swing phase, maximum horizontal (anterior-posterior) push-off force, step length) 100% generalisation accuracy is achieved. Thus feature selection improves accuracy. Performance of BR remains unaltered over threshold range (-0.4 to 0.4).	ANFIS showed better performance for walking ambulation 4 and 5 km/h. FURIA and linear SVM have highest classification accuracies for 6 km/h and balance test. Overall best performance by FURIA for all activities.	Determination of hip, knee, ankle and fall (unable to recover) strategy depending upon applied force and DoM. Hierarchical fuzzy techniques can streamline complex behaviour. Fuzzy inference system behaviour streamline complex computationally intensive can handle forces from all directions	LSTM learns the gait features with reduced mean square error and gradually decreasing cost and reaching a stable convergence. DRGL can be used to
Method	regularisation) LDA	Fuzzy dustering techniques (ANFIS, FURIA) SVM Hybrid CBR	Fuzzy logic based learning approach	LSTM DRGL (deep rehabilitation gait learning)
Parameters/ activity performed		Walking at 4 km/h, 5 km/h, 6 km/h Single leg balance test	Force applied Direction of motion (DoM))	Knee joint trajectories
Pathology		ACL reconstructed (3 months to 1 year) after injury	1	1
Participants		6H, 13P	5H (5 left handed dataset, 5 right handed dataset)	20Н
Measurement techniques		Kinematics (IMU at shank and thigh) EMG data (vastus medialis, vastus lateralis, semittendinosus, biceps femoris, gastrocnemius)	Joint angles (hip, knee, ankle)	SIAT exoskeleton
Objective of the study		To monitor and classify the recovery status of ACL reconstructed subject.	To employ push recovery in humanoid robots. To simplify its complex nonlinear behaviour to linear using ML techniques.	To determine the gait trajectories of wearer for modelling knee joints of lower limb exoskeleton using LSTM and DRGL for
Source of study		out asymnety aetection 34. Arosha Senanayake et al., 2014 [12]	2014 [32]	Liu et al., 2016 [25]
S. no.		34.	35.	36.

_	Ŀ
τ	3
d	į
Ξ	3
2	=
Ψ	5
Ċ	Ξ
c	2
L	J
C	ر
	ر :
C	į
	ניייי
	ייייייייייייייייייייייייייייייייייייייי
	פעייי
	ממוע אי


S. no.	Source of study	Objective or the study	Measurement techniques	Participants	Pathology	ralameters/ activity performed	Method	Outcomes	Limitations	Future scope
37. F	Paulo et al., De 2017 [24]	modelling knee joints of lower limb exoskeleton. Development of ISR- AIWALKER for lower limb rehabilitation in HMI and gait analysis system.	Vision-based approach – leap motion controller-kinematic model of lower limb using RGB and depth map	HMI safety system – 10H (8M, 2F) HMI operation – SH Gait analysis system- 10 (9M,1F)	ı	Robotic walker HMI – leap motion controller Gait analysis system	SVM (cubic kemel)	predict and correct abnormal knee joint trajectories depending on wearers other joints. Gait asymmety pattern detection 88.3%. 94.76% classification accuracy of user's stability safety system, i.e., incorrect gripping trajectories are discovered and the system stops	Leap motion sensor is based on IR reflection Undetectable gait pattern with no leg separation Gait analysis in linear motion	Gait analysis system to perceive the shifting gait patterns over time. To develop a tool for clinicians to identify gait changes and for changes and chang
38.	Sobral et al., To 2018 [49]	To develop gait indices – NGI and AGI (Normal and Abnormal Gait lindex) to access the recovery status after ACL reconstruction.	Vertical ground reaction force	28H (28M), 5P (5M)	Anterior cruciate ligament (ACL) reconstruction	8 force sensors in instrumented shoes for walking conditions – very slow, normal, fast and very fast	ELM	New gait indices discriminated healthy and impaired gait patterns based on symmetry (SI) and difference between 16 different parameters from VGRF for gait error (GE). A normal gait has GE and [SI] lower than 1	Lacks GE calculation variables from centre of pressure and fore-aft and medial-lateral GRF.	therapeutic interventions. To determine GE implementation of PCA to compute variance of all variables to reduce the number of variables is needed.
39.	Qiu et al, Ga 2019 [29]	Gait quality assessment for clinical decision support using multi-sensor fusion and decision making algorithms.	Kinematics data from both ankles Spatio-temporal Parameters Nonlinear parameters	20H, 20P (13M, 7F)	Neurological diseases	Spatio-temporal gait parameters Walking Speed Step length Stride Time Foot elevation Gait phase detection Heel strike Foot flat Heel off Foot swing Non-linear parameters Gait asymmetry	WWH	and ideally equal to zero. Based on gait abnormality, a decision support tool using BSN. Ankle rigidity can lead to abnormality. Higher gait symmetry score leads to abnormality. Gait regularity of spatiotemporal parameters leads to determine degree of motion	Implementation is offline and needs to be tested for clinical settings	To ensure low power consumption by BSN hardware and real-time monitoring.
Neurolog 40. H	Neurological effects on gait 40. Hortal et al., To 2015 [50]	To identify the voluntary initiation and gait cycle stop.	32 channel EEG signal, IMUs and Wireless motion analysis system	3Н (3М)	I	Gait regularity Gait complexity 32 electrodes EEG cap IMUs – lumbar, right thigh, leg and foot, left thigh, leg and foot	SVM (RBF kernel)	abnormality. Good detection accuracy for rest and walking state, i.e., 70.5% and 75%. Low accuracy for start	High false positive rate limits its real- time application.	To reduce false positives, other features and classifiers are to be tested.

Table 2. Continued.

S. no.	Source of study	Objective of the study	Measurement techniques	Participants	Pathology	Parameters/ activity performed	Method	Outcomes	Limitations	Future scope
	Costa et al., 2016 [22]	To study the cognitive mechanisms depending on user's gait attention.	Treadmill walk and EEG	12H (4F, 8M), 3P (3M)	Incomplete SCI	32 channel EEG recording	Classification based on attentional tasks using LDA, SVM, KNN, NB, DTL	and stop state, i.e., 30.6% and 17%. Y (gamma) band provides attention mechanism. Attention to visual tasks causes a decrease in gait consideration. Patients' attention level to gait increases during non-attentional tasks. 67% success rate for healthy using SVM and 59% for patients using LDA and KNN achieved.	To interpret the low- frequency band cortical information.	To determine attention level during lower limb rehabilitation in real-time.
	2018 [23]	Decoding cortical processes under varied walking conditions using EEG signal for gait determination. A SSRL topology of deep neural network to learn spatial spectral representation of multi-channel EEG signal at the time	EEG at walking (free walking, exoskeleton assisted walking)	27Н	1	Normal walk without exoskeleton support Walking with exoskeleton At zero force At low assistive force At high assistive force	RF-FS SVM-PCA SVM-FS SSRL (raw)	ture racy of brain on left right h the brain or right care.	Increase in computational cost with the increase in no. of parameters.	Improvement in topology of SSRL for cross-subject gait pattern classification. To enhance abnormal gait diagnosis.
ш.	Park et al., 2019 [51]	To reduce delay time at the time of controlling exoskeletons with EEG decoders by considering short segment EEG data (0.2 s) based on gait/stand intention that is 0.5 s prior to actual gait/ stand movements.	32 channel wireless EEG system in walking condition based on cue signs	3H, 8P (5 subacute, 3 chronic stroke patients)	Stroke	19 channels for healthy and subacute stroke patients, 32 channels for chronic stroke patients	ONN	related brain regions. A spatio-spectral CNN model with accuracies 85.2% for chronic, 85.9% for subacute, and 77.4% for healthy on gait recognition. Gait and stand intention recognition average accuracy of 77.3% and 77.7%, respectively.	19 EEG channels data not enough to reflect spatial features.	Stroke patient's gait intention would lead to reactive exoskeleton.

abs: absolute; ANFIS: adaptive neuro-fuzzy inference system; AR: average rule fusion algorithm; BSN: body sensor network; CART: classification and regression tree: CBR: case based reasoning; DCT: discrete wavelet transform; ELM: extreme learning machine; FLDA: Fisher's linear discriminant analysis; FoG: freezing of gait; FURIA: fuzzy unordered rule induction algorithm; GRBM: Gaussian restricted Boltzmann machines; GRF: ground reaction force; IC: initial contact; IEMG: integrated electromyography; LF-MAS: leader follower multi-agent system; LLE: locally linear embedding; LR: locally linear embedding; LS: locally linear expression linear locally linear expression locally linear expression locally linear expression locally linear locally linear expression locally linear locally linear locally linear locally

No. of Papers using different analysis approaches for gait analysis



■ Vision Based ■ Sensor Based ■ Hybrid Based

Figure 2. Literature classification of studies based on the human gait analysis approach.

the field of computer processing, the integration of learning frameworks with computing had led the researchers to find new directions to propose smart systems capable of inferring and handling bulks of data. This automated diagnosis will pave the way to plan rehabilitation interventions using robotic-assisted therapies, motor rehabilitation, etc. The following section discusses ML in gait detection under five categories that report the major findings in each domain and the use of ML techniques for gait rehabilitation.

4.1. Assessed outcomes

The most commonly used human gait analysis approaches are shown in Figure 2. The popularity of using sensor-based approaches is portability, reliability and unrestricted environmental conditions.

4.1.1. Machine learning in human gait analysis

The selected studies were separated on the basis of recent application areas in the human gait domain. Figure 3 shows the number of studies used for GAD, GED, gait disorder detection (a) due to pathology (GDDP) (b) due to aging (GDDA), GASD and GNE.

4.1.1.1. Gait activity detection. The basic study that explored the ML in gait was to demonstrate recognition of normal gait for performing various activities. This led to the detection of motion from motion capture systems and inertial sensors to identify various activities ranging from levelled surfaces (level ground walk), transition phase (sit to stand, stand to sit) to unlevelled surfaces (stair ascending/descending, up/down slope). The worth noting point from Table 2 is that GAD carried out using supervised ML techniques shows good classification accuracies using a sensor-based approach Figure 4.

No. of Studies using different Gait Detection Approaches




Figure 3. Classification of studies as per the literature review.




Figure 4. Taxonomy of Machine Learning Application.

These studies were reported on both pathological as well as healthy groups, activity tracking being valuable to determine patient outcomes [39]. Further, the research carried out by Lau et al. [7] demonstrated SVM as the best predictor for activity classification. Thus, a single sensor placed either at the shank or foot can accurately classify the activities. Together with kinematics, muscle movement was useful for robotics, exoskeleton control and rehabilitation applications. The study [36,38] discussed that reduction in the dimensionality of data helps to achieve high recognition performance. In [37], researchers discussed the need for a global patient model containing particular deformity in the training phase to evaluate the efficacy of models on test datasets. The model trained on healthy subjects was not considered capable enough to draw inferences on the patient group, although personal model training was not required. Considerable work was reported by Devanne and Berretti [20] to detect walking and stair climbing using depth sensors for abnormity detection along with person re-identification by considering shape variations of motion trajectories. The study by Leightley et al. [11] proposed a need for the models free from the human interpretation and also, representative at the same time to assist clinicians in the decision-making process. The development of technological advances and computationally efficient techniques has aided the

professionals to refine the physical and drug therapies for patient's improvement. Sok et al. [39] classified the activities of incomplete spinal cord injury (SCI) patients using dynamic classifiers which proved better than static ones. The studies related to the biomechanics of gait were useful to detect the impairments in gait and articulation of the treatment plans but the interaction between numerous inputs of a locomotor system was rather difficult to characterise. Thus, ML models are used for analysing implicit interactions between the locomotor system inputs for healthy and impaired subjects with accuracy levels of more than 80% for depicting multiple activities. The developed models for optimisation have reduced the analysis time and can assist the clinicians in selecting the key parameter for the intended pathological group.

The analysis showed that activity detection was a crucial design parameter for gait assistive walking systems and control of wearable orthotic device like exoskeleton. Moreover, different classifiers were tested by researchers but clear evidence is not available to test a particular classifier for levelled and unlevelled terrains. To conclude, a technique of ML that is supervised learning worked efficiently in this domain, and further research is required on the patient's gait. This type of model, once trained can do much better on a similar group of patients and can assist the physiotherapists. The precision, objectivity and reliability in outcome evaluation will definitely help the clinician's plan or refine the rehabilitation to improve the patient's mobility. From the above discussion, the limitations reported are less generalisation capability of the ML model, reliability needs to be strengthened by the inclusion of more data samples and evaluation of ML models for unlevelled terrains or real-life scenarios which is an area of scope for further research.

4.1.1.2. Gait event detection. The fundamental study to detect gait phases and sub-phases refer to as GED. The gait phase detection mainly needs to determine two characteristic time points, i.e., when the foot is in contact with the ground (heel strike) and when it is off the ground (toe-off). However, the sub-phases of gait, i.e., loading, push off, swing and terminal swing provides insight into individual variations and dynamic assessment of one's gait [5]. Accurate GED decides about control strategy to use for rehabilitation robots to avoid harm to patients. Also, in functional electrical stimulation (FES), real-time orthotics control, and gait rehabilitation, gait phase detection is crucial [5]. Electromyography (EMG) signals can also be useful in gait phase recognition as they provide activation of electrical indicators of nerves and muscles during relaxation and contraction states [43] together with force-sensitive resistors (FSRs). Some studies (Table 2) have used joint angles and ground reaction force for event detection by using multi or hybrid supervised classifiers. The authors Jung et al. [41] studied gait phase recognition for robotic exoskeletons using NN and the intent detection method for rehabilitation of stroke patients. Further, Yoo et al. [16] demonstrated the association between changes in movement and prognosis of knee osteoarthritis (OA). Another study by Jiang et al. [42] was successful in designing of gait recognition system using an ensemble learning framework for the classification of patterns in therapeutic interventions. Using the NN classifier, human recognition with 99% accuracy rates was reported [8]. Another researcher Farah et al. [5] identified four gait phases across different walking conditions using local sensor signals from thigh and knee through a logistic model DT. As in conventional rehabilitation, the monitoring and follow up process is slow, continuous and entirely dependent on the professional's experience; therefore, Paulo et al. [3] used one-class SVM to automatically detect shifts in gait pattern. Reinforcement and adaptive learning were utilised for adaptive controllers with hemiplegia patients [30].

Thus, it can be interpreted that ML provides the basis for automatic gait phases detection required for rehabilitation, FES to uniquely recognise human gait to plan diagnosis and therapy that considers individual needs. The supervised and RL techniques proved to be beneficial in this domain because RL helps to track one's capability particularly suited for the rehabilitation phase. For, real-time implementation, more rigorous trials are required to attain 100% accuracy before deployment. It will be beneficial for automatic clinical analysis tool development and therapeutic interventions. The shortcomings can be articulated as small sample size, monitoring of gait during activities of daily living and to detect the quantifiable shift in gait pattern.

4.1.1.3. Gait disorder detection. Gait is a person's pattern of walking. Walking involves balance and coordination of muscles so that the body is propelled forward in a rhythm. The deviation from normal walking leads to gait abnormalities. The causes of gait disorders include neurological conditions (e.g., sensory or motor impairments), orthopaedic problems (e.g., OA and skeletal deformities) and medical conditions (illness or injuries). To diagnose these disorders is referred to as gait disorder detection. Earlier, the normalcy index [52] criterion was used to quantify deterioration in one's gait. Till now the wide ranges of studies include disorder detection that is embarked upon by ML approaches to determine abnormality in gait with high prediction accuracies. Research in healthy gait is essential to determine the deviation for disorder detection. In recent years [47], integration of the joint angle approach, ground reaction forces and electrophysiological phenomena for data collection during walking had significantly aided the automatic gait study with high performance (98.1%). The concurrent recognition of multimodal data is desirable for qualitative and quantitative gait abnormality assessment in clinical practices and such methods performed better than single models [2].

Gait disorders are sub-divided as antalgic gait patterns arising from pathologies in lower extremities such as knee OA [16], rheumatoid OA [34] and patellofemoral pain syndrome [33]. These patterns were characterised by pain in weight-bearing structure and resemble limping movements. The variability in gait disorders due to stroke [38,41,47], Parkinson's disease [11], cerebral palsy (CP) [19] and multiple sclerosis [19] leads to apraxic gait resulting in loss of ability to initiate and coordinate movement. Another type being ataxic gait in which a person suffers a loss of proprioception. Elderly related gait disorders are senile gait due to aging and sensory ataxic when the brain does not receive sensory inputs and myopathic due to muscle atrophy. These disorders cause fall-related injuries [6,21] that lead to high medical expenses.

To detect disorders, ML is an emerging paradigm in biomedical for pathology determination. It has been observed through the extensive survey; the choice of a classifier largely depends on features and their correlation with outcomes. However, some studies have reported gait disorders are due to traumatic brain injuries, and then their rehabilitation required sub-classification to predict functional implications [28]. Apart from depicting abnormality only, the level-based approach is essential. It has also been validated by another study [46] conducted for scoliosis. The ML techniques provided a solution to the problem with the inclusion of multi-class support. Another study by Sok et al. [39] determined that accuracy increases with hybrid classification.

Early work by Lai et al. [33] applied SVM to classify deviation in gait patterns due to patellofemoral pain syndrome on the right knee. Studies in knee pathologies mainly focussed on the detection and quantification of biomechanical factors. The study marked an initial attempt to explain biomechanical relationships of optimal selected features. The study by Zeng et al. [18] used NN to detect anterior cruciate ligament (ACL) in impaired individuals using kinematic signals. It was demonstrated by Nair et al. [34] that EMG inputs also play a crucial role in arthritis detection. Another study [45] used the EMG signal to distinguish a foot drop from healthy subjects. Many neurological conditions as studied in [13,19,27,28,46-48] are used to detect toe walking, a disorder indicative of more serious pathologies with unknown aetiology such as CP, neuropathy and myopathy making it difficult to identify. In CP, the challenge is to detect the type and severity of gait [13] such as type's I-IV, hemiplegia, diplegia and triplegia so timely measures can be applied (Table 2).

To conclude, earlier efforts to carry out gait diagnosis were performed through observations for measuring gait deterioration. The reassurances of patient revival and rehabilitation remained complicated without quantifications after surgical interventions. Recently, research is investigating the possibility of ML models to identify and evaluate the treatment measures. Monitoring and screening of at-risk gait are significant as detection of initial signs would enable protective actions such as physical movements and dietary regimes. Also, the development of disorders like knee OA, and balance impairments in the elderly occurs slowly over time to measure improvements. However, this progression needs to be tracked with the application of ML techniques towards prediction, and extrapolation interpolation investigations. Moreover, clinicians' effective diagnostic tools may consist of multimodal data for better predictions to assist physicians. These results reported are encouraging for decision making, rehabilitation planning and therapeutic interventions. As seen through analysis that SVM is a frequently used classifier among various classifiers used in the studies for pathology determination. The primary reason for using SVM in maximum studies is due to its better generalisation ability for the newer dataset by reducing empirical risks and can be used to detect movement disorders and their severity levels. Table 3 describes the accuracy statistics of using SVM classifier for gait disorder detection. Using descriptive statistics, parameters obtained are mean 87.55, standard error 1.98 and deviation of 7.42. Despite an aggregation of different studies, results reported were obtained using different sensors for varying conditions and involving various gait pathologies, SVM has shown remarkable capabilities for pathology determination. The future perspective is to devise a tool that entails different gait pathologies rather than a single one.

Table 3. Statistical analysis.

Gait disorder detection using SVM classifier

	Su	bjects		
Study	Healthy	Pathology	Parameters	Accuracy (%)
Lai et al., 2007 [33]	14	13	GRF	85.19
			Kinematics	74.07
			GRF and kinematics	88.88
Williams et al., 2015 [28]		102	Kinematics (unaffected limbs)	82.35
,			Kinematics (affected limbs)	76.47
Adil et al., 2016 [45]	5	5	sEMG signal	93
Cho et al., 2018 [46]	24	18	Kinematics (using feature selection algorithm)	95.2
			Kinematics	85.7
			Kinematics (without feature selection algorithm)	90.5
			Kinematics (with severity levels)	81
Guo et al., 2017 [27]	95	6	Foot pressure data	94.36
Cui et al., 2018 [47]	21	21	MT and GRF	95.83
			MT, GRF and EMG	98.21
Zhang and Ma, 2019 [48]		102	Kinematics	85

4.1.1.4. Gait asymmetry detection. The detection of small differences between the step motions of two lower limbs, results in asymmetric gait. The geriatric studies mainly emphasised on balance or postural controls to determine fall-risk. The higher asymmetries in lower limb dysfunction are indicative of fall-risk. Such type of evaluation finds applications particularly in rehabilitation [24], identification and determination of diseases that relates to reduced balance stability [25,53]. People with balance difficulties were analysed using inertial sensors (IMUs) or force plates. But the inclusion of EMG signals aids in fall prediction [53] as standing balance involves significant muscle usage. Even to access the recovery of the patients, gait indexes were developed to access the symmetry between dominant and non-dominant limb [50]. The results presented were encouraging and promising as highlighted in Table 2.

Arosha Senanayake et al. [12] combined fuzzy techniques with other classifiers for the development of the learning models to determine the recovery stages of ACL reconstructed subjects. This ML model consisting of kinematics and neuromuscular signals can be used by physiatrists, physiotherapists and clinicians for monitoring patient progress and recovery. To handle the more complex dataset, simple fuzzy rules were not enough. Thus, Semwal et al. [32] proposed hierarchical fuzzy systems for complex behaviour that leads to an exponential increase in a number of rules. Further, research was carried by Paulo et al. [24] to test the gait asymmetry using SVM and human machine interface (HMI) and overall pattern accuracy was about 88.3%. Another study by Liu et al. [25] used long short term memory (LSTM) to learn the correlations between gait features using gait trajectories from healthy subjects imitating knee injuries. The research by Qiu et al. [29] proposed health care monitoring using multi-sensor data fusion through ML and wearable devices.

The analysis shows that ML coupled with wearable sensors can be used to diagnose postural instabilities. Age, injury or illness can be the potential causes of balance difficulties resulting in fall-related injuries. The ML models developed using kinematics, kinetics and neuromuscular signals can aid physicians, physiotherapists in diagnosis, and monitoring recovery status. However, the analysis reported through these studies is offline. The ongoing research is using LSTM and DNNs to work on captured signals to draw inferences for clinical perspectives.

4.1.1.5. Neurological studies in gait. To determine participant's gait by studying the cognitive mechanisms [22], decoding cortical processes under different walking conditions [23], gait intention recognition [39] referred to as neurological gait studies. Gait is no longer considered related to motor functions only but also an activity that requires attention and executive functions. The decline in cognitive functions leads to gait dysfunction. Thus, cognitive analysis plays an important role in gait intervention studies. A study by Goh et al. [23] discussed that electroencephalogram (EEG) has unfolded the concept of event-related synchronisation/desynchronisation (ERS/ERD) at times of motor imagery of gait, ankle movements or lower limb movements. Further work by, Costa et al. [22] determined that the severity level and duration of a patient's recovery, by the cognitive mechanism through ML techniques were helpful in the determination of rehabilitation strategy. Another implementation by Hortal et al. [50] described the gait initiation and stopping using EEG signals that could prove to be beneficial in designing assistive devices for patients with motor disabilities.

The results of combining ML, gait, and the cognitive mechanism will be beneficial in the development of real-time systems at the time of rehabilitation processes, exoskeleton designs and clinical diagnosis. However, the accuracies achieved were around 70-80% using different supervised learning classifiers. This was due to difficulty in studying the cognitive mechanism using EEG and modelling it with gait. More precision is needed during data collection and in feature engineering to make robust systems.

4.1.2. Machine learning for gait rehabilitation

Rehabilitation helps to improve the patient's lost functions due to neurological disease, injury or illness. Gait rehabilitation can lead to mobility of bedbound, restoration in gait [54] and gait improvement [55,56]. The basic understanding of the biomechanical system has resulted in the simulation of prosthetic controls and robotic rehabilitation. Now, together with the advent of the applications of ML, the shift can be made from just modelling of partial differential equations to intelligent control. The various control strategies for gait rehabilitation were developed [57,58]. The study by Dallali et al. [57] applied ANN for control of ankle-footprostheses and able to predict the impedance with good prediction accuracy (~90%). Using the EMG signal of lower extremity, the method was able to control ankle impedance in dorsi-plantar and inversion-eversion directions. The intent to start and stop the gait based on EEG signal using SVM classifier had been tested on both healthy and patient groups with minimum latencies in real-time. To avoid the latencies, Park et al. [51] predicted the gait intention prior to the actual event using CNN. Thus, such systems can be used for the development of lower limb exoskeletons for SCI and stroke patients [50,51]. Even, the trained general regression neural network (GRNN) based model utilised in [59] for robotic gait rehabilitation depicts that the trained model showed better results in comparison to gait parameter equations methods. The purpose of robotic rehabilitation is to generate repetitive and progressive movements according to motor disabilities [59] for gait improvement. The performance of ML lies in its learning capability from a great amount of variability hidden in human characteristics. Further research by Huang et al. [30] had defined the trajectory approach in which movement of the healthy side is followed by the abnormal side and which helps in increasing the motivation level of an individual to generate active movements. Therefore, achieving better tracking performance required for paraplegic and hemiplegic patients. The RL framework was used for optimal policy control for the adaptation of patients to different walking conditions. By the use of LSTM and deep rehabilitation gait learning (DRGL) study carried out by Liu et al. [25] predicted and corrected abnormal joint trajectories of the knee based on normal joints of the wearers. The rehabilitation outcome will depend entirely on patients' ability to adapt to the speed and load variations determined by the device. Another study where ML has been successful is in predicting rehabilitation length and discharge schedule of the patient after an injury or illness using a sensor-based approach and on-going research is to plan more precise predictions. A study by Pla et al. [15] predicted the rehabilitation schedule for lower limb fractured subjects using SVM with accuracy statistics as $87.69 \pm 4.07\%$. This will enable improvement in patient's quality of life and reduction in cost after surgery.

Thus, if control strategies are successfully applied in dynamic walking conditions, they will assist the amputee population when interacting with the prostheses allowing for more natural gait based on their intent to walk. Recently, another on-going area in this field is in the use of brain-machine interfaces for gait prostheses using ML techniques that may lead to benefits in deriving the relationship between diverse signals and movement. Moreover, the robotic rehabilitation used for intervention is complex in nature involving multiple degrees-of-freedom (DOFs) causing it to be very expensive. Since this process involves continuous monitoring in a duration of a few months or years. A need of the hour is to develop affordable and portable devices to train individuals in home environments [58,60]. The home-based therapy is a promising solution. The robots when used in combination with rehabilitation approaches such as proprioceptive and haptic simulation will have the potential to recover cognitive ability together with motor functions.

4.2. Future perspectives

The joint efforts of ML and biomedical engineering have shown greater potential in gait detection such as activity detection, event detection, disorder detection, asymmetry detection and control of rehabilitation devices. They used data, algorithms, patterns and learning theory to arrive at the model. The greatest asset is saving time by circumventing multiple visits to medical professionals and avoiding long treatment duration. The review by Devanne and Berretti [20] proposed a need for the inclusion of statistical methods to handle noisy and missing data to ensure reliability. Strengthening of the tool can be achieved by the inclusion of all joint angles and planes required for gait diagnosis [13]. Another major concern is to make its applicability worldwide by overcoming its black box nature and to derive means to visualise the relationship between gait disorders and ML models.

Undoubtedly, further research in this area is not to doubt on the classifiers' accuracy. But the need is to handle feature engineering using domain knowledge. The gait analysis is diverse due to its non-stationary, temporal and complex nature. The contemporary state of the art is pathology detection at different levels because it needs a high degree of robustness to participants' variability. Therefore, much of the research lies in feature analysis methods and how to quantify the underlying biomechanics in learning models. However, the complex and tedious process of feature engineering is overcome by deep learning (a complex form of NN) and ongoing research is in the determination of the basis of decision making. Another burning area is in depicting the optimal classifier for a given dataset. Progress made during the last few decades in gait research has been established and for expansion, the future directions are highlighted in Table 2 with their current limitations.

A great need is to make the journey from offline analysis, i.e., laboratory-based experiments to real-time scenarios, i.e., clinician's efficient intelligent tools. A much needful is required to deal with patients' gait rather than normal gait though a basic understanding will pave way for abnormality detection. The generalisation of the model with different conditions, more population size, and transfer learning approach is needed, as this model once trained can be used on patients with similar impairments [37]. The ambitious application areas for gait analysis such as prediction of onset of diseases, the progression of the disease, rehabilitation rates and planning trajectories identical to human natural gait for controlling rehabilitation devices will get benefitted using AI in near future. However, technology development has proven useful by providing portability, continuous monitoring and cost-effective solution to gait study domains by the development of sensor-based approaches. The development of these approaches requires long-term data collection, monitoring of patient state, determining rehabilitation modules, and systematic data storage for training and validation of learning techniques. However, these techniques may possibly be applicable as assisting devices for physicians in clinical diagnosis and control of gait rehabilitation devices that can help physiotherapists in therapeutic interventions being adaptive to different conditions and entails different gait pathologies.

5. Conclusions

In this review, the progress made in gait studies with Al applications during the past few decades has been reported. The popularity of ML techniques in the gait domain is due to the fact that it can provide accurate, robust and fast classification by extracting simple features using highly temporal and nonlinear biomechanical data. The variability, instinctive, flexibility and adaptability of such dynamic approaches give them uptake over conventional ones based on thresholding. static detection and observational methods. Moreover, it simplifies the walking evaluation. The observational gait assessment is subjective in nature and lacks validity, reliability, sensitivity and specificity.

After literature screening, it has been depicted that the current studies are mainly focussing on sensorbased approach (86%); around 76.7% when using sensor approach alone and an additional 9.3% when used with other approaches thus, determining a shift from a laboratory-based environment to real walking environment.

The accuracy rates found for gait detection are: detection (\sim 85–100%), event detection $(\sim 95\%)$, abnormality detection $(\sim 90-98\%)$, asymmetry detection (\sim 85–95%) and GNE (\sim 67–86%). Though, these studies were carried out with different methods in varying conditions and for distinct domains. For activity detection, an accelerometer is the most used sensor for physical activity recognition (~85–95%) but for activities such as stair ascent and descent, performance improves with the addition of gyroscopic data for binary classification tasks. The combination of sensors at foot, shank and thigh is considered to be the best position to determine all types of activities. The GED has been successfully computed using ML, the best accuracies are reported using IMUs and FSRs; for three event detection (nearly 100%), four events $(\sim 98\%)$, five events $(\sim 83-97\%)$ and seven events a further decrease in accuracy was reported. The success rate of IMUs in abnormality detection is around 95% and KPCA is the best dimensionality reduction technique for non-linear data. The IMUs have predicted gait abnormality with an asymmetry score close to 1 for neurological disorders. Further, from the analysis, it can be concluded that these predictive models can



support healthcare professionals with this cost-effective solution but until now little has been done for its practical implementation as a diagnostics tool.

The coming years are looking forward to continuous monitoring, routine diagnostics and generating feedback for controlling techniques in gait analysis in real-time. Thereby, it can assist the rehabilitation by providing information to medical professionals and even to the patients. Hence, more exhaustive research needs to be done before using inertial sensors and ML techniques as a standardised clinical diagnostic tool. As a step towards standardisation of the method, it has to be tested more rigorously on the patient group as well. In consequence, the future intellectual systems for gait analysis will be based on sensor technology consisting of automated monitoring with biofeedback for walking detection in indoor and outdoor environments. Thus, it will enable benefits for a healthy quality of life.

Acknowledgements

The authors would like to present their sincere gratitude for the support and contribution provided by Director, CSIR-Central Scientific Instrument Organisation, and researchers in Biomedical Instrumentation Unit, CSIR-CSIO, Chandigarh, India. PK acknowledges UGC, India for supporting her Ph.D. through its national fellowship programme. A wholehearted thanks to independent reviewer Ms. Vibhuti (VI) for her contribution.

Disclosure statement

The authors declare no conflict of interest.

References

- Dolatabadi E, Taati B, Mihailidis A. An automated clas-[1] sification of pathological gait using unobtrusive sensing technology. IEEE Trans Neural Syst Rehabil Eng. 2017;25(12):2336-2346.
- [2] Lai DTH, Begg RK, Palaniswami M. Computational intelligence in gait research: a perspective on current applications and future challenges. IEEE Trans Inf Technol Biomed. 2009;13(5):687-701.
- Paulo J, Peixoto P, Amorim P. Trajectory-based gait pattern shift detection for assistive robotics applications. Intel Serv Robot. 2019;12(3):255-264.
- [4] Kutilek P, Hozman J. Prediction of lower extremities movement using characteristics of angle-angle diagrams and artificial intelligence. Proceedings of the 3rd International Conference on E-Health and Bioengineering - EHB: 2011.
- [5] Farah JD, Baddour N, Lemaire ED. Design, development, and evaluation of a local sensor-based gait phase recognition system using a logistic model

- decision tree for orthosis-control. J Neuroeng Rehabil. 2019;16(1):22.
- Begg R, Kamruzzaman J. Neural networks for detec-[6] tion and classification of walking pattern changes due to ageing. Aust Phys Eng Sci Med. 2006;29(2): 188-195.
- Lau HY, Tong KY, Zhu H. Support vector machine for classification of walking conditions using miniature kinematic sensors. Med Biol Eng Comput. 2008;46(6): 563-573.
- Souza AM, Stemmer MR. Extraction and classification [8] of human body parameters for gait analysis. J Control Autom Electr Syst. 2018;29(5):586-604.
- Hyeon-Min S, Sangmin L. Multi-channel electromyography pattern classification using deep belief networks for enhanced user experience. J Central South Univ. 2015;22(5):1801-1808.
- [10] Wei W, McElroy C, Dey S. Action understanding and movement error identification for the treatment of patients with Parkinson's disease. IEEE International Conference on Healthcare Informatics: 2018.
- [11] Leightley D, McPhee JS, Yap MH. Automated analysis and quantification of human mobility using a depth sensor. IEEE J Biomed Health Inform. 2017;21(4): 939-948.
- [12] Arosha Senanayake SMN, Malik OA, Iskandar PM, et al. A knowledge-based intelligent framework for anterior cruciate ligament rehabilitation monitoring. Appl Soft Comput. 2014;20:127-141.
- [13] Gestel LV, Laet TD, Lello ED, et al. Probabilistic gait classification in children with cerebral palsy: a Bayesian approach. Res Dev Disabil. 2011;32(6): 2542-2552.
- [14] Mu T, Pataky TC, Findlow AH, et al. Automated nonlinear feature generation and classification of foot pressure lesions. IEEE Trans Inf Technol Biomed. 2010; 14(2):418-424.
- [15] Pla A, Mordvanyuk N, Lopez M, et al. Bag-of-steps: predicting lower-limb fracture rehabilitation length. Neurocomputing. 2017;268:109-115.
- [16] Yoo TK, Kim SK, Choi SB, et al. Interpretation of movement during stair ascent for predicting severity and prognosis of knee osteoarthritis in elderly women using support vector machine. Proceedings of International Conference of the IEEE EMBS; 2013. p. 192-196.
- [17] Xiong B, Zeng N, Li H, et al. Intelligent prediction of human lower extremity joint moment: an artificial neural network approach. IEEE Access Spec Sect Data Enab Intell Dig Health. 2019;7:29973-29980.
- [18] Zeng W, Ismail SA, Pappas E. Classification of gait patterns in patients with unilateral anterior cruciate ligament deficiency based on phase space reconstruction. Soft Comput. 2020;24(3):1851-1868.
- [19] Alaqtash M, Sarkodie-Gyan T, Yu H, et al. Automatic classification of pathological gait patterns using ground reaction forces and machine learning algorithms. Proceedings of International Conference of the IEEE-EMBS; Boston, MA; 2011. p. 453-457.
- [20] Devanne M, Berretti S. Learning shape variations of motion trajectories for gait analysis. 23rd International

- Conference on Pattern Recognition (ICPR); 2016. p. 895-900.
- [21] Begg RK, Palaniswami M, Owen B. Support vector machines for automated gait classification. IEEE Trans Biomed Eng. 2005;52(5):828-838.
- [22] Costa A, Ianez E, Ubeda A, et al. Decoding the attentional demands of gait through EEG gamma band features. PLoS One. 2016;11(4):e0154136.
- [23] Goh SK, Abbass HA, Tan KC, et al. Spatio-spectral representation learning for electroencephalographic gaitpattern classification. IEEE Trans Neural Syst Rehabil Eng. 2018;26(9):1858-1867.
- [24] Paulo J, Peixoto P, Nunes UJ. ISR-AIWALKER: robotic walker for intuitive and safe mobility assistance and gait analysis. IEEE Trans Hum Mach Syst. 2017;47(6):
- [25] Liu DX, Du W, Wu X, et al. Deep rehabilitation gait learning for modelling knee joints of lower limb exoskeleton. Proceedings of IEEE International Conference on Robotics and Biomimetics; 2016. p. 1058-1063.
- [26] Thongsook A, Nunthawarasilp T, Kraypet P, et al. C4.5 decision tree against neural network on gait phase recognition for lower limp exoskeleton. First International Symposium on Instrumentation, Control, Artificial Intelligence, and Robotics (ICA-SYMP); 2019. p. 69-72.
- Guo G, Guffey K, Chen W, et al. Classification of normal and pathological gait in young children based on foot pressure data. Neuroinformatics. 2017;15(1): 13-24.
- [28] Williams G, Lai D, Schache A, et al. Classification of gait disorders following traumatic brain injury. J Head Trauma Rehabil. 2015;30(2):E13-E23.
- [29] Qiu S, Liu L, Wang Z, et al. Body sensor networkbased gait quality assessment for clinical decisionsupport via multi-sensor fusion. IEEE Access Spec Sect Wireless Body Area Netw. 2019;7:59884-59894.
- [30] Huang R, Peng Z, Cheng H, et al. Learning-based walking assistance control strategy for a lower limb exoskeleton with hemiplegia patients. IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS); 2018. p. 2280-2285.
- [31] Muniz AMS, Manfio EF, Andrade MC, et al. Principal component analysis of vertical ground reaction force: a powerful method to discriminate normal and abnormal gait and assess treatment. Proceedings of International Conference of the IEEE Engineering in Medicine and Biology Society; 2006. p. 2683-2686.
- [32] Semwal VB, Chakraborty P, Nandi GC. Less computationally intensive fuzzy logic (type-1)-based controller for humanoid push recovery. Rob Auton Syst. 2015; 63:122-135.
- [33] Lai DTH, Levinger P, Begg RK, et al. Identification of patellofemoral pain syndrome using a support vector machine approach. Proceedings of the 29th Annual International Conference of the IEEE EMBS; 2007. p. 3144-3147.
- [34] Nair SS, French RM, Laroche D, et al. The application of machine learning algorithms to the analysis of electromyographic patterns from arthritic patients.

- IEEE Trans Neural Syst Rehabil Eng. 2010;18(2): 174-184.
- [35] Moher D, Liberati A, Tetzlaff J, et al. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. PLoS Med. 2009;6(7):e1000097.
- [36] Pham H. Kawanishi M. Narikivo T. A LLE-HMM-based framework for recognizing human gait movement from EMG. IEEE International Conference on Robotics and Automation (ICRA); 2015. p. 2997-3002.
- [37] Lonini L, Gupta A, Kording K, et al. Activity recognition in patients with lower limb impairments: do we need training data from each patient? International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC); 2016.
- [38] Recher F, Banos O, Nikamp CDM, et al. Optimizing activity recognition in stroke survivors for wearable exoskeletons. 7th IEEE International Conference on Biomedical Robotics and Biomechatronics (Biorob): 2018. p. 173-178.
- [39] Sok P, Xiao T, Azeze Y, et al. Activity recognition for incomplete spinal cord injury subjects using hidden Markov models. IEEE Sens J. 2018;18(15):6369-6374.
- [40] Lau H, Tong K. The reliability of using accelerometer and gyroscope for gait event identification on persons with dropped foot. Gait Posture. 2008;27(2): 248-257.
- [41] Jung JY, Chae MG, Jang IH, et al. A hybrid control method of an exoskeleton robot for intention-driven walking rehabilitation of stroke patients. 9th International Conference on Ubiquitous Robots and Ambient Intelligence (URAI); 2012. p. 58-60.
- [42] Jiang S, Pang Y, Wang D, et al. Gait recognition based on graphene porous network structure pressure sensors for rehabilitation therapy. Proceedings of IEEE International Conference on Electron Devices and Solid State Circuits (EDSSC); 2018.
- [43] Nazmi N, Rahman MAA, Ariff MHM. Generalization of ANN model in classifying stance and swing phases of gait using EMG signals. IEEE-EMBS Conference on Biomedical Engineering and Sciences; 2018. p. 461-466.
- [44] Lee SI, Ozsecen MY, Toffola LD, et al. Activity detection in uncontrolled free living conditions using a single accelerometer. Proceedings of IEEE International Conference on Wearable and Implantable Body Sensor Networks (BSN); 2015.
- [45] Adil S, Anwar T, Jumaily AA. Extreme learning machine based sEMG for drop-foot after stroke detection. Proceedings of International Conference on Information Science and Technology; 2016. p. 18–22.
- Cho J, Cho YS, Moon SB, et al. Scoliosis screening through a machine learning based gait analysis test. Int J Precis Eng Manuf. 2018;19(12):1861-1872.
- Cui C, Bian GB, Hou ZG, et al. Simultaneous recogni-[47] tion and assessment of post-stroke hemiparetic gait by fusing kinematic, kinetic, and electrophysiological data. IEEE Trans Neural Syst Rehabil Eng. 2018;26(4): 856-864.
- [48] Zhang Y, Ma Y. Application of supervised machine learning algorithms in the classification of sagittal gait patterns of cerebral palsy children with spastic diplegia. Comput Biol Med. 2019;106:33-39.



- [49] Sobral H, Ferreira JP, Vieira A, et al. Two new indices to assess gait disturbances applied to anterior cruciate ligament reconstructed knees. Proceedings of IEEE 8th Annual International Conference on CYBER Technology in Automation, Control, and Intelligent Systems; 2018. p. 701-706.
- [50] Hortal E, Ubeda A, Ianez E, et al. Using EEG signals to detect the intention of walking initiation and stop. Artif Comput Biol Med (IWINAC). 2015;278-287.
- [51] Park S, Park FC, Choi J, et al. EEG-based gait state and gait intention recognition using spatio-spectral convolutional neural network. IEEE International Winter Conference on Brain-Computer Interface (BCI): 2019.
- Schutte LM, Narayanan U, Stout JL, et al. An index for [52] quantifying deviations from normal gait. Gait Posture. 2000;11(1):25-31.
- [53] Ghasemzadeh H, Jafari R, Prabhakaran B. A body sensor network with electromyogram and inertial sensors: multimodal interpretation of muscular activities. IEEE Trans Inf Technol Biomed. 2010;14(2):198-206.
- [54] Levinger P, Lai DTH, Begg RK, et al. The application of support vector machines for detecting recovery from knee replacement surgery using spatio-temporal gait parameters. Gait Posture. 2009;29(1):91-96.
- [55] Alagtash M, Yu H, Brower R, et al. Application of wearable sensors for human gait analysis using fuzzy

- computational algorithm. Eng Appl Artif Intell. 2011; 24(6):1018-1025.
- [56] Ferreira JP, Vieira A, Ferreira P, et al. Human knee joint walking pattern generation using computational intelligence techniques. Neural Comput Appl. 2018; 30(6):1701-1713.
- Dallali H, Knop L, Castelino L, et al. Estimating the multivariable human ankle impedance in dorsi-plantarflexion and inversion-eversion directions using EMG signals and artificial neural networks. Int J Intell Robot Appl. 2017;1(1):19-31.
- Khor KX, Chin PJH, Hisyam AR, et al. Development of [58] CR2-Haptic: a compact and portable rehabilitation robot for wrist and forearm training. EEE Conference on Biomedical Engineering and Sciences (IECBES); 2014.
- [59] Luu TP, Lim HB, Hoon KH, et al. Subject specific gait parameters prediction for robotic gait rehabilitation generalized regression neural Proceedings of IEEE International Conference on Robotics and Biomimetics; 2011. p. 914-919.
- [60] Gholami M, Ejupi A, Rezaei A, et al. Estimation of knee joint angle using a fabric-based strain sensor and machine learning: a preliminary investigation. 7th IEEE International Conference on Biomedical Robotics and Bio mechatronics (Biorob); 2018. p. 589-594.