TEST RECORD JnJ Assy No. 0100-8945

BES-13701-03 REV. (Use latest revision)

JOB#: 1004202 R4

		QDF6: (Pdecay< 0.5 PSI/min)					QDM14: (Pdecay < 0.5 PSI/min)					QDF2: (Pdecay< 0.5 PSI/min)					QDF4: (Pdecay< 0.5 PSI/min)				
DATE CODE	SERIAL NUMBER	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL	START	STOP		PASS	FAIL
09/14/26	1004202-01	87.2	87.2	0	✓		86.5	86.5	0	✓		27.5	27.5	0	✓		27.3	27.3	0	✓	
3926	-02	87.2	87.2	0	✓		87.0	87.0	0	✓		27.5	27.5	0	✓		27.3	27.3	0	✓	
01-	-03	87.3	87.3	0	✓		86.6	86.6	0	✓		27.8	27.8	0	✓		27.7	27.7	0	✓	
	-04	87.2	87.2	0	✓		86.6	86.5	0.1	✓		27.3	27.3	0	✓		27.1	27.1	0	✓	
	-05	87.3	87.3	0	✓		87.1	87.0	0.1	✓		27.5	27.5	0	✓		27.4	27.4	0	✓	
	-06	87.2	87.1	0.1	✓		86.4	86.4	0	✓		26.3	26.3	0	✓		26.2	26.2	0	✓	
	-07	87.2	87.2	0	✓		86.8	86.8	0	✓		27.3	27.3	0	✓		27.2	27.2	0	✓	
	-08	86.2	86.2	0	✓		85.6	85.6	0	✓		27.3	27.3	0	✓		27.2	27.2	0	✓	
	-09	86.1	86.0	.1	✓		85.4	85.3	.1	✓		26.9	26.9	0	✓		26.9	26.9	0	✓	
	-10	86.1	86.0	0.1	✓		86.0	85.9	0.1	✓		26.2	26.2	0	✓		26.1	26.1	0	✓	
	11	86.0	86.0	0	✓		85.3	85.3	0	✓		27.4	27.4	0	✓		27.2	27.2	0	✓	
	12	86.1	86.1	0	✓		85.5	85.5	0	✓		27.4	27.4	0	✓		27.3	27.2	0.1	✓	
	13	85.9	85.7	0.2	✓		85.2	85.0	0.2	✓		26.7	26.7	0	✓		26.5	26.5	0	✓	
	14	85.9	85.9	0	✓		85.2	85.1	0.1	✓		27.5	27.5	0	✓		27.4	27.4	0	✓	
	15	86.1	86.1	0	✓		85.4	85.4	0	✓		27.5	27.5	0	✓		27.4	27.4	0	✓	
	16	86.0	86.0	0	✓		85.4	85.3	0.1	✓		26.1	26.1	0	✓		25.9	25.9	0	✓	
	17	86.0	85.9	0.1	✓		85.2	85.2	0	✓		26.1	26.1	0	✓		26.0	26.0	0	✓	
	18	85.8	85.7	0.1	✓		85.3	85.3	0	✓		25.7	25.7	0	✓		25.5	25.5	0	✓	
	19	85.9	85.9	0	✓		85.4	85.4	0	✓		26.9	26.8	0.1	✓		26.7	26.6	0.1	✓	
	20	85.8	85.8	0	✓		85.3	85.2	0.1	✓		27.6	27.5	0.1	✓		27.2	27.1	0.1	✓	
	21	86.9	86.9	0	✓		86.6	86.6	0	✓		26.5	26.5	0	✓		26.0	26.0	0	✓	
	22	86.9	86.8	0.1	✓		86.2	86.2	0	✓		26.0	26.0	0	✓		25.5	25.5	0	✓	
	23	87.5	87.5	0	✓		86.8	86.7	0.1	✓		27.2	27.2	0	✓		27.1	27.1	0	✓	
	24	87.4	87.4	0	✓		86.7	86.6	0.1	✓		27.2	27.1	0.1	✓		26.9	26.9	0	✓	
	25	87.1	87.1	0	✓		86.4	86.3	0.1	✓		26.7	26.7	0	✓		26.4	26.4	0	✓	
	26	87.0	87.0	0	✓		86.2	86.2	0	✓		25.9	25.9	0	✓		25.7	25.7	0	✓	
	27																				

TEST RECORD JnJ Assy No. 0100-8945

BES-13701-03 REV. (Use latest revision)

JOB#: 1004202

		CV5: (Pdecay < 0.15 PSI/min)					QDM5: (Pdecay< 2mmHg/min) ER1					QDF22: (Pdecay < 2mmHg/min)					QDF3:(Pdecay < 5mmHg/min)				
DATE CODE	SERIAL NUMBER	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL
	-01	3.05	3.05	0	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-693	-693	0	✓	
	-02	2.99	3.00	0.01	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-692	-692	0	✓	
	-03	3.03	3.03	0	✓		22.2	22.2	0	✓		22.2	22.2	0	✓		-694	-694	0	✓	
	-04	3.04	3.04	0	✓		22.5	22.5	0	✓		22.5	22.5	0	✓		-688	-688	0	✓	
	-05	3.04	3.04	0	✓		22.3	22.3	0	✓		22.3	22.3	0	✓		-695	-695	0	✓	
	-06	3.04	3.03	0.01	✓		22.2	22.2	0	✓		22.2	22.2	0	✓		-693	-693	0	✓	
	-07	3.03	3.03	0	✓		22.6	22.6	0	✓		22.6	22.6	0	✓		-686	-686	0	✓	
	-08	3.00	2.98	0.02	✓		22.5	22.5	0	✓		22.5	22.5	0	✓		-687	-687	0	✓	
	-09	3.01	3.01	0	✓		22.7	22.7	0	✓		22.7	22.7	0	✓		-694	694	0	✓	
	-10	3.01	3.01	0	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-707	706	1	✓	
	11	2.94	2.94	0	✓		22.5	22.5	0	✓		22.5	22.5	0	✓		-697	-697	0	✓	
	12	3.00	2.99	0.01	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-692	-692	0	✓	
	13	2.93	2.93	0	✓		22.7	22.7	0	✓		22.7	22.7	0	✓		-688	-688	0	✓	
	14	3.02	3.01	0.01	✓		22.7	22.7	0	✓		22.7	22.7	0	✓		-694	-694	0	✓	
	15	2.95	2.94	0.01	✓		21.6	21.6	0	✓		21.6	21.6	0	✓		-694	-694	0	✓	
	16	3.04	3.03	0.01	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-699	-694	0	✓	
	17	2.97	2.97	0	✓		22.2	22.2	0	✓		22.2	22.2	0	✓		-692	-692	0	✓	
	18	3.00	2.99	0.01	✓		22.3	22.3	0	✓		22.3	22.3	0	✓		-687	-687	0	✓	
	19	2.99	2.99	0	✓		22.7	22.7	0	✓		22.7	22.7	0	✓		-689	-689	0	✓	
	20	3.02	3.02	0	✓		22.5	22.5	0	✓		22.5	22.5	0	✓		-689	-689	0	✓	
	21	2.96	2.96	0	✓		22.2	22.2	0	✓		22.2	22.2	0	✓		-699	-699	0	✓	
	22	3.07	3.07	0.0	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-701	-701	0	✓	
	23	2.98	2.98	0	✓		22.6	22.6	0	✓		22.4	22.4	0	✓		-691	-691	0	✓	
	24	3.02	3.01	0.01	✓		22.5	22.5	0	✓		22.5	22.5	0	✓		-690	-690	0	✓	
	25	3.00	3.01	0.01	✓		22.4	22.4	0	✓		22.4	22.4	0	✓		-694	-694	0	✓	
	26	3.00	3.00	0	✓		22.5	22.5	0	✓		22.5	22.5	0	✓		-691	-691	0	✓	
	27																				

TEST RECORD JnJ Assy No. 0100-8945

BES-13701-03 REV. (Use latest revision)

JOB#: 1004202 2/10/20 DC

7/10/20 VDC

		QDF18: (Pdecay < 0.5 PSI/min)					QDM7: (Pdecay < 0.5 PSI/min)					QDM1: (Pdecay < 5mmHg/min) ER2					QDF19: (Pdecay < 5mmHg/min)				
DATE CODE	SERIAL NUMBER	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL	START	STOP	DIFF	PASS	FAIL	START			PASS	FAIL
	-01	86.5	86.4	0.1	✓		86.5	86.5	0	✓		-686	-		✓		-686			✓	
	-02	86.7	86.7	0	✓		86.8	86.7	0.1	✓		-681			✓		-681			✓	
	-03	86.5	86.5	0	✓		86.4	86.4	0	✓		-684			✓		-684			✓	
	-04	86.3	86.2	0.1	✓		86.3	86.2	0.1	✓		-682			✓		-682			✓	
	-05	87.0	87.0	0	✓		86.9	86.9	0	✓		-687			✓		-687			✓	
	-06	86.6	86.6	0	✓		86.6	86.6	0	✓		-685			✓		-685			✓	
	-07	86.9	86.9	0	✓		86.9	86.9	0	✓		-680			✓		-680			✓	
	-08	85.7	85.6	.1	✓		85.7	85.7	0	✓		-680			✓		-680			✓	
	-09	85.4	85.3	.1	✓		85.4	85.3	.1	✓		-688			✓		-688			✓	
	-10	85.8	85.6	.2	✓		85.7	85.5	.2	✓		-701			✓		-701			✓	
	11	85.1	85.1	0	✓		85.2	85.2	0	✓		-690			✓		-690			✓	
	12	85.6	85.5	0.1	✓		85.6	85.5	0.1	✓		-686			✓		-686			✓	
	13	85.2	85.2	0	✓		85.2	85.1	0.1	✓		-684			✓		-684			✓	
	14	85.2	85.2	0	✓		85.2	85.1	0.1	✓		-689			✓		-689			✓	
	15	85.2	85.2	0	✓		85.2	85.1	0.1	✓		-687			✓		-687			✓	
	16	85.2	85.2	0	✓		85.2	85.2	0	✓		-691			✓		-691			✓	
	17	85.1	85.1	0	✓		85.2	85.2	0	✓		-684			✓		-684			✓	
	18	85.2	85.2	0	✓		85.1	85.1	0	✓		-678			✓		-678			✓	
	19	85.6	85.6	0	✓		85.6	85.5	0.1	✓		-682			✓		-682			✓	
	20	84.7	84.7	0	✓		84.9	84.9	0	✓		-683			✓		-683			✓	
	21	86.0	86.0	0	✓		86.1	86.1	0	✓		-682			✓		-682			✓	
	22	85.9	85.9	0	✓		86.0	86.0	0	✓		-690			✓		-690			✓	
	23	86.9	86.9	0	✓		86.9	86.9	0	✓		-674			✓		-674			✓	
	24	86.4	86.4	0	✓		86.4	86.4	0	✓		-687			✓		-687			✓	
	25	86.2	86.2	0	✓		86.2	86.2	0	✓		-685			✓		-685			✓	
	26	86.1	86.1	0	✓		86.1	86.1	0	✓		-684			✓		-684			✓	
	27																				

TEST RECORD JnJ Assy No. 0100-8945

BES-13701-03 REV. (Use latest revision) JOB#:

Serial No.	Failure Notes:	Repair Notes:
	ER2 x 1 vac hi at 0v	
ER1: 111	ER1 x 1 No pressure at QDM5	
ER2: 111	ER2 x 1 tuned & passed. unstable vac.	
	ER2 had high vac at 1,3,5 v.	(-88, -238, -388)
	ER1 Failed console.	tested twice. Unstable graph.
	ER1 Failed console.	

09/21/26

Test Record JnJ Assy No. 0100-8945

BENCH TEST VACCUUM TEST

BES-13701-03 REV L

JOB#: 1004202

DATE CODE	S/N	ORIG CV5	CV3	0V	1V	3V	5V	8V	10V
	-01	3.23	0.43	0	-74	-224	-375	-600	-686
3826	-02	3.90	0.42	0	-61	-215	-369	-599	-681
09-21-26	-03	3.50	0.43	0	-63	-216	-369	-599	-684
	04	3.22	0.42	0	-68	-221	-372	-598	-682
	05	3.86	0.42	0	-64	-217	-370	-598	-687
	06	4.42	0.43	0	-81	-235	-386	-613	-685 ~~##~~
	07	3.83	0.42	0	-71	-221	-371	-595	-680
	08	3.60	0.40	0	-75	-226	-377	-600	-680
	09	4.10	0.45	0	-62	-218	-371	-602	-688
	10	3.81	0.47	0	-65	-216	-368	-594	-701
	11	1.62	0.42	0	-81	-235	-386	-611	-690
	12	3.54	0.43	0	-79	-230	-380	-604	-686
	13	4.40	0.45	0	-77	-228	-379	-603	-684
	14	4.17	0.42	0	-77	-230	-382	-605	-689
	15	3.82	0.42	0	-77	-227	-385	-606	-687
	16	4.12	0.44	0	-73	-224	-375	-600	-691
	17	3.75	0.40	0	-71	-224	-377	-605	-684
	18	3.69	0.42	0	-69	-223	-376	-604	-678
	19	3.86	0.43	0	-78	-228	-376	-599	-682
	20	3.65	0.43	0	-71	-225	-377	-604	-683
	21	3.43	0.41	0	-75	-229	-381	-606	-682
	22	3.90	0.42	0	-76	-228	-380	-604	-690
	23	3.67	0.41	0	-78	-228	-378	-601	-674
	24	3.45	0.42	0	-74	-224	-374	-598	-687
	25	3.89	0.42	0	-69	-223	-376	-604	-685
	26	3.73	0.42	0	-74	-227	-381	-610	-684
	27								

= tuned ER1

= tuned ER2