PHILIPS

Document Information
Document Identification: D000992888
Document Name: ISP13 DRR validation results
Version: C.2
Life cycle state: Approved 08-Dec-2022

Cover Sheet

Document Classification
Design Category: 6. Design Validation
DMR Group Number: -

Signature Information		
Role	**Assignee**	**Date**
Approver (Accountable)	Afek, Keren	03-Dec-2022
Approver (User)	Kavirayar, Chockalingam	03-Dec-2022
Approver (User)	Singh, Jai	05-Dec-2022
Approver (User)	Aryev, Anna	08-Dec-2022
Approver (User)	Eyal Biran	03-Dec-2022

Security Information		
IP Sensitivity	**Strategic Good Indicator**	**ECCN**
Internal	3 Controlled, listed in the US EAR US-origin/subject to the US EAR	EAR99

Note: The stamp below is identifying the following elements
ID number; Version; State; Document Name; DMR Group Number

D000992888;C.2;Released;ISP13 DRR validation results;-

DESIGN REVIEW RECORD

1 Document Introduction 3
1.1 Purpose....3
1.2 Scope 3
1.2.1 VAL1: CT ASPECTS....3
1.2.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMRQflow , CMR Functional Long Axis .3
1.2.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping 3
1.3 References....3
1.4 Abbreviations and Definitions 3
2 Design Review Material 4
2.1 VAL1: CT ASPECTS....4
2.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMR Qflow, CMR Functional Long Axis 4
2.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping 4
3 Design Review Meeting 4
3.1 Meeting date and participants 4
3.2 Meeting date and participants 4
3.3 Meeting date and participants 5
3.4 Validation results review 5
3.4.1 VAL1: CT ASPCTS....5
3.4.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMRQflow , CMR Functional Long Axis .6
3.4.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping 8
3.5 Validation exit criteria 10
3.5.1 VAL1: CT ASPECTS....10
3.5.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMRQflow , CMR Functional Long Axis 10
3.5.1 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping 11
4 Actions 12
4.1 VAL1: CT ASPECTS....12
4.2 VAL2: CT Liver, CT CCS, CMR Viewer, CMR Spatial, CMR Qflow , CMR Functional Long Axis....12
4.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping 12
4.4 VAL1: Validation dates and evaluators....12
4.4.1 CT ASPECTS....12
4.5 VAL2: Validation dates and evaluators....12
4.5.1 CT Liver 12
4.5.2 CT CCA 12
4.5.3 CMR Viewer....12
4.5.4 CMR Spatial....13
4.5.5 CMR Qflow 13
4.5.6 CMR Functional Long Axis 13
4.6 VAL3: Validation dates and evaluators....14
4.6.1 CT Auto Calcium Scoring (CAA)....14
4.6.2 CT sAVA 14
4.6.3 MR Longitudinal Brain Analysis (MR CVS)....14

Printed copies are uncontrolled unless authenticated

4.6.4 CMR Functional LV RV 14
4.6.5 CMR Mapping 14
4.6.6 CMR Temporal 15

5 Conclusion 15

5.1 VAL1: CT ASPECTS 15
5.2 VAL2: CT Liver, CT CCA, CMR viewer, CMR Spatial, CMR Qflow, CMR Functional Long Axis 15
5.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping 16

6 Record history 16

7 Record Review and Approval 16

8 Appendix A: DRR presentation 17

Printed copies are uncontrolled unless authenticated

1 Document Introduction

1.1 Purpose

Purpose of this document is to approve the Advance Visualization Workspace v15.0 results as summarized in the Validation Report D000926820.

To ensure the correctness, completeness and traceability of the validation results relative to the user needs. Ensure that the user needs and product design are consistent and complete

Ensure that all Readiness activities for design transfer activities as identified in the design transfer plan

1.2 Scope

1.2.1 VAL1: CT ASPECTS

The scope of this design review is validation of product ISP13 VAL1 and Design transfer Readiness Review.

1.2.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMR Qflow , CMR Functional Long Axis

The scope of this design review is validation of product ISP13 VAL2 and Design transfer Readiness Review

1.2.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping

The scope of this design review is validation of product ISP13 VAL3 and Design transfer Readiness Review

1.3 References

Reference Number	Document Title	Document iD
In windchill	ISP13 Validation Report	D000926820
	ISP13 Design Validation Plan	DHF380226
	ISP13 PED	D000926821
	ISP13 Users Requirements Specification (URS)	DHF380082
	ISP13 traceability matrix for CT ASPECTS	D000889687
	ISP13_MR Cardiac Suite Traceability Matrix	D000996539
	ISP13 Traceability Matrix CT Cardiac	D000946874
	Traceability Matrix CT Liver	D001044264

1.4 Abbreviations and Definitions

Terminology & Abbreviations	Description/Definition
TRR	Technical Review Record
PED	Production Equivalent Device

Printed copies are uncontrolled unless authenticated

2 Design Review Material

2.1 VAL1: CT ASPECTS

Document Reference	Title	Rev
D000926820	ISP13 Validation Report	A

2.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMR Qflow, CMR Functional Long Axis

Document Reference	Title	Rev
D000926820	ISP13 Validation Report	B

2.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping

Document Reference	Title	Rev
D000926820	ISP13 Validation Report	C

3 Design Review Meeting

3.1 Meeting date and participants

VAL1 - Date: 19-JAN-2022

Name	Role	Represented Function	Independent
Sheela Biradar	Verification Lead	R&D	No
Shipra Shivangi	Quality Engineer	Q&R	No
Jai Singh	Independent Reviewer	Clinical Excellence	Yes
Sujesh Paravatty	Clinical Applications Lead	Clinical Excellence	No
Keren Afek	Regulatory Affairs Specialist	Q&R	No
Hagit Davidesco	Clinical Validation Lead	Clinical Solutions(Review Owner)	No
Offer Barziali	Product Manager	Marketing	No
Chockalingam Kavirayar	System architect	R&D	No
Yuval Kopel*	NPI Project Leader (Manufacturing, Customer Service and supply chain)	OSS	No
Adi Ron*	Design Transfer Lead	OSS	No

* OSS representative do not need to attend as this is an intermediate validation results review only

3.2 Meeting date and participants

VAL2 - Date: 28-JUL-2022

Name	Role	Represented Function	Independent
Aashish Kumar	Verification Lead	R&D	No

Printed copies are uncontrolled unless authenticated

Anna Aryev	Quality Engineer	Q&R	No
Jai Singh	Independent Reviewer	Clinical Excellence	Yes
Sujesh Paravatty	Clinical Applications Lead	Clinical Excellence	No
Keren Afek	Regulatory Affairs Specialist	Q&R	No
Hagit Davidesco	Clinical Validation Lead	Clinical Solutions (Review Owner)	No
Offer Barziali	Product Manager	Marketing	No
Chockalingam Kavirayar	System architect	R&D	No
Yuval Kopel*	NPI Project Leader (Manufacturing, Customer Service and supply chain)	OSS	No
Adi Ron*	Design Transfer Lead	OSS	No

* OSS representative do not need to attend as this is an intermediate validation results review only

3.3 Meeting date and participants

VAL3 - Date: 21-NOV-2022

Name	Role	Represented Function	Independent
Aashish Kumar	Verification Lead	R&D	No
Anna Aryev	Quality Engineer	Q&R	No
Jai Singh	Independent Reviewer	Clinical Excellence	Yes
Sujesh Paravatty	Clinical Applications Lead	Clinical Excellence	No
Keren Afek	Regulatory Affairs Specialist	Q&R	No
Hagit Davidesco	Clinical Validation Lead	Clinical Solutions (Review Owner)	No
Efrat Kliper Goldin	Product Manager	Marketing	No
Chockalingam Kavirayar	System architect	R&D	No
Yuval Kopel	NPI Project Leader (Manufacturing, Customer Service and supply chain)	OSS	No
Adi Ron	Design Transfer Lead	OSS	No

3.4 Validation results review

3.4.1 VAL1: CT ASPCTS

	Item	Yes / No	Evidence
Validation results review:	Have the resulting changes from product defect fixes resulted in consistent requirements updates to design input and product design(URS,PRS,SDS, DRS-DDS (if applicable)	No	See section conclusion in this doc.
	Are the user needs still fulfilled?	Yes	DHF380226 ISP13 Validation Plan section traceability to user requirements
	Do the validation results cover the user needs?	Yes	DHF380226 ISP13 Validation Plan section traceability to user requirements
	Is product labeling validated and ready for release?	Yes	Build screenshots are available in D000926821 ISP13 PED

Printed copies are uncontrolled unless authenticated

	Are the risks properly mitigated? Is the risk analysis done and mitigated for all the hazardous defects noticed during validation? (Either clinically or by design)	NA	No risks were found during validation
	Product Defect Summary from validation presented (All the defects identified from validation need to be resolved/verified before the product release)	Yes	D000926820 ISP13 validation report defect section
	Have the Validation plan and Validation report, Validation results are reviewed and approved appropriately?	Yes	• DHF380226 ISP13 Validation plan • D000926820 ISP13 validation report section training forms • D000926820 ISP13 validation report section evaluation forms
	Does the validation results requires a change in requirements?	No	no change in requirements is needed
	Is the link between DHF & DMR appropriate?	NA	All activities will be completed towards the production ECO
	Is the traceability evidenced as appropriate for this stage?	Yes	D000889687 ISP13 traceability matrix for CT ASPECTS
Design Transfer readiness review:	Are approved design outputs used for Design transfer?	NA	will be part as of ECO
	Are all the Design transfer activities completed as planned? E.g. Creation of manufacturing SOP's, completion of process validation etc.	NA	All activities will be completed towards the production ECO
	Check the readiness of functions such as Manufacturing, Service, Customer support for product release	NA	The readiness will be checked at ECO
	Are there any other prerequisites to start Design Transfer activities?	NA	All activities will be completed towards the production ECO
	Is the link between DHF & DMR appropriate? Provide evidences for the same.	NA	All activities will be completed towards the production ECO

3.4.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMR Qflow , CMR Functional Long Axis

	Item	Yes / No	Evidence
Validation results review:	Have the resulting changes from product defect fixes resulted in consistent requirements updates to design input and product design(URS,PRS,SDS, DRS-DDS (if applicable)	No	See section conclusion in this doc.

Printed copies are uncontrolled unless authenticated

	Are the user needs still fulfilled?	Yes	DHF380226 ISP13 Validation Plan section traceability to user requirements
	Do the validation results cover the user needs?	Yes	DHF380226 ISP13 Validation Plan section traceability to user requirements
	Is product labeling validated and ready for release?	Yes	Build screenshots are available in D000926821 ISP13 PED
	Are the risks properly mitigated? Is the risk analysis done and mitigated for all the hazardous defects noticed during validation? (Either clinically or by design)	NA	No risks were found during validation
	Product Defect Summary from validation presented (All the defects identified from validation need to be resolved/verified before the product release)	Yes	D000926820 ISP13 validation report defect section D000926809 ISP13 Validation Site Management defect section
	Have the Validation plan and Validation report, Validation results are reviewed and approved appropriately?	Yes	• DHF380226 ISP13 Validation plan • D000926820 ISP13 validation report section training forms • D000926820 ISP13 validation report section evaluation forms
	Does the validation results requires a change in requirements?	No	no change in user requirements is needed
	Is the link between DHF & DMR appropriate?	NA	All activities will be completed towards the production ECO
	Is the traceability evidenced as appropriate for this stage?	Yes	• D000996539 ISP13_MR Cardiac Suite Traceability Matrix Rev_C • D000946874 ISP13 Traceability Matrix CT Cardiac • D001044264_Traceability Matrix CT Liver
Design Transfer readiness review:	Are approved design outputs used for Design transfer?	NA	will be part as of ECO
	Are all the Design transfer activities completed as planned? E.g. Creation of manufacturing SOP's, completion of process validation etc.	NA	All activities will be completed towards the ECO
	Check the readiness of functions such as Manufacturing, Service, Customer support for product release	NA	The readiness will be checked at ECO
	Are there any other prerequisites to start Design Transfer activities?	NA	All activities will be completed towards the ECO

Printed copies are uncontrolled unless authenticated

	Is the link between DHF & DMR appropriate? Provide evidences for the same.	NA	All activities will be completed towards the ECO

3.4.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping

	Item	Yes / No	Evidence
Validation results review:	Have the resulting changes from product defect fixes resulted in consistent requirements updates to design input and product design(URS,PRS,SDS, DRS-DDS (if applicable)	No	See section conclusion in this doc.
	Are the user needs still fulfilled?	Yes	DHF380226 ISP13 Validation Plan section traceability to user requirements
	Do the validation results cover the user needs?	Yes	DHF380226 ISP13 Validation Plan section traceability to user requirements
	Is product labeling validated and ready for release?	Yes	Build screenshots are available in D000926821 ISP13 PED
	Are the risks properly mitigated? Is the risk analysis done and mitigated for all the hazardous defects noticed during validation? (Either clinically or by design)	NA	No risks were found during validation
	Product Defect Summary from validation presented (All the defects identified from validation need to be resolved/verified before the product release)	Yes	D000926820 ISP13 validation report defect section – NOTE no defects were opened against VAL3 validation forms D000926809 ISP13 Validation Site Management defect section
	Have the Validation plan and Validation report, Validation results are reviewed and approved appropriately?	Yes	• DHF380226 ISP13 Validation plan • D000926820 ISP13 validation report section training forms • D000926820 ISP13 validation report section evaluation forms
	Does the validation results requires a change in requirements?	No	no change in user requirements is needed
	Is the link between DHF & DMR appropriate?	NA	All activities will be completed towards the production ECO
	Is the traceability evidenced as appropriate for this stage?	Yes	• CT ASPECTS (VAL1) - D001157485 Traceability Matrix CT Aspects • CT Liver (VAL2) - D001044264_Traceability Matrix CT Liver

Printed copies are uncontrolled unless authenticated

Printed copies are uncontrolled unless authenticated

			• CT CCA (VAL2) - D000946874 ISP13 Traceability Matrix CT Cardiac • CMR Viewer (VAL2) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CMR Qflow (VAL2) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CMR Spatial (VAL2) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CMR Functional Long Axis (VAL2) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CMR Functional LV RV (VAL3) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CMR Mapping (VAL3) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CMR Temporal (VAL3) - D000996539, ISP13 MR Cardiac Suite Traceability Matrix • CT CAA (CT Auto. CS VAL3) - D000993523_Traceability Matrix CT Apps • Spectral AVA (VAL3) - D000993523_Traceability Matrix CT Apps
Design Transfer readiness review:	Are approved design outputs used for Design transfer?	NA	will be part as of ECO
	Are all the Design transfer activities completed as planned? E.g. Creation of manufacturing SOP's, completion of process validation etc.	NA	All activities will be completed towards the ECO
	Check the readiness of functions such as Manufacturing, Service, Customer support for product release	NA	The readiness will be checked at ECO
	Are there any other prerequisites to start Design Transfer activities?	NA	All activities will be completed towards the ECO

	Is the link between DHF & DMR appropriate? Provide evidences for the same.	NA	All activities will be completed towards the ECO

3.5 Validation exit criteria

3.5.1 VAL1: CT ASPECTS

Item	Yes / No	Evidence
The Validation test protocol(s), as identified by the Plan, have been completed with at least a preliminary Test Report submitted and agreed upon by the Project Team	Yes	D000926820 ISP13 validation report
Validation plan has been executed. Deviations are captured in the Validation report	Yes	D000926820 ISP13 validation report section Deviations from the Validation Plan
Successful completion of the required number of patient studies as listed in this plan along with associated evaluation forms	Yes	D000926820 ISP13 validation report
A minimum of one internal review was conducted.	Yes	D000926820 ISP13 validation report
All documentation, electronic help, and related materials listed in the Technical Publications Validation section of the validation plan have been validated	Yes	product labeling is being verified only build screenshots are available in D000926821 ISP13 PED
Plan is in place to upgrade external validation sites to production equivalent systems per the site agreement	NA	internal validation only
All defects identified against ISP13 to date have been reviewed and assigned an appropriate severity, priority, risk analysis, and mitigation by the project team has been completed for each, per operating procedure or work instruction for defect tracking	Yes	D000926820 ISP13 validation report defect section
Validation site summary letter was send to each external validation test site and a receipt of acknowledgment from the site has been received	NA	internal validation only
Validation design review results review is passed with all identified actions closed	Yes	D000992888 DRR of the ISP13 validation results – all actions identified are closed

3.5.2 VAL2: CT Liver, CT CCA, CMR Viewer, CMR Spatial, CMR Qflow , CMR Functional Long Axis

Item	Yes / No	Evidence
The Validation test protocol(s), as identified by the Plan, have been completed with at least a preliminary Test Report submitted and agreed upon by the Project Team	Yes	D000926820 ISP13 validation report
Validation plan has been executed. Deviations are captured in the Validation report	Yes	D000926820 ISP13 validation report section Deviations from the Validation Plan
Successful completion of the required number of patient studies as listed in this plan along with associated evaluation forms	Yes	D000926820 ISP13 validation report
A minimum of one internal review was conducted.	Yes	D000926820 ISP13 validation report

Printed copies are uncontrolled unless authenticated

All documentation, electronic help, and related materials listed in the Technical Publications Validation section of the validation plan have been validated	Yes	product labeling is being verified only build screenshots are available in D000926821 ISP13 PED
Plan is in place to upgrade external validation sites to production equivalent systems per the site agreement	NA	Due for RFD
All defects identified against ISP13 to date have been reviewed and assigned an appropriate severity, priority, risk analysis, and mitigation by the project team has been completed for each, per operating procedure or work instruction for defect tracking	Yes	D000926820 ISP13 validation report defect section D000926809 ISP13 Validation Site Management defect section
Validation site summary letter was send to each external validation test site and a receipt of acknowledgment from the site has been received	NA	Due for RFD
Validation design review results review is passed with all identified actions closed	Yes	D000992888 DRR of the ISP13 validation results – all actions identified are closed

3.5.1 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping

Item	Yes / No	Evidence
The Validation test protocol(s), as identified by the Plan, have been completed with at least a preliminary Test Report submitted and agreed upon by the Project Team	Yes	D000926820 ISP13 validation report
Validation plan has been executed. Deviations are captured in the Validation report	Yes	D000926820 ISP13 validation report section Deviations from the Validation Plan
Successful completion of the required number of patient studies as listed in this plan along with associated evaluation forms	Yes	D000926820 ISP13 validation report
A minimum of one internal review was conducted.	Yes	D000926820 ISP13 validation report
All documentation, electronic help, and related materials listed in the Technical Publications Validation section of the validation plan have been validated	Yes	product labeling is being verified only build screenshots are available in D000926821 ISP13 PED
Plan is in place to upgrade external validation sites to production equivalent systems per the site agreement	NA	Due for RFD
All defects identified against ISP13 to date have been reviewed and assigned an appropriate severity, priority, risk analysis, and mitigation by the project team has been completed for each, per operating procedure or work instruction for defect tracking	Yes	D000926820 ISP13 validation report defect section D000926809 ISP13 Validation Site Management defect section
Validation site summary letter was send to each external validation test site and a receipt of acknowledgment from the site has been received	NA	Due for RFD
Validation design review results review is passed with all identified actions closed	Yes	D000992888 DRR of the ISP13 validation results – all actions identified are closed

Printed copies are uncontrolled unless authenticated

4 Actions

4.1 VAL1: CT ASPECTS

Id	Action	Status	Expected Closure date	Reference
1.	NA – there are no actions needed from the review	NA	NA	NA

4.2 VAL2: CT Liver, CT CCS, CMR Viewer, CMR Spatial, CMR Qflow , CMR Functional Long Axis

Id	Action	Status	Expected Closure date	Reference
1.	NA – there are no actions needed from the review	NA	NA	NA

4.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping

Id	Action	Status	Expected Closure date	Reference
1.	NA – there are no actions needed from their review	NA	NA	NA

4.4 VAL1: Validation dates and evaluators

4.4.1 CT ASPECTS

08-Nov-2021 Dr. Reuven Shreiber
09-Nov-2021 Dr. David Jacobs
12-Nov-2021 Dr. Harsha

4.5 VAL2: Validation dates and evaluators

4.5.1 CT Liver

14-Apr-2022 Dr. Karthik
14-Apr-2022 Dr. Richa
23-May-2022 Dr. Amitai

4.5.2 CT CCA

14-Apr-2022 Dr. Karthik
14-Apr-2022 Dr. Richa
7-May-2022 Dr. Goitein

4.5.3 CMR Viewer

10-May-2022 Dr. Richa
10-May-2022 Dr. Karthik
18-May-2022 Dr. Castellaccio
20-May-2022 Dr. Bunk

Printed copies are uncontrolled unless authenticated

20-May-2022	Dr. Sokolowski
23-May-2022	Dr. Goitein
31-May-2022	Dr. Karthik
31-May-2022	Dr. Vimal
31-May-2022	Dr. Hamdan
7-Jun-2022	Prof. Douek
7-Jun-2022	Prof. Boussel
9-Jun-2022	Dr. Hashemi
9-Jun-2022	Dr. Kelle
9-Jun-2022	Dr. Doeblin

4.5.4 CMR Spatial

10-May-2022	Dr. Richa
10-May-2022	Dr. Karthik
18-May-2022	Dr. Castellaccio
20-May-2022	Dr. Bunk
20-May-2022	Dr. Sokolowski
20-May-2022	Dr. Penning
23-May-2022	Dr. Goitein
31-May-2022	Dr. Karthik
31-May-2022	Dr. Vimal
31-May-2022	Dr. Hamdan
7-Jun-2022	Prof. Douek
7-Jun-2022	Prof. Boussel
9-Jun-2022	Dr. Hashemi
9-Jun-2022	Dr. Kelle
9-Jun-2022	Dr. Doeblin
20-Jun-2022	Dr. Aviram
27-Jun-2022	Dr. Salem

4.5.5 CMR Qflow

10-May-2022	Dr. Richa
10-May-2022	Dr. Karthik
18-May-2022	Dr. Castellaccio
20-May-2022	Dr. Bunk
20-May-2022	Dr. Sokolowski
20-May-2022	Dr. Penning
23-May-2022	Dr. Goitein
31-May-2022	Dr. Karthik
31-May-2022	Dr. Vimal
31-May-2022	Dr. Hamdan
7-Jun-2022	Prof. Douek
7-Jun-2022	Prof. Boussel
9-Jun-2022	Dr. Hashemi
9-Jun-2022	Dr. Kelle
9-Jun-2022	Dr. Doeblin
20-Jun-2022	Dr. Aviram
27-Jun-2022	Dr. Salem

4.5.6 CMR Functional Long Axis

10-May-2022	Dr. Richa
10-May-2022	Dr. Karthik
18-May-2022	Dr. Castellaccio

Printed copies are uncontrolled unless authenticated

20-May-2022	Dr. Bunk
20-May-2022	Dr. Sokolowski
20-May-2022	Dr. Penning
23-May-2022	Dr. Goitein
31-May-2022	Dr. Karthik
31-May-2022	Dr. Vimal
31-May-2022	Dr. Hamdan
7-Jun-2022	Prof. Douek
7-Jun-2022	Prof. Boussel
9-Jun-2022	Dr. Hashemi
9-Jun-2022	Dr. Kelle
9-Jun-2022	Dr. Doeblin

4.6 VAL3: Validation dates and evaluators

4.6.1 CT Auto Calcium Scoring (CAA)

16-SEP-2022	Dr. Karthik
16-SEP2022	Dr. Richa
20-SEP-2022	Pedro Abaso (3D Tech)

4.6.2 CT sAVA

19-SEP-2022	Dr. Galit Aviram
20-SEP-2022	Pedro Abaso (3D Tech)
26-SEP-2022	Dr. Antonio Castellaccio

4.6.3 MR Longitudinal Brain Analysis (MR CVS)

02-OCT-2022	Dr. Karthik
22-OCT-2022	Dr. Harsha krishna
03-NOV-2022	Dr. Reuven Shrieber

4.6.4 CMR Functional LV RV

07-SEP-2022	Dr. Hamdan
07-SEP-2022	Dr. Wolak
09-SEP-2022	Dr. Karthik
09-SEP-2022	Dr. Richa
12-SEP-2022	Dr. Leb
12-SEP-2022	Dr. Vucic
13-SEP-2022	Dr. Hamdan
14-SEP-2022	Prof. Douek
14-SEP-2022	Prof. Boussel
19-SEP-2022	Dr. Aviram
20-SEP-2022	Dr. Molinelli
20-SEP-2022	Dr. Vano
20-SEP-2022	Dr. Castellaccio
24-OCT-2022	Dr. Goitein

4.6.5 CMR Mapping

07-SEP-2022	Dr. Hamdan
07-SEP-2022	Dr. Wolak
09-SEP-2022	Dr. Karthik

Printed copies are uncontrolled unless authenticated

09-SEP-2022	Dr. Richa
12-SEP-2022	Dr. Leb
12-SEP-2022	Dr. Vucic
13-SEP-2022	Dr. Hamdan
14-SEP-2022	Prof. Douek
14-SEP-2022	Prof. Boussel
19-SEP-2022	Dr. Aviram
20-SEP-2022	Dr. Molinelli
20-SEP-2022	Dr. Vano
20-SEP-2022	Dr. Castellaccio
26-SEP-2022	Dr. Doeblin
24-OCT-2022	Dr. Goitein

4.6.6 CMR Temporal

07-SEP-2022	Dr. Hamdan
09-SEP-2022	Dr. Karthik
12-SEP-2022	Dr. Leb
12-SEP-2022	Dr. Vucic
13-SEP-2022	Dr. Hamdan
14-SEP-2022	Prof. Douek
17-SEP-2022	Dr. Aneja
17-SEP-2022	Dr. Al Kindi
19-SEP-2022	Dr. Aviram
20-SEP-2022	Dr. Castellaccio
26-SEP-2022	Dr. Doeblin
12-OCT-2022	Dr. Richa
12-OCR-2022	Dr. Karthik
24-OCT-2022	Dr. Goitein
28-OCT-2022	Dr. Vimal

Printed copies are uncontrolled unless authenticated

5 Conclusion

5.1 VAL1: CT ASPECTS

It has been concluded that the ISP13 **CT ASPECTS Application** was validated as described in DHF380226 ISP13 Validation Plan .Validation Results Design review is approved by all stake holders and ready for release (CT ASPECTS).

The design validation results provide the objective evidence that the product conforms to defined user needs and intended use.

Review Result: **Approved**

5.2 VAL2: CT Liver, CT CCA, CMR viewer, CMR Spatial, CMR Qflow, CMR Functional Long Axis

It has been concluded that the ISP13 **CT Liver, CT CCA, CMR viewer, CMR Spatial, CMR Qflow, CMR Functional Long Axis applications** were validated as described in DHF380226 ISP13 Validation Plan .Validation Results Design review is approved by all stake holders and ready for release.

The design validation results provide the objective evidence that the product conforms to defined user needs and intended use.

Review Result: **Approved**

5.3 VAL3: CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping

It has been concluded that the ISP13 **CT Calcium Automated Analysis (Auto CS), Spectral AVA, MR Longitudinal Brain Analysis (CVS), CMR Functional LV RV, CMR Temporal and CMR Mapping applications** were validated as described in DHF380226 ISP13 Validation Plan .Validation Results Design review is approved by all stake holders and ready for release.

The design validation results provide the objective evidence that the product conforms to defined user needs and intended use.

Review Result: **Approved**

Printed copies are uncontrolled unless authenticated

6 Record history

Rev	Release Date	Author	Description of changes
A	In WindChill	Hagit Davidesco	Review the validation results - VAL1
B	In WindChill	Hagit Davidesco	Review the validation results - VAL2
C	In WindChill	Hagit Davidesco	Review the validation results - VAL3

7 Record Review and Approval

Name Author	Function Author
Hagit Davidesco	Validation Lead

Name Reviewer(s)	Function Reviewer(s)	Review Date(s) *DD-MMM-YYYY*	Signature(s)
Keren Afek	Regulatory Affairs	In Windchill	In Windchill

Name Approver(s)	Function Approver (s)	Review Date(s) DD-MMM-YYYY	Signature(s)
Chockalingam Kavirayar	System architect	In Windchill	In Windchill
Anna Aryev	Quality representative	In Windchill	In Windchill
Jai Singh	Independent reviewer	In Windchill	In Windchill
Eyal Biran	Program manager	In Windchill	In Windchill

8 Appendix A: DRR presentation

ISP13 validation Results review VAL1.%20Results%20revie%20

ISP13%20validation Results%20revie

ISP13%20validation Results%20revie

End of Document

Printed copies are uncontrolled unless authenticated