PHILIPS

Document Information
Document Identification: D001045958
Document Name: ISP13 Design Review Record
Version: D.2
Life cycle state: Released 29-Jun-2023

Cover Sheet

Document Classification
Design Category: 3. Design Output
DMR Group Number: -

Signature Information		
Role	**Assignee**	**Date**
Approver (User)	Kavirayar, Chockalingam	27-Jun-2023
Approver (User)	Halliyal, Shilpa	27-Jun-2023
Approver (Accountable)	Kumar, Aashish	27-Jun-2023
Approver (User)	J, Vivek Kumar	29-Jun-2023
Approver (User)	Kamaluddin, Nisha	28-Jun-2023

Security Information		
IP Sensitivity	**Strategic Good Indicator**	**ECCN**
Internal	3 Controlled, listed in the US EAR US-origin/subject to the US EAR	EAR99

Note: The stamp below is identifying the following elements
ID number; Version; State; Document Name; DMR Group Number

D001045958;D.2;Released;ISP13 Design Review Record;-

Verification Design Review Record

For AVW15, CT ASPECTS & NM Suite RC Release

Document ID: D001045958
Document Revision: D

1 Document Introduction 2
1.1 Purpose 2
1.2 Scope 2
1.2.1 ISP13 VAL-2 Release: 2
1.2.2 ISP13 VAL-3 Release: 2
1.2.3 AVW15, CT ASPECTS & NM Suite RFD Release: 2
1.2.4 AVW15, CT ASPECTS & NM Suite RC Release: 3
1.3 References 3
1.4 Abbreviations and Definitions 4
2 Design Review Material 5
3 Design Review Meeting 5
3.1.1 VAL-2 Release 5
3.1.2 VAL-3 Release 6
3.1.3 AVW15, CT ASPECTS & NM Suite RFD Release 7
3.1.4 AVW15, CT ASPECTS & NM Suite RC Release 7
4 Actions 8
4.1.1 VAL-2 Release 8
4.1.2 VAL-3 Release 8
4.1.3 AVW15, CT ASPECTS & NM Suite RFD Release 9
4.1.4 AVW15, CT ASPECTS & NM Suite RC Release 9
5 Conclusion 9
5.1.1 VAL-2 Release 9
5.1.2 VAL-3 Release 10
5.1.3 AVW15, CT ASPECTS & NM Suite RFD Release 11
5.1.4 AVW15, CT ASPECTS & NM Suite RC Release 12
6 Record history 13
7 Record Review and Approval 13

Printed copies are uncontrolled unless authenticated

1 Document Introduction

1.1 Purpose

This document reports the results of the systematic review to evaluate the ability of the output of design activities to meet the requirements and to document the actions required to address the anomalies found.

ISP13 VAL-2 Release: The Purpose of this meeting is to review the correctness, completeness and traceability of the verification results relative to the product requirements, so that the design output meets the design input. Also to ensure that the product requirements and product design are consistent and complete.

ISP13 VAL-3 Release: The Purpose of this meeting is to review the correctness, completeness and traceability of the verification results relative to the product requirements, so that the design output meets the design input. Also to ensure that the product requirements and product design are consistent and complete.

AVW15, CT ASPECTS & NM Suite RFD Release: The Purpose of this meeting is to review the correctness, completeness and traceability of the verification results relative to the product requirements, so that the design output meets the design input. Also to ensure that the product requirements and product design are consistent and complete.

AVW15, CT ASPECTS & NM Suite RC Release: The Purpose of this meeting is to review the correctness, completeness and traceability of the verification results relative to the product requirements, so that the design output meets the design input. Also to ensure that the product requirements and product design are consistent and complete.

1.2 Scope

1.2.1 ISP13 VAL-2 Release:

Scope of this Design Review is to Review the Verification Results for 'VAL-2 Release' Scope.
Please refer the VRR slides for the high level scope.
For detailed information refer the ISP13 Design Verification Plan D000995566.

TYPE of design review as per ICAP-P-030005: Design Review Procedure.
Verification Results Review for ISP13 VAL-2 Release.

1.2.2 ISP13 VAL-3 Release:

Scope of this Design Review is to Review the Verification Results for 'VAL-3 Release' Scope.
Please refer the VRR slides for the high level scope.
For detailed information refer the ISP13 Design Verification Plan D000995566.

TYPE of design review as per ICAP-P-030005: Design Review Procedure.
Verification Results Review for ISP13 VAL-3 Release.

1.2.3 AVW15, CT ASPECTS & NM Suite RFD Release:

Scope of this Design Review is to Review the Verification Results for 'RFD Release' Scope.
Please refer the VRR slides for the high level scope.
For detailed information refer "D000995566" the AVW15, CT ASPECTS & NM Suite RFD_Design Verification Plan_Rev F.

TYPE of design review as per ICAP-P-030005: Design Review Procedure.
Verification Results Review for AVW15, CT ASPECTS & NM Suite RFD Release.

Printed copies are uncontrolled unless authenticated

1.2.4 AVW15, CT ASPECTS & NM Suite RC Release:

Scope of this Design Review is to Review the Verification Results for 'RC Release' Scope.
Please refer the VRR slides for the high level scope.
For detailed information refer "D000995566" the AVW15, CT ASPECTS & NM Suite RC_Design Verification Plan.

TYPE of design review as per ICAP-P-030005: Design Review Procedure.
Verification Results Review for AVW15, CT ASPECTS & NM Suite RC Release.

1.3 References

Reference Number	Document Title	Document iD
VAL-2 Release		
REF – 1	ISP13 Design Verification Plan	D000995566 (Rev A)
REF – 2	ISP13 Design Verification Report	D000996308 (Rev A)
REF – 3	ISP13 Product Defect Status Report	D001025480 (Rev A)
REF – 4	ISP13 Expert Test Report	D001032014 (Rev A)
REF – 5	ISP13 System Malware Scan Verification Report	D001025479 (Rev A)
REF – 6	ISP13 RISK MANAGEMENT MATRIX CARDIAC SUITE	D000879224 (Rev A)
REF – 7	ISP13 MR Cardiac Suite Traceability Matrix	D000996539 (Rev B)
REF – 8	ISP13 RMM General CTApps	D000892550 (Rev C)
REF – 9	Traceability Matrix CTapps	D000993523 (Rev B)
REF – 10	ISP13 RMM CT Cardiac	D000946860 (Rev A)
REF – 11	ISP13 Traceability Matrix CT Cardiac	D000946874 (Rev B)
REF – 12	ISP13 Site Management	D000926809 (Rev A)
VAL-3 Release		
REF – 1	ISP13 Design Verification Plan	D000995566 (Rev C)
REF – 2	ISP13 Design Verification Report	D000996308 (Rev B)
REF – 3	ISP13 Product Defect Status Report	D001025480 (Rev B)
REF – 4	ISP13 Expert Test Report	D001032014 (Rev B)
REF – 5	ISP13 System Malware Scan Verification Report	D001025479 (Rev B)
REF – 6	ISP13 RISK MANAGEMENT MATRIX CARDIAC SUITE	D000879224 (Rev C)
REF – 7	ISP13 MR Cardiac Suite Traceability Matrix	D000996539 (Rev D)
REF – 8	Risk Management Matrix for ISP13,MR Applications and MMV	D001104871 (Rev B)
REF – 9	ISP13_RMM_MRAPPS & MMV	D001065108 (Rev A)
REF – 10	ISP13 Traceability Matrix LoBI	D001105932 (Rev B)
REF – 11	ISP13 RMM General CTApps	D000892550 (Rev D)
REF – 12	Traceability Matrix CTapps	D000993523 (Rev D)
REF – 13	ISP13 RMM CT Cardiac	D000946860 (Rev A)
REF – 14	ISP13 Traceability Matrix CT Cardiac	D000946874 (Rev C)
REF – 15	ISP13 RMM SpectralCT Apps	D000996304 (Rev B)
REF – 16	ISP13 Traceability Matrix for CTSpectral Applications	D001097887 (Rev B)
REF – 17	ISP13 System RMM	D000996286 (Rev A)
REF – 18	ISP13 Site Management	D000926809 (Rev A)
AVW15, CT ASPECTS & NM Suite RFD Release		
REF – 1	ISP13 Design Verification Plan	D000995566 (Rev F)
REF – 2	AVW15 Design Verification Report	D000996308 (Rev D)

Printed copies are uncontrolled unless authenticated

REF – 3	Design Verification Report_Aspect	D000889689 (Rev B)
REF – 4	NM Suite_Design Verification Report	D001246601 (Rev A)
REF – 5	ISP13 Product Defect Status Report	D001025480 (Rev C)
REF – 6	ISP13 Expert Test Report	D001032014 (Rev C)
REF – 7	ISP13 System Malware Scan Verification Report	D001025479 (Rev C)
REF – 8	ISP13 RISK MANAGEMENT MATRIX CARDIAC SUITE	D000879224 (Rev E)
REF – 9	ISP13 MR Cardiac Suite Traceability Matrix	D000996539 (Rev F)
REF – 10	Risk Management Matrix for ISP13,MR Applications and MMV	D001104871 (Rev E)
REF – 11	ISP13_RMM_MRAPPS & MMV	D001065108 (Rev A)
REF – 12	ISP13 RMM General CTApps	D000892550 (Rev F)
REF – 13	Traceability Matrix CTapps	D000993523 (Rev F)
REF – 14	ISP13 RMM CT Cardiac	D000946860 (Rev B)
REF – 15	ISP13 Traceability Matrix CT Cardiac	D000946874 (Rev G)
REF – 16	ISP13 RMM SpectralCT Apps	D000996304 (Rev D)
REF – 17	ISP13 Traceability Matrix for CTSpectral Applications	D001097887 (Rev D)
REF – 18	ISP13 System RMM	D000996286 (Rev C)
REF – 19	AVW15_Traceability_SystemFeatures-BLR & Serviceability	D001164937 (Rev B)
AVW15, CT ASPECTS & NM Suite RC Release		
REF – 1	ISP13 Design Verification Plan	D000995566 (Rev H)
REF – 2	AVW15 Design Verification Report	D000996308 (Rev F)
REF – 3	Design Verification Report_Aspect	D001157473 (Rev B)
REF – 4	NM Suite_Design Verification Report	D001246601 (Rev B)
REF – 5	ISP13 Product Defect Status Report	D001025480 (Rev D)
REF – 6	ISP13 Expert Test Report	D001032014 (Rev D)
REF – 7	ISP13 System Malware Scan Verification Report	D001025479 (Rev D)
REF – 8	ISP13 RISK MANAGEMENT MATRIX CARDIAC SUITE	D000879224 (Rev F)
REF – 9	ISP13 MR Cardiac Suite Traceability Matrix	D000996539 (Rev I)
REF – 10	Risk Management Matrix for ISP13,MR Applications and MMV	D001104871 (Rev E)
REF – 11	ISP13_RMM_MRAPPS & MMV	D001065108 (Rev A)
REF – 12	ISP13 RMM General CTApps	D000892550 (Rev G)
REF – 13	Traceability Matrix CTapps	D000993523 (Rev I)
REF – 14	ISP13 RMM CT Cardiac	D000946860 (Rev B)
REF – 15	ISP13 Traceability Matrix CT Cardiac	D000946874 (Rev I)
REF – 16	ISP13 RMM SpectralCT Apps	D000996304 (Rev E)
REF – 17	ISP13 Traceability Matrix for CTSpectral Applications	D001097887 (Rev F)
REF – 18	ISP13 System RMM	D000996286 (Rev E)
REF – 19	AVW15_Traceability_SystemFeatures-BLR & Serviceability	D001164937 (Rev D)
REF- 20	Traceability Matrix System HFA	D001159554(Rev D)
REF- 21	PVT_AVW15_ Installation and Upgrades _Traceability Matrix	D001157062(Rev D)

1.4 Abbreviations and Definitions

Terminology & Abbreviations	Description/Definition
VRR	Verification Results Review
SVer	Start Verification
RFD	Release For Delivery

Printed copies are uncontrolled unless authenticated

Terminology & Abbreviations	Description/Definition
AVW	Advanced Visualization Workspace

2 Design Review Material

Document Reference	Title	Version
D001045958_ISP13 Design Review Reco	Design Review-Verification for VAL-2 Release Milestone Presentation	A
D001045958_ISP13 Design Review Reco	Design Review-Verification for VAL-3 Release Milestone Presentation	B
D001045958_AVW1 5,CT ASPECTS & NM	Design Review-Verification for AVW15, CT ASPECTS & NM Suite RFD Release Milestone Presentation	C
D001045958_AVW1 5,CT ASPECTS & NM	Design Review-Verification for AVW15, CT ASPECTS & NM Suite RC Release Milestone Presentation	D

3 Design Review Meeting

3.1.1 VAL-2 Release

Date: 12-Apr-2022

Name	Role	Represented Function	Independent
Aashish Kumar	Global Verification Lead	Verification (Review Owner)	N
Srinivas Arakere	Global Verification Manager	R&D	N
Chockalingam Kavirayar	System Architect	R&D (Mandatory)	N
Byamma Munichandrappa	System Engineer	R&D	N
J, Vivek Kumar	Quality Representative	Q&R	N
Anna Aryev	Quality Representative	Q&R	N
Suhas Bharadwaj	Quality Representative	Q&R (Offline)	N
Shipra Shivangi	Quality Representative	Q&R (Mandatory)	N
Keren Afek	Regulatory Representative	Q&R (Mandatory)	N
Nisha Kamaluddin	Project Leader II, CPE	Independent Reviewer (Mandatory)	Y
Ruth Katz	Clinical representative	Clinical	N
K, Jayaprakash	Clinical representative	Clinical	N
Narayan Srikrishna Suru	Clinical representative	Clinical	N
Sajimon Emmanuel	Clinical representative	Clinical (Offline)	N
Sujesh P Paravatty	Clinical representative	Clinical (Mandatory)	N
Offer Barzilai	Product Manager	Marketing (Offline)	N

Printed copies are uncontrolled unless authenticated

K Sugesh	Senior Product Support Specialist – II	Customer Service	N
Hagit Davidesco	Validation Lead & Beta Sites Coordinator	Clinical	N
Adi Ron	Design Transfer	OSS	N
Raghavendra Prasad Halaharvi	RTE	R&D	N
Shilpa Halliyal	RTE	R&D	N
Roman Grekin	Product Security Officer	R&D (Offline)	N
Dror Orell	Program Manager	PMO	N
Mustaq Mohammed	Project Manager	R&D	N
Abhishek GN	Development Manager	R&D	N
Ophir Harran	Development Manager	R&D	N

3.1.2 VAL-3 Release

Date: 24-Aug-2022

Name	Role	Represented Function	Independent
Aashish Kumar	Global Verification Lead	Verification (Review Owner)	N
Manjunath Hunsur Adishesha	Verification Manager	R&D	N
Chockalingam Kavirayar	System Architect	R&D (Mandatory)	N
Byamma Munichandrappa	System Engineer	R&D	N
J, Vivek Kumar	Quality Representative	Q&R	N
Anna Aryev	Quality Representative	Q&R	N
Suhas Bharadwaj	Quality Representative	Q&R	N
Shipra Shivangi	Quality Representative	Q&R (Mandatory)	N
Keren Afek	Regulatory Representative	Q&R (Mandatory)	N
Nisha Kamaluddin	Project Leader II, CPE	Independent Reviewer (Mandatory)	Y
Ruth Katz	Clinical representative	Clinical (Offline)	N
Rishov Majumdar	Clinical representative	Clinical	N
Narayan Srikrishna Suru	Clinical representative	Clinical	N
Sajimon Emmanuel	Clinical representative	Clinical	N
Sujesh P Paravatty	Clinical representative	Clinical (Mandatory)	N
Offer Barzilai	Product Manager	Marketing (Offline)	N
K Sugesh	Senior Product Support Specialist – II	Customer Service	N
Hagit Davidesco	Validation Lead & Beta Sites Coordinator	Clinical	N
Yuval Kopel	Design Transfer	OSS (Offline)	N
Raghavendra Prasad Halaharvi	RTE	R&D	N
Shilpa Halliyal	RTE	R&D	N
Roman Grekin	Product Security Officer	R&D (Offline)	N
Dror Orell	Program Manager	PMO	N
Mustaq Mohammed	Project Manager	R&D	N
Abhishek GN	Development Manager	R&D (Offline)	N

Printed copies are uncontrolled unless authenticated

Sravan Kumar Prathap Sekar Reddy	Development Manager	R&D (Offline)	N
Ophir Harran	Development Manager	R&D	N

3.1.3 AVW15, CT ASPECTS & NM Suite RFD Release

Date: 08-Dec-2022

Name	Role	Represented Function	Independent
Santhosh H	Global Verification Lead	Verification (Review Owner)	N
Aashish Kumar	Verification Manager	R&D	N
Chockalingam Kavirayar	System Architect	R&D (Mandatory)	N
Byamma Munichandrappa	System Engineer	R&D	N
J, Vivek Kumar	Quality Representative	Q&R	N
Anna Aryev	Quality Representative	Q&R	N
Suhas Bharadwaj	Quality Representative	Q&R	N
Shipra Shivangi	Quality Representative	Q&R (Mandatory)	N
Keren Afek	Regulatory Representative	Q&R (Mandatory)	N
Nisha Kamaluddin	Project Leader II, CPE	Independent Reviewer (Mandatory)	Y
Ruth Katz	Clinical representative	Clinical (Offline)	N
Rishov Majumdar	Clinical representative	Clinical	N
Narayan Srikrishna Suru	Clinical representative	Clinical	N
Sajimon Emmanuel	Clinical representative	Clinical	N
Sujesh P Paravatty	Clinical representative	Clinical (Mandatory)	N
Efrat Kliper Goldin	Product Manager	Marketing	N
K Sugesh	Senior Product Support Specialist – II	Customer Service	N
Hagit Davidesco	Validation Lead & Beta Sites Coordinator	Clinical	N
Adi Ron	Design Transfer	OSS	N
Raghavendra Prasad Halaharvi	RTE	R&D (Offline)	N
Shilpa Halliyal	RTE	R&D	N
Roman Grekin	Product Security Officer	R&D (Offline)	N
Eyal Biran	Program Manager	PMO	N
Mustaq Mohammed	Project Manager	R&D	N
Abhishek GN	Development Manager	R&D	N
Sravan Kumar Prathap Sekar Reddy	Development Manager	R&D	N
Chaitra VS	Development Manager	R&D	N
Ophir Harran	Development Manager	R&D	N

3.1.4 AVW15, CT ASPECTS & NM Suite RC Release

Date: 26-June-2023

Name	Role	Represented Function	Independent
Santhosh H	Global Verification Lead	Verification (Review Owner)	N

Printed copies are uncontrolled unless authenticated

Aashish Kumar	Verification Manager	R&D	N
Chockalingam Kavirayar	System Architect	R&D (Mandatory)	N
Natalia Kogan	System Engineer	R&D	N
J, Vivek Kumar	Quality Representative	Q&R	N
Anna Aryev	Quality Representative	Q&R	N
Suhas Bharadwaj	Quality Representative	Q&R	N
Shipra Shivangi	Quality Representative	Q&R (Mandatory)	N
Keren Afek	Regulatory Representative	Q&R (Mandatory)	N
Nisha Kamaluddin	Project Leader II, CPE	Independent Reviewer (Mandatory)	Y
Ruth Katz	Clinical representative	Clinical	N
Rishov Majumdar	Clinical representative	Clinical	N
Narayan Srikrishna Suru	Clinical representative	Clinical	N
Sujesh P Paravatty	Clinical representative	Clinical (Mandatory)	N
Efrat Kliper Goldin	Product Manager	Marketing	N
K Sugesh	Senior Product Support Specialist – II	Customer Service	N
Hagit Davidesco	Validation Lead & Beta Sites Coordinator	Clinical	N
Adi Ron	Design Transfer	OSS	N
Shilpa Halliyal	RTE	R&D	N
Roman Grekin	Product Security Officer	R&D	N
Eyal Biran	Program Manager	PMO	N
Mustaq Mohammed	Project Manager	R&D	N
Abhishek GN	Development Manager	R&D	N
Sravan Kumar Prathap Sekar Reddy	Development Manager	R&D	N
Ophir Harran	Development Manager	R&D	N

4 Actions

4.1.1 VAL-2 Release

Id	Action	Status	Expected Closure date	Reference
1	None			

4.1.2 VAL-3 Release

Id	Action	Status	Expected Closure date	Reference
1	All the relevant documents to be released.	Closed	25-Aug-2022	All the below documents are released in Windchill: 1. Verification Report – D000996308. 2. Traceability Matrix – D000996539, D001105932, D000993523, D000946874 & D001097887.

Printed copies are uncontrolled unless authenticated

4.1.3 AVW15, CT ASPECTS & NM Suite RFD Release

Id	Action	Status	Expected Closure date	Reference
1	None			

4.1.4 AVW15, CT ASPECTS & NM Suite RC Release

Id	Action	Status	Expected Closure date	Reference
1	None			

5 Conclusion

5.1.1 VAL-2 Release

Verification Results Review Summary	Yes/No/NA	Evidence/ Rationale
Have the resulting changes from product defect fixes resulted in consistent requirements, product design?	No Design Changes	Please refer the Slides embedded for the VRR
Do the verification results cover the product requirements?	Yes, The Verification report	D000996308 – Section 3
Is product labeling verified and ready for release?	Yes, The Verification report	D000996308 – Section 9.3
Is the Product Defect Status/Summary presented?	Yes, as a part of the review its presented	Please refer the Slides embedded for the VRR
Does the verification results requires a change in any of the requirements?	No changes to requirements	Please refer the Slides embedded for the VRR
Does the verification plan, verification report, verification results documents reviewed and released in DHF?	Yes	D000995566, D000996308
Are the safety and security aspects from Risk management tested and results are documented with appropriate evidences?	Safety – Yes Security – N/A, no requirement.	D000879224, D000892550, D000946860
Is all the Validation sites identified for the limited site deployment & testing?	Yes	D000926809
Are all the product defects identified during verification are resolved and verified?	Yes	Please refer the Slides embedded for the VRR
Is all the instructions in IFU are verified and traceable to test IDs?	N/A	Not Applicable as Software for validation will be Installed on Demo laptop or installed on separate test server not for clinical use. Validation will be performed only in the presence of a trained clinical team after they train the evaluators.
Is the risk analysis done for all the hazardous defects noticed?	Yes	1374260, 1379135
Is the traceability evidenced as appropriate for this stage?	Yes	D000996539, D000993523, D000946874

Printed copies are uncontrolled unless authenticated

Overall conclusion of VRR for ISP13 VAL-2 Release
The verification results fulfil, after systematic examination, the product requirements and suit validation requirements for the planned content of 'VAL-2 Release' scope.

Design Verification has passed the exit criteria ensuring the product is safe and effective for use.
All functions agreed the VRR as passed and agreed to proceed for Validation.

Overall conclusion is **APPROVED**

5.1.2 VAL-3 Release

Verification Results Review Summary	Yes/No/NA	Evidence/ Rationale
Have the resulting changes from product defect fixes resulted in consistent requirements, product design?	No Design Changes	Please refer the Slides embedded for the VRR
Do the verification results cover the product requirements?	Yes, The Verification report	D000996308 – Section 3
Is product labeling verified and ready for release?	Yes, The Verification report	D000996308 – Section 9.3
Is the Product Defect Status/Summary presented?	Yes, as a part of the review its presented	Please refer the Slides embedded for the VRR
Does the verification results requires a change in any of the requirements?	No changes to requirements	Please refer the Slides embedded for the VRR
Does the verification plan, verification report, verification results documents reviewed and released in DHF?	No	D000995566, D000996308
Are the safety and security aspects from Risk management tested and results are documented with appropriate evidences?	Safety – Yes Security – N/A, no requirement.	D000879224, D001104871, D001065108, D000892550, D000946860, D000996304, D000996286
Is all the Validation sites identified for the limited site deployment & testing?	Yes	D000926809
Are all the product defects identified during verification are resolved and verified?	Yes	Please refer the Slides embedded for the VRR
Is all the instructions in IFU are verified and traceable to test IDs?	N/A	Not Applicable as Software for validation will be Installed on Demo laptop or installed on separate test server not for clinical use. Validation will be performed only in the presence of a trained clinical team after they train the evaluators.
Is the risk analysis done for all the hazardous defects noticed?	Yes	1431581, 1431899
Is the traceability evidenced as appropriate for this stage?	No	D000996539, D001105932, D000993523, D000946874, D001097887

Overall conclusion of VRR for ISP13 VAL-3 Release
The verification results fulfil, after systematic examination, the product requirements and suit validation requirements for the planned content of 'VAL-3 Release' scope.

Printed copies are uncontrolled unless authenticated

Design Verification has passed the exit criteria ensuring the product is safe and effective for use.
All functions agreed the VRR as passed and agreed to proceed for Validation.

Overall conclusion is **APPROVED**

5.1.3 AVW15, CT ASPECTS & NM Suite RFD Release

Verification Results Review Summary	Yes/No/NA	Evidence/ Rationale
Have the resulting changes from product defect fixes resulted in consistent requirements, product design?	No Design Changes	Please refer the Slide #14 embedded for the VRR
Do the verification results cover the product requirements?	Yes, The Verification report	D000996308 – Section 3
Is product labeling verified and ready for release?	Yes, The Verification report	D000996308 – Section 9.3
Is the Product Defect Status/Summary presented?	Yes, as a part of the review its presented	Please refer the Slide #13 embedded for the VRR
Does the verification results requires a change in any of the requirements?	No changes to requirements	Please refer the Slide #14 embedded for the VRR
Does the verification plan, verification report, verification results documents reviewed and released in DHF?	Yes	D000995566, D000996308
Are the safety and security aspects from Risk management tested and results are documented with appropriate evidences?	Yes	D000879224, D001104871, D001065108, D000892550, D000946860, D000996304, D000996286 Safety Test Results: D000996308 – Sections 3.1, 3.2, 3.3, 3.5, 3.6, 3.7, 3.8, 3.9 & 3.11. Security Test Results: D000996308 – Section 3.10 & Section 9.10
Is all the Validation sites identified for the limited site deployment & testing?	N/A	Not Applicable for this verification cycle (RFD Release) as the validation is already performed.
Are all the product defects identified during verification are resolved and verified?	Yes	D000996308 – Section 4.
Is all the instructions in IFU are verified and traceable to test IDs?	Yes	D000996308, D000889689 D001246601
Is the risk analysis done for all the hazardous defects noticed?	Yes	D001025480 – Section 3.
Is the traceability evidenced as appropriate for this stage?	Yes	D000996539, D000993523, D000946874, D001097887, D001164937

<u>Overall conclusion of VRR for AVW15, CT ASPECTS & NM Suite RFD Release</u>
The verification results fullfill, after systematic examination, the product requirements and ready for release.
Design Verification has passed the exit criteria ensuring the product is safe and effective for use.
All functions agreed the VRR as passed and agreed to proceed for release.

Printed copies are uncontrolled unless authenticated

Overall conclusion is **APPROVED**

5.1.4 AVW15, CT ASPECTS & NM Suite RC Release

Verification Results Review Summary	Yes/No/NA	Evidence/ Rationale
Have the resulting changes from product defect fixes resulted in consistent requirements, product design?	No Design Changes	Please refer the Slide #14 embedded for the VRR
Do the verification results cover the product requirements?	Yes, The Verification report	D000996308 – Section 3
Is product labeling verified and ready for release?	Yes, The Verification report	D000996308 – Section 9.3
Is the Product Defect Status/Summary presented?	Yes, as a part of the review its presented	Please refer the Slide #13 embedded for the VRR
Does the verification results requires a change in any of the requirements?	No changes to requirements	Please refer the Slide #14 embedded for the VRR
Does the verification plan, verification report, verification results documents reviewed and released in DHF?	Yes	D000995566, D000996308
Are the safety and security aspects from Risk management tested and results are documented with appropriate evidences?	Yes	D000879224, D001104871, D001065108, D000892550, D000946860, D000996304, D000996286 Safety Test Results: D000996308 – Sections 3.1, 3.2, 3.3, 3.5, 3.6, 3.7, 3.8, 3.9 & 3.11. Security Test Results: D000996308 – Section 3.10 & Section 9.10
Is all the Validation sites identified for the limited site deployment & testing?	N/A	Not Applicable for this verification cycle (RC Release) as the validation is already performed.
Are all the product defects identified during verification are resolved and verified?	Yes	D000996308 – Section 4.
Is all the instructions in IFU are verified and traceable to test IDs?	Yes	D000996308, D000889689 D001246601
Is the risk analysis done for all the hazardous defects noticed?	Yes	D001025480 – Section 3.
Is the traceability evidenced as appropriate for this stage?	Yes	D000996539, D000993523, D000946874, D001097887, D001164937

Overall conclusion of VRR for AVW15, CT ASPECTS & NM Suite RC Release

The verification results fullfill, after systematic examination, the product requirements and ready for release.
Design Verification has passed the exit criteria ensuring the product is safe and effective for use.
All functions agreed the VRR as passed and agreed to proceed for release.

Overall conclusion is **APPROVED**

Printed copies are uncontrolled unless authenticated

6 Record history

Rev	Release Date	Author	Description of changes
A	In Windchill	Aashish Kumar	ISP13 Verification Result Review Record for VAL-2 Release. • Updated all sections of this document with the changes required for ISP13 VAL-2 Release milestone.
B	In Windchill	Aashish Kumar	ISP13 Verification Result Review Record for VAL-3 Release. • Updated all sections of this document with the changes required for ISP13 VAL-3 Release milestone.
C	In Windchill	Santhosh H	AVW15, CT ASPECTS & NM Suite Verification Result Review Record for RFD Release. • Updated all sections of this document with the changes required for AVW15, CT ASPECTS & NM Suite RFD Release milestone.
D	In Windchill	Santhosh H	AVW15, CT ASPECTS & NM Suite Verification Result Review Record for RC Release. Updated all sections of this document with the changes required for AVW15, CT ASPECTS & NM Suite RC Release milestone.

7 Record Review and Approval

Name Author	Function Author
Santhosh H	ICAP Global Verification Lead

Name Reviewer(s)	Function Reviewer(s)	Review Date(s) *DD-MMM-YYYY*	Signature(s)
Aashish Kumar	Verification Manager	In Windchill	In Windchill

Name Approver	Function Approver	Date *DD-MMM-YYYY*	Signature
Vivek Kumar J	Process QA/QE	In Windchill	In Windchill
Chockalingam Kavirayar	System Architect	In Windchill	In Windchill
Nisha Kamaluddin	Independent Reviewer	In Windchill	In Windchill
Shilpa Halliyal	Project Manager	In Windchill	In Windchill

End of Document

Printed copies are uncontrolled unless authenticated