PHILIPS

Document Information
Document Identification: DHF381571
Document Name: ISP13 DRR
Version: LE.3
Life cycle state: Released 29-Mar-2023

Cover Sheet

Document Classification
Design Category: 4. Design Review
DMR Group Number: -

Signature Information		
Role	**Assignee**	**Date**
Approver (User)	Mustaq, Mohammed	28-Mar-2023
Approver (User)	Gharpure, Mukul	29-Mar-2023
Approver (User)	Kamaluddin, Nisha	28-Mar-2023
Approver (User)	J, Vivek Kumar	28-Mar-2023

Security Information		
IP Sensitivity	**Strategic Good Indicator**	**ECCN**
Internal	3 Controlled, listed in the US EAR US-origin/subject to the US EAR	EAR99

Note: The stamp below is identifying the following elements
ID number; Version; State; Document Name; DMR Group Number

DHF381571;LE.3;Released;ISP13 DRR;-

Design Plan and Input Review Record

IntelliSpace Portal 13/Advanced Visualization workspace AVW15 ,CT ASPECTS v15, & NM Suite v15

Document Identification: DHF381571

Document Revision:LE

Printed copies are uncontrolled unless authenticated

1 Document Introduction 3
1.1 Purpose 3
1.2 Scope 3
1.2.1 PRC 3
1.2.2 VAL2 3
1.2.3 VAL3 3
1.2.4 RFD 3
1.2.5 RC 4
1.3 References 4
1.3.1 PRC 4
1.3.2 VAL2 5
1.3.3 VAL3 5
1.3.4 RFD 5
1.3.5 RC 5
1.4 Abbreviations and Definitions 6
2 Design Review Material 7
2.1 PRC 7
2.2 VAL2 7
2.3 VAL3 7
2.4 RFD 8
2.5 RC 9
3 Design Review Meeting 10
3.1 PRC 10
3.2 VAL2 14
3.3 VAL3 18
3.4 RFD 21
3.5 RC 25
4 Actions 30
4.1 PRC 30
4.2 VAL2 30
4.3 VAL3 30
4.4 RFD 30
4.5 RC 30
5 Conclusion 31
5.1 PRC 31
5.2 VAL2 31
5.3 VAL3 31
5.4 RFD 31
5.5 RC 31
6 Record history 31
7 Record Review and Approval 32

Printed copies are uncontrolled unless authenticated

1 Document Introduction

1.1 Purpose

The purpose of the review is to ensure,

- All the applicable plans are available as needed, reviewed and approved for design and development phase.
- Completeness and consistency of the product requirements and appropriate mitigation of high level release risks.
- Completeness, consistency, and traceability of product design related to the product requirements.
- the risk-related elements of the product design are adhering to the overall product requirements and product design

- Note: IntelliSpace Portal is rebranded to Advanced Visualization Workspace (AVW) at RfD , therefore, throughout the document the name “IntelliSpace Portal V13 (ISP13)” is equivalent to the rebranded new name “Advanced Visualization Workspace V15 (AVW15)”

1.2 Scope

1.2.1 PRC

The scope of the Design Plan and Input Review at PRC is Plans, Requirements and System design for IntelliSpace Portal 13.

1.2.2 VAL2

The scope of the Design Plan and Input Review is Plans, Requirements and System design for IntelliSpace Portal 13 VAL2 milestone. Major features are

- CT Liver enhancements
- MR Cardiac Suite Applications
- CCA enhancements

1.2.3 VAL3

The scope of the Design Plan and Input Review is Plans, Requirements and System design for IntelliSpace Portal 13 VAL3 milestone. Major features are

- MR Cardiac Suite[FCMR,Temporal,Mapping]
- Spectral AVA
- Auto CS
- LoBI CVS

1.2.4 RFD

The scope of the Design Plan and Input Review is Plans, Requirements and System design for the following medical devices for RFD milestone.

- Advanced Visualization workspace AVW15
- ASPECTS (Alberta stroke program early CT score) V15
- Nuclear Medicine Suite (NM Suite) V15

Major features are

- Discreet reporting
- MR Cardiac Suite apps
- Spectral Enhancements
- Sys Enhancements
  - performance, reliability, security
  - Dynacad upgrade

Printed copies are uncontrolled unless authenticated

- EOL and EOS apps
- Labelling Changes
- Serviceability
- Branding Changes
- ZFP
- NM Suite
- CT ASPECTS

1.2.5 RC

The scope of the Design Plan and Input Review is Plans, Requirements and System design for the following medical devices for RC milestone.

- Advanced Visualization workspace AVW15
- ASPECTS (Alberta stroke program early CT score) V15
- Nuclear Medicine Suite (NM Suite) V15

Major features for

- Advanced Visualization workspace AVW15
  - Cardiac Suite changes
  - CCA changes
  - AVA Changes
  - LOBI changes
  - CT Brain Perfusion changes
  - CT Viewer changes for Angio CT support
  - sMGoP 3D changes.
  - 3rd party upgrades
    - Autoquant 2017.22
    - Corridor4DM 2023
    - Cortech 3.0.2 + patch
    - Dynacad 5.1 (Not for sVer)
  - System Features
    - Adoption IoT infra for Remote serviceability
    - Enable Encryption option on PMT
    - Client auto upgrade – grace period configuration
  - HW upgrades
    - Support for 3rd Gen CPU
    - Support for A2 GPU
  - KnowledgeScape updates
- ASPECTS (Alberta stroke program early CT score) V15
  - No Changes
- Nuclear Medicine Suite (NM Suite) V15
  - No Changes

1.3 References

1.3.1 PRC

Reference Number	Document Title	Document iD
REF-1	Product Release Plan	DHF380218
REF-2	User Requirement Specification (URS)	DHF380082
REF-3	Product Requirement Specification (PRS)	DHF380211
REF-4	System Design Specification	DHF380234
REF-5	System Risk Management Analysis	DHF380158
REF-6	Configuration Management Plan	DHF380530
REF-7	Software Tool List	DHF379868
REF-8	3rd Party items assessment checklist	DHF366496

Printed copies are uncontrolled unless authenticated

REF-9	Risk Management Analysis-Purchased part safety classification RMA	DHF380157

1.3.2 VAL2

Reference Number	Document Title	Document iD
REF-*1*	Regulatory and Compliance Plan	DHF380065
REF-*2*	Usability Engineering Plan	DHF380228
REF-*3*	Product Safety Risk Management Plan	DHF380156

1.3.3 VAL3

Reference Number	Document Title	Document iD
REF-*1*	Regulatory and Compliance Plan	DHF380065
REF-*2*	Usability Engineering Plan	DHF380228
REF-*3*	Product Safety Risk Management Plan	DHF380156

1.3.4 RFD

Reference Number	Document Title	Document iD
REF-*1*	Regulatory and Compliance Plan	DHF380065
REF-*2*	Usability Engineering Plan	DHF380228
REF-*3*	Product Safety Risk Management Plan	DHF380156
REF-*4*	ISP NM Suite Device Description	D001027712
REF-*5*	CT ASPECT Scoring Device Description	D001010314
REF-*6*	ISP13 Device Description	DHF380253

1.3.5 RC

Reference Number	Document Title	Document ID
REF-1	Regulatory and Compliance Plan	DHF380065
REF-2	Usability Engineering Plan	DHF380228
REF-3	Product Safety Risk Management Plan	DHF380156
REF-4	ISP NM Suite Device Description	D001027712
REF-5	CT ASPECT Scoring Device Description	D001010314
REF-6	ISP13 Device Description	DHF380253
REF-7	Product Release Plan	DHF380218
REF-8	AVW Regulatory Plan	DHF380065
REF-9	ASPECTS regulatory plan	D000997926

Printed copies are uncontrolled unless authenticated

REF-10	AVW NM Suite Regulatory plan	D001157509
REF-11	Third party assessment	DHF380838
REF-12	URS	DHF380082
REF-13	PRS	DHF380211
REF-14	SDS	DHF380234
REF-15	Usability Engineering Plan	DHF380228
REF-16	Reliability Plan	D000933496
REF-17	ISP13 RMA for ISP System	DHF380158
REF-18	Product Risk Management Plan	DHF380156
REF-19	Software Tool List	DHF379868
REF-20	ISP13 Traceability Matrix Cardiac Suite	D000996539
REF-21	AVW15_Traceability_SystemFeatures-BLR & Serviceability	D001164937
REF-22	ISP13 Traceability Matrix CT Cardiac	D000946874
REF-23	ISP13 CTViewer Traceability Matrix	D000999072
REF-24	Traceability Matrix CT Apps	D000993523
REF-25	Traceability Matrix System HFA rev A	D001159554
REF-26	AVW15 Traceability Matrix for CTSpectral Applications	D001097887
REF-27	AVW15 Traceability Matrix_NFT	D001163855
REF-28	ISP13 URS-PRS Traceability	D001105259

1.4 Abbreviations and Definitions

Terminology & Abbreviations	Description/Definition
ISP	IntelliSpace Portal
PRC	Product Release Committed
PRS	Product Requirement Specification
RFD	Release for delivery
RMP	Risk Management Plan
RMM	Risk Management Matrix
SDS	System Design Specification
SVer	Start of Verification
URS	User Requirement Specification
DRR	Design Review Record

Printed copies are uncontrolled unless authenticated

2 Design Review Material

2.1 PRC

Document Reference	Title	Version
DHF380218	Product Release Plan	1.0
DHF380082	User Requirement Specification (URS)	1.0
DHF380211	Product Requirement Specification (PRS)	1.0
DHF380234	System Design Specification	1.0
DHF380158	System Risk Management Analysis	1.0
DHF380530	Configuration Management Plan	1.0
DHF379868	Software Tool List	1.0
DHF366496	3rd Party items assessment checklist	1.0
DHF380157	Risk Management Analysis-Purchased part safety classification RMA	1.0

2.2 VAL2

Document Reference	Title	Version
DHF380218	Product Release Plan	LC
DHF380082	User Requirement Specification (URS)	LD
DHF380211	Product Requirement Specification (PRS)	LC
DHF380234	System Design Specification	LB
DHF380530	Configuration Management Plan	LB
DHF379868	Software Tool List	LD
DHF380838	3rd Party items assessment checklist	A
DHF380158	System Risk Management Analysis	LB
DHF380156	System Risk Management Plan	LB
DHF380157	Risk Management Analysis-Purchased part safety classification RMA	LB
D000996286	System RMM	A
D000992371	Code Review Record	A
D000941165	Label Specification	B
D000993080	Design Input review – BlrApps (MR Cardiac Suite)	A
D000998369	Design Input Review – HFA Apps (CT Liver)	A
D000999005	Design Input Review – CCA	A

2.3 VAL3

Printed copies are uncontrolled unless authenticated

Document Reference	Title	Version
DHF380218	Product Release Plan	LD
DHF380082	User Requirement Specification (URS)	LD
DHF380211	Product Requirement Specification (PRS)	LD
DHF380234	System Design Specification	LC
DHF381138	PSRA	LA
DHF380065	Regulatory Plan	LB
DHF380228	Usability engineering plan	LB
DHF380530	Configuration Management Plan	LB
DHF379868	Software Tool List	LD
DHF380838	3rd Party items assessment checklist	B
DHF380157	Risk Management Analysis-Purchased part safety classification RMA	LB
D000992371	Code Review record	B
D000996286	System RMM	A
DHF380158	System RMA	LC
D000941165	Label Specs	B
DHF380156	Risk Management Plan	LB
DHF380216	Quality Plan	LC
D000993080	Design Input review – BlrApps (Cardiac Suite and LOBI)	B
D000998369	Design Input Review – HFA Apps (Auto CS, sAVA)	B
DHF381571	System Design input review	LC

2.4 RFD

Document Reference	Title	Version
DHF380218	Product Release Plan	LE
DHF380082	User Requirement Specification (URS)	LE
DHF380211	Product Requirement Specification (PRS)	LF
DHF380234	System Design Specification	LD
DHF381138	PSRA	LB
DHF380065	Regulatory and compliance Plan	LD
D000997926	CT ASPECT Regulatory and Compliance Plan	A
D001157509	ISP NM Suite Regulatory Plan	A
DHF380228	Usability engineering plan	LC
DHF380530	Configuration Management Plan	LC
DHF379868	Software Tool List	LE
DHF380838	3rd Party items assessment checklist	C
DHF380216	Quality Plan	LD
D001169806	CT ASPECT Quality Plan	A
D001162088	AVW NM Suite Quality Plan	A

Printed copies are uncontrolled unless authenticated

DHF380157	Risk Management Analysis-Purchased part safety classification RMA	LC
D000992371	Code Review record	C
D000996286	System RMM	B
DHF380158	System RMA	LD
DHF381090	ISP13 Labeling Plan	A
D001158353	ISP NM Suite Labeling Plan	A
D000996803	CT ASPECT Labeling Plan	A
D000941165	Label Specs	C
DHF380156	Risk Management Plan	LD
D000933496	ISP13 Reliability plan	B
D001145384	ISP NM Suite Labeling Spec	A
D000996565	ASPECTS Label Specification	A
D001149241	Advanced Visualization Workspace Enterprise Label Specification	A
D001105259	ISP13 URS-PRS Traceability	A
D001157240	ISP13 System Performance DRS	A
D001149811	ISP13 Installation DRS	A
D000993080	Design Input review – Cardiac Suite and NM Suite and System Blr	C
D000998369	Design Input Review – HFA Apps & Backbone Hfa	C
D000999005	Design Input Review - Shina	B
D001155045	AVW15 Design Input Review Record (Spectral) Rev A.docx	A
D001163855	AVW15 Traceability Matrix_NFT	A
D001027712	ISP NM Suite Device Description	A
D001010314	CT ASPECT Scoring Device Description	A
DHF380253	ISP13 Device Description	LB
DHF381571	System Design input review	LD

Printed copies are uncontrolled unless authenticated

2.5 RC

ID	Description	Version
D001311796	AVW15 Change Request for RC	B
DHF380082	ISP13 URS	LG
DHF380211	ISP13 PRS	LI
DHF380530	ISP13 CMP	LD
DHF380234	ISP13 SDS	LE
DHF380218	ISP13 Product Release Plan	LF
DHF380228	Usability engineering plan for ISP13	LE
DHF380065	ISP13 Regulatory and Compliance Plan	LG
D000997926	CT ASPECT Regulatory and Compliance Plan	C
D001157509	AVW NM Suite Regulatory Plan	D
DHF381090	ISP13 Labeling Plan	LE
D000996803	CT ASPECT Labeling Plan	C

D001158353	AVW NM Suite Labeling Plan	D
DHF380216	ISP13 Quality Plan	LF
D001169806	CT ASPECT Quality Plan	B
D000992371	ISP13 Code Review Record - Consolidated	E
DHF379868	ISP13 Tool List	LH
DHF380838	ISP13 3rd Party items assessment checklist	E
DHF380158	ISP13 RMA for ISP System	LH
D001157240	ISP13 System Performance DRS	B
D001149811	AVW15_DRS_System_Installation	B
D001163851	AVW15 NFT Performance Test Specification	B
D001163852	AVW15_Automation Test Script Validation Report_NFT_Performance	B
D001163854	AVW15_Consolidated Code TRR for NFTPerformanceAutomation	B
D001163855	AVW15 Traceability Matrix_NFT	B
D001105259	ISP13 URS-PRS Traceability	C
D001163817	Risk Management Analysis for AVW product labels	C
D000996565	ASPECTS Label Specification	C
D001145384	ISP NM Suite Labeling Spec	C
D000995566	ISP13_Design Verification Plan	G
D001115079	AVW15_PVT_Installation & Upgrade _TestSpec	F
D001058943	ISP13_PVT_TestSpec	C
D000998369	ISP13 DRR HFA	D
D000999005	ISP13 Design Review Shina Apps	C
D000993080	Design Input review – Cardiac Suite and NM Suite and System Blr	D
D001155045	AVW15 Design Input Review Record (Spectral)	B
DHF381571	ISP13 System DRR	LE

Printed copies are uncontrolled unless authenticated

3 Design Review Meeting

3.1 PRC

Date: 23-July-2020

Participants

Name	Role	Represented Function	Independent	Present
Dror Orell	Program Manager	PMO	N	Y
Daniel Broiman	Project Manager/Release Manager	R&D	N	Y
Sheela Biradar	Verification Lead	R&D Verification	N	Y
Sandhya Kaiprath	Quality Engineer	Q&R	N	Y
Keren Afek	Regulatory Affairs Specialist	Q&R	N	Y
Sujesh Paravatty	Senior Manager II	ICAP Clinical Science	N	Y
Srinivas Arakere	Project Manager	Independent reviewer	Y	Y

Copel, Zeev	ICAP Portal Backbone R&D Director	R&D ICAP	N	Y
Moshe Katz	Portal Architecture & System Engineering Director	R&D System Architecture	N	Y
Bornstein, Carmit	Systems Engineer	System Engineering	N	Y
Shapira, Alina	Product Manager	Marketing	N	Y
Davidesco, Hagit	Validation	ICAP Clinical Science	N	Y(Offline)
Kohn, Eyal	Lead Systems Engineer	System Engineering	N	Y
Hunsur Adishesha, Manjunath	Devops Project Manager/ Configuration Manager	R&D DevOps	N	Y
Afota, Mark Mordechi	ICAP OSS NPI Lead	Customer Service	N	Y
Amir, Arik	DevOps Specialist	R&D DevOps	N	Y
Byamma,munichandrappa	Systems Engineer	System Engineering	N	Y
Tripurana, Syamala	Project Manager	R&D Operations	N	Y
Tsikun, Andrey	Design Transfer Lead	Customer Service	N	Y
Yanovich, Zohar	Platform SW Sr. manager	R&D Spectral	N	Y
Yavin, Amir	DeveOps Specialist	R&D DevOps	N	Y

Discussions from Design review meetings and questions asked:

Design plan and input checks	**Evidence**	**Remarks**
Have all the software development planning aspects considered in your appropriate plan?	Yes, Product Release Plan - DHF380218	NA
Are applicable regulations and standards for compliance considered as part of Regulatory and Compliance Plan?	Yes, Regulatory Plan - DHF381138	NA
Have you planned for legacy software?	Yes, Product Release Plan - DHF380218	NA
Have you identified and planned for SOUP and third party software and their integration?	Third party assessment - DHF366496 Risk Management Analysis- Purchased part safety classification RMA DHF380157	Assessment done based on the initial release for PRC. This will be updated based on incremental plan of ISP13.
Are all SOUP items third party software assessed for their risks?	Yes : Risk Management Analysis- Purchased part safety classification RMA	Initial assessment done based on ISP12. Further improvements will be

Printed copies are uncontrolled unless authenticated

	DHF380157	handled post-PRC during incremental development.
Are the user needs understood and the intended use clearly articulated?	Yes, URS - DHF380082	
Is there at least one product requirement for each user need?	Yes ,PRS - DHF380211	ISP13 specific URS keys PRS keys will be updated Post PRC.
Are product requirements unambiguous and testable? Do they cover both functional and nonfunctional requirements?	Yes ,PRS - DHF380211	Will be further updates as part of ISP13 incremental development.
Are all the requirements compliant with clinical relevance, applicable laws and regulations?	Yes ,PRS - DHF380211	Will be same as ISP12 and will be updated incrementally during ISP13 incremental development
Does the User Requirement Specification (URS) capture User needs, Intended use, Indications for use, Primary Operating Functions (i.e. s afety related use scenarios), Intended u sers' profile, intended use environment, target markets and commercial claims?	Yes, URS - DHF380082	
Does Product Requirement Specification (PRS) ad equately translate the user and patient needs into a set of product requirements and covers all the relevant aspects such as functional and non-functional requirements	Yes ,PRS - DHF380211	PRS at higher level. These will be handled post-PRC during incremental development.
Are the design input requirements been assessed for all attributes required for the static verification?	Yes ,PRS - DHF380211	These will be handled post-PRC during incremental development.
Is the architecture and detailed design (if applicable) been assessed for static verification?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in subsequent reviews via updated SDS and/or DDS as applicable.
Product-level issues to be discussed have been listed, focusing on proposed design	Yes, URS - DHF380082	Customer inputs and RFCs is one of the inputs for URS. Additional list of RFCs to be considered is prepared and reviewed. It will be handled post-PRC during incremental development.
Are all the clinical algorithms defined and covered?	NA	These will be handled post-PRC during incremental development.

Printed copies are uncontrolled unless authenticated

Does the product design cover the product requirements?	Yes, SDS - DHF380234	SDS - Section 2.2 covers product design at high level. It will be updated post-PRC during incremental development.
Have the Usability and Human interfaces requirements been considered?	Yes - Usability Engineering Plan - DHF380228	Usability Engineering Plan covers at high level. More specific requirements will be defined post-PRC during incremental development.
Have the Reliability requirements been considered?	Yes ,PRS - DHF380211	PRS covers reliability requirements at high level. More specific requirements will be defined and handled post-PRC during incremental development.
Have Service and Maintainability requirements been considered?	Yes, URS - DHF380082	URS covers service requirements at high level. More specific requirements will be defined and handled post-PRC during incremental development.
Do the product requirements cover Interoperability, Reusability, Portability attributes?	Yes ,PRS - DHF380211	All these attributes will be equivalent of ISP12. Additional improvements will be addressed as part of ISP13 incremental development
Are the technical risks understood and articulated?	Yes, SDS - DHF380234	Higher level Technical risk will be covered as part of SDS which is equivalent of ISP12. Detailed risk will be covered later in respective RMM, DDS as part of incremental development of ISP13.
Are the component requirements appropriate for the assigned behaviors / responsibilities / interfaces?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in subsequent reviews via updated SDS and/or DDS as applicable.
Do the component requirements meet the product requirements allocated?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in subsequent reviews via updated SDS and/or DDS as applicable.

Printed copies are uncontrolled unless authenticated

Do the component requirements and design comply with the product requirements and design?	Yes, SDS - DHF380234	SDS covers at high level. Details will be specified and handled post-PRC during incremental development.
Are the risks(safety,Security,Privacy) addressed and the mitigations covered in Requirements, Design and Risk Management	Yes, Product Risk Management Plan - DHF380156, URS - DHF380082	Safety risk management plan is prepared. Specific requirements will be defined post-PRC during incremental development. Security and privacy user requirements are defined. The risks will be monitored and addressed during incremental development.
Are all design constraints and restrictions understood and covered appropriately? Ex: Some requirements would require a specific programming languages, specific algorithms, specific communication protocols, etc.	Yes, SDS -DHF380234	More will be handled Post PRC via SDS, DDS
Are all the application layers, class diagrams, sequence diagram/flow diagram, required actors, entity-relationship diagrams, interfacing systems identified as part of design(as appropriate)?	Yes, SDS -DHF380234	Post PRC will be handled via SDS,DDS updates
Are all the tools/technologies/methodologies identified for the development of the software?	Yes, SDS -DHF380234 Product Release Plan - DHF380218 Software Tool List - DHF379868	More details will be added post-PRC during incremental development.
Is the new applications design aligned with HSRA architecture and tool checklist.	Yes SDS- DHF380234	More details will be added post-PRC during incremental development.
Is the traceability evidenced as appropriate for this stage?	NA	These will be handled post-PRC during incremental development.

3.2 VAL2

Date: 07-Feb-2022

Participants

Printed copies are uncontrolled unless authenticated

Name	Role	Represented Function	Independent	Present
Dror Orell	Program Manager	PMO	N	Y
Daniel Broiman	Project Manager/Release Manager	R&D	N	Y
Aashish Kumar	Verification Lead	R&D Verification	N	Y
Shivangi, Shipra	Quality Engineer	Q&R	N	Y
Keren Afek	Regulatory Affairs Specialist	Q&R	N	Y
Sujesh Paravatty	Clinical Lead	ICAP Clinical Science	N	Y
Srinivas Arakere	Independent reviewer	Independent reviewer	Y	Y
Barzilai, Offer	Product Manager	Marketing	N	Y
Davidesco, Hagit	Validation	ICAP Clinical Science	N	Y
Rakesh C	Devops Lead	R&D DevOps	N	Y
Gross, Shahar	ICAP OSS NPI Lead	Customer Service	N	Y
Byamma,munichandrappa	Systems Engineer	System Engineering	N	Y
Baruah, Abhijit	R&D Manager BLR	R&D Development	N	Y
Harran, Ophir	R&D Manager HFA	R&D Development	N	Y

Discussions from Design review meetings and questions asked:

Design plan and input checks	Evidence	Remarks
Have all the software development planning aspects considered in your appropriate plan?	Yes, Product Release Plan - DHF380218	NA
Are applicable regulations and standards for compliance considered as part of Regulatory and Compliance Plan?	Yes, Regulatory Plan - DHF380065	NA
Have you planned for legacy software?	Yes, Product Release Plan - DHF380218	NA
Have you identified and planned for SOUP and third party software and their integration?	Yes: Third party assessment - DHF380838	NA
Are all SOUP items third party software assessed for their risks?	Yes: Third party assessment - DHF380838	NA
Are the user needs understood and the intended use clearly articulated?	Yes, URS - DHF380082	NA
Is there at least one product requirement for each user need?	Yes , PRS - DHF380211	NA

Printed copies are uncontrolled unless authenticated

	URS - DHF380082	
Are product requirements unambiguous and testable? Do they cover both functional and nonfunctional requirements?	Yes ,PRS - DHF380211	NA.
Are all the requirements compliant with clinical relevance, applicable laws and regulations?	Yes ,PRS - DHF380211 Regulatory and Compliance Plan - DHF380065	NA
Does the User Requirement Specification (URS) capture User needs, Intended use, Indications for use, Primary Operating Functions (i.e. safety related use scenarios), Intended users' profile, intended use environment, target markets and commercial claims?	Yes, URS - DHF380082	NA
Does Product Requirement Specification (PRS) ad equately translate the user and patient needs into a set of product requirements and covers all the relevant aspects such as functional and non-functional requirements	Yes ,PRS - DHF380211	NA
Are the design input requirements been assessed for all attributes required for the static verification?	Yes ,PRS - DHF380211 SDS - DHF380234	NA
Is the architecture and detailed design (if applicable) been assessed for static verification?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in sub systems reviews via updated SDS and/or DDS as applicable.
Product-level issues to be discussed have been listed, focusing on proposed design	Yes, URS - DHF380082 Yes, PRS - DHF380211	
Are all the clinical algorithms defined and covered?	YEs	algorithm requirements are covered under subsystem requirement documents.
Does the product design cover the product requirements?	Yes, SDS - DHF380234	NA
Have the Usability and Human interfaces requirements been considered?	Yes - Usability Engineering Plan - DHF380228	NA
Have the Reliability requirements been considered?	Yes ,PRS - DHF380211	NA
Have Service and Maintainability requirements been considered?	Yes, URS - DHF380082	URS covers service requirements at high level. for VAL2 specific service and maintainability not applicable.

Printed copies are uncontrolled unless authenticated

Do the product requirements cover Interoperability, Reusability, Portability attributes?	Yes ,PRS - DHF380211	NA
Are the technical risks understood and articulated?	Yes, SDS - DHF380234	Higher level Technical risk will be covered as part of SDS , Detailed risk will be covered respective RMM, DDS of sub systems.
Are the component requirements appropriate for the assigned behaviors / responsibilities / interfaces?	Yes, SDS - DHF380234	NA
Do the component requirements meet the product requirements allocated?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Do the component requirements and design comply with the product requirements and design?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are the risks(safety,Security,Privacy) addressed and the mitigations covered in Requirements, Design and Risk Management	Yes, Product Risk Management Plan - DHF380156, URS - DHF380082	NA NA
Are all design constraints and restrictions understood and covered appropriately? Ex: Some requirements would require a specific programming languages, specific algorithms, specific communication protocols, etc.	Yes, SDS -DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are all the application layers, class diagrams, sequence diagram/flow diagram, required actors, entity-relationship diagrams, interfacing systems identified as part of design(as appropriate)?	Yes, SDS -DHF380234	SDS covers high level design. Details design are covered in DDS of sub systems
Are all the tools/technologies/methodologies identified for the development of the software?	Yes, SDS -DHF380234	NA.

Printed copies are uncontrolled unless authenticated

	Product Release Plan - DHF380218 Software Tool List - DHF379868	
Is the new applications design aligned with HSRA architecture and tool checklist.	Yes SDS- DHF380234	NA
Is the traceability evidenced as appropriate for this stage?	Yes	Covered in the traceability of each sub systems. D000996539 – Cardiac Suite Traceability matrix D000993523 – Ct Apps traceability matrix D000946874 – CT Cardiac traceability matrix

3.3 VAL3

Date: 04-July-2022

Participants

Name	Role	Represented Function	Independent	Present
Dror Orell	Program Manager	PMO	N	Y
Mustaq, Mohammed	Project Manager/Release Manager	R&D	N	Y
Srinivas Arakere	Verification Lead	R&D Verification	N	Y
Shivangi, Shipra	Quality Engineer	Q&R	N	Y
Keren Afek	Regulatory Affairs Specialist	Q&R	N	Y
Sujesh Paravatty	Clinical Lead	ICAP Clinical Science	N	Y
Baruah, Abhijit	Independent reviewer	Independent reviewer	Y	Y
Barzilai, Offer	Product Manager	Marketing	N	N
Davidesco, Hagit	Validation	ICAP Clinical Science	N	Y
Rakesh C	Devops Lead	R&D DevOps	N	Y
Gross, Shahar	ICAP OSS NPI Lead	Customer Service	N	Y
Byamma,munichandrappa	Systems Engineer	System Engineering	N	Y
Patil, Vighnesh Kumar A	R&D Manager BLR	R&D Development	N	Y
Harran, Ophir	R&D Manager HFA	R&D Development	N	Y
Halaharvi, Raghavendra Prasad	RTE , BlrApps	PMO	N	Y
Halliyal, Shilpa	RTE , HFAApps	PMO	N	Y

Printed copies are uncontrolled unless authenticated

Discussions from Design review meetings and questions asked:

Design plan and input checks	Evidence	Remarks
Have all the software development planning aspects considered in your appropriate plan?	Yes, Product Release Plan - DHF380218	NA
Are applicable regulations and standards for compliance considered as part of Regulatory and Compliance Plan?	Yes, Regulatory Plan - DHF380065	NA
Have you planned for legacy software?	Yes, Product Release Plan - DHF380218	NA
Have you identified and planned for SOUP and third party software and their integration?	Yes: Third party assessment - DHF380838	NA
Are all SOUP items third party software assessed for their risks?	Yes: Third party assessment - DHF380838	NA
Are the user needs understood and the intended use clearly articulated?	Yes, URS - DHF380082	NA
Is there at least one product requirement for each user need?	Yes , PRS - DHF380211 URS - DHF380082	NA
Are product requirements unambiguous and testable? Do they cover both functional and nonfunctional requirements?	Yes ,PRS - DHF380211	NA.
Are all the requirements compliant with clinical relevance, applicable laws and regulations?	Yes ,PRS - DHF380211 Regulatory and Compliance Plan - DHF380065	NA
Does the User Requirement Specification (URS) capture User needs, Intended use, Indications for use, Primary Operating Functions (i.e. s afety related use scenarios), Intended u sers' profile, intended use environment, target markets and commercial claims?	Yes, URS - DHF380082	NA
Does Product Requirement Specification (PRS) ad equately translate the user and patient needs into a set of product requirements and covers all the relevant aspects such as functional and non-functional requirements	Yes ,PRS - DHF380211	NA
Are the design input requirements been assessed for all attributes required for the static verification?	Yes ,PRS - DHF380211 SDS - DHF380234	NA

Printed copies are uncontrolled unless authenticated

Is the architecture and detailed design (if applicable) been assessed for static verification?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in sub systems reviews via updated SDS and/or DDS as applicable.
Product-level issues to be discussed have been listed, focusing on proposed design	Yes, URS - DHF380082 Yes, PRS - DHF380211	
Are all the clinical algorithms defined and covered?	YEs	algorithm requirements are covered under subsystem requirement documents.
Does the product design cover the product requirements?	Yes, SDS - DHF380234	NA
Have the Usability and Human interfaces requirements been considered?	Yes - Usability Engineering Plan - DHF380228	NA
Have the Reliability requirements been considered?	Yes ,PRS - DHF380211	NA
Have Service and Maintainability requirements been considered?	Yes, URS - DHF380082	URS covers service requirements at high level. for VAL3 specific service and maintainability not applicable.
Do the product requirements cover Interoperability, Reusability, Portability attributes?	Yes ,PRS - DHF380211	NA
Are the technical risks understood and articulated?	Yes, SDS - DHF380234	Higher level Technical risk will be covered as part of SDS , Detailed risk will be covered respective RMM, DDS of sub systems.
Are the component requirements appropriate for the assigned behaviors / responsibilities / interfaces?	Yes, SDS - DHF380234	NA
Do the component requirements meet the product requirements allocated?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Do the component requirements and design comply with the product requirements and design?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are the risks(safety,Security,Privacy) addressed and the mitigations covered in Requirements, Design and Risk Management	Yes, Product Risk Management Plan - DHF380156, URS - DHF380082	NA NA

Printed copies are uncontrolled unless authenticated

Printed copies are uncontrolled unless authenticated

Are all design constraints and restrictions understood and covered appropriately? Ex: Some requirements would require a specific programming languages, specific algorithms, specific communication protocols, etc.	Yes, SDS -DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are all the application layers, class diagrams, sequence diagram/flow diagram, required actors, entity-relationship diagrams, interfacing systems identified as part of design(as appropriate)?	Yes, SDS -DHF380234	SDS covers high level design. Details design are covered in DDS of sub systems
Are all the tools/technologies/methodologies identified for the development of the software?	Yes, SDS -DHF380234 Product Release Plan - DHF380218 Software Tool List - DHF379868	NA.
Is the new applications design aligned with HSRA architecture and tool checklist.	Yes SDS- DHF380234	NA
Is the traceability evidenced as appropriate for this stage?	Yes	Covered in the traceability of each sub systems. D001105932– Tracebility matrix LOBI D000996539–traceability matrix Cardiac Suite D000993523– Traceability matrix CT Apps

3.4 RFD

Date: 15-Sep-2022

Participants

Name	Role	Represented Function	Independent	Present
Dror Orell	Program Manager	PMO	N	Y

Mustaq, Mohammed	Project Manager	R&D	N	Y
Srinivas Arakere	Verification Lead	R&D Verification	N	Y
Shivangi, Shipra	Quality Engineer	Q&R	N	Y
Keren Afek	Regulatory Affairs Specialist	Q&R	N	Y (Offline)
Sujesh Paravatty	Clinical Lead	Clinical Excellence	N	Y
Kamaluddin, Nisha	Independent reviewer	Independent reviewer	Y	Y
Barzilai, Offer	Product Manager	Marketing	N	Y (Offline)
Davidesco, Hagit	Validation Lead	Clinical Excellence	N	Y
Rakesh C	Devops Lead	R&D DevOps	N	Y
Afota, Mark Mordechi	ICAP OSS NPI Lead	Customer Service	N	Y
Byamma,munichandrappa	Systems Engineer	System Engineering	N	Y
Parakal, Elizabeth	Global R&D lead	R&D management.	N	Y
Halaharvi, Raghavendra Prasad	RTE , BlrApps	PMO	N	Y
Halliyal, Shilpa	RTE , HFAApps	PMO	N	Y

Discussions from Design review meetings and questions asked:

Design plan and input checks	**Evidence**	**Remarks**
Have all the software development planning aspects considered in your appropriate plan?	Yes, Product Release Plan - DHF380218	NA
Are applicable regulations and standards for compliance considered as part of Regulatory and Compliance Plan?	Yes, AVW Regulatory Plan - DHF380065 ASPECTS regulatory plan - D000997926 NM Suite Regulatory plan - D001157509	NA
Have you planned for legacy software?	Yes, Product Release Plan - DHF380218	NA
Have you identified and planned for SOUP and third party software and their integration?	Yes: Third party assessment - DHF380838	NA
Are all SOUP items third party software assessed for their risks?	Yes: Third party assessment - DHF380838	NA
Are the user needs understood and the intended use clearly articulated?	Yes, URS - DHF380082	NA

Printed copies are uncontrolled unless authenticated

Is there at least one product requirement for each user need?	Yes , PRS - DHF380211 URS - DHF380082	NA
Are product requirements unambiguous and testable? Do they cover both functional and nonfunctional requirements?	Yes ,PRS - DHF380211	NA.
Are all the requirements compliant with clinical relevance, applicable laws and regulations?	Yes ,PRS - DHF380211 AVW Regulatory Plan - DHF380065 ASPECTS regulatory plan - D000997926 NM Suite Regulatory plan - D001157509	NA
Does the User Requirement Specification (URS) capture User needs, Intended use, Indications for use, Primary Operating Functions (i.e. safety related use scenarios), Intended users' profile, intended use environment, target markets and commercial claims?	Yes, URS - DHF380082	NA
Does Product Requirement Specification (PRS) ad equately translate the user and patient needs into a set of product requirements and covers all the relevant aspects such as functional and non-functional requirements	Yes ,PRS - DHF380211	NA
Are the design input requirements been assessed for all attributes required for the static verification?	Yes ,PRS - DHF380211 SDS - DHF380234	NA
Is the architecture and detailed design (if applicable) been assessed for static verification?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in sub systems reviews via updated SDS and/or DDS as applicable.
Product-level issues to be discussed have been listed, focusing on proposed design	Yes, URS - DHF380082 Yes, PRS - DHF380211 Yes, SDS - DHF380234	
Are all the clinical algorithms defined and covered?	Yes	algorithm requirements are covered under subsystem requirement documents.
Does the product design cover the product requirements?	Yes, SDS - DHF380234	NA

Printed copies are uncontrolled unless authenticated

Have the Usability and Human interfaces requirements been considered?	Yes - Usability Engineering Plan - DHF380228	NA
Have the Reliability requirements been considered?	Yes ,PRS - DHF380211 Reliability Plan - D000933496	NA
Have Service and Maintainability requirements been considered?	Yes, URS - DHF380082 PRS - DHF380211	NA
Do the product requirements cover Interoperability, Reusability, Portability attributes?	Yes ,PRS - DHF380211	NA
Are the technical risks understood and articulated?	Yes, SDS - DHF380234	Higher level Technical risk will be covered as part of SDS , Detailed risk will be covered respective RMM, DDS of sub systems.
Are the component requirements appropriate for the assigned behaviors / responsibilities / interfaces?	Yes, SDS - DHF380234	NA
Do the component requirements meet the product requirements allocated?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Do the component requirements and design comply with the product requirements and design?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are the risks(safety,Security,Privacy) addressed and the mitigations covered in Requirements, Design and Risk Management	Yes, Product Risk Management Plan - DHF380156, URS - DHF380082	NA NA
Are all design constraints and restrictions understood and covered appropriately? Ex: Some requirements would require a specific programming languages, specific algorithms, specific communication protocols, etc.	Yes, SDS -DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are all the application layers, class diagrams, sequence diagram/flow diagram, required actors, entity-relationship diagrams, interfacing systems identified as part of design(as appropriate)?	Yes, SDS -DHF380234	SDS covers high level design. Details design are covered in DDS of sub systems

Printed copies are uncontrolled unless authenticated

Are all the tools/technologies/methodologies identified for the development of the software?	Yes, SDS -DHF380234 Product Release Plan - DHF380218 Software Tool List - DHF379868	NA.
Is the new applications design aligned with HSRA architecture and tool checklist.	Yes SDS- DHF380234 Software Tool List - DHF379868 Third party assessment - DHF380838	NA
Is the traceability evidenced as appropriate for this stage?	Yes	Covered in the traceability of each sub systems. D000996539- ISP13 Traceability Matrix Cardiac Suite D001164937 - AVW15_Traceability_SystemFeatures-BLR & Serviceability D000946874 ISP13 Traceability Matrix CT Cardiac D000993523 Traceability Matrix CT Apps D001159554 Traceability Matrix System HFA rev A D001097887 AVW15 Traceability Matrix for CTSpectral Applications D001163855 AVW15 Traceability Matrix_NFT D001105259 ISP13 URS-PRS Traceability

3.5 RC

Date: 28-Mar-2023

Participants

Printed copies are uncontrolled unless authenticated

Name	Role	Represented Function	Independent	Present
Biran, Eyal	Program Manager	PMO	N	Y
Mustaq, Mohammed	Project Manager	R&D	N	Y
Srinivas Arakere	Verification Lead	R&D Verification	N	Y
Vivek Kumar J	Quality Engineer	Q&R	N	Y
Keren Afek	Regulatory Affairs Specialist	Q&R	N	Y (Offline)
Sujesh Paravatty	Clinical Lead	Clinical Excellence	N	Y
Kamaluddin, Nisha	Independent reviewer	Independent reviewer	Y	Y
Kliper Goldin, Efrat	Product Manager	Marketing	N	Y
Davidesco, Hagit	Validation Lead	Clinical Excellence	N	Y
Rakesh C	Devops Lead	R&D DevOps	N	Y(Offline)
Sinha, Rohit Kumar	ICAP OSS NPI Lead	Customer Service	N	Y
Kogan, Natalia	Systems Engineer	System Engineering	N	Y
Parakal, Elizabeth	Global R&D lead	R&D management.	N	Y (Offline)
Halliyal, Shilpa	RTE , HFAApps	PMO	N	Y
Prathap Sekar Reddy, Sravan Kumar	R&D Lead	R&D management.	N	Y(Offline)

Discussions from Design review meetings and questions asked:

Design plan and input checks	Evidence	Remarks
Have all the software development planning aspects considered in your appropriate plan?	Yes, Product Release Plan - DHF380218	NA
Are applicable regulations and standards for compliance considered as part of Regulatory and Compliance Plan?	Yes, AVW Regulatory Plan - DHF380065 ASPECTS regulatory plan - D000997926 NM Suite Regulatory plan - D001157509	NA
Have you planned for legacy software?	Yes, Product Release Plan - DHF380218	NA
Have you identified and planned for SOUP and third party software and their integration?	Yes: Third party assessment - DHF380838	NA
Are all SOUP items third party software assessed for their risks?	Yes: Third party assessment - DHF380838	NA

Printed copies are uncontrolled unless authenticated

Are the user needs understood and the intended use clearly articulated?	Yes, URS - DHF380082	NA
Is there at least one product requirement for each user need?	Yes , PRS - DHF380211 URS - DHF380082 PRS-URS tracebility D001105259	NA
Are product requirements unambiguous and testable? Do they cover both functional and nonfunctional requirements?	Yes ,PRS - DHF380211 PRS-URS tracebility D001105259	NA.
Are all the requirements compliant with clinical relevance, applicable laws and regulations?	Yes ,PRS - DHF380211 AVW Regulatory Plan - DHF380065 ASPECTS regulatory plan - D000997926 NM Suite Regulatory plan - D001157509	NA
Does the User Requirement Specification (URS) capture User needs, Intended use, Indications for use, Primary Operating Functions (i.e. safety related use scenarios), Intended users' profile, intended use environment, target markets and commercial claims?	Yes, URS - DHF380082	NA
Does Product Requirement Specification (PRS) ad equately translate the user and patient needs into a set of product requirements and covers all the relevant aspects such as functional and non-functional requirements	Yes ,PRS - DHF380211	NA
Are the design input requirements been assessed for all attributes required for the static verification?	Yes ,PRS - DHF380211 SDS - DHF380234	NA
Is the architecture and detailed design (if applicable) been assessed for static verification?	Yes, SDS - DHF380234	SDS covers high level assignments. Details will follow in sub systems reviews via updated SDS and/or DDS as applicable.
Product-level issues to be discussed have been listed, focusing on proposed design	Yes, URS - DHF380082 PRS - DHF380211 SDS - DHF380234	
Are all the clinical algorithms defined and covered?	Yes	algorithm requirements are covered under subsystem

Printed copies are uncontrolled unless authenticated

	PRS - DHF380211	requirement and design documents.
Does the product design cover the product requirements?	Yes, SDS - DHF380234	NA
Have the Usability and Human interfaces requirements been considered?	Yes - Usability Engineering Plan - DHF380228	NA
Have the Reliability requirements been considered?	Yes, URS - DHF380082 PRS - DHF380211 Reliability Plan - D000933496	NA
Have Service and Maintainability requirements been considered?	Yes, URS - DHF380082 PRS - DHF380211	NA
Do the product requirements cover Interoperability, Reusability, Portability attributes?	Yes ,PRS - DHF380211 SDS - DHF380234	NA
Are the technical risks understood and articulated?	Yes, SDS - DHF380234 System RMA - DHF380158	Higher level Technical risk will be covered as part of SDS , Detailed risk will be covered respective RMM, DDS of sub systems.
Are the component requirements appropriate for the assigned behaviors / responsibilities / interfaces?	Yes, SDS - DHF380234	NA
Do the component requirements meet the product requirements allocated?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Do the component requirements and design comply with the product requirements and design?	Yes, SDS - DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are the risks(safety,Security,Privacy) addressed and the mitigations covered in Requirements, Design and Risk Management	Yes, Product Risk Management Plan - DHF380156, URS - DHF380082	NA NA
Are all design constraints and restrictions understood and covered appropriately? Ex: Some requirements would require a specific programming languages, specific algorithms, specific communication protocols, etc.	Yes, SDS -DHF380234	SDS covers high level design. Details are covered in DDS of sub systems
Are all the application layers, class diagrams, sequence diagram/flow diagram,	Yes, SDS -DHF380234	SDS covers high level design. Details design are

Printed copies are uncontrolled unless authenticated

required actors, entity-relationship diagrams, interfacing systems identified as part of design(as appropriate)?		covered in DDS of sub systems
Are all the tools/technologies/methodologies identified for the development of the software?	Yes, SDS -DHF380234 Product Release Plan - DHF380218 Software Tool List - DHF379868	NA.
Is the new applications design aligned with HSRA architecture and tool checklist.	Yes SDS- DHF380234 Software Tool List - DHF379868 Third party assessment - DHF380838	NA
Is the traceability evidenced as appropriate for this stage?	Yes	Covered in the traceability of each sub systems. D000996539- ISP13 Traceability Matrix Cardiac Suite D001164937 - AVW15_Traceability_SystemFeatures-BLR & Serviceability D000946874 ISP13 Traceability Matrix CT Cardiac D000999072 ISP13 CTViewer Traceability Matrix D000993523 Traceability Matrix CT Apps D001159554 Traceability Matrix System HFA rev A D001097887 AVW15 Traceability Matrix for CTSpectral Applications D001163855 AVW15 Traceability Matrix_NFT D001105259 ISP13 URS-PRS Traceability

Printed copies are uncontrolled unless authenticated

4 Actions

4.1 PRC

No Actions.

Id	Action	Status	Expected Closure date	Reference

4.2 VAL2

No Actions.

Id	Action	Status	Expected Closure date	Reference

4.3 VAL3

No Actions.

Id	Action	Status	Expected Closure date	Reference

4.4 RFD

No Actions.

Id	Action	Status	Expected Closure date	Reference

4.5 RC

No Actions.

Id	Action	Status	Expected Closure date	Reference

Printed copies are uncontrolled unless authenticated

Printed copies are uncontrolled unless authenticated

5 Conclusion

5.1 PRC

The user requirements fulfil and suit, after systematic examination, manufacturing, installation and service stakeholders. The relevant plans are ready to start next phase of project development.

Review Result: Approved

5.2 VAL2

The product requirements fulfill, after systematic examination, the user requirements and business requirements; and suit installation and service stakeholders.

The system and detailed designs fullfill, after systematic examination, the product requirements; and suite installation and service stakeholders

Review Result: Approved

5.3 VAL3

The product requirements fulfill, after systematic examination, the user requirements and business requirements; and suit installation and service stakeholders.

The system and detailed designs fullfill, after systematic examination, the product requirements; and suite installation and service stakeholders

Review Result: Approved

5.4 RFD

The product requirements fulfill, after systematic examination, the user requirements and business requirements; and suit installation and service stakeholders.

The system and detailed designs fullfill, after systematic examination, the product requirements; and suite installation and service stakeholders

Review Result: Approved

5.5 RC

The product requirements fulfill, after systematic examination, the user requirements and business requirements; and suit installation and service stakeholders.

The system and detailed designs fullfill, after systematic examination, the product requirements; and suite installation and service stakeholders

Review Result: Approved

6 Record history

Rev	Release Date	Author	Description of changes

1.0	In Agile	Chockalingam Kavirayar	Review record for ISP13 Design plan and input review
LB	In WindChill	Chockalingam Kavirayar	Review record for ISP13 Design plan and input review for VAL2
LC	In WindChill	Chockalingam Kavirayar	Review record for ISP13 Design plan and input review for VAL3 Updated Section 5.2 for correcting typographical error.
LD	In WindChill	Chockalingam Kavirayar	Review record for AVW15 Design plan and input review for RFD
LE	In WindChill	Chockalingam Kavirayar	Review record for AVW15 Design plan and input review for RC

7 Record Review and Approval

Name Author	Function Author
Chockalingam Kaviraryar	System Architect

Name Approver	Function Approver	Date *DD-MMM-YYYY*	Signature
Gharpure, Mukul	Product Architect	In WindChill	In WindChill
J, Vivek Kumar	Quality representative	In WindChill	In WindChill
Kamaluddin, Nisha	Independent reviewer	In WindChill	In WindChill
Mustaq, Mohammed	Project manager	In WindChill	In WindChill

End of Document

Printed copies are uncontrolled unless authenticated