PHILIPS

Document Information
Document Identification: D000992888
Document Name: ISP13 DRR validation results
Version: C.2
Life cycle state: Released 08-Dec-2022

Cover Sheet

Document Classification
Design Category: 6. Design Validation
DMR Group Number: -

Signature Information		
Role	**Assignee**	**Date**
Approver (Accountable)	Afek, Keren	03-Dec-2022
Approver (User)	Kavirayar, Chockalingam	03-Dec-2022
Approver (User)	Singh, Jai	05-Dec-2022
Approver (User)	Aryev, Anna	08-Dec-2022
Approver (User)	Eyal Biran	03-Dec-2022

Security Information		
IP Sensitivity	**Strategic Good Indicator**	**ECCN**
Internal	3 Controlled, listed in the US EAR US-origin/subject to the US EAR	EAR99

Note: The stamp below is identifying the following elements
ID number; Version; State; Document Name; DMR Group Number

D000992888;C.2;Released;ISP13 DRR validation results;-

ISPv13 Validation Results Review

VAL3

Hagit Davidesco

Nov-2022

PHILIPS

Validation plan

Validation Cycle	Feature/Objective	Where we validate ?
VAL1	• CT ASPECTS	• Validation Passed – internal
VAL2	• CMR suite (Viewer, Qflow, Functional Long Axis and Spatial) • CT CCA • CT Liver	• Internal / External
VAL3	• CT CS • MR CVS • sAVA • CMR suite (Functional, Mapping and Temporal)	• Internal / External

PHILIPS

D000992888;C.2;Released;ISP13 DRR validation results;-

User Validation Results

Validation Cycle	Application	Total No. of Forms required - 3	Total No. of cases reviewed per evaluator	Pass/Fail	Test
VAL1	ASPECTS	3	2	Passed	All steps have passed
VAL2	CT CCA	3	3	Passed	All steps have passed
VAL2	CT Liver	3	2	Passed	All steps have passed
VAL3	CT Auto Calcium Scoring (CAA)	3	1	Passed	All steps have passed
VAL3	CT sAVA	3	1	Passed	All steps have passed
VAL3	MR Longitudinal Brain Analysis (MR CVS	3	1	Passed	All steps have passed

PHILIPS

D000992888;C.2;Released;ISP13 DRR validation results;-

Validation Results – CMR VAL2









PHILIPS

D000992888;C.2;Released;ISP13 DRR validation results;-

Validation Results – CMR VAL3







PHILIPS

D000992888;C.2;Released;ISP13 DRR validation results;-

Validation Results – forms collection per site



D000992888;C.2;Released;ISP13 DRR validation results;-

Defects assigned to poor scoring

VAL2 CMR Spatial

Comment may by	Description	Defect / RFC ID	Safety Related	Defect Score	RFC	Plan fix	status
Dr. Boussel - 2 bad scoring on 2 different cases read	MRC Spacial :could not correct contours and segmentation in an easy way	1425115	No	3- Normal	Yes	NA	accepted

Defects from Trip report are in section defects in the D000926809 ISP13 Site Management

PHILIPS

D000992888;C.2;Released;ISP13 DRR validation results;-

D000992888;C.2;Released;ISP13 DRR validation results;-