

Validation Master Plan (VMP)

Odoo Enterprise — Mixed GAMP 5 Classification · MDR Devices Medizintechnik GmbH

1 Scope & Classification

This system is mixed-classification, and the validation plan is scaled accordingly — this precision matters to a CIO evaluating audit burden, and overstating a single blanket classification would not survive scrutiny.

GAMP 5 CATEGORY 4 — CONFIGURED PRODUCT

The native core (credential gating, quarantine automation, the analytic spine) runs on Studio field configuration and native automated rules — no bespoke Python modules. Category 4 treatment: Configuration Specification and risk-based testing, not full source-code-level validation.

GAMP 5 CATEGORY 5 — CUSTOM APPLICATION (SCOPED)

The MedTech extension layer — structured UDI compliance-reporting fields, recall/consignment reporting depth, multi-payer & rebate settlement (payer-split invoicing two-way and multi-tier, plus automated rebate match-and-trigger), reusable-instrument sterilization-cycle tracking, and the two-step QA dual-authentication release gate — receives Category 5 rigor because it goes beyond native configuration, even though it remains implemented as Studio-based fields and automated actions. Isolating this scope keeps the validation burden proportional: the small extension footprint gets the higher scrutiny without dragging the native core into Category 5 unnecessarily.

2 Validation Lifecycle

STAGE	DELIVERABLE	OWNER
User Requirements Specification (URS)	Documented business/regulatory requirements per department	MDR Devices business owners
Functional Risk Assessment	Requirement scored by patient-safety and compliance impact; classification (Cat 4 vs. Cat 5) assigned per requirement	much. Consulting / QA
Configuration Specification	Exact field, automation, and workflow configuration mapped to each requirement	much. Consulting
Traceability Matrix	Requirement → Configuration → Test Case → GAMP Category, fully cross-referenced	much. Consulting / QA
Installation Qualification (IQ)	Verified tenant configuration baseline matches Configuration Specification	much. Consulting
Operational Qualification (OQ)	Confirms each automated rule fires correctly in isolation (credential block, quarantine release gate, UDI parse)	QA
Performance Qualification (PQ)	Confirms full lifecycle executes correctly under production-representative data volume	QA / MDR Devices
Validation Summary Report	Signed closure document, referenced at every subsequent system change	QA Lead

rebuild: Studio-based fields and automated actions — including those in the extension layer — carry forward across major version upgrades far more reliably than external custom code, though Odoo's upgrade process does not guarantee zero adjustment for complex automated rules, and the plan budgets time for that review rather than assuming it away.

3 Audit Trail Controls (21 CFR Part 11 · EU GMP Annex 11, EudraLex Vol. 4 — applied as the accepted GxP benchmark for computerised systems, alongside ISO 13485)

Odoo's native `mail.thread` chatter combined with field-level `tracking=True` configuration provides a timestamped, user-attributed, non-editable change log across all in-scope models. Critical models (`stock.lot` , `quality.check` , `account.move`) are configured with archive-only deletion policy — no hard delete path exists for records within the validated boundary, satisfying the permanent audit trail requirement without additional tooling.

4 Electronic Signature Controls

Odoo does not natively provide dual-authentication electronic signatures for quality release actions — a known platform boundary, not an oversight, matching the experience of other regulated device manufacturers running Odoo. It is addressed as a designed two-step release gate, within the Category 5 extension scope, satisfying the dual-control intent of EU GMP Annex 11 and 21 CFR Part 11 §11.10/11.50 without overstating native product capability:

STEP	CONTROL	MECHANISM
1	First authentication	Quality Check recorded as Pass by the assigned quality inspector (credentialed user, logged in chatter).
2	Second authentication	A separately permissioned “QA Release” activity, requiring re-entry of credentials by a distinct authorized role, must complete before the automated quarantine-release action executes. Both steps independently timestamped and user-attributed.

5 Continuous Regression Protection

To protect validated state across upgrade cycles, a defined regression test suite executes against the Configuration Specification on every release candidate, prior to production deployment:

CASE	CORE COMPLIANCE CONTROL
TC-01	Sale order blocks confirmation — expired HCP certification
TC-02	Sale order blocks confirmation — uncertified product category
TC-03	UDI parses correctly from GS1-formatted barcode scan NATIVE
TC-04	UDI sub-fields (batch/serial/expiry) populate correctly EXTENSION LAYER
TC-05	Stock move blocked — quarantine hold not yet elapsed
TC-06	Stock move blocked — quality check not passed
TC-07	Stock move permitted — both hold and check conditions satisfied
TC-08	Consumption move triggers correct payer-split invoice set — two-way case EXTENSION LAYER
TC-09	Analytic distribution tags correctly inherited across split invoices NATIVE
TC-10	Audit trail entries immutable and correctly attributed post-transaction
TC-11	QA Release gate requires distinct second-user authentication EXTENSION LAYER
TC-12	Multi-tier payer claim (3+ parties or rebate netting on one event) reconciles correctly EXTENSION LAYER
TC-13	Quality Check failure — before or after hold timer elapsed — routes to Disposition and blocks movement until Scrap/RTV/Extended-Hold is logged
TC-14	Reusable instrument tray blocked from hospital dispatch once sterilization cycle count reaches its limit EXTENSION LAYER
TC-15	Consignment return (Stage 4B) preserves lot history and correctly triggers credit note only when the original delivery was already billed
TC-16	Sale order against a Pending Verification partner routes to manual review queue rather than hard-blocking
TC-17	<code>x_return_disposition</code> correctly routes a returned lot to Re-Quarantine, Scrap, or Return-to-Vendor per reviewer selection

much. Consulting | MDR Devices Medizintechnik GmbH — Pitch Package v2

Page 3 of 6

TC-17b	An unscannable/unmatched return correctly creates an <i>Under Investigation</i> record routed to Quality, rather than being dropped or force-matched to the wrong lot
TC-18	Non-GS1-conforming label triggers the manual-entry UDI fallback and flags the record <code>x_udi_source = Manual</code>
TC-19	Vendor rebate threshold crossing generates the credit-note claim against the originating purchase chain EXTENSION LAYER

Each test case ties directly back to a Traceability Matrix line item and its assigned GAMP category — failure of any test case blocks promotion to production and triggers a documented deviation review before re-validation.

6 Legacy Data Cutover & Parallel Operation

Open device history records, active complaints, and in-flight CAPAs in the current QMS/LIMS do not get abandoned at go-live, and this plan does not treat cutover as an afterthought discovered in week six.

RECORD STATE AT CUTOVER	TREATMENT
Closed lot, closed complaint	Migrates as reference data; no further transactions expected against it.
Open device history record	Runs in a defined parallel-operation window in the legacy system until closed, then migrates as closed reference data — never opened fresh mid-cutover in the new system.
Active complaint / in-flight CAPA	Remains owned in the legacy QMS through resolution; only closes out and migrates once the legacy record is formally closed.
New activity against an already-migrated device	Opens directly in the new system against the existing migrated lot — the legacy system does not reopen for a device once it has migrated. Only records already open before cutover stay in the legacy system through the parallel run.

Proposed default parallel-run window: the shorter of 90 days or the natural closure of the oldest open record at cutover — a starting position for discovery to confirm against MDR Devices' actual open-record volume, not a fixed promise. Once the parallel window closes, the legacy system moves to **read-only archive status, not decommissioning** — GxP retention obligations on historical records outlive the system being replaced. A cutover reconciliation log (source-to-target record counts and checksums) proves which system was authoritative for which record on which date, retained as a Validation Summary Report attachment.

End-state of the LIMS itself, stated as a position: quality process workflows consolidate into Odoo Quality; laboratory instrumentation data capture likely remains in the LIMS, integrated rather than replaced — confirmed as a discovery decision, so the pitch has a defensible answer rather than silence when asked “what happens to the LIMS after go-live?”

7 Deployment Tier & Change-Control Fit

FIRM RECOMMENDATION: ODOO.SH, NOT ODOO ONLINE

The validated-state claims throughout this plan depend on being able to test a change against a copy of production before it reaches production. Odoo Online applies platform-level upgrades automatically on Odoo's own schedule — the timing of the underlying infrastructure change is not fully within MDR Devices' control, which sits in tension with a change-control regime requiring review-before-promotion. Odoo.sh provides a staging branch tested against a copy of production data before a controlled promotion — the mechanism this plan's “configuration review before each release” step depends on to be more than a paper policy. If a technical CIO asks which tier is proposed, the answer is Odoo.sh, and this is why. Version proposal: current stable Odoo Enterprise at project start, upgraded on a controlled cadence through the Odoo.sh staging mechanism — not an older release held back indefinitely.

8 Data Protection (GDPR) — Distinct From, and Additional To, MDR/GxP

This plan's scope has been MDR/GxP throughout — device traceability, quality process control, audit trail. [res.partner](#) credential records and CRM entries also carry patient-linked information, which is special-category health data under GDPR Article 9, governed by a separate legal regime. This section names the shape of that work without pretending to complete it — a proper GDPR design is its own workstream, owned by MDR Devices' legal/DPO function:

WORKSTREAM	SHAPE OF THE QUESTION
Legal basis and minimization	Which lawful basis covers patient-linked fields in the CRM (a legal determination, not an architectural one), and whether every patient-linked field needs to live in Odoo versus being referenced by an anonymized identifier.
Retention distinct from MDR's clock	MDR/GxP retention (often 10–15 years post-manufacture for implantables) and GDPR data-minimization can pull in different directions — the device history record may need

	to persist while the patient-identifying fields attached to it may not. Two retention rules layered on one record, not one.
Hospital data-processing relationships	If hospitals enter patient-linked data into MDR Devices' CRM, that is a controller/processor relationship needing its own agreement — confirm who is controller and who is processor.
DPIA likelihood	Given the volume and special-category sensitivity, a Data Protection Impact Assessment is likely required before go-live — a named workstream in the project plan, not an afterthought.

Presenting MDR/GxP rigor without acknowledging GDPR sits alongside it as a separate, live requirement is the kind of gap a compliance-literate panel member notices by its absence.