



COMSATS University Islamabad (CUI)
Applied Thermal Physics Laboratory (ATPL)
Department of Physics

Date: April 18, 2022

ATPL/CUI/04/2022/01-03

Thermal transport data of the sample are measured by using state of the art, Hot disk-TPS1500S Thermal Data Measurement System.

Sample Name 1	Thermal Conductivity (W/m-K)
RD-19C/TC/01-03	63.32
	64.14
	64.36
Mean Values	63.94

Sample Name 2	Thermal Conductivity (W/m-K)
RD-19C/TC/01-03	64.11
	64.33
	64.50
Mean Values	64.31

Sample Name 3	Thermal Conductivity (W/m-K)
RD-19C/TC/01-03	63.66
	64.07
	65.11
Mean Values	64.28



Pa
18/4/2022